

Verein für  
Association pour le  
Associazione per il



medizinische Qualitätskontrolle  
contrôle de qualité médical  
controllo di qualità medico

# Survey Report

## 2013 - 3

## Survey specimens

The homogeneity and stability of all specimens was tested before and during shipment and no irregularities were observed.

The following survey specimens were produced by subcontract specifically for MG:

B1 Strep A Test, B2 Uricult, H1 hematology, H4 parasitic hematology, K14 tumor markers.

## Determination of Target Values

For method groups with more than 10 participants, the target values were determined as "consensus values." We use the average value after correction for runaways.


In smaller method groups with a homogeneous distribution, the median value is used as target value. If the distribution of the results does not allow determination of a "consensus value," target values are calculated based on data from production or determined by a specialized laboratory.

## QUALAB Tolerances

For all mandatory analyzes, the Qualab tolerances are used ([www.qualab.ch](http://www.qualab.ch), external quality control).

## Graphs

The results are graphically displayed as follows:



## Comparison of Devices

The data in this report allows you to compare the performance of different devices. However, remember to consider the following:


- The chemical control K1 is a ready-to-use commercial control serum. Even if the specimen is of human origin, it is possible that matrix effects occur. These are device-specific and result in different target values.
- Only one specimen was measured. Since the scatter of the results is dependent on the nature of the specimen (matrix effects) and on the signal strength, the determined coefficient of variations (CV in %) cannot be applied generally.
- A large number of runaways is due to administrative errors (wrong unit, results mixed up) or to operator errors (wrong sample, not correctly taken up in solution, not mixed well) and has nothing to do with the type of device.

Zurich, 09/24/2013

Dr. R. Fried  
Survey Director

*Publication of this report or any portion thereof without our prior written consent is not permitted. The original is archived at [www.mqzh.ch](http://www.mqzh.ch)*

# Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Automat	62	100.0	0.0	0.0	128.3	2.8
2	Cyanmethemoglobin	77	97.4	1.3	1.3	128.8	3.5
3	Sysmex XT/XE/XS	38	97.4	0.0	2.6	130.8	1.5
4	Sysmex K1000	7	85.7	0.0	14.3	129.0	1.3
5	ABX Pentra	13	100.0	0.0	0.0	127.1	1.5
6	Reflotron	109	82.6	11.0	6.4	125.6	5.7
7	Hemocue	313	94.8	1.0	4.2	128.2	2.4
8	Dr. Lange	31	87.1	3.2	9.7	128.5	4.6
9	Hemocontrol	9	88.9	11.1	0.0	132.0	3.9
10	Eurolyser Smart	5	80.0	20.0	0.0	128.0	6.6

# Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Abx Micros	1040	96.7	1.1	2.2	126	2.9
2	Microsemi	49	100.0	0.0	0.0	125	2.5
3	Sysmex KX21	490	96.7	0.4	2.9	128	1.7
4	Sysmex PochH - 100i	213	99.1	0.0	0.9	128	2.0
5	Sysmex XP 300	43	95.3	0.0	4.7	132	1.5
6	Mythic	268	95.5	1.5	3.0	123	3.2
7	Swelab	62	93.6	1.6	4.8	130	2.7
8	MS4	10	90.0	10.0	0.0	125	5.9
9	Abacus Junior	14	100.0	0.0	0.0	131	2.4
10	Medonic	22	95.5	4.5	0.0	127	3.4
11	Nihon Kohden Celltac	23	95.7	0.0	4.3	124	3.7
12	Samsung HC10	14	85.8	7.1	7.1	129	3.8

## Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Automat	52	90.4	5.8	3.8	0.38	4.7
2	Centrifuge	21	81.0	19.0	0.0	0.40	5.3
3	Sysmex XT/XE/XS	37	97.3	0.0	2.7	0.40	2.6
4	Sysmex K1000	7	85.7	0.0	14.3	0.37	3.7
5	ABX Pentra	13	100.0	0.0	0.0	0.36	2.6

## Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Abx Micros	1040	95.6	2.2	2.2	0.36	3.8
2 Microsemi	48	100.0	0.0	0.0	0.36	2.6
3 Sysmex KX21	489	97.4	0.4	2.2	0.37	2.1
4 Sysmex PochH - 100i	212	98.6	0.9	0.5	0.39	2.6
5 Sysmex XP 300	43	93.0	0.0	7.0	0.37	1.7
6 Mythic	268	94.4	3.7	1.9	0.37	3.9
7 Swelab	62	95.2	0.0	4.8	0.38	4.0
8 MS4	10	100.0	0.0	0.0	0.38	4.2
9 Abacus Junior	14	92.9	0.0	7.1	0.38	4.4
10 Medonic	22	95.5	4.5	0.0	0.38	3.7
11 Nihon Kohden Celltac	23	95.7	0.0	4.3	0.38	3.5
12 Samsung HC10	14	92.9	0.0	7.1	0.40	3.9

# Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Automat	48	100.0	0.0	0.0	4.30	4.2
2	Microscopic	14	100.0	0.0	0.0	4.26	10.5
3	Sysmex XT/XE/XS	38	97.4	2.6	0.0	4.41	5.6
4	Sysmex K1000	7	85.7	0.0	14.3	4.29	2.5
5	ABX Pentra	13	100.0	0.0	0.0	4.26	1.7

# Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Abx Micros	1040	98.6	0.7	0.7	4.2	4.6
2	Microsemi	49	100.0	0.0	0.0	4.2	2.2
3	Sysmex KX21	490	98.4	0.4	1.2	4.3	3.2
4	Sysmex PochH - 100i	213	99.5	0.5	0.0	4.5	3.3
5	Sysmex XP 300	43	97.7	0.0	2.3	4.3	3.7
6	Mythic	268	97.4	1.5	1.1	4.2	5.3
7	Swelab	62	100.0	0.0	0.0	4.3	5.0
8	MS4	10	100.0	0.0	0.0	4.4	4.2
9	Abacus Junior	14	100.0	0.0	0.0	4.4	5.5
10	Medonic	22	100.0	0.0	0.0	4.3	3.0
11	Samsung HC10	14	100.0	0.0	0.0	4.3	5.6
12	Nihon Kohden Celltac	23	95.7	0.0	4.3	4.4	4.6

## Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Automat	44	100.0	0.0	0.0	5.74	7.3
2	Microscopic	101	88.1	6.9	5.0	5.37	14.1
3	Sysmex XT/XE/XS	38	100.0	0.0	0.0	5.93	3.4
4	Sysmex K1000	7	85.7	0.0	14.3	5.70	1.8
5	ABX Pentra	13	100.0	0.0	0.0	5.76	5.1

# Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Abx Micros	1040	99.4	0.2	0.4	5.27	4.9
2	Microsemi	49	100.0	0.0	0.0	5.26	3.2
3	Sysmex KX21	490	99.6	0.0	0.4	5.64	3.7
4	Sysmex PochH - 100i	213	100.0	0.0	0.0	5.59	3.6
5	Sysmex XP 300	43	97.7	2.3	0.0	5.95	4.8
6	Mythic	266	98.8	0.8	0.4	5.53	5.1
7	Nihon Kohden Celltac	23	100.0	0.0	0.0	5.67	4.4
8	Swelab	62	100.0	0.0	0.0	5.85	6.5
9	MS4	10	100.0	0.0	0.0	5.83	6.4
10	Abacus Junior	14	92.9	0.0	7.1	6.48	8.0
11	Medonic	22	100.0	0.0	0.0	5.77	4.5
12	Samsung HC10	14	100.0	0.0	0.0	5.47	6.8

# Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Automat	40	100.0	0.0	0.0	175.6	7.9
2	Microscopic	66	92.5	3.0	4.5	177.7	12.3
3	Sysmex XT/XE/XS	38	97.4	2.6	0.0	175.8	6.4
4	Sysmex K1000	7	100.0	0.0	0.0	176.0	7.0
5	Advia 120	4	100.0	0.0	0.0	175.5	8.1
6	ABX Pentra	13	100.0	0.0	0.0	191.8	4.1

# Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Abx Micros	1040	97.4	1.3	1.3	177.0	8.7
2	Microsemi	49	100.0	0.0	0.0	180.4	5.8
3	Sysmex KX21	490	99.6	0.2	0.2	175.7	4.9
4	Sysmex PochH - 100i	210	99.5	0.5	0.0	179.4	4.7
5	Sysmex XP 300	43	97.7	0.0	2.3	181.3	4.1
6	Mythic	268	96.7	2.2	1.1	182.0	10.1
7	Swelab	62	93.5	6.5	0.0	166.4	10.7
8	MS4	10	70.0	0.0	30.0	175.0	9.9
9	Abacus Junior	14	85.8	7.1	7.1	196.2	12.9
10	Medonic	22	100.0	0.0	0.0	167.8	5.8
11	Nihon Kohden Celltac	23	91.3	0.0	8.7	177.7	7.6
12	Samsung HC10	14	92.9	7.1	0.0	182.1	14.6

## Hematocrite - QBC




QUALAB tolerance : 9 %

Hematocrite - QBC (I/I)


No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 QBC	10	100.0	0.0	0.0	0.32	3.0

## Hemoglobin - QBC



No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 QBC	11	100.0	0.0	0.0	106.3	1.5

## Leucocytes - QBC




QUALAB tolerance : 25 %

Leucocytes - QBC (G/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 QBC	11	100.0	0.0	0.0	16.72	15.4

## Thrombocytes - QBC




QUALAB tolerance : 25 %

Thrombocytes - QBC (G/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 QBC	10	100.0	0.0	0.0	200.7	13.6


## CRP



No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Afinion	949	98.2	1.1	0.7	43.0	8.0
2	NycoCard SingleTest-	658	85.1	5.3	9.6	46.4	11.3
3	Abx Micros	231	90.9	1.3	7.8	45.0	9.3
4	ABX Micros CRP200	346	88.8	7.2	4.0	45.8	11.1
5	QuikRead (Vollblut)	235	96.6	1.7	1.7	77.1	7.5
6	Quick Read go	43	97.7	0.0	2.3	46.1	7.4
7	Turbidimetry	35	100.0	0.0	0.0	46.3	8.0
8	Cobas	13	100.0	0.0	0.0	45.4	9.1
9	Fuji Dri-Chem	22	95.5	0.0	4.5	41.8	9.8
10	Eurolyser Smart	102	77.4	5.9	16.7	42.3	10.9
11	AQT 90 FLEX	7	100.0	0.0	0.0	42.5	13.4
12	Spotchem D-Concept	5	100.0	0.0	0.0	51.1	6.1
13	Spotchem SI-3510	10	100.0	0.0	0.0	44.7	8.1

# I2 Plasmaproteins

## IgG




QUALAB tolerance : 25 %

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Turbidimetry	9	100.0	0.0	0.0	14.9	2.6
2	Nephelometry	7	100.0	0.0	0.0	15.5	6.9

# I2 Plasmaproteins

## IgA




QUALAB tolerance : 25 %

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Turbidimetry	10	100.0	0.0	0.0	2.2	6.4
2	Nephelometry	7	100.0	0.0	0.0	2.4	3.6

# I2 Plasmaproteins

## IgM




QUALAB tolerance : 25 %

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Nephelometry	7	100.0	0.0	0.0	1.2	5.3
2	Cobas Integra 800/40	6	100.0	0.0	0.0	1.1	4.3


# I2 Plasmaproteins

## IgE



No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	all Participants	9	100.0	0.0	0.0	101	6.7

## Alpha-1-Antitrypsin




QUALAB tolerance : 25 %

Alpha-1-Antitrypsin (g/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Nephelometry	4	100.0	0.0	0.0	1.13	3.1

## Anti-streptolysin-O




QUALAB tolerance : 25 %

Anti-streptolysin-O (kIU/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	10	100.0	0.0	0.0	145	4.9

## Complement C3




QUALAB tolerance : 25 %

Complement C3 (g/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	11	100.0	0.0	0.0	1.41	5.1

## Complement C4




QUALAB tolerance : 25 %

Complement C4 (g/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	10	100.0	0.0	0.0	0.29	7.4

# Haptoglobin




QUALAB tolerance : 25 %

Haptoglobin (g/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	11	100.0	0.0	0.0	1.22	3.5

## Transferrin




QUALAB tolerance : 25 %

Transferrin (g/l)


No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	14	100.0	0.0	0.0	2.22	4.0

## Präalbumin




No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	6	100.0	0.0	0.0	198.0	6.2

## Albumin



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	12	100.0	0.0	0.0	37	5.2
2	Cobas	16	100.0	0.0	0.0	41	4.4
3	Fuji Dri-Chem	135	99.3	0.0	0.7	50	3.8
4	Spotchem/Ready	59	89.8	6.8	3.4	42	6.4
5	Spotchem D-Concept	47	100.0	0.0	0.0	49	4.6
6	Piccolo	15	100.0	0.0	0.0	38	2.3
7	Abx Mira	6	100.0	0.0	0.0	40	2.3
8	Hitachi S40/M40	4	100.0	0.0	0.0	42	0.6

## Alkaline phosphatase




QUALAB tolerance : 21 %

Alkaline phosphatase (U/l)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 IFCC/SGKC 37°C	7	100.0	0.0	0.0	109	8.3
2 Cobas	22	100.0	0.0	0.0	93	6.3
3 Reflotron	758	97.7	1.2	1.1	145	7.4
4 Fuji Dri-Chem	567	99.6	0.2	0.2	114	6.1
5 Spotchem/Ready	153	98.0	2.0	0.0	140	8.4
6 Spotchem D-Concept	90	97.8	0.0	2.2	129	8.1
7 Hitachi S40/M40	9	100.0	0.0	0.0	89	5.1
8 Olympus	5	100.0	0.0	0.0	128	8.0
9 Piccolo	16	100.0	0.0	0.0	125	5.4
10 Abx Mira	18	94.4	5.6	0.0	105	10.2

## Amylase




QUALAB tolerance : 18 %

Amylase (U/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	IFCC EPS liquid 37°C	9	100.0	0.0	0.0	176	5.6
2	Cobas	8	100.0	0.0	0.0	170	2.5
3	Reflotron	204	97.5	2.5	0.0	141	7.9
4	Fuji Dri-Chem	433	99.0	0.5	0.5	122	5.3
5	Spotchem/Ready	94	82.9	12.8	4.3	63	11.9
6	Spotchem D-Concept	72	100.0	0.0	0.0	151	7.1
7	Piccolo	13	100.0	0.0	0.0	152	2.3
8	Abx Mira	8	100.0	0.0	0.0	156	5.5
9	Hitachi S40/M40	6	100.0	0.0	0.0	194	6.7

## Pancreatic amylase




QUALAB tolerance : 18 %

Pancreatic amylase (U/l)


No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	IFCC EPS liquid 37°C	14	92.9	7.1	0.0	141	8.5
2	Cobas	13	100.0	0.0	0.0	145	3.2
3	Reflotron	481	97.3	2.3	0.4	134	7.7
4	wet chemistry, other	4	100.0	0.0	0.0	140	2.2

## Bilirubin



No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	9	100.0	0.0	0.0	30.0	12.3
2	Cobas	19	100.0	0.0	0.0	27.9	7.3
3	Reflotron	553	93.5	3.4	3.1	34.6	7.8
4	Fuji Dri-Chem	411	94.0	4.1	1.9	34.0	8.7
5	Spotchem/Ready	116	87.1	9.5	3.4	34.0	11.7
6	Spotchem D-Concept	74	93.2	4.1	2.7	28.8	8.3
7	Beckman/Olympus	6	83.3	16.7	0.0	37.2	13.4
8	Piccolo	15	80.0	13.3	6.7	27.1	12.1
9	Abx Mira	18	83.3	16.7	0.0	28.1	11.4
10	Hitachi S40/M40	8	87.5	0.0	12.5	32.5	8.9

## Bilirubin direct




QUALAB tolerance : 18 %

Bilirubin direct (umol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Fuji Dri-Chem	28	64.3	10.7	25.0	20.3	11.7

## Calcium




QUALAB tolerance : 12 %

Calcium (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	15	93.3	0.0	6.7	2.06	2.6
2	Cobas	16	100.0	0.0	0.0	2.06	2.3
3	Fuji Dri-Chem	291	98.6	0.0	1.4	1.99	4.1
4	Spotchem/Ready	61	93.4	6.6	0.0	1.73	5.7
5	Spotchem D-Concept	44	95.4	2.3	2.3	1.66	6.3
6	Piccolo	14	100.0	0.0	0.0	2.01	2.8
7	Abx Mira	13	92.3	0.0	7.7	2.05	3.9
8	Hitachi S40/M40	5	100.0	0.0	0.0	1.95	3.2
9	ISE	4	100.0	0.0	0.0	0.92	1.4

# Chloride




QUALAB tolerance : 6 %

Chloride (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	ISE	11	100.0	0.0	0.0	99	1.8
2	Cobas	12	83.4	8.3	8.3	97	2.9
3	Fuji Dri-Chem	490	98.4	0.6	1.0	111	2.1
4	Spotchem D-Concept	80	82.5	7.5	10.0	124	3.3
5	Spotchem EL-SE 1520	120	72.5	10.8	16.7	135	3.8
6	Piccolo	8	100.0	0.0	0.0	97	2.4
7	iStat Chem8	4	100.0	0.0	0.0	107	3.1

## Cholesterol total




QUALAB tolerance : 10 %

Cholesterol total (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	19	100.0	0.0	0.0	4.8	2.7
2	Cobas	20	100.0	0.0	0.0	4.7	3.6
3	Reflotron	970	97.2	1.6	1.2	4.5	4.2
4	Fuji Dri-Chem	608	98.7	0.8	0.5	4.6	3.7
5	Spotchem/Ready	180	92.8	5.0	2.2	4.7	5.1
6	Spotchem D-Concept	93	97.8	1.1	1.1	5.0	4.4
7	Piccolo	12	91.7	8.3	0.0	4.8	3.8
8	Cholestech LDX	197	95.9	3.6	0.5	4.6	4.9
9	Abx Mira	16	93.7	6.3	0.0	4.9	5.3
10	Lange	4	75.0	0.0	25.0	4.6	5.8
11	Hitachi S40/M40	11	100.0	0.0	0.0	4.7	2.5

## Cholesterin HDL




QUALAB tolerance : 21 %

Cholesterin HDL (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Wet chemistry, direc	15	80.0	6.7	13.3	1.3	12.2
2	Cobas	19	100.0	0.0	0.0	1.7	4.3
3	Reflotron	735	75.6	12.4	12.0	1.2	13.9
4	Fuji Dri-Chem	567	99.6	0.2	0.2	1.9	5.2
5	Spotchem/Ready	163	82.2	8.6	9.2	0.8	11.0
6	Spotchem D-Concept	90	82.2	8.9	8.9	0.9	12.2
7	Piccolo	12	100.0	0.0	0.0	1.2	9.8
8	Cholestech LDX	197	92.4	6.1	1.5	1.5	10.5
9	Abx Mira	16	74.9	6.3	18.8	1.4	13.1
10	Hitachi S40/M40	10	100.0	0.0	0.0	1.6	4.1

## Creatine kinase




QUALAB tolerance : 18 %

Creatine kinase (U/l)


No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	IFCC/SGKC 37'C	15	93.3	0.0	6.7	225	4.3
2	Cobas	18	100.0	0.0	0.0	226	2.9
3	Reflotron	483	95.2	3.1	1.7	181	6.8
4	Fuji Dri-Chem	362	97.8	0.8	1.4	232	4.6
5	Spotchem/Ready	74	89.1	9.5	1.4	199	10.9
6	Spotchem D-Concept	59	71.1	15.3	13.6	177	13.7
7	Abx Mira	16	93.7	6.3	0.0	252	7.7
8	Hitachi S40/M40	5	100.0	0.0	0.0	175	5.5

# Iron



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	6	100.0	0.0	0.0	35	5.1
2	Cobas	12	100.0	0.0	0.0	37	2.2
3	Abx Mira	4	100.0	0.0	0.0	37	4.9

## Gamma-glutamyltransferase




QUALAB tolerance : 18 %

Gamma-glutamyltransferase (U/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 IFCC/SGKC 37'C	5	100.0	0.0	0.0	51	10.3
2 Cobas	19	100.0	0.0	0.0	49	7.1
3 Reflotron	993	97.7	1.0	1.3	63	6.0
4 Fuji Dri-Chem	620	99.5	0.3	0.2	67	6.0
5 Spotchem/Ready	177	95.4	2.3	2.3	61	7.6
6 Spotchem D-Concept	99	60.6	15.2	24.2	72	13.6
7 Vitros/Ektachem	5	80.0	20.0	0.0	78	10.7
8 DGKC 37'C	8	100.0	0.0	0.0	50	9.1
9 Piccolo	15	100.0	0.0	0.0	46	3.5
10 Abx Mira	18	100.0	0.0	0.0	53	4.4
11 Hitachi S40/M40	11	100.0	0.0	0.0	61	3.1

## Glucose




QUALAB tolerance : 10 %

Glucose (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	32	87.5	9.4	3.1	7.3	5.7
2	Cobas	20	100.0	0.0	0.0	7.5	3.5
3	Reflotron	1015	91.4	7.0	1.6	6.9	5.6
4	Fuji Dri-Chem	582	99.5	0.0	0.5	6.9	2.1
5	Spotchem/Ready	162	88.9	8.0	3.1	7.8	5.2
6	Spotchem D-Concept	95	100.0	0.0	0.0	7.5	4.0
7	Vitros/Ektachem	4	100.0	0.0	0.0	7.4	3.4
8	Piccolo	17	100.0	0.0	0.0	7.6	1.7
9	Cholestech LDX	160	95.0	3.1	1.9	7.1	4.2
10	Abx Mira	18	94.4	5.6	0.0	7.6	4.4
11	Lange	9	77.8	0.0	22.2	7.1	5.2
12	Hitachi S40/M40	11	90.9	0.0	9.1	7.4	2.4

# Glucose




QUALAB tolerance : 10 %

Glucose (mmol/l)


No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Accu-Chek Aviva	572	97.2	1.6	1.2	7.8	3.4
2	Accu-Chek Inform 2	49	100.0	0.0	0.0	8.0	2.5
3	Accu-Chek Mobile	5	100.0	0.0	0.0	7.9	3.2
4	Bayer Contour 2 (5s)	355	77.2	6.5	16.3	5.8	5.1
5	Bayer Contour XT/NEX	616	98.7	1.1	0.2	7.4	3.8
6	Bayer Breeze 2	18	94.4	5.6	0.0	12.7	5.8
7	Glucocard	9	100.0	0.0	0.0	10.6	6.1
8	Hemocue (Plasma)	49	96.0	2.0	2.0	8.9	4.0
9	mylife Pura	48	95.8	2.1	2.1	9.0	4.2
10	Hemocue RT	13	92.3	0.0	7.7	8.8	2.1
11	Freestyle Freedom li	12	91.7	0.0	8.3	8.0	3.0
12	Sanofi BG Star	4	100.0	0.0	0.0	9.4	4.4

## Glucose B



No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Bayer Elite	14	92.9	7.1	0.0	8.7	4.9
2	Hemocue	68	92.7	4.4	2.9	8.6	5.8
3	AccuChek Sensor	7	85.7	14.3	0.0	7.8	5.6
4	OneTouch Ultra	33	90.9	0.0	9.1	11.6	4.3
5	OneTouch Verio	4	100.0	0.0	0.0	7.5	2.3
6	AccuChek Compact	7	85.7	14.3	0.0	7.7	4.9
7	Bayer Contour (15s)	127	68.5	13.4	18.1	9.0	6.3
8	Wellion Smart	4	75.0	25.0	0.0	9.2	7.8
9	Healthpro	9	100.0	0.0	0.0	13.1	2.9

## Uric Acid




QUALAB tolerance : 15 %

Uric Acid (umol/l)


No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	18	100.0	0.0	0.0	282	5.7
2	Cobas	15	100.0	0.0	0.0	281	3.2
3	Reflotron	881	98.3	0.3	1.4	311	4.5
4	Fuji Dri-Chem	583	99.8	0.0	0.2	311	2.4
5	Spotchem/Ready	154	93.5	6.5	0.0	267	7.0
6	Spotchem D-Concept	90	98.9	0.0	1.1	271	4.6
7	Piccolo	13	92.3	0.0	7.7	234	2.9
8	Abx Mira	16	100.0	0.0	0.0	274	6.2
9	Lange	4	50.0	0.0	50.0	268	5.6
10	Hitachi S40/M40	11	100.0	0.0	0.0	278	4.4

## Urea



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	14	92.9	7.1	0.0	10.0	6.7
2	Cobas	19	100.0	0.0	0.0	9.7	3.9
3	Reflotron	382	97.4	0.5	2.1	9.8	4.9
4	Fuji Dri-Chem	357	100.0	0.0	0.0	10.0	2.4
5	Spotchem/Ready	107	89.8	6.5	3.7	9.4	8.4
6	Spotchem D-Concept	64	85.9	9.4	4.7	8.6	8.4
7	Piccolo	17	94.1	0.0	5.9	8.9	3.0
8	Abx Mira	9	100.0	0.0	0.0	10.0	5.6
9	Hitachi S40/M40	8	100.0	0.0	0.0	9.5	3.0
10	iStat Chem8	4	100.0	0.0	0.0	13.9	2.4

# Potassium




QUALAB tolerance : 6 %

Potassium (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	ISE	21	90.5	0.0	9.5	3.58	1.2
2	Cobas	20	100.0	0.0	0.0	3.60	2.2
3	Reflotron	907	89.1	8.3	2.6	3.90	3.4
4	Fuji Dri-Chem	613	98.3	1.0	0.7	3.50	2.0
5	Spotchem D-Concept	91	97.8	1.1	1.1	3.35	2.3
6	Vitros/Ektachem	6	66.7	33.3	0.0	3.70	6.4
7	Spotchem EL-SE 1520	127	95.2	2.4	2.4	3.15	2.6
8	Piccolo	9	88.9	0.0	11.1	4.10	3.2
9	Abx Mira	6	100.0	0.0	0.0	3.63	3.0
10	iStat Chem8	4	100.0	0.0	0.0	3.50	1.4

## Creatinine




QUALAB tolerance : 20 %

Creatinine (umol/l)


No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	7	100.0	0.0	0.0	202	7.2
2	Cobas	21	100.0	0.0	0.0	202	4.5
3	Reflotron	1092	98.4	0.8	0.8	222	5.8
4	Fuji Dri-Chem	649	99.1	0.9	0.0	194	4.7
5	Vitros/Ektachem	6	100.0	0.0	0.0	222	9.4
6	Jaffé	8	100.0	0.0	0.0	201	8.2
7	Enzymatic	6	100.0	0.0	0.0	201	5.1
8	Piccolo	15	100.0	0.0	0.0	198	8.6
9	Statsensor i / Nova	20	85.0	0.0	15.0	347	10.0
10	Abx Mira	18	100.0	0.0	0.0	196	6.6
11	Hitachi S40/M40	10	90.0	0.0	10.0	190	2.4
12	iStat Chem8	6	100.0	0.0	0.0	200	4.1
13	ABL700/800 Radiomete	10	100.0	0.0	0.0	213	3.0

# LDH



No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 IFCC	17	100.0	0.0	0.0	149	5.4
2 Cobas	11	100.0	0.0	0.0	314	1.8
3 Fuji Dri-Chem	138	97.9	0.7	1.4	125	6.0
4 Spotchem/Ready	46	97.8	2.2	0.0	240	9.4
5 Spotchem D-Concept	29	62.1	24.1	13.8	85	18.6
6 Abx Mira	9	100.0	0.0	0.0	156	5.2

## Magnesium




QUALAB tolerance : 12 %

Magnesium (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	10	90.0	0.0	10.0	0.81	3.9
2	Cobas	13	100.0	0.0	0.0	0.82	2.3
3	Fuji Dri-Chem	104	96.2	1.9	1.9	0.87	4.2
4	Spotchem D-Concept	20	95.0	5.0	0.0	0.61	6.0
5	Spotchem/Ready	21	100.0	0.0	0.0	0.65	4.2

## Sodium




QUALAB tolerance : 6 %

Sodium (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	ISE	19	100.0	0.0	0.0	143	1.4
2	Cobas	20	100.0	0.0	0.0	144	1.6
3	Fuji Dri-Chem	565	98.1	1.4	0.5	146	1.4
4	Spotchem D-Concept	86	100.0	0.0	0.0	138	1.8
5	Vitros/Ektachem	4	100.0	0.0	0.0	152	4.0
6	Spotchem EL-SE 1520	127	96.9	3.1	0.0	136	2.2
7	Piccolo	9	100.0	0.0	0.0	145	2.6
8	Abx Mira	6	100.0	0.0	0.0	148	1.8
9	iStat Chem8	4	100.0	0.0	0.0	142	0.7

## Phosphate




QUALAB tolerance : 15 %

Phosphate (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	10	100.0	0.0	0.0	1.2	5.2
2	Cobas	13	100.0	0.0	0.0	1.2	1.7
3	Fuji Dri-Chem	70	100.0	0.0	0.0	1.2	2.9
4	Spotchem D-Concept	13	100.0	0.0	0.0	1.3	2.6
5	Spotchem/Ready	12	100.0	0.0	0.0	1.2	4.2
6	Piccolo	5	100.0	0.0	0.0	1.5	5.7
7	Abx Mira	4	100.0	0.0	0.0	1.4	1.4

## Protein total




QUALAB tolerance : 12 %

Protein total (g/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	13	100.0	0.0	0.0	62.0	2.0
2	Cobas	15	100.0	0.0	0.0	61.2	2.5
3	Fuji Dri-Chem	155	98.8	0.6	0.6	60.1	3.3
4	Spotchem/Ready	58	94.9	3.4	1.7	63.1	5.2
5	Spotchem D-Concept	43	90.6	4.7	4.7	69.3	7.6
6	Piccolo	14	100.0	0.0	0.0	63.9	1.7
7	Abx Mira	6	100.0	0.0	0.0	60.0	5.0

## Aspartate aminotransferase




QUALAB tolerance : 18 %

Aspartate aminotransferase (U/l)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 IFCC with Pyridox 37	13	100.0	0.0	0.0	97	4.7
2 Cobas	20	100.0	0.0	0.0	96	5.2
3 Reflotron	1012	98.2	1.1	0.7	128	6.4
4 Fuji Dri-Chem	625	99.6	0.2	0.2	90	3.4
5 Spotchem/Ready	189	96.8	1.1	2.1	70	5.7
6 Spotchem D-Concept	96	96.9	3.1	0.0	77	6.8
7 Vitros/Ektachem	5	100.0	0.0	0.0	123	9.2
8 IFCC with Pyridox 37	6	83.3	16.7	0.0	98	12.0
9 Piccolo	16	100.0	0.0	0.0	90	3.3
10 Abx Mira	18	94.4	5.6	0.0	96	9.0
11 Hitachi S40/M40	12	91.7	0.0	8.3	96	2.9

## Alanine aminotransferase




QUALAB tolerance : 18 %

Alanine aminotransferase (U/l)


No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 IFCC with Pyridox 37	14	100.0	0.0	0.0	92	5.0
2 Cobas	21	100.0	0.0	0.0	90	3.8
3 Reflotron	1042	98.0	1.2	0.8	82	5.2
4 Fuji Dri-Chem	636	99.3	0.2	0.5	89	3.5
5 Spotchem/Ready	190	97.9	2.1	0.0	68	6.8
6 Spotchem D-Concept	99	99.0	0.0	1.0	61	7.1
7 Vitros/Ektachem	4	75.0	25.0	0.0	106	12.8
8 IFCC with Pyridox 37	4	100.0	0.0	0.0	91	2.9
9 Piccolo	17	94.1	0.0	5.9	86	4.7
10 Abx Mira	18	100.0	0.0	0.0	93	5.5
11 Hitachi S40/M40	12	91.7	0.0	8.3	93	2.0

## Triglycerides




No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	21	90.4	4.8	4.8	2.21	9.1
2	Cobas	20	100.0	0.0	0.0	2.19	2.3
3	Reflotron	857	98.5	0.6	0.9	2.88	5.0
4	Fuji Dri-Chem	583	99.5	0.3	0.2	2.65	3.8
5	Spotchem/Ready	166	100.0	0.0	0.0	1.46	6.0
6	Spotchem D-Concept	93	100.0	0.0	0.0	1.63	3.8
7	Vitros/Ektachem	4	100.0	0.0	0.0	2.69	12.2
8	Hitachi S40/M40	10	100.0	0.0	0.0	0.60	7.5
9	Piccolo	10	100.0	0.0	0.0	2.52	1.3
10	Cholestech LDX	197	100.0	0.0	0.0	2.25	5.5
11	Abx Mira	15	100.0	0.0	0.0	2.11	4.9

## Fructosamine




No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Spotchem/Ready	5	100.0	0.0	0.0	601	3.5

# Lipase



No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Abx Mira	4	100.0	0.0	0.0	42.0	12.0
2 Beckman/Olympus	5	60.0	40.0	0.0	52.0	13.1
3 Standard chemistry	8	100.0	0.0	0.0	42.5	6.2

## LDL Cholesterin




QUALAB tolerance : 25 %

LDL Cholesterin (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Abx Mira	10	100.0	0.0	0.0	2.4	5.6
2	Standard chemistry	5	100.0	0.0	0.0	2.5	15.9
3	Roche, Cobas	5	100.0	0.0	0.0	3.3	3.3
4	Hitachi S40/M40	4	100.0	0.0	0.0	1.8	4.1

# Lithium




QUALAB tolerance : 15 %

Lithium (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	all Participants	12	100.0	0.0	0.0	1.26	6.2

# Lactate




QUALAB tolerance : 15 %

Lactate (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	8	87.5	0.0	12.5	2.59	7.3

## Creatinine SP




QUALAB tolerance : 20 %

Creatinine SP (umol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Spotchem/Ready	188	95.2	2.7	2.1	237	8.6
2	Spotchem D-Concept	97	98.0	1.0	1.0	230	7.5

## HbA1c ample A




QUALAB tolerance : 9 %

HbA1c ample A (%)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas b101	6	100.0	0.0	0.0	5.2	3.9
2	Afinion	541	97.2	2.2	0.6	5.4	3.1
3	Hemocue HbA1c 501	4	25.0	50.0	25.0	5.3	3.2
4	NycoCard	183	85.2	7.7	7.1	5.5	5.2
5	DCA2000/Vantage	207	95.6	3.9	0.5	5.3	3.6
6	Others	8	87.5	0.0	12.5	5.3	3.5
7	HPLC	5	80.0	0.0	20.0	5.2	1.6
8	Roche, Cobas	21	81.0	9.5	9.5	5.3	4.1
9	Hitado Super D	6	100.0	0.0	0.0	5.2	5.0
10	A1c Now	27	92.6	3.7	3.7	5.3	4.6

## HbA1c sample B




QUALAB tolerance : 9 %

HbA1c sample B (%)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Afinion	426	96.5	2.1	1.4	5.2	3.4
2 A1c Now	27	92.6	3.7	3.7	5.1	4.4
3 Hemocue HbA1c 501	4	75.0	25.0	0.0	5.2	8.3
4 NycoCard	169	76.3	16.6	7.1	5.3	5.8
5 DCA2000/Vantage	177	98.3	0.0	1.7	5.2	3.3
6 Others	5	100.0	0.0	0.0	5.2	5.0
7 Roche, Cobas	17	88.2	0.0	11.8	5.2	4.3
8 Hitado Super D	8	100.0	0.0	0.0	5.4	3.3

## pCO2




QUALAB tolerance : 15 %

pCO2 (kPa)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Roche (OMNI/AVL)	12	100.0	0.0	0.0	1.61	3.8
2	iStat	34	97.1	0.0	2.9	1.78	5.6
3	EPOC	7	57.1	14.3	28.6	1.11	10.6

# K4 Blood gases


## pO2



No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Roche (OMNI/AVL)	9	66.7	11.1	22.2	10.85	10.8
2	iStat	34	94.1	5.9	0.0	11.29	6.5
3	EPOC	7	85.7	0.0	14.3	9.75	5.6


# K4 Blood gases

## pH



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Roche (OMNI/AVL)	11	100.0	0.0	0.0	7.58	0.2
2	iStat	35	100.0	0.0	0.0	7.66	0.1
3	EPOC	7	100.0	0.0	0.0	7.70	0.3

## Glucose BG




QUALAB tolerance : 10 %

Glucose BG (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 iStat	12	100.0	0.0	0.0	13.7	1.0
2 EPOC	4	100.0	0.0	0.0	15.2	3.1

## Potassium BG




QUALAB tolerance : 6 %

Potassium BG (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Roche (OMNI/AVL)	9	100.0	0.0	0.0	5.4	2.1
2	iStat	18	100.0	0.0	0.0	5.3	0.9
3	EPOC	5	100.0	0.0	0.0	4.9	1.1

## Sodium BG




QUALAB tolerance : 6 %

Sodium BG (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Roche (OMNI/AVL)	10	100.0	0.0	0.0	124.1	1.8
2	iStat	18	100.0	0.0	0.0	126.6	0.6
3	EPOC	5	100.0	0.0	0.0	116.0	1.4

## Calcium-BG




QUALAB tolerance : 12 %

Calcium-BG (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Roche (OMNI/AVL)	7	85.7	0.0	14.3	0.38	6.0
2	iStat	10	100.0	0.0	0.0	0.37	1.1
3	EPOC	5	100.0	0.0	0.0	0.31	5.0

## Lactate-BG




QUALAB tolerance : 10 %

Lactate-BG (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 EPOC	5	60.0	0.0	40.0	13.80	4.7
2 iStat	5	100.0	0.0	0.0	10.87	2.2

## Kreatinin BG




QUALAB tolerance : 20 %

Kreatinin BG (umol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 iStat	4	100.0	0.0	0.0	45	7.8

## Calcium - Urine




QUALAB tolerance : 12 %

Calcium - Urine (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Standard chemistry	12	100.0	0.0	0.0	2.26	1.7

## Chloride - Urine




QUALAB tolerance : 6 %

Chloride - Urine (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Standard chemistry	6	100.0	0.0	0.0	143	2.0

## Glucose - Urine




QUALAB tolerance : 10 %

Glucose - Urine (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Standard chemistry	12	100.0	0.0	0.0	9.2	2.2

## Magnesium - Urine




QUALAB tolerance : 12 %

Magnesium - Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Standard chemistry	6	100.0	0.0	0.0	3.6	5.3

## Osmolality - Urine




QUALAB tolerance : 10 %

Osmolality - Urine (mosm/kg)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Cryoskopie	5	100.0	0.0	0.0	620	0.7

## Phosphate - Urine




QUALAB tolerance : 15 %

Phosphate - Urine (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Standard chemistry	12	100.0	0.0	0.0	12.4	3.5

## Potassium - Urine




QUALAB tolerance : 15 %

Potassium - Urine (mmol/l)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	16	100.0	0.0	0.0	50	4.4

## total Protein - Urine




QUALAB tolerance : 20 %

total Protein - Urine (mg/l)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Cobas/Roche	8	100.0	0.0	0.0	329.5	2.0
2 Other methods	4	100.0	0.0	0.0	421.0	7.9

## Sodium - Urine




QUALAB tolerance : 15 %

Sodium - Urine (mmol/l)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	16	100.0	0.0	0.0	125	1.9

## Urea - Urine




QUALAB tolerance : 15 %

Urea - Urine (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Standard chemistry	12	100.0	0.0	0.0	203	5.0

## Uric Acid - Urine




QUALAB tolerance : 15 %

Uric Acid - Urine (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Standard chemistry	11	100.0	0.0	0.0	0.74	5.5

## Specific Gravity - Urine




QUALAB tolerance : 5 %

Specific Gravity - Urine ()

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Refractometer	8	100.0	0.0	0.0	1.019	0.1

# G1 Coagulation INR

## INR




QUALAB tolerance : 15 %

INR ()


No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Thromborel S	7	85.7	0.0	14.3	1.73	5.1
2	Neoplastin Plus	4	100.0	0.0	0.0	2.08	3.6
3	Innovin	19	100.0	0.0	0.0	1.63	6.6
4	Other methods	5	80.0	0.0	20.0	1.68	3.5
5	Neoplastin R	9	88.9	11.1	0.0	1.77	7.5

## Fibrinogen OA




No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Other methods	10	90.0	10.0	0.0	1.03	9.5
2	Siemens Thrombin	8	100.0	0.0	0.0	1.00	6.9
3	Stago/STA	7	100.0	0.0	0.0	1.09	4.1

## Activated Prothrombin Time



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Other methods	15	100.0	0.0	0.0	46.1	9.1
2	Actin FS	11	90.9	9.1	0.0	50.4	13.2
3	Pathromtin SL	4	100.0	0.0	0.0	65.3	5.8
4	Stago/STA	5	100.0	0.0	0.0	50.4	3.0

## Prothrombin time NT




QUALAB tolerance : 15 %

Prothrombin time NT (%)


No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Neoplastin R	6	100.0	0.0	0.0	86	1.4
2	Innovin	15	100.0	0.0	0.0	92	6.3
3	all Participants	4	100.0	0.0	0.0	96	4.8
4	Recombiplastin IL	4	100.0	0.0	0.0	102	3.3

## Fibrinogen N



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Siemens Thrombin	9	100.0	0.0	0.0	2.23	9.5
2	Other methods	13	100.0	0.0	0.0	2.28	6.3
3	Stago/STA	7	100.0	0.0	0.0	2.37	3.7

## aPTT N




QUALAB tolerance : 25 %

aPTT N (Sek)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Actin FS	12	100.0	0.0	0.0	28.2	8.2
2 Other methods	13	100.0	0.0	0.0	28.0	9.9
3 Stago/STA	5	100.0	0.0	0.0	32.4	4.9

## Prothrombin time HT




QUALAB tolerance : 15 %

Prothrombin time HT (%)


No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Neoplastin R	5	100.0	0.0	0.0	92	3.2
2	Innovin	13	92.3	7.7	0.0	85	8.7
3	Recombiplastin IL	4	100.0	0.0	0.0	99	3.6

## Fibrinogen H



No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Siemens Thrombin	8	87.5	12.5	0.0	3.06	8.6
2	Other methods	12	83.3	16.7	0.0	3.16	10.2
3	Stago/STA	5	100.0	0.0	0.0	3.32	4.9

## aPTT H




QUALAB tolerance : 25 %

aPTT H (Sek)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Actin FS	9	100.0	0.0	0.0	150.0	10.7
2 Other methods	10	90.0	10.0	0.0	140.1	12.0

## Troponin I




QUALAB tolerance : 24 %

Troponin I (ug/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Vidas	9	100.0	0.0	0.0	2.4	6.4
2	AQT 90 FLEX	5	100.0	0.0	0.0	1.0	4.1
3	Advia Centaur	6	100.0	0.0	0.0	5.7	4.9

# Troponin T




QUALAB tolerance : 24 %

Troponin T (ug/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas hs STAT	6	100.0	0.0	0.0	1.20	10.2

# Myoglobin




QUALAB tolerance : 30 %

Myoglobin (ug/l)


No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Advia Centaur	4	100.0	0.0	0.0	138.0	5.7

## D-dimer



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas (Zitratplasma)	5	100.0	0.0	0.0	1.78	6.4
2	STA Liatest	9	100.0	0.0	0.0	1.66	2.5
3	Eurolyser Smart	13	84.6	7.7	7.7	0.67	13.0
4	AQT 90 FLEX	7	100.0	0.0	0.0	1.14	4.8
5	Vidas	12	91.7	8.3	0.0	1.73	7.9

## D-Dimer NC




QUALAB tolerance : 21 %

D-Dimer NC (mg/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 NycoCard	154	64.3	12.3	23.4	1.12	15.1

# TSH




QUALAB tolerance : 20 %

TSH (mU/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas E / Elecsys	11	100.0	0.0	0.0	14.5	4.3
2	Advia Centaur	5	100.0	0.0	0.0	11.8	4.3
3	Architect	5	100.0	0.0	0.0	11.0	3.7
4	Vidas	10	90.0	10.0	0.0	14.5	8.3
5	Other methods	5	80.0	20.0	0.0	11.4	12.2

# K6 Hormones


## T3



No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas E / Elecsys	4	100.0	0.0	0.0	3.7	5.8

# K6 Hormones

## FT3




QUALAB tolerance : 20 %

FT3 (pmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas E / Elecsys	10	100.0	0.0	0.0	11.4	5.3
2	Advia Centaur	4	100.0	0.0	0.0	8.9	5.1
3	Architect	4	100.0	0.0	0.0	8.8	6.7
4	Vidas	4	100.0	0.0	0.0	10.0	10.3


# K6 Hormones

## FT4



No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Cobas E / Elecsys	11	100.0	0.0	0.0	36.0	6.0
2 Advia Centaur	4	100.0	0.0	0.0	25.3	5.0
3 Architect	5	100.0	0.0	0.0	32.2	4.7
4 Vidas	6	100.0	0.0	0.0	35.5	4.7
5 Other methods	4	75.0	0.0	25.0	32.6	10.4

# Cortisol




QUALAB tolerance : 20 %

Cortisol (nmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas E / Elecsys	5	100.0	0.0	0.0	751	2.3

## Luteinizing hormone




QUALAB tolerance : 24 %

Luteinizing hormone (U/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Other methods	6	100.0	0.0	0.0	35.9	5.5


## Follicle-stimulating hormone



QUALAB tolerance : 24 % Follicle-stimulating hormone (U/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Other methods	6	100.0	0.0	0.0	29.1	10.0

# Prolactine




QUALAB tolerance : 24 %

Prolactine (ug/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Other methods	4	100.0	0.0	0.0	21.6	14.6

## Troponin T CR




QUALAB tolerance : 24 %

Troponin T CR (ug/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas h 232	503	89.5	8.5	2.0	0.22	13.8
2	Cardiac Reader	106	87.8	11.3	0.9	0.20	15.7

## Troponin I WB




QUALAB tolerance : 24 %

Troponin I WB (ug/l)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 iStat	4	100.0	0.0	0.0	1.47	15.0

## D-dimer CR




QUALAB tolerance : 21 %

D-dimer CR (mg/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas h 232	508	90.8	5.3	3.9	0.72	11.5
2	Cardiac Reader	97	82.5	13.4	4.1	0.65	14.1

## proBNP




QUALAB tolerance : 27 %

proBNP (ng/l)


No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas h 232	318	89.9	6.3	3.8	2408	13.8
2	Cardiac Reader	32	81.2	18.8	0.0	2302	18.5

## CKMB - K8



No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Cobas h 232	7	100.0	0.0	0.0	6.7	6.7

## PCO2 CCA




QUALAB tolerance : 15 %

PCO2 CCA (kPa)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 OPTI CCA	13	100.0	0.0	0.0	6.19	1.6

## PO2 CCA




QUALAB tolerance : 15 %

PO2 CCA (kPa)


No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 OPTI CCA	13	100.0	0.0	0.0	14.05	7.1

# pH CCA



No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 OPTI CCA	13	100.0	0.0	0.0	7.41	0.1

## Potassium CCA




QUALAB tolerance : 6 %

Potassium CCA (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 OPTI CCA	7	100.0	0.0	0.0	4.9	0.9

## Sodium CCA




QUALAB tolerance : 6 %

Sodium CCA (mmol/l)


No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 OPTI CCA	6	100.0	0.0	0.0	144.1	0.4

## Ferritin



No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	all Participants	7	71.4	14.3	14.3	219.00	15.2
2	Advia Centaur	4	100.0	0.0	0.0	213.85	5.1
3	Cobas E / Elecsys	11	100.0	0.0	0.0	226.96	4.8
4	Mini Vidas	5	100.0	0.0	0.0	214.00	8.1
5	Eurolyser Smart	12	100.0	0.0	0.0	157.10	11.0

## Vitamin B12




QUALAB tolerance : 20 %

Vitamin B12 (pmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Advia Centaur	6	100.0	0.0	0.0	342.50	10.7
2	Cobas E / Elecsys	8	87.5	0.0	12.5	436.00	4.9

# Folate




QUALAB tolerance : 20 %

Folate (nmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Advia Centaur	4	100.0	0.0	0.0	12.13	7.6
2	Cobas E / Elecsys	8	100.0	0.0	0.0	16.20	9.2

# BNP




No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Biosite, Triage	28	78.6	10.7	10.7	2219.2	16.4

## G10 INR INRatio

# INR INRatio

Unispital Zürich




INRatio Participants

G10 is a split-sample survey. We compare INR-values from our participants with the corresponding plasma INR from University Hospital Zuerich.

Nr.	Device	Total	% good	% insufficient	% outlier
1	INRatio	69	84.1	11.6	4.6

## Bilirubin total Neo




QUALAB tolerance : 18 %

Bilirubin total Neo (umol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	all Participants	12	100.0	0.0	0.0	167	6.9

## Bilirubin direct




QUALAB tolerance : 18 %

Bilirubin direct (umol/l)


No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	all Participants	12	91.7	0.0	8.3	88	5.3

## Bilirubin neonatal




No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	6	83.3	16.7	0.0	216	8.6

## CK-MB



No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Fuji Dri-Chem	31	93.6	3.2	3.2	24.1	13.8

# PSA




QUALAB tolerance : 25 %

PSA (ug/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas E / Elecsys	8	100.0	0.0	0.0	12.10	2.4
2	Advia Centaur	5	100.0	0.0	0.0	10.80	3.4
3	Beckman, Access	4	100.0	0.0	0.0	13.60	2.4
4	Vidas	4	100.0	0.0	0.0	12.10	5.8
5	Architect	5	100.0	0.0	0.0	11.50	4.8

### free PSA




QUALAB tolerance : 25 %

free PSA (ug/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas E / Elecsys	5	100.0	0.0	0.0	2.68	2.5
2	Architect	4	100.0	0.0	0.0	2.84	7.5


# K14 Tumor Markers

## CEA



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas E / Elecsys	5	100.0	0.0	0.0	10.5	7.9
2	Advia Centaur	4	100.0	0.0	0.0	12.8	4.9
3	Beckman, Access	4	100.0	0.0	0.0	13.4	8.4
4	Architect	4	100.0	0.0	0.0	16.4	8.9

# CA 125




QUALAB tolerance : 25 %

CA 125 (kIU/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Architect	4	100.0	0.0	0.0	92.5	5.7

## CA 15-3




QUALAB tolerance : 25 %

CA 15-3 (kIU/l)


No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Architect	4	100.0	0.0	0.0	65.4	5.7

## HCG qn



No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas E / Elecsys	6	100.0	0.0	0.0	97	1.7
2	Advia Centaur	4	100.0	0.0	0.0	45	7.8
3	Vidas	6	100.0	0.0	0.0	12	3.8

## pH OR




QUALAB tolerance : 1 %

pH OR ()

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	ABL700/800 Radiomete	76	100.0	0.0	0.0	7.58	0.0
2	Radiometer NPT-7	7	100.0	0.0	0.0	7.61	0.1
3	ABL 90	17	100.0	0.0	0.0	7.59	0.1
4	ABL 80 / Coox	7	100.0	0.0	0.0	7.62	0.4
5	ABL 5	6	100.0	0.0	0.0	7.58	0.1

## pCO2 OR




QUALAB tolerance : 15 %

pCO2 OR (kPa)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 ABL700/800 Radiomete	76	97.4	2.6	0.0	1.61	4.4
2 Radiometer NPT-7	7	71.4	14.3	14.3	1.70	10.3
3 ABL 90	17	100.0	0.0	0.0	1.67	3.5
4 ABL 80 / Coox	7	71.4	28.6	0.0	1.53	13.2
5 ABL 5	6	100.0	0.0	0.0	1.66	4.3

## pO2 OR




QUALAB tolerance : 15 %

pO2 OR (kPa)


No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 ABL700/800 Radiomete	77	93.5	2.6	3.9	10.12	5.2
2 Radiometer NPT-7	7	100.0	0.0	0.0	8.65	3.4
3 ABL 90	17	76.5	0.0	23.5	7.03	4.8
4 ABL 80 / Coox	7	71.4	14.3	14.3	8.10	9.1
5 ABL 5	6	83.3	16.7	0.0	8.27	9.2

## ctHb OR



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	ABL700/800 Radiomete	63	96.8	3.2	0.0	192.7	3.4
2	Radiometer NPT-7	7	100.0	0.0	0.0	294.0	1.2
3	ABL 90	16	100.0	0.0	0.0	194.3	0.6
4	ABL 80 / Coox	7	100.0	0.0	0.0	195.5	2.0

## sO2 OR




QUALAB tolerance : 20 %

sO2 OR (%)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 ABL700/800 Radiomete	51	100.0	0.0	0.0	70.373	0.1
2 Radiometer NPT-7	6	100.0	0.0	0.0	70.200	0.2
3 ABL 90	16	100.0	0.0	0.0	70.413	0.1
4 ABL 80 / Coox	7	100.0	0.0	0.0	70.500	0.3

## FO2Hb OR




QUALAB tolerance : 20 %

FO2Hb OR (%)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	ABL700/800 Radiomete	47	100.0	0.0	0.0	48.623	2.7
2	Radiometer NPT-7	7	100.0	0.0	0.0	49.100	0.2
3	ABL 90	16	100.0	0.0	0.0	48.800	0.2
4	ABL 80 / Coox	7	100.0	0.0	0.0	48.650	0.3

## FCOHb OR




QUALAB tolerance : 20 %

FCOHb OR (%)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 ABL700/800 Radiomete	48	97.9	0.0	2.1	21.090	0.8
2 Radiometer NPT-7	7	100.0	0.0	0.0	20.100	0.4
3 ABL 90	16	100.0	0.0	0.0	20.750	0.8
4 ABL 80 / Coox	7	100.0	0.0	0.0	20.950	1.8

## FMetHb OR




QUALAB tolerance : 20 %

FMetHb OR (%)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 ABL700/800 Radiomete	49	98.0	0.0	2.0	10.104	1.1
2 Radiometer NPT-7	7	100.0	0.0	0.0	9.900	0.5
3 ABL 90	16	100.0	0.0	0.0	9.956	0.5
4 ABL 80 / Coox	7	100.0	0.0	0.0	10.000	1.1

# FHbF OR




QUALAB tolerance : 20 %

FHbF OR (%)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 ABL 90	6	100.0	0.0	0.0	60.500	5.0

## Bilirubin OR




QUALAB tolerance : 18 %

Bilirubin OR (umol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	ABL700/800 Radiomete	7	100.0	0.0	0.0	453.0	3.6
2	ABL 90	10	100.0	0.0	0.0	447.7	0.8

## Potassium OR




QUALAB tolerance : 6 %

Potassium OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	ABL700/800 Radiomete	64	100.0	0.0	0.0	5.5	1.2
2	ABL 90	17	100.0	0.0	0.0	5.5	0.7
3	ABL 80 / Coox	4	100.0	0.0	0.0	5.1	4.1

## Sodium OR




QUALAB tolerance : 6 %

Sodium OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	ABL700/800 Radiomete	61	100.0	0.0	0.0	127.9	0.8
2	ABL 90	17	100.0	0.0	0.0	128.1	0.2

## Calcium OR




QUALAB tolerance : 12 %

Calcium OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	ABL700/800 Radiomete	63	96.8	1.6	1.6	0.45	3.2
2	ABL 90	17	100.0	0.0	0.0	0.47	0.5

## Choride OR




QUALAB tolerance : 6 %

Choride OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	ABL700/800 Radiomete	51	100.0	0.0	0.0	67.20	1.7
2	ABL 90	16	100.0	0.0	0.0	65.50	1.8
3	ABL 80 / Coox	4	75.0	0.0	25.0	67.00	3.8

## Glucose OR




QUALAB tolerance : 10 %

Glucose OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	ABL700/800 Radiomete	63	100.0	0.0	0.0	14.4	1.9
2	ABL 90	17	100.0	0.0	0.0	13.8	4.2

## Lactate OR




QUALAB tolerance : 10 %

Lactate OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	ABL700/800 Radiomete	65	95.4	4.6	0.0	10.57	4.4
2	ABL 90	17	100.0	0.0	0.0	10.15	5.0

## BNP Plasma




QUALAB tolerance : 27 %

BNP Plasma (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Advia Centaur	5	100.0	0.0	0.0	3503.0	11.1

## NT-proBNP




QUALAB tolerance : 27 %

NT-proBNP (ng/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	AQT 90 FLEX	5	100.0	0.0	0.0	7935.0	7.9
2	Cobas E / Elecsys	13	100.0	0.0	0.0	2056.2	4.9

## Cholesterin PTS




QUALAB tolerance : 10 %

Cholesterin PTS (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 CardioChek	6	83.3	0.0	16.7	5.2	7.1

## Cholesterin HDL PTS




QUALAB tolerance : 21 %

Cholesterin HDL PTS (mmol/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 CardioChek	6	66.6	16.7	16.7	2.6	11.1

## Triglyceride PTS




QUALAB tolerance : 20 %

Triglyceride PTS (mmol/l)


No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 CardioChek	6	66.6	16.7	16.7	2.42	10.5

## Albumin U




No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Siemens Clinitek	10	60.0	10.0	30.0	80.0	15.0
2	Afinion	254	98.0	0.4	1.6	70.2	10.7
3	NycoCard	29	79.3	6.9	13.8	70.4	15.1
4	Turbidimetry	12	100.0	0.0	0.0	75.7	3.3
5	DCA2000/Vantage	97	98.0	1.0	1.0	78.4	7.6

## Kreatinin Urin



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	DCA2000/Vantage	97	99.0	0.0	1.0	7.8	4.8
2	Afinion	253	98.0	0.8	1.2	6.9	5.3
3	Standard chemistry	23	95.7	0.0	4.3	7.4	6.4
4	Siemens Clinitek	8	75.0	0.0	25.0	8.8	0.2

## INR CCXS




QUALAB tolerance : 15 %

INR CCXS ()


No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 CoaguChek XS	2276	99.0	0.9	0.1	1.8	4.5

# INR HC



No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Hemochron j.	35	97.1	2.9	0.0	4.8	9.3

# Osmolality




QUALAB tolerance : 20 %

Osmolality (mosm/kg)


No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Cryoskopy	9	100.0	0.0	0.0	311	2.4

## Procalcitonin



No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	14	100.0	0.0	0.0	3.07	13.9
2 Mini Vidas	8	100.0	0.0	0.0	3.81	2.0

## Troponin I DP




QUALAB tolerance : 24 %

Troponin I DP (ng/ml)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 DXpress Reader	58	63.8	13.8	22.4	0.95	16.0

## D-dimer DP




QUALAB tolerance : 21 %

D-dimer DP (ng/ml)


No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 DXpress Reader	59	52.5	10.2	37.3	732.90	14.6

## NT-proBNP DP



No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	DXpress Reader	43	67.4	9.3	23.3	2923	15.7

# Digoxin




QUALAB tolerance : 24 %

Digoxin (nmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Other methods	8	100.0	0.0	0.0	2.50	7.8

## Hemoglobin BG




QUALAB tolerance : 9 %

Hemoglobin BG (g/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 iStat	4	100.0	0.0	0.0	109.0	1.4

# Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 iStat	4	100.0	0.0	0.0	0.32	1.6

## Troponin Triage




QUALAB tolerance : 24 %

Troponin Triage (ug/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Triage Meter	18	50.0	5.6	44.4	0.6	16.9

## D-dimer Triage




QUALAB tolerance : 21 %

D-dimer Triage (ng/ml)

No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Triage Meter	16	100.0	0.0	0.0	1600.94	5.1

## CK-MB Triage




QUALAB tolerance : 40 %

CK-MB Triage (ug/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Triage Meter	10	100.0	0.0	0.0	4.6	15.8

## Myoglobin Triage




QUALAB tolerance : 30 %

Myoglobin Triage (ug/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Triage Meter	10	100.0	0.0	0.0	172.8	10.0

# Ethanol




QUALAB tolerance : 25 %

Ethanol (mmol/l)


No. Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 all Participants	4	100.0	0.0	0.0	11.6	8.6

# Calprotectin




No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Bühlmann	8	87.5	0.0	12.5	168	18.1

## Cholesterol total Af/b101



No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas b101	5	100.0	0.0	0.0	4.6	2.2
2	Afinion	144	99.3	0.0	0.7	4.8	3.3

## Cholesterol HDL Af/b101




QUALAB tolerance : 21 %

Cholesterol HDL Af/b101 (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas b101	5	100.0	0.0	0.0	1.1	5.8
2	Afinion	144	97.9	0.0	2.1	1.0	5.8

## Tryglycerides Af/b101




QUALAB tolerance : 20 %

Tryglycerides Af/b101 (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	target value	CV%
1	Cobas b101	5	100.0	0.0	0.0	1.65	4.7
2	Afinion	143	100.0	0.0	0.0	1.70	3.9

## Tropnin I S



No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Samsung LABGEO IB10	21	95.2	4.8	0.0	0.77	13.5

## D-dimer qn S




QUALAB tolerance : 21 %

D-dimer qn S (mg/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Samsung LABGEO IB10	25	100.0	0.0	0.0	991.73	8.3

## NT-pro BNP S




QUALAB tolerance : 27 %

NT-pro BNP S (ng/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Samsung LABGEO IB10	17	100.0	0.0	0.0	814.3	9.6

## NR MI




QUALAB tolerance : 15 %

NR MI ()

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 microINR	32	65.6	9.4	25.0	1.8	9.6

# CRP



QUALAB tolerance : 21 %

CRP (mg/l)

No.Methode	Total	% good	% insuff.	% outlier	target value	CV%
1 Microsemi	47	95.7	4.3	0.0	50.0	10.1