

Verein für
Association pour le
Associazione per il



medizinische Qualitätskontrolle
contrôle de qualité médical
controllo di qualità medico

Survey Report

2016 - 2

Survey Specimens

The homogeneity and stability of all specimens were checked before and/or during shipment and no irregularities were noted. The suitability tests were performed by the laboratories of the Universitätsspital Zürich (University Hospital Zurich) (<http://www.uzl.usz.ch>).

The following survey specimens were produced specifically for MQ by a sub-contractor:
B1 Strep A Test, B2 Uricult, H4 Parasitic Hematology, K14 tumor marker

Determination of target values

For each target value, the type of determination per ISO17043: 2010 B2.1 is indicated (column "type"):

- a Value known due to production.
- b Certified reference value for use with special specimens
- c Reference value determined by analysis
- d Consensus values of expert laboratories
- e Consensus values of the participants

For methods groups with more than 9 participants, consensus values of the participants ("e") are generally determined.

In order to calculate the target values, we use the mean value of the method group. Values that differ more than 1.5 times the QUALAB-tolerance are outliers and are not used to calculate the target value. Starting point for the elimination of outliers are the values of our suitability tests.

In order to provide all participants with target values that are as meaningful as possible, other methods may also be applied for smaller method groups.

Uncertainty of the determined target values

The standard uncertainty (u_x) is calculated using the following formula (ISO13528):

$$u_x = (\text{target value}/100) * (1.25/\text{square root of "number of participants"}) * \%CV$$

- u_x has the same unit as the target value
- u_x can be compared with the standard deviation of the participants' collective ($SD = \text{target value} * \%CV / 100$)
- For participant numbers >18, the standard uncertainty (u_x) is significantly lower than the scatter of the collective participants and can be neglected.


QUALAB and MQ tolerances

For all mandatory analyzes, QUALAB tolerances are used (www.qualab.ch, external quality control). For non-mandatory analyzes, the tolerances are specified by MQ's survey specimen leader.

If the determined uncertainty, u_x , of the target value is greater than 15% of the QUALAB or MQ tolerance, the letter indicating the type of target detection is marked with an additional star (example "e*"). Thereby, we are alerting the participants to the fact that the uncertainty of the target value can have an impact on the evaluation.

Graphics

The results are shown graphically as follows:



Comparison of Devices

The data in this report allows you to compare the performance of different devices. However, remember to consider the following:


- The chemical control K1 is a ready-to-use commercial control serum. Even if the specimen is of human origin, it is possible that matrix effects occur. These are device-specific and result in different target values.
- Only one specimen was measured. Since the scatter of the results is dependent on the nature of the specimen (matrix effects) and on the signal strength, the determined coefficient of variations (CV in %) cannot be applied generally.
- A large number of runaways is due to administrative errors (wrong unit, results mixed up) or to operator errors (wrong sample, not correctly taken up in solution, not mixed well) and has nothing to do with the type of device.

Zürich, 4.7.2016

Dr. R. Fried
Survey Director

Publication of this report or any portion thereof without our prior written consent is not permitted. The original is archived at www.mqzh.ch

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	DiaSpect	6	100.0	0.0	0.0	132.0	4.3	e*
2	Automat	34	100.0	0.0	0.0	125.6	2.7	e
3	Cyanmethemoglobin	45	91.1	2.2	6.7	125.2	4.1	e
4	Sysmex X	39	97.4	0.0	2.6	127.4	1.3	e
5	ABX Pentra	10	100.0	0.0	0.0	125.2	1.4	e
6	Reflotron	64	82.8	10.9	6.3	129.3	5.1	e
7	Hemocue	350	95.7	1.4	2.9	123.8	2.5	e
8	Dr. Lange	19	89.5	0.0	10.5	126.4	4.0	e
9	Hemocontrol	13	100.0	0.0	0.0	126.4	2.2	e
10	Eurolyser	5	100.0	0.0	0.0	130.0	2.9	e*
11	Celldyn	4	100.0	0.0	0.0	125.5	1.1	e

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Micros	739	96.2	1.1	2.7	123.0	3.1	e
2	Microsemi	366	99.2	0.3	0.5	124.9	2.1	e
3	Sysmex KX21	397	95.7	0.8	3.5	126.0	2.2	e
4	Sysmex PochH - 100i	206	95.2	2.4	2.4	125.0	2.7	e
5	Sysmex XP 300	261	96.5	0.8	2.7	127.2	2.0	e
6	Mythic	242	94.6	1.7	3.7	121.2	3.6	e
7	Swelab	68	97.0	1.5	1.5	127.2	3.1	e
8	Abacus Junior	12	100.0	0.0	0.0	127.5	3.2	e
9	Medonic	15	93.3	0.0	6.7	125.9	2.1	e
10	Nihon Kohden Celltac	35	97.1	0.0	2.9	125.6	2.1	e
11	Samsung HC10	45	93.4	4.4	2.2	126.4	3.2	e
12	Norma Icon 3	26	100.0	0.0	0.0	125.7	2.9	e

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Automat	28	85.7	10.7	3.6	0.35	5.4	e
2	Centrifuge	13	100.0	0.0	0.0	0.38	4.7	e*
3	Sysmex X	38	97.4	0.0	2.6	0.38	2.6	e
4	ABX Pentra	10	100.0	0.0	0.0	0.34	2.8	e
5	Celldyn	4	100.0	0.0	0.0	0.36	2.8	e*

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Micros	739	94.4	2.4	3.2	0.34	3.8	e
2	Microsemi	363	98.6	0.8	0.6	0.34	2.4	e
3	Sysmex KX21	397	95.7	0.8	3.5	0.34	2.5	e
4	Sysmex PochH - 100i	206	98.0	1.5	0.5	0.37	3.1	e
5	Sysmex XP 300	257	98.0	1.2	0.8	0.34	2.4	e
6	Mythic	242	88.4	3.3	8.3	0.35	4.0	e
7	Swelab	68	98.5	0.0	1.5	0.35	3.8	e
8	Abacus Junior	12	100.0	0.0	0.0	0.37	3.6	e
9	Medonic	15	93.3	0.0	6.7	0.35	3.0	e
10	Nihon Kohden Celltac	36	94.4	0.0	5.6	0.37	3.0	e
11	Samsung HC10	45	93.4	4.4	2.2	0.38	4.0	e
12	Norma Icon 3	26	96.2	3.8	0.0	0.35	4.7	e

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Automat	26	100.0	0.0	0.0	4.26	3.4	e
2	Microscopic	9	100.0	0.0	0.0	4.20	12.3	e*
3	Sysmex X	40	97.5	2.5	0.0	4.34	4.8	e
4	ABX Pentra	10	100.0	0.0	0.0	4.26	2.1	e
5	Celldyn	4	100.0	0.0	0.0	4.24	4.0	e

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Micros	739	98.3	0.3	1.4	4.23	4.3	e
2	Microsemi	365	99.2	0.5	0.3	4.16	3.6	e
3	Sysmex KX21	397	97.7	0.0	2.3	4.27	2.7	e
4	Sysmex Poch - 100i	206	98.5	0.5	1.0	4.48	3.3	e
5	Sysmex XP 300	259	98.8	0.0	1.2	4.29	2.0	e
6	Mythic	242	97.5	0.0	2.5	4.03	5.2	e
7	Swelab	68	100.0	0.0	0.0	4.30	3.9	e
8	Abacus Junior	12	100.0	0.0	0.0	4.46	3.8	e
9	Medonic	15	100.0	0.0	0.0	4.31	4.4	e
10	Nihon Kohden Celltac	35	97.1	2.9	0.0	4.44	5.1	e
11	Samsung HC10	45	100.0	0.0	0.0	4.32	4.7	e
12	Norma Icon 3	26	100.0	0.0	0.0	4.07	3.8	e

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Automat	25	100.0	0.0	0.0	6.92	6.1	e
2	Microscopic	48	91.6	4.2	4.2	5.98	12.6	e
3	Sysmex X	39	100.0	0.0	0.0	7.03	5.4	e
4	ABX Pentra	10	100.0	0.0	0.0	6.92	5.5	e
5	Celldyn	4	100.0	0.0	0.0	6.75	5.6	e

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Micros	739	99.2	0.3	0.5	6.41	5.5	e
2	Microsemi	366	100.0	0.0	0.0	6.59	4.0	e
3	Sysmex KX21	397	99.2	0.3	0.5	6.71	4.2	e
4	Sysmex PochH - 100i	206	100.0	0.0	0.0	6.73	4.3	e
5	Sysmex XP 300	261	99.2	0.4	0.4	6.93	4.1	e
6	Mythic	239	97.0	1.7	1.3	6.30	6.9	e
7	Swelab	68	98.5	1.5	0.0	7.01	5.9	e
8	Abacus Junior	12	100.0	0.0	0.0	8.15	6.3	e
9	Medonic	15	100.0	0.0	0.0	7.11	6.0	e
10	Nihon Kohden Celltac	35	100.0	0.0	0.0	7.06	4.2	e
11	Samsung HC10	45	97.8	2.2	0.0	6.44	6.9	e
12	Norma Icon 3	26	100.0	0.0	0.0	5.88	8.1	e

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Automat	23	100.0	0.0	0.0	245.5	9.0	e
2	Microscopic	28	92.8	3.6	3.6	225.6	14.2	e
3	Sysmex X	39	100.0	0.0	0.0	246.7	4.7	e
4	ABX Pentra	10	100.0	0.0	0.0	260.4	4.5	e
5	Celldyn	4	100.0	0.0	0.0	249.0	5.9	e

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Micros	739	97.6	1.6	0.8	248.1	8.5	e
2	Microsemi	365	99.4	0.3	0.3	252.1	5.9	e
3	Sysmex KX21	397	99.2	0.5	0.3	253.0	4.9	e
4	Sysmex PochH - 100i	206	100.0	0.0	0.0	257.8	4.5	e
5	Sysmex XP 300	260	100.0	0.0	0.0	265.4	5.1	e
6	Mythic	242	95.5	3.3	1.2	232.5	10.7	e
7	Swelab	68	97.0	1.5	1.5	223.6	8.8	e
8	Abacus Junior	12	100.0	0.0	0.0	279.9	8.1	e
9	Medonic	15	100.0	0.0	0.0	228.4	8.8	e
10	Nihon Kohden Celltac	35	97.1	0.0	2.9	266.1	5.5	e
11	Samsung HC10	45	100.0	0.0	0.0	250.5	7.2	e
12	Norma Icon 3	26	100.0	0.0	0.0	219.1	7.0	e

CRP



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Piccolo	6	100.0	0.0	0.0	59.5	4.5	e
2	Cobas	13	100.0	0.0	0.0	47.2	10.0	e*
3	Turbidimetry	27	100.0	0.0	0.0	49.9	7.3	e
4	Abx Micros	95	96.8	1.1	2.1	48.4	9.2	e
5	ABX Micros CRP200	287	96.5	2.8	0.7	46.0	9.2	e
6	Afinion	1249	99.3	0.6	0.1	50.9	7.4	e
7	NycoCard SingleTest-	347	82.7	4.9	12.4	52.7	10.7	e
8	Quick Read go	150	96.0	1.3	2.7	49.3	6.5	e
9	Eurolyser	128	78.9	3.9	17.2	64.2	9.2	e
10	Fuji Dri-Chem	27	92.6	3.7	3.7	49.3	11.0	e
11	Autolyser/DiaSys	9	88.9	0.0	11.1	43.0	10.2	e*

CRP




QUALAB tolerance : 21 %

CRP (mg/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	QuikRead (Vollblut)	143	99.3	0.0	0.7	82.8	6.0	e

CRP



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	AQT 90 FLEX	8	100.0	0.0	0.0	53.0	6.4	e
2	Spotchem D-Concept	6	100.0	0.0	0.0	56.3	5.8	e
3	Spotchem SI-3510	5	100.0	0.0	0.0	46.1	9.9	e*
4	Other methods	5	80.0	0.0	20.0	47.1	7.3	e*

CRP




QUALAB tolerance : 21 %

CRP (mg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Microsemi	362	95.0	3.6	1.4	52.3	9.8	e

I2 Plasmaproteins

IgG




QUALAB tolerance : 25 %

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Turbidimetry	11	100.0	0.0	0.0	14.9	2.8	e
2	Nephelometry	7	100.0	0.0	0.0	15.5	7.4	e

I2 Plasmaproteins

IgA




QUALAB tolerance : 25 %

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Turbidimetry	10	100.0	0.0	0.0	3.0	5.9	e
2	Nephelometry	7	100.0	0.0	0.0	3.3	6.0	e

I2 Plasmaproteins

IgM




QUALAB tolerance : 25 %

IgM (g/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Turbidimetry	6	100.0	0.0	0.0	1.6	7.3	e
2	Nephelometry	7	100.0	0.0	0.0	1.8	7.5	e
3	Cobas Integra 800/40	5	100.0	0.0	0.0	1.6	5.0	e

IgE




QUALAB tolerance : 20 %

IgE (kU/L)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	9	77.8	0.0	22.2	161	8.2	e*

Alpha-1-Antitrypsin



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Nephelometry	4	100.0	0.0	0.0	1.53	19.1	a


Anti-Streptolysin-Antibodies



QUALAB tolerance : 25 % Anti-Streptolysin-Antibodies (kIU/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 all Participants	11	100.0	0.0	0.0	160	7.7	e

Complement C3




QUALAB tolerance : 25 %

Complement C3 (g/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	13	100.0	0.0	0.0	2.13	7.7	e

Complement C4



No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 all Participants	12	100.0	0.0	0.0	0.46	9.5	e

Haptoglobin




QUALAB tolerance : 25 %

Haptoglobin (g/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	13	100.0	0.0	0.0	1.90	2.2	e

Transferrin




QUALAB tolerance : 25 %

Transferrin (g/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	16	100.0	0.0	0.0	3.08	3.9	ex

Ceruloplasmin




QUALAB tolerance : 25 %

Ceruloplasmin (mg/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	4	100.0	0.0	0.0	396.5	3.4	e

Präalbumin



No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 all Participants	7	100.0	0.0	0.0	284.0	4.3	e

Albumine




QUALAB tolerance : 12 %

Albumine (g/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	23	95.7	0.0	4.3	34	5.9	e
2	Cobas	12	100.0	0.0	0.0	37	3.1	e
3	Fuji Dri-Chem	193	100.0	0.0	0.0	44	3.1	e
4	Spotchem/Ready	43	88.4	9.3	2.3	37	6.2	e
5	Spotchem D-Concept	83	96.4	3.6	0.0	44	5.5	e
6	Piccolo	31	100.0	0.0	0.0	34	3.7	e
7	Abx Mira	5	100.0	0.0	0.0	35	3.9	e*
8	Hitachi S40/M40	9	100.0	0.0	0.0	37	4.1	e
9	Autolyser/DiaSys	5	100.0	0.0	0.0	36	1.3	e

Alkaline phosphatase




QUALAB tolerance : 18 %

Alkaline phosphatase (U/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC	11	100.0	0.0	0.0	189	9.4	e*
2 Cobas	17	100.0	0.0	0.0	170	3.7	e
3 Reflotron	632	95.9	2.4	1.7	335	7.3	e
4 Fuji Dri-Chem	704	99.3	0.0	0.7	201	5.1	e
5 Spotchem/Ready	109	90.9	7.3	1.8	244	9.0	e
6 Spotchem D-Concept	151	98.0	0.7	1.3	219	6.5	e
7 Hitachi S40/M40	16	100.0	0.0	0.0	143	5.8	e
8 DGKC 37'C	4	100.0	0.0	0.0	280	8.5	e*
9 Beckman	9	100.0	0.0	0.0	245	6.7	e*
10 Piccolo	29	100.0	0.0	0.0	233	2.7	e
11 Abx Mira	8	87.5	12.5	0.0	187	10.6	e*
12 Autolyser/DiaSys	14	100.0	0.0	0.0	204	5.5	e

Amylase




QUALAB tolerance : 18 %

Amylase (U/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	IFCC	12	100.0	0.0	0.0	267	8.2	e*
2	Cobas	5	100.0	0.0	0.0	248	2.7	e
3	Reflotron	179	98.3	1.1	0.6	261	6.5	e
4	Fuji Dri-Chem	522	99.2	0.0	0.8	213	3.0	e
5	Spotchem/Ready	73	84.9	4.1	11.0	135	9.7	e
6	Spotchem D-Concept	115	99.1	0.0	0.9	212	6.4	e
7	Piccolo	27	100.0	0.0	0.0	234	2.9	e
8	Abx Mira	5	80.0	20.0	0.0	234	10.0	e*
9	Hitachi S40/M40	9	88.9	0.0	11.1	279	7.4	e*
10	Autolyser/DiaSys	4	100.0	0.0	0.0	237	3.5	e

Pancreatic amylase




QUALAB tolerance : 18 %

Pancreatic amylase (U/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC	11	100.0	0.0	0.0	233	6.8	e
2 Cobas	13	100.0	0.0	0.0	223	2.3	e
3 Reflotron	425	97.9	1.4	0.7	242	7.5	e
4 Autolyser/DiaSys	9	100.0	0.0	0.0	225	3.3	e

Bilirubin




QUALAB tolerance : 18 %

Bilirubin (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	16	93.7	6.3	0.0	50.9	10.5	e*
2	Cobas	15	100.0	0.0	0.0	46.1	4.8	e
3	Reflotron	466	95.7	2.4	1.9	50.5	6.9	e
4	Fuji Dri-Chem	534	98.8	0.6	0.6	51.5	4.7	e
5	Spotchem/Ready	90	97.8	2.2	0.0	56.0	7.4	e
6	Spotchem D-Concept	117	98.3	0.0	1.7	43.7	4.9	e
7	Beckman	8	100.0	0.0	0.0	60.4	7.4	e*
8	Piccolo	31	96.8	0.0	3.2	45.5	5.8	e
9	Abx Mira	7	85.7	14.3	0.0	48.0	11.2	e*
10	Hitachi S40/M40	14	100.0	0.0	0.0	53.3	5.9	e
11	Autolyser/DiaSys	14	92.9	0.0	7.1	47.3	4.1	e

Bilirubin direct




QUALAB tolerance : 18 %

Bilirubin direct (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Fuji Dri-Chem	34	91.2	2.9	5.9	31.0	8.0	e

Calcium




QUALAB tolerance : 12 %

Calcium (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	24	95.8	0.0	4.2	2.52	3.6	e
2	Cobas	12	100.0	0.0	0.0	2.54	2.1	e
3	Fuji Dri-Chem	361	97.7	0.6	1.7	2.67	4.4	e
4	Spotchem/Ready	40	95.0	5.0	0.0	2.35	6.0	e
5	Spotchem D-Concept	70	94.3	5.7	0.0	1.91	6.1	e
6	Piccolo	30	100.0	0.0	0.0	2.48	2.3	e
7	Abx Mira	6	100.0	0.0	0.0	2.46	1.4	e
8	Hitachi S40/M40	12	83.4	8.3	8.3	2.44	6.1	e*
9	Autolysér/DiaSys	6	100.0	0.0	0.0	2.48	4.5	e*

Calcium ISE




QUALAB tolerance : 12 %

Calcium ISE (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	iStat Chem8	4	100.0	0.0	0.0	0.99	1.3	e

Chloride




QUALAB tolerance : 6 %

Chloride (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ISE	17	100.0	0.0	0.0	97	1.9	e
2	Cobas	7	100.0	0.0	0.0	91	1.5	e
3	Fuji Dri-Chem	637	96.9	2.2	0.9	110	2.2	e
4	Spotchem D-Concept	139	95.0	1.4	3.6	115	2.3	e
5	Standard chemistry	4	100.0	0.0	0.0	103	2.9	a
6	Spotchem EL-SE 1520	112	90.2	2.7	7.1	121	3.1	e
7	Piccolo	16	100.0	0.0	0.0	95	2.6	e
8	iStat Chem8	4	100.0	0.0	0.0	104	0.0	e

Cholesterol total




QUALAB tolerance : 10 %

Cholesterol total (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	22	100.0	0.0	0.0	3.99	3.6	e
2	Cobas	15	100.0	0.0	0.0	3.97	1.8	e
3	Reflotron	699	98.1	0.9	1.0	3.67	3.4	e
4	Fuji Dri-Chem	715	99.4	0.0	0.6	3.77	3.3	e
5	Spotchem/Ready	134	97.0	1.5	1.5	4.04	4.7	e
6	Spotchem D-Concept	156	97.5	1.9	0.6	3.98	3.6	e
7	Piccolo	19	100.0	0.0	0.0	4.03	3.0	e
8	Cholestech LDX	190	94.7	3.2	2.1	3.77	4.3	e
9	Abx Mira	8	100.0	0.0	0.0	3.91	2.6	e
10	Hitachi S40/M40	16	100.0	0.0	0.0	3.98	2.4	e
11	Autolyser/DiaSys	13	100.0	0.0	0.0	3.87	4.5	e*

Cholesterin HDL




QUALAB tolerance : 21 %

Cholesterin HDL (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Wet chemistry, direc	15	100.0	0.0	0.0	1.29	5.9	e
2	Cobas	14	100.0	0.0	0.0	1.67	3.5	e
3	Reflotron	520	77.6	13.7	8.7	0.99	14.0	e
4	Fuji Dri-Chem	680	99.3	0.4	0.3	1.53	3.8	e
5	Spotchem/Ready	120	89.1	6.7	4.2	0.72	11.2	e
6	Spotchem D-Concept	152	90.1	9.2	0.7	0.84	12.4	e
7	Piccolo	17	94.1	0.0	5.9	1.22	4.5	e
8	Cholestech LDX	190	94.7	3.2	2.1	1.29	8.8	e
9	Abx Mira	8	87.5	12.5	0.0	1.23	17.0	e*
10	Hitachi S40/M40	15	100.0	0.0	0.0	1.43	5.8	e
11	Autolyser/DiaSys	13	100.0	0.0	0.0	1.34	5.9	e

Creatine kinase




QUALAB tolerance : 18 %

Creatine kinase (U/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	IFCC	19	100.0	0.0	0.0	361	7.1	e
2	Cobas	14	100.0	0.0	0.0	365	4.1	e
3	Reflotron	394	93.6	1.8	4.6	334	6.5	e
4	Fuji Dri-Chem	451	99.4	0.2	0.4	368	4.5	e
5	Spotchem/Ready	52	94.2	5.8	0.0	390	8.0	e
6	Spotchem D-Concept	97	95.9	3.1	1.0	397	7.1	e
7	Piccolo	11	100.0	0.0	0.0	376	3.3	e
8	Abx Mira	6	100.0	0.0	0.0	362	3.0	e
9	Hitachi S40/M40	10	90.0	0.0	10.0	323	3.5	e
10	Autolyser/DiaSys	12	100.0	0.0	0.0	374	3.2	e

Iron




QUALAB tolerance : 20 %

Iron (µmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	7	100.0	0.0	0.0	28	2.0	e
2	Cobas	9	100.0	0.0	0.0	28	2.8	e
3	Abx Mira	5	80.0	0.0	20.0	28	1.8	e

Gamma-glutamyltransferase




QUALAB tolerance : 18 %

Gamma-glutamyltransferase (U/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC	11	90.9	0.0	9.1	78	6.8	e
2 Cobas	15	100.0	0.0	0.0	73	6.8	e
3 Reflotron	833	98.3	1.1	0.6	98	6.0	e
4 Fuji Dri-Chem	769	99.0	0.7	0.3	101	5.4	e
5 Spotchem/Ready	140	97.2	2.1	0.7	88	6.8	e
6 Spotchem D-Concept	172	97.1	2.3	0.6	86	6.9	e
7 DGKC 37'C	10	100.0	0.0	0.0	77	4.9	e
8 Piccolo	28	96.4	3.6	0.0	68	6.2	e
9 Abx Mira	9	100.0	0.0	0.0	79	7.3	e*
10 Hitachi S40/M40	18	100.0	0.0	0.0	84	2.7	e
11 Autolyser/DiaSys	14	100.0	0.0	0.0	80	3.1	e

Glucose




QUALAB tolerance : 10 %

Glucose (mmol/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	30	96.7	3.3	0.0	11.1	4.4	e
2	Cobas	16	100.0	0.0	0.0	11.2	2.4	e
3	Reflotron	843	91.0	5.6	3.4	10.7	5.0	e
4	Fuji Dri-Chem	728	99.2	0.1	0.7	10.6	2.6	e
5	Spotchem/Ready	124	93.6	5.6	0.8	11.3	5.0	e
6	Spotchem D-Concept	160	98.1	1.9	0.0	10.5	4.4	e
7	Piccolo	36	100.0	0.0	0.0	10.8	1.0	e
8	Cholestech LDX	153	96.7	3.3	0.0	10.4	5.0	e
9	Abx Mira	9	100.0	0.0	0.0	11.1	3.0	e
10	Lange	4	100.0	0.0	0.0	10.1	3.6	e*
11	Hitachi S40/M40	19	100.0	0.0	0.0	11.4	4.5	e
12	Autolyser/DiaSys	14	100.0	0.0	0.0	10.9	3.8	e
13	iStat Chem8	5	100.0	0.0	0.0	10.5	1.5	e

Glucose



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Accu-Chek Aviva	385	93.5	2.3	4.2	12.0	3.5	e
2	Accu-Chek Inform 2	258	98.8	0.4	0.8	12.1	2.9	e
3	Accu-Chek Mobile	4	100.0	0.0	0.0	10.6	4.7	e*
4	Bayer Contour 2 (5s)	44	88.6	2.3	9.1	9.1	4.3	e
5	Bayer Contour XT/NEX	1130	97.5	1.3	1.2	11.1	4.0	e
6	Bayer Breeze 2	13	92.3	7.7	0.0	20.3	5.1	e*
7	Glucocard	11	100.0	0.0	0.0	17.1	5.1	e*
8	Hemocue 201+ P-equiv	87	96.6	1.1	2.3	12.7	3.5	e
9	Hemocue 201RT P-equiv	42	97.6	2.4	0.0	12.5	3.1	e
10	FreeStyle Precision	8	62.5	25.0	12.5	12.3	6.4	e*
11	Freestyle Freedom li	8	100.0	0.0	0.0	11.7	2.3	e
12	Sanofi BG Star	5	100.0	0.0	0.0	15.8	3.4	e*

Glucose B




QUALAB tolerance : 10 %

Glucose B (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Bayer Elite	9	77.8	11.1	11.1	12.9	6.2	e*
2	Hemocue 201+ (alt)	51	96.0	2.0	2.0	12.2	5.4	e
3	OneTouch Ultra	7	71.4	28.6	0.0	16.7	7.4	e*
4	OneTouch Verio	28	100.0	0.0	0.0	11.3	3.2	e
5	Bayer Contour (15s)	11	100.0	0.0	0.0	15.7	3.8	e
6	Healthpro	14	92.9	0.0	7.1	20.2	4.4	e
7	Mylife UNIO	7	100.0	0.0	0.0	12.8	3.5	e*
8	mylife Pura	59	84.7	3.4	11.9	13.9	4.8	e
9	Omnitest	17	94.1	5.9	0.0	21.1	4.9	e
10	Alpha Check	4	100.0	0.0	0.0	17.3	6.0	e*

Uric Acid




QUALAB tolerance : 12 %

Uric Acid (µmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	25	100.0	0.0	0.0	455	4.5	e
2	Cobas	11	100.0	0.0	0.0	448	2.7	e
3	Reflotron	736	98.5	1.1	0.4	484	4.7	e
4	Fuji Dri-Chem	729	99.6	0.1	0.3	496	2.0	e
5	Spotchem/Ready	117	96.5	2.6	0.9	410	5.0	e
6	Spotchem D-Concept	156	98.7	1.3	0.0	420	3.9	e
7	Piccolo	23	100.0	0.0	0.0	376	4.2	e
8	Abx Mira	8	100.0	0.0	0.0	418	5.6	e*
9	Hitachi S40/M40	17	100.0	0.0	0.0	445	2.2	e
10	Autolyser/DiaSys	13	100.0	0.0	0.0	431	4.7	e

Urea




QUALAB tolerance : 15 %

Urea (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	24	95.8	4.2	0.0	13.6	7.6	e
2	Cobas	14	100.0	0.0	0.0	13.9	3.2	e
3	Reflotron	329	96.4	3.0	0.6	14.4	6.0	e
4	Fuji Dri-Chem	446	99.6	0.2	0.2	14.5	2.3	e
5	Spotchem/Ready	80	93.7	6.3	0.0	13.4	8.5	e
6	Spotchem D-Concept	95	90.5	6.3	3.2	14.6	8.2	e
7	Piccolo	34	100.0	0.0	0.0	12.8	3.0	e
8	Abx Mira	6	100.0	0.0	0.0	13.7	5.7	e*
9	Hitachi S40/M40	13	92.3	0.0	7.7	13.9	3.3	e
10	Autolyser/DiaSys	7	100.0	0.0	0.0	13.7	4.0	e
11	iStat Chem8	6	100.0	0.0	0.0	18.3	2.7	e

Potassium




QUALAB tolerance : 6 %

Potassium (mmol/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ISE	33	97.0	0.0	3.0	4.42	1.9	e
2	Cobas	16	100.0	0.0	0.0	4.47	1.4	e
3	Reflotron	752	85.5	9.6	4.9	4.52	3.7	e
4	Fuji Dri-Chem	764	97.7	1.3	1.0	4.33	2.0	e
5	Spotchem D-Concept	158	97.5	0.0	2.5	3.87	2.0	e
6	Spotchem EL-SE 1520	116	97.4	0.9	1.7	3.80	2.8	e
7	Piccolo	22	72.8	13.6	13.6	4.54	4.4	e*
8	iStat Chem8	6	100.0	0.0	0.0	4.30	1.0	e

Creatinine



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	14	92.9	0.0	7.1	343	4.5	e
2	Cobas	16	100.0	0.0	0.0	327	3.7	e
3	Reflotron	937	98.3	1.0	0.7	378	5.7	e
4	Fuji Dri-Chem	794	98.8	0.3	0.9	326	4.1	e
5	Spotchem/Ready	135	81.5	1.5	17.0	195	6.1	e
6	Spotchem D-Concept	157	96.9	0.6	2.5	207	5.2	e
7	Spotchem old test	19	94.7	0.0	5.3	322	5.3	e
8	Enzymatic	7	100.0	0.0	0.0	334	2.3	e
9	Piccolo	34	100.0	0.0	0.0	331	4.3	e
10	Abx Mira	10	80.0	10.0	10.0	335	7.9	e*
11	Hitachi S40/M40	18	100.0	0.0	0.0	321	2.1	e
12	Autolyser/DiaSys	14	100.0	0.0	0.0	337	2.2	e

Creatinine E




QUALAB tolerance : 18 %

Creatinine E (µmol/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Statsensor i / Nova	35	85.7	11.4	2.9	514	10.5	e
2	iStat Chem8	10	100.0	0.0	0.0	318	4.5	e
3	ABL700/800	9	100.0	0.0	0.0	351	2.5	e

eGFR CKD-EPI



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	56	85.8	7.1	7.1	20	14.4	e
2	Spotchem/Ready	111	86.5	8.1	5.4	36	14.6	e
3	Reflotron	339	95.3	0.9	3.8	17	8.0	e
4	Fuji Dri-Chem	327	94.1	2.8	3.1	21	8.7	e

eGFR Cockcroft-Gault




QUALAB tolerance : 30 %

eGFR Cockcroft-Gault ()


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Spotchem/Ready	7	57.1	14.3	28.6	35	19.3	e*
2	Reflotron	25	96.0	4.0	0.0	24	12.3	e
3	Fuji Dri-Chem	25	96.0	0.0	4.0	25	15.1	e

eGFR MDRD



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Reflotron	14	71.5	21.4	7.1	18	18.7	e*
2	Fuji Dri-Chem	4	100.0	0.0	0.0	19	10.6	e*

LDH




QUALAB tolerance : 18 %

LDH (U/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	IFCC	20	100.0	0.0	0.0	210	4.1	e
2	Cobas	9	100.0	0.0	0.0	435	2.5	e
3	Fuji Dri-Chem	144	97.9	0.7	1.4	172	5.1	e
4	Spotchem/Ready	31	96.8	3.2	0.0	143	9.8	e
5	Spotchem D-Concept	45	77.8	11.1	11.1	132	10.8	e
6	Abx Mira	6	100.0	0.0	0.0	199	7.3	e*
7	Hitachi S40/M40	6	83.3	0.0	16.7	191	3.8	e
8	Autolyser/DiaSys	7	100.0	0.0	0.0	196	3.8	e

Magnesium




QUALAB tolerance : 12 %

Magnesium (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	17	100.0	0.0	0.0	1.11	2.2	e
2	Cobas	10	100.0	0.0	0.0	1.11	1.8	e
3	Fuji Dri-Chem	119	98.3	0.0	1.7	1.09	3.8	e
4	Spotchem D-Concept	33	100.0	0.0	0.0	0.83	4.4	e
5	Spotchem/Ready	15	100.0	0.0	0.0	0.90	3.5	e
6	Piccolo	5	100.0	0.0	0.0	1.07	1.7	e

Sodium




QUALAB tolerance : 6 %

Sodium (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ISE	32	90.6	9.4	0.0	138	3.0	e
2	Cobas	15	100.0	0.0	0.0	136	1.4	e
3	Fuji Dri-Chem	709	98.9	0.7	0.4	139	1.5	e
4	Spotchem D-Concept	150	100.0	0.0	0.0	134	1.8	e
5	Spotchem EL-SE 1520	116	97.4	0.9	1.7	131	1.9	e
6	Piccolo	23	95.7	4.3	0.0	133	2.3	e
7	iStat Chem8	5	100.0	0.0	0.0	134	0.8	e

Phosphate




QUALAB tolerance : 15 %

Phosphate (mmol/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	16	100.0	0.0	0.0	1.7	2.8	e
2	Cobas	10	100.0	0.0	0.0	1.7	2.2	e
3	Fuji Dri-Chem	81	98.8	1.2	0.0	1.7	3.7	e
4	Spotchem D-Concept	17	100.0	0.0	0.0	1.7	3.5	e
5	Spotchem/Ready	9	100.0	0.0	0.0	1.6	5.4	e
6	Piccolo	4	100.0	0.0	0.0	1.9	1.2	e
7	Abx Mira	4	100.0	0.0	0.0	1.7	5.0	e*

Protein total



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	24	100.0	0.0	0.0	54.9	2.9	e
2	Cobas	11	100.0	0.0	0.0	53.2	2.4	e
3	Fuji Dri-Chem	184	99.5	0.5	0.0	54.7	2.3	e
4	Spotchem/Ready	38	97.4	0.0	2.6	56.4	5.7	e
5	Spotchem D-Concept	69	91.4	7.2	1.4	58.5	6.7	e
6	Piccolo	25	100.0	0.0	0.0	54.2	1.7	e
7	Abx Mira	5	100.0	0.0	0.0	55.7	5.4	e*
8	Hitachi S40/M40	7	85.7	14.3	0.0	54.0	6.0	e*
9	Autolysers/DiaSys	4	75.0	25.0	0.0	54.6	8.2	e*

Aspartate aminotransferase




QUALAB tolerance : 18 %

Aspartate aminotransferase (U/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC with PP	19	100.0	0.0	0.0	159	4.8	e
2 Cobas	16	100.0	0.0	0.0	150	6.3	e
3 Reflotron	838	97.1	2.1	0.8	193	7.0	e
4 Fuji Dri-Chem	768	99.5	0.1	0.4	145	3.2	e
5 Spotchem/Ready	147	99.3	0.0	0.7	125	4.8	e
6 Spotchem D-Concept	166	98.8	1.2	0.0	127	4.6	e
7 IFCC without PP	4	75.0	25.0	0.0	158	12.0	e*
8 Piccolo	34	97.1	2.9	0.0	137	6.1	e
9 Abx Mira	9	88.9	11.1	0.0	148	11.5	e*
10 Hitachi S40/M40	20	100.0	0.0	0.0	153	3.2	e
11 Autolyser/DiaSys	14	100.0	0.0	0.0	148	4.4	e
12 Other methods	4	100.0	0.0	0.0	157	0.9	e

Alanine aminotransferase




QUALAB tolerance : 18 %

Alanine aminotransferase (U/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC with PP	20	100.0	0.0	0.0	130	5.7	e
2 Cobas	17	100.0	0.0	0.0	124	1.9	e
3 Reflotron	871	98.5	0.7	0.8	112	5.9	e
4 Fuji Dri-Chem	788	99.1	0.3	0.6	125	4.2	e
5 Spotchem/Ready	150	98.0	0.7	1.3	107	6.0	e
6 Spotchem D-Concept	171	99.4	0.0	0.6	92	4.5	e
7 Piccolo	35	97.1	2.9	0.0	115	4.9	e
8 Abx Mira	8	100.0	0.0	0.0	131	7.5	e*
9 Hitachi S40/M40	20	95.0	0.0	5.0	134	2.8	e
10 Autolyser/DiaSys	14	100.0	0.0	0.0	124	3.4	e
11 Other methods	5	100.0	0.0	0.0	132	5.0	e*

Triglycerides




QUALAB tolerance : 20 %

Triglycerides (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	19	100.0	0.0	0.0	1.92	6.9	e
2	Cobas	16	100.0	0.0	0.0	1.85	2.8	e
3	Reflotron	607	97.7	0.8	1.5	2.50	5.2	e
4	Fuji Dri-Chem	696	98.6	0.4	1.0	2.35	3.4	e
5	Spotchem/Ready	127	98.4	1.6	0.0	1.24	5.5	e
6	Spotchem D-Concept	150	100.0	0.0	0.0	1.40	4.2	e
7	Hitachi S40/M40	15	93.3	0.0	6.7	0.50	8.2	e
8	Piccolo	17	94.1	0.0	5.9	2.14	1.4	e
9	Cholestech LDX	190	99.5	0.0	0.5	1.90	5.1	e
10	Abx Mira	9	100.0	0.0	0.0	1.85	7.9	e*
11	Autolyser/DiaSys	13	100.0	0.0	0.0	1.78	6.4	e

LDL Cholesterin




QUALAB tolerance : 25 %

LDL Cholesterin (mmol/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Mira	6	100.0	0.0	0.0	2.1	13.4	e*
2	Standard chemistry	6	83.3	16.7	0.0	2.5	12.3	e*
3	Roche, Cobas	4	100.0	0.0	0.0	2.7	2.4	e
4	Hitachi S40/M40	8	87.5	0.0	12.5	1.5	4.0	e
5	Autolyser/DiaSys	12	100.0	0.0	0.0	2.1	9.7	e

Lithium



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	14	100.0	0.0	0.0	1.62	6.2	e

Lactate




QUALAB tolerance : 15 %

Lactate (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	7	85.7	14.3	0.0	4.10	7.7	e*

HbA1c sample A




QUALAB tolerance : 9 %

HbA1c sample A (%)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b101	34	100.0	0.0	0.0	5.1	3.8	e
2	Afinion	665	98.7	1.1	0.2	5.1	2.7	e
3	Eurolyser	16	100.0	0.0	0.0	5.6	3.9	e
4	Hemocue HbA1c 501	8	87.5	12.5	0.0	5.5	4.3	e*
5	NycoCard	74	81.0	4.1	14.9	5.4	4.6	e
6	DCA2000/Vantage	204	97.5	2.0	0.5	5.2	3.5	e
7	Others	8	100.0	0.0	0.0	5.2	3.2	a
8	HPLC	6	100.0	0.0	0.0	5.2	3.3	a
9	Roche, Cobas	18	88.9	11.1	0.0	5.1	5.7	e*

HbA1c sample B




QUALAB tolerance : 9 %

HbA1c sample B (%)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b101	32	96.9	3.1	0.0	5.2	4.4	e
2	Afinion	579	98.7	1.0	0.3	5.1	2.5	e
3	Eurolyser	20	100.0	0.0	0.0	5.7	3.2	e
4	A1c Now	4	100.0	0.0	0.0	5.3	4.9	a
5	Hemocue HbA1c 501	11	90.9	0.0	9.1	5.1	5.3	e*
6	NycoCard	73	89.1	6.8	4.1	5.5	5.1	e
7	DCA2000/Vantage	225	98.2	0.9	0.9	5.5	3.2	e
8	Others	4	100.0	0.0	0.0	5.3	3.8	e*
9	HPLC	5	100.0	0.0	0.0	5.3	2.8	a
10	Roche, Cobas	13	92.3	7.7	0.0	5.1	5.2	e*

pCO2




QUALAB tolerance : 15 %

pCO2 (kPa)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b121/123/221	16	100.0	0.0	0.0	3.00	3.8	e
2	iStat	36	100.0	0.0	0.0	2.69	3.6	e
3	EPOC	24	91.7	0.0	8.3	2.27	5.4	e

pO2




QUALAB tolerance : 15 %

pO2 (kPa)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b221	6	100.0	0.0	0.0	12.16	1.9	e
2	Cobas b121/123	7	57.1	14.3	28.6	8.71	11.3	e*
3	iStat	36	97.2	0.0	2.8	10.67	6.2	e
4	EPOC	24	75.0	12.5	12.5	8.46	9.7	e*

K4 Blood gases

pH




QUALAB tolerance : 1 %

pH ()

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b121/123/221	15	100.0	0.0	0.0	7.58	0.2	e
2	iStat	36	100.0	0.0	0.0	7.66	0.1	e
3	EPOC	24	95.8	0.0	4.2	7.67	0.1	e

Glucose BG




QUALAB tolerance : 10 %

Glucose BG (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b121/123/221	8	87.5	0.0	12.5	14.5	4.1	e*
2	iStat	11	100.0	0.0	0.0	13.4	0.8	e
3	EPOC	16	93.7	0.0	6.3	15.3	3.2	e

Potassium BG




QUALAB tolerance : 6 %

Potassium BG (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b121/123/221	15	100.0	0.0	0.0	5.4	2.5	e
2	iStat	21	100.0	0.0	0.0	5.3	0.4	e
3	EPOC	19	94.7	0.0	5.3	5.0	0.9	e

Sodium BG




QUALAB tolerance : 6 %

Sodium BG (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b121/123/221	16	93.7	6.3	0.0	122.2	3.0	e*
2	iStat	21	100.0	0.0	0.0	124.4	0.6	e
3	EPOC	19	94.7	0.0	5.3	115.9	0.8	e

Chlorid-BG




QUALAB tolerance : 6 %

Chlorid-BG (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b121/123/221	6	83.3	0.0	16.7	71.0	2.0	e*

Calcium-BG




QUALAB tolerance : 12 %

Calcium-BG (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b121/123/221	10	80.0	20.0	0.0	0.29	8.7	e*
2	iStat	10	100.0	0.0	0.0	0.28	1.5	e
3	EPOC	18	94.4	0.0	5.6	0.26	2.4	e

Lactate-BG




QUALAB tolerance : 10 %

Lactate-BG (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b121/123/221	6	83.3	0.0	16.7	11.80	5.7	e*
2	EPOC	20	95.0	0.0	5.0	11.70	4.8	e
3	iStat	8	100.0	0.0	0.0	10.79	1.7	e

sO2




QUALAB tolerance : 20 %

sO2 (%)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	iStat	6	100.0	0.0	0.0	98.000	0.4	e

Calcium - Urine




QUALAB tolerance : 12 %

Calcium - Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	13	100.0	0.0	0.0	2.17	2.9	e

Chloride - Urine




QUALAB tolerance : 6 %

Chloride - Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	6	83.3	0.0	16.7	142	2.7	e*
2	ISE	4	100.0	0.0	0.0	150	3.4	e*

Glucose - Urine




QUALAB tolerance : 10 %

Glucose - Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	13	92.3	7.7	0.0	7.9	4.6	e*

Magnesium - Urine




QUALAB tolerance : 12 %

Magnesium - Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	8	100.0	0.0	0.0	3.4	4.6	e*

Osmolality - Urine




QUALAB tolerance : 10 %

Osmolality - Urine (mosm/kg)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Cryoscopy	7	100.0	0.0	0.0	627	0.9	e

Phosphate - Urine




QUALAB tolerance : 15 %

Phosphate - Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	13	100.0	0.0	0.0	12.1	2.4	e

Potassium - Urine




QUALAB tolerance : 15 %

Potassium - Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	19	100.0	0.0	0.0	48	3.3	e

total Protein - Urine




QUALAB tolerance : 20 %

total Protein - Urine (mg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas/Roche	13	100.0	0.0	0.0	321.6	4.5	e
2	Standard chemistry	5	100.0	0.0	0.0	408.0	4.8	e

Sodium - Urine



QUALAB tolerance : 15 %

Sodium - Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	19	94.7	0.0	5.3	123	1.5	e

Urea - Urine




QUALAB tolerance : 15 %

Urea - Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	14	100.0	0.0	0.0	202	3.7	e

Uric Acid - Urine




QUALAB tolerance : 15 %

Uric Acid - Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	13	100.0	0.0	0.0	0.91	3.6	e

Specific Gravity - Urine




QUALAB tolerance : 5 %

Specific Gravity - Urine ()

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Refractometer	7	100.0	0.0	0.0	1.018	0.2	e

INR CoaguChek




QUALAB tolerance : 15 %

INR CoaguChek ()

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	CoaguChek Pro II	15	93.3	0.0	6.7	2.3	4.5	e


G1 Coagulation INR

INR




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Neoplastin Plus	6	100.0	0.0	0.0	1.89	7.6	e*
2	Innovin	17	94.1	5.9	0.0	1.57	7.0	e
3	Recombiplastin 2G	10	100.0	0.0	0.0	1.56	3.7	e
4	Eurolyser	6	83.3	16.7	0.0	1.52	13.8	a
5	Neoplastin R	7	100.0	0.0	0.0	1.70	1.7	e

Fibrinogen OA



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Siemens Thrombin	6	100.0	0.0	0.0	1.07	5.6	e*
2	Stago/STA	9	100.0	0.0	0.0	1.19	5.0	e
3	Fibrinogen Q.F.A.	4	75.0	25.0	0.0	1.27	12.6	a

Activated Prothrombin Time




QUALAB tolerance : 25 %

Activated Prothrombin Time (Sek)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Actin FS	8	100.0	0.0	0.0	45.5	12.1	e*
2	Pathromtin SL	5	80.0	20.0	0.0	64.5	17.8	e*
3	Stago/STA	6	100.0	0.0	0.0	48.5	5.6	e
4	aPTT-SP	8	87.5	12.5	0.0	39.6	13.2	e*

Prothrombin time NT




QUALAB tolerance : 15 %

Prothrombin time NT (%)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Neoplastin R	4	100.0	0.0	0.0	91	5.4	e*
2	Neoplastin Plus	4	100.0	0.0	0.0	94	4.6	e*
3	Innovin	11	100.0	0.0	0.0	97	3.8	e
4	Recombiplastin 2G	11	100.0	0.0	0.0	100	0.0	e

Fibrinogen N




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Siemens Thrombin	7	100.0	0.0	0.0	2.60	2.1	e
2	Stago/STA	9	100.0	0.0	0.0	2.89	5.3	e
3	Fibrinogen Q.F.A.	4	100.0	0.0	0.0	2.85	2.1	e
4	Fib Clauss (IL)	5	100.0	0.0	0.0	2.97	5.0	e*

aPTT N



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Actin FS	7	100.0	0.0	0.0	27.0	7.4	e
2	Other methods	4	100.0	0.0	0.0	25.7	4.3	e
3	Stago/STA	7	100.0	0.0	0.0	33.0	2.5	e
4	aPTT-SP	9	100.0	0.0	0.0	25.8	3.4	e

Prothrombin time HT




QUALAB tolerance : 15 %

Prothrombin time HT (%)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Neoplastin R	5	100.0	0.0	0.0	94	4.1	e*
2	Innovin	9	100.0	0.0	0.0	94	4.4	e
3	Recombiplastin 2G	9	100.0	0.0	0.0	100	3.5	e

Fibrinogen H




QUALAB tolerance : 15 %

Fibrinogen H (g/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Siemens Thrombin	6	100.0	0.0	0.0	3.05	2.6	e
2	Stago/STA	7	100.0	0.0	0.0	3.20	4.9	e*
3	Fib Clauss (IL)	5	100.0	0.0	0.0	3.53	7.1	e*

aPTT H




QUALAB tolerance : 25 %

aPTT H (Sek)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Actin FS	6	100.0	0.0	0.0	56.4	11.7	e*
2	Other methods	4	75.0	25.0	0.0	74.7	14.5	e*
3	Stago/STA	5	60.0	20.0	20.0	64.0	17.8	e*
4	aPTT-SP	6	83.3	16.7	0.0	73.9	12.1	e*

Troponin I




QUALAB tolerance : 24 %

Troponin I (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Vidas hs	4	100.0	0.0	0.0	762.2	8.6	e*
2	Architect High Sensi	6	100.0	0.0	0.0	99.6	3.0	e
3	AQT 90 FLEX	6	100.0	0.0	0.0	10.0	0.0	e
4	ADVIA Centaur XP/CP	5	100.0	0.0	0.0	92.4	9.8	e*
5	Eurolyser	14	100.0	0.0	0.0	1000.0	0.0	e

Troponin T




QUALAB tolerance : 24 %

Troponin T (ng/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas hs	5	100.0	0.0	0.0	25.50	14.3	e*
2	Cobas hs STAT	6	100.0	0.0	0.0	24.50	1.3	e

Myoglobin




QUALAB tolerance : 30 %

Myoglobin (µg/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	4	100.0	0.0	0.0	42.6	6.3	e

D-dimer



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	STA Liatest	7	100.0	0.0	0.0	1.32	4.6	e
2	Eurolyser	27	77.8	3.7	18.5	0.63	9.7	e
3	ACL	5	100.0	0.0	0.0	2.10	3.9	e
4	AQT 90 FLEX	8	100.0	0.0	0.0	0.85	6.8	e
5	Vidas	12	100.0	0.0	0.0	1.36	6.3	e

D-Dimer NC




QUALAB tolerance : 21 %

D-Dimer NC (mg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	NycoCard	25	64.0	0.0	36.0	0.60	11.7	e

TSH




QUALAB tolerance : 20 %

TSH (mU/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	10	100.0	0.0	0.0	2.5	4.7	e
2	ADVIA Centaur XP/CP	6	100.0	0.0	0.0	2.1	2.7	e
3	Architect	8	100.0	0.0	0.0	2.1	2.9	e
4	Vidas	13	100.0	0.0	0.0	2.6	5.6	e
5	Qualigen	5	80.0	0.0	20.0	1.7	7.8	e*

K6 Hormones


FT3



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	9	100.0	0.0	0.0	6.2	5.8	e
2	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	5.3	4.8	e
3	Architect	7	100.0	0.0	0.0	4.8	5.1	e
4	Vidas	6	100.0	0.0	0.0	4.8	5.6	e

K6 Hormones

FT4




QUALAB tolerance : 20 %

FT4 (pmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Qualigen	4	75.0	0.0	25.0	22.7	19.7	e*
2	Cobas E / Elecsys	10	100.0	0.0	0.0	24.8	1.9	e
3	ADVIA Centaur XP	4	100.0	0.0	0.0	18.1	7.7	e*
4	Architect	8	100.0	0.0	0.0	17.1	6.0	e
5	Vidas	7	100.0	0.0	0.0	21.2	3.3	e

Testosterone




QUALAB tolerance : 30 %

Testosterone (nmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	7	3.7	e

Cortisol




QUALAB tolerance : 20 %

Cortisol (nmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	6	100.0	0.0	0.0	278	10.3	e*
2	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	315	10.0	e*

Luteinizing hormone




QUALAB tolerance : 24 %

Luteinizing hormone (U/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	9.0	1.9	e


Follicle-stimulating hormone



QUALAB tolerance : 24 % Follicle-stimulating hormone (U/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	11.3	8.3	e*

Prolactine




QUALAB tolerance : 24 %

Prolactine (µg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	8.3	4.8	e

Troponin T CR




QUALAB tolerance : 24 %

Troponin T CR (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas h 232	869	93.2	5.5	1.3	424.00	12.5	e
2	Cardiac Reader	51	88.3	7.8	3.9	416.45	13.1	e

Troponin I WB




QUALAB tolerance : 24 %

Troponin I WB (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	iStat	4	100.0	0.0	0.0	5.13	5.0	e

D-dimer CR




QUALAB tolerance : 21 %

D-dimer CR (mg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas h 232	898	93.9	3.9	2.2	0.31	9.4	e
2	Cardiac Reader	41	92.7	7.3	0.0	0.29	12.7	e

CKMB - K8




QUALAB tolerance : 40 %

CKMB - K8 (µg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas h 232	10	100.0	0.0	0.0	9.5	10.1	e

proBNP CR




QUALAB tolerance : 27 %

proBNP CR (ng/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas h 232	543	95.2	1.7	3.1	2439	11.5	e
2	Cardiac Reader	13	76.9	23.1	0.0	2329	19.0	e*

PCO2 CCA



No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 OPTI CCA	13	84.6	15.4	0.0	5.44	6.2	e*

PO2 CCA




QUALAB tolerance : 15 %

PO2 CCA (kPa)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	OPTI CCA	13	92.3	0.0	7.7	13.36	6.7	e*

pH CCA



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	OPTI CCA	13	100.0	0.0	0.0	7.43	0.2	e

Potassium CCA




QUALAB tolerance : 6 %

Potassium CCA (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	OPTI CCA	8	100.0	0.0	0.0	4.7	1.5	e

Sodium CCA




QUALAB tolerance : 6 %

Sodium CCA (mmol/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	OPTI CCA	7	100.0	0.0	0.0	143.7	1.6	e

Ferritin



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Beckman	5	80.0	20.0	0.0	74.00	16.4	e*
2	all Participants	4	75.0	0.0	25.0	72.28	7.5	e*
3	Cobas E / Elecsys	8	100.0	0.0	0.0	86.29	10.4	e*
4	Architect	4	100.0	0.0	0.0	115.48	3.1	e
5	Mira/DiaSys	5	80.0	0.0	20.0	66.00	10.6	e*
6	Mini Vidas	6	100.0	0.0	0.0	67.65	9.1	e*
7	Eurolyser	17	94.1	0.0	5.9	61.49	6.1	e

Vitamin B12




QUALAB tolerance : 20 %

Vitamin B12 (pmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	226.00	5.3	e*
2	Cobas E / Elecsys	7	100.0	0.0	0.0	278.50	5.4	e
3	Architect	6	100.0	0.0	0.0	262.00	5.4	e

Folate




QUALAB tolerance : 20 %

Folate (nmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas alt	4	100.0	0.0	0.0	11.08	8.4	a
2	Architect	5	100.0	0.0	0.0	10.40	9.7	e*

BNP




QUALAB tolerance : 27 %

BNP (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Triage	43	86.1	2.3	11.6	397.1	13.4	e

NT-Pro-BNP



QUALAB tolerance : 27 %


NT-Pro-BNP (ng/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Triage	9	33.3	11.1	55.6	353	19.4	e*

G10 INR INRatio

INR INRatio

University Hospital Zuerich




INRatio Participants

G10 is a split-sample survey. We compare INR-values from our participants with the corresponding plasma INR from University Hospital Zuerich.

Nr.	Device	Total	% good	% insufficient	% outlier
1	INRatio	50	82.00	8.00	10.00

Bilirubin total Neo




QUALAB tolerance : 18 %

Bilirubin total Neo (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	14	100.0	0.0	0.0	196	5.8	e

Bilirubin direct




QUALAB tolerance : 18 %

Bilirubin direct (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	12	91.7	8.3	0.0	99	9.7	e*

Bilirubin neonatal




QUALAB tolerance : 18 %

Bilirubin neonatal (µmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	6	100.0	0.0	0.0	272	3.8	e
2	ABL700/800	4	25.0	0.0	75.0	272	0.0	a

CK-MB




QUALAB tolerance : 30 %

CK-MB (U/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Fuji Dri-Chem	44	100.0	0.0	0.0	72.7	6.0	e

PSA




QUALAB tolerance : 25 %

PSA (µg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	9	100.0	0.0	0.0	6.60	4.5	a
2	ADVIA Centaur XP/CP	5	100.0	0.0	0.0	4.85	3.1	a
3	Architect	6	100.0	0.0	0.0	5.09	5.0	a
4	Qualigen	6	83.3	0.0	16.7	7.08	6.3	a

free PSA




QUALAB tolerance : 25 %

free PSA (µg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	5	100.0	0.0	0.0	1.59	2.3	a
2	Architect	5	100.0	0.0	0.0	1.47	9.8	a

CEA




QUALAB tolerance : 25 %

CEA (µg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	6	100.0	0.0	0.0	24.8	7.4	a
2	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	33.0	8.1	a
3	Architect	5	100.0	0.0	0.0	38.1	4.9	a

CA 125




QUALAB tolerance : 25 %

CA 125 (kIU/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Architect	4	100.0	0.0	0.0	247.0	7.7	a

CA 15-3




QUALAB tolerance : 25 %

CA 15-3 (kIU/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Architect	4	100.0	0.0	0.0	86.0	5.6	a

AFP




QUALAB tolerance : 25 %

AFP (µg/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	4	100.0	0.0	0.0	12	6.0	e*
2	Architect	4	100.0	0.0	0.0	12	1.4	e

HCG qn



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	5	100.0	0.0	0.0	22	2.7	a
2	Vidas	4	100.0	0.0	0.0	10	9.5	a

pH OR




QUALAB tolerance : 1 %

pH OR ()


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	83	100.0	0.0	0.0	7.58	0.1	e
2	Radiometer NPT-7	4	100.0	0.0	0.0	7.60	0.1	e
3	ABL 90	33	100.0	0.0	0.0	7.59	0.0	e
4	ABL 80 / Coox	22	100.0	0.0	0.0	7.61	0.3	e
5	ABL 5	5	100.0	0.0	0.0	7.57	0.1	e

pCO2 OR




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	82	100.0	0.0	0.0	2.87	2.5	e
2	Radiometer NPT-7	4	100.0	0.0	0.0	2.99	7.2	e*
3	ABL 90	33	100.0	0.0	0.0	2.87	2.3	e
4	ABL 80 / Coox	22	86.4	4.5	9.1	2.83	6.6	e
5	ABL 5	5	100.0	0.0	0.0	2.93	3.6	e*

pO2 OR



No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 ABL700/800	82	96.4	2.4	1.2	9.99	5.8	e
2 Radiometer NPT-7	4	75.0	0.0	25.0	8.84	4.4	e*
3 ABL 90	33	75.8	3.0	21.2	7.22	6.3	e
4 ABL 80 / Coox	22	68.2	13.6	18.2	7.92	12.1	e*
5 ABL 5	5	80.0	20.0	0.0	8.11	9.8	e*

ctHb OR




QUALAB tolerance : 9 %

ctHb OR (g/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	71	100.0	0.0	0.0	190.3	2.3	e
2	ABL 90	33	93.9	0.0	6.1	191.2	0.3	e
3	ABL 80 / Coox	15	100.0	0.0	0.0	189.4	3.2	e

sO2 OR



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	57	100.0	0.0	0.0	70.153	0.1	e
2	ABL 90	31	96.8	0.0	3.2	70.103	0.1	e
3	ABL 80 / Coox	14	92.9	0.0	7.1	70.162	0.2	e

FO2Hb OR




QUALAB tolerance : 20 %

FO2Hb OR (%)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	52	100.0	0.0	0.0	48.773	0.2	e
2	ABL 90	31	100.0	0.0	0.0	48.929	0.2	e
3	ABL 80 / Coox	15	100.0	0.0	0.0	48.893	0.1	e

FCOHb OR




QUALAB tolerance : 20 %

FCOHb OR (%)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	52	100.0	0.0	0.0	20.427	0.7	e
2	ABL 90	31	100.0	0.0	0.0	20.190	0.6	e
3	ABL 80 / Coox	15	100.0	0.0	0.0	20.327	1.0	e

FMetHb OR




QUALAB tolerance : 20 %

FMetHb OR (%)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	54	100.0	0.0	0.0	10.041	1.2	e
2	ABL 90	31	100.0	0.0	0.0	9.997	0.5	e
3	ABL 80 / Coox	15	100.0	0.0	0.0	10.013	0.6	e

FHHb




QUALAB tolerance : 20 %

FHHb (%)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	ABL 80 / Coox	5	100.0	0.0	0.0	20.900	0.3	e

FHbF OR




QUALAB tolerance : 20 %

FHbF OR (%)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	ABL 90	8	100.0	0.0	0.0	54.500	4.0	e

Bilirubin OR




QUALAB tolerance : 18 %

Bilirubin OR (µmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	5	100.0	0.0	0.0	441.0	2.9	e
2	ABL 90	12	100.0	0.0	0.0	441.1	0.3	e

Potassium OR




QUALAB tolerance : 6 %

Potassium OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	68	100.0	0.0	0.0	5.5	0.9	e
2	ABL 90	33	100.0	0.0	0.0	5.5	0.7	e
3	ABL 80 / Coox	10	100.0	0.0	0.0	5.1	1.4	e

Sodium OR




QUALAB tolerance : 6 %

Sodium OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	65	100.0	0.0	0.0	127.1	0.9	e
2	ABL 90	33	100.0	0.0	0.0	127.1	0.5	e
3	ABL 80 / Coox	8	100.0	0.0	0.0	122.0	1.2	e

Calcium OR




QUALAB tolerance : 12 %

Calcium OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	66	98.5	1.5	0.0	0.38	4.1	e
2	ABL 90	33	100.0	0.0	0.0	0.38	1.8	e
3	ABL 80 / Coox	8	62.5	12.5	25.0	0.30	8.2	a

Choride OR




QUALAB tolerance : 6 %

Choride OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	55	96.4	3.6	0.0	67.15	2.6	e
2	ABL 90	33	100.0	0.0	0.0	65.00	1.9	e
3	ABL 80 / Coox	7	71.4	0.0	28.6	68.00	2.2	e*

Glucose OR




QUALAB tolerance : 10 %

Glucose OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	67	98.5	1.5	0.0	14.0	2.3	e
2	ABL 90	33	100.0	0.0	0.0	13.6	3.3	e

Lactate OR




QUALAB tolerance : 10 %

Lactate OR (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	71	100.0	0.0	0.0	10.28	4.0	e
2	ABL 90	33	97.0	3.0	0.0	10.54	2.8	e

BNP Plasma




QUALAB tolerance : 27 %

BNP Plasma (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ADVIA Centaur XP/CP	5	100.0	0.0	0.0	519.0	7.9	e*

NT-proBNP




QUALAB tolerance : 27 %

NT-proBNP (ng/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	AQT 90 FLEX	7	100.0	0.0	0.0	1740.0	12.4	e*
2	Vidas	4	100.0	0.0	0.0	650.5	16.0	e*
3	Cobas E / Elecsys	12	100.0	0.0	0.0	537.6	7.2	e

Cholesterin PTS




QUALAB tolerance : 10 %

Cholesterin PTS (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	CardioChek	7	85.7	0.0	14.3	2.59	0.0	e

Cholesterin HDL PTS




QUALAB tolerance : 21 %

Cholesterin HDL PTS (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	CardioChek	7	71.4	0.0	28.6	1.37	3.8	e

Triglyceride PTS




QUALAB tolerance : 20 %

Triglyceride PTS (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	CardioChek	7	71.4	14.3	14.3	0.64	11.2	e*

Creatinine U




QUALAB tolerance : 30 %

Creatinine U (mg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Siemens Clinitek	15	73.3	6.7	20.0	30.0	9.4	a
2	Afinion	336	96.1	1.8	2.1	41.6	11.2	e
3	NycoCard	10	90.0	0.0	10.0	40.0	17.8	e*
4	Turbidimetry	17	100.0	0.0	0.0	43.7	5.8	e
5	DCA2000/Vantage	124	97.6	0.0	2.4	43.9	6.0	e

Creatinin Urin




QUALAB tolerance : 21 %

Creatinin Urin (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	DCA2000/Vantage	123	94.3	1.6	4.1	5.4	6.0	e
2	Afinion	336	98.5	0.3	1.2	4.6	5.2	e
3	Standard chemistry	31	100.0	0.0	0.0	5.0	4.2	e
4	Siemens Clinitek	14	42.9	0.0	57.1	8.8	0.2	e

INR CCXS




QUALAB tolerance : 15 %

INR CCXS ()

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	CoaguChek XS	2325	99.4	0.3	0.3	1.9	4.4	e

INR HC




QUALAB tolerance : 15 %

INR HC ()

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Hemochron j.	21	66.7	23.8	9.5	4.1	12.3	e*

Osmolality




QUALAB tolerance : 6 %

Osmolality (mosm/kg)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Cryoskopie	13	100.0	0.0	0.0	313	1.2	e

Potassium - K22




QUALAB tolerance : 6 %

Potassium - K22 (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	ISE	9	100.0	0.0	0.0	3.9	1.0	e

Sodium - K22




QUALAB tolerance : 6 %

Sodium - K22 (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	ISE	9	100.0	0.0	0.0	141	0.5	e

Glucose - K22




QUALAB tolerance : 10 %

Glucose - K22 (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	9	100.0	0.0	0.0	3.9	3.0	e

Urea - K22




QUALAB tolerance : 15 %

Urea - K22 (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	9	100.0	0.0	0.0	5.1	2.7	e

Osmotic Gap




QUALAB tolerance : 20 %

Osmotic Gap (mmol/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Formel 1 (2Na+K+Glu+	6	100.0	0.0	0.0	17.9	9.9	e*

Procalcitonin



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Other methods	5	100.0	0.0	0.0	1.99	6.4	e
2	Cobas	8	100.0	0.0	0.0	1.84	6.9	e
3	Mini Vidas	9	100.0	0.0	0.0	2.23	4.9	e
4	Liason	4	100.0	0.0	0.0	3.38	3.5	e

Parathyroid hormone




QUALAB tolerance : 24 %

Parathyroid hormone (pmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas PTH STAT	5	100.0	0.0	0.0	7.9	5.8	e

25-OH Vitamin D




QUALAB tolerance : 27 %

25-OH Vitamin D (nmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Qualigen	5	80.0	0.0	20.0	91.4	13.7	a
2	Other methods	5	80.0	0.0	20.0	71.0	3.9	a
3	Cobas	5	100.0	0.0	0.0	81.4	7.5	e*
4	Architect	7	100.0	0.0	0.0	68.9	4.2	e

Digoxin




QUALAB tolerance : 24 %

Digoxin (nmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Other methods	10	100.0	0.0	0.0	2.06	10.7	e*

Valproat




QUALAB tolerance : 24 %

Valproat (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	4	100.0	0.0	0.0	533.5	3.6	e

Cystatin C




QUALAB tolerance : 24 %

Cystatin C (mg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	4	100.0	0.0	0.0	0.7	13.2	e*

Hemoglobin BG




QUALAB tolerance : 9 %

Hemoglobin BG (g/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	iStat	4	75.0	0.0	25.0	178.5	4.1	e*

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	iStat	5	100.0	0.0	0.0	0.55	3.8	e*
2	EPOC	5	80.0	0.0	20.0	0.51	6.8	a

Troponin Triage




QUALAB tolerance : 24 %

Troponin Triage (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Triage Next Gen	30	93.4	3.3	3.3	4592.90	14.9	e
2	Triage SOB/Cardiac	23	78.3	13.0	8.7	2497.62	14.0	e*

D-dimer Triage




QUALAB tolerance : 21 %

D-dimer Triage (ng/ml)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Triage	49	91.8	4.1	4.1	284.89	10.9	e

CK-MB Triage




QUALAB tolerance : 40 %

CK-MB Triage (µg/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Triage	21	100.0	0.0	0.0	6.8	12.6	e

Myoglobin Triage




QUALAB tolerance : 30 %

Myoglobin Triage (µg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Triage	19	94.7	5.3	0.0	264.8	11.7	e

Lipoprotein (a)




QUALAB tolerance : 25 %

Lipoprotein (a) (nmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	4	100.0	0.0	0.0	35	13.3	a

Ethanol




QUALAB tolerance : 25 %

Ethanol (mmol/l)


No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 all Participants	13	100.0	0.0	0.0	53.7	3.1	e

Calprotectin




No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Bühlmann	7	100.0	0.0	0.0	167	9.0	e

Cholesterol total Af/b101



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b101	44	100.0	0.0	0.0	4.04	2.5	e
2	Afinion	275	98.9	0.7	0.4	4.29	3.9	e

Cholesterol HDL Af/b101




QUALAB tolerance : 21 %

Cholesterol HDL Af/b101 (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b101	44	95.5	0.0	4.5	1.56	4.6	e
2	Afinion	263	96.9	2.3	0.8	1.92	7.1	e

Tryglycerides Af/b101




QUALAB tolerance : 20 %

Tryglycerides Af/b101 (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b101	44	84.1	2.3	13.6	1.31	6.3	e
2	Afinion	272	88.6	1.1	10.3	1.22	6.7	e

Troponin I S




QUALAB tolerance : 24 %

Troponin I S (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Samsung LABGEO IB10	67	88.1	10.4	1.5	1441.89	14.9	e

D-dimer qn S




QUALAB tolerance : 21 %

D-dimer qn S (mg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Samsung LABGEO IB10	81	98.8	0.0	1.2	1.14	8.9	e

NT-pro BNP S




QUALAB tolerance : 27 %

NT-pro BNP S (ng/l)


No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Samsung LABGEO IB10	53	71.7	20.8	7.5	709.7	19.7	e

INR MI




No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 MicroINR	72	88.9	1.4	9.7	1.9	7.5	e

Lipase




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Autolyser/DiaSys	4	100.0	0.0	0.0	23.0	9.9	e*
2	Architect	4	75.0	0.0	25.0	18.0	6.5	e*
3	Beckman	5	80.0	0.0	20.0	18.0	2.8	e
4	Cobas	9	100.0	0.0	0.0	20.0	6.9	e*
5	Fuji Dri-Chem	88	98.9	0.0	1.1	38.0	4.4	e

INR Xprecia



No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Xprecia	21	100.0	0.0	0.0	1.2	5.8	e

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	32	100.0	0.0	0.0	145.1	1.1	e

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	32	96.9	3.1	0.0	0.55	4.1	e

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	32	100.0	0.0	0.0	6.99	1.5	e

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	32	100.0	0.0	0.0	16.71	3.5	e

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	32	100.0	0.0	0.0	334.5	8.1	e

Neutrophils




QUALAB tolerance : 25 %

Neutrophils (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	32	100.0	0.0	0.0	14.38	4.2	e

Lymphocytes




QUALAB tolerance : 25 %

Lymphocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	32	100.0	0.0	0.0	1.22	6.5	e

Monocytes




QUALAB tolerance : 25 %

Monocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	32	81.3	3.1	15.6	0.35	11.6	e

Eosinophils




QUALAB tolerance : 50 %

Eosinophils (G/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Sysmex	32	100.0	0.0	0.0	0.46	8.0	e

Basophiles




QUALAB tolerance : 80 %

Basophiles (G/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Sysmex	32	100.0	0.0	0.0	0.29	18.6	a

Reticulocytes




QUALAB tolerance : 25 %

Reticulocytes (G/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Sysmex	16	81.2	12.5	6.3	38.0	29.8	a

Hämolyseindex Probe A




QUALAB tolerance : 15 %

Hämolyseindex Probe A ()

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Cobas	14	92.9	7.1	0.0	18.0	8.1	e*

Hämolyseindex Probe B



QUALAB tolerance : 15 %

Hämolyseindex Probe B ()

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Cobas	14	100.0	0.0	0.0	41.0	5.3	e