

Report U03 Drugs of Abuse

Amphetamine

	Target Value	Participants	Conform	% Conform
Architect	negative	1	1	100.00
Cobas	negative	3	3	100.00
Syva Rapid Test	negative	1	1	100.00
Intex	negative	5	5	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	15	15	100.00
Others	negative	54	54	100.00
Total		80	80	100.00

Methamphetamine

	Target Value	Participants	Conform	% Conform
Intex	negative	2	2	100.00
Syva Rapid Test	negative	4	4	100.00
Others	negative	25	24	96.00
Total		31	30	97.00

Ecstasy pos/neg

	Target Value	Participants	Conform	% Conform
Others	negative	15	15	100.00

Barbiturates

	Target Value	Participants	Conform	% Conform
Architect	negative	1	1	100.00
Cobas	negative	3	3	100.00
Intex	negative	1	1	100.00
Other quantitativ	negative	10	10	100.00
Others	negative	24	24	100.00
Total		39	39	100.00

Benzodiazepines

	Target Value	Participants	Conform	% Conform
Architect	negative	1	1	100.00
Cobas	negative	4	4	100.00
Syva Rapid Test	negative	1	1	100.00
Intex	negative	10	10	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	16	16	100.00
Others	negative	60	60	100.00
Total		93	93	100.00

Cannabinoïdes

	Target Value	Participants	Conform	% Conform
Architect	negative	1	1	100.00
Cobas	negative	4	4	100.00
Syva Rapid Test	negative	5	5	100.00
Intex	negative	11	11	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	18	18	100.00
Others	negative	66	66	100.00
Total		106	106	100.00

Cocaine

	Target Value	Participants	Conform	% Conform
Architect	positive	1	1	100.00
Cobas	positive	4	4	100.00
Syva Rapid Test	positive	4	4	100.00
Intex	positive	8	8	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	positive	17	17	100.00
Others	positive	64	63	98.00
Total		99	98	99.00

Methadone

	Target Value	Participants	Conform	% Conform
Architect	negative	1	1	100.00
Cobas	negative	4	4	100.00
Intex	negative	10	10	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	16	16	100.00
Others	negative	50	49	98.00
Total		82	81	99.00

EDDP

	Target Value	Participants	Conform	% Conform
all Participants	negative	2	2	100.00

Opiate

	Target Value	Participants	Conform	% Conform
Architect	positive	1	1	100.00
Cobas	positive	4	4	100.00
Syva Rapid Test	positive	4	4	100.00
Intex	positive	11	11	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	positive	17	17	100.00
Others	positive	63	62	98.00
Total		101	100	99.00

Buprenorphin

	Target Value	Participants	Conform	% Conform
all Participants	negative	3	3	100.00

6-Acetylmorphin

	Target Value	Participants	Conform	% Conform
all Participants	negative	2	2	100.00

Phencyclidin

	Target Value	Participants	Conform	% Conform
Others	positive	7	6	86.00

TCA

	Target Value	Participants	Conform	% Conform
all Participants	negative	15	15	100.00

LSD

	Target Value	Participants	Conform	% Conform
Other quantitativ	negative	1	1	100.00

Commentary U3