

G3/G4 Coagulation advanced 2018-1

Instructions

- The numbers show how many participants had a result in a class.
- The fields with a box show in which class the results of your lab were.
- Target values are calculated as consensus values after elimination of outliers.
- For thrombintime, we evaluate only the interpretation

Interpretation of the results

No therapy		Normal	Limit	Abnormal
G3	PTT	19	1	0
G3	Thrombintime	22	0	0

Heparin therapy		No Heparin	Subtherap.	Therap.	To much
G4	PTT	0	0	14	3
G4	Thrombintime	0	0	4	10

14.3.2018

Distribution of aPTT results

Due to the large scatter of the quantitative results, it is difficult to calculate target values. As an additional help to interpret the results, we provide the following histograms that show the distribution of all values.

