

Report U03 Drugs of Abuse

Amphetamine

	Target Value	Participants	Conform	% Conform
Architect	negative	2	2	100.00
Cobas	negative	3	3	100.00
Syva Rapid Test	negative	1	1	100.00
Intex	negative	4	4	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	16	16	100.00
Others	negative	55	53	96.00
Total		82	80	98.00

Methamphetamine

	Target Value	Participants	Conform	% Conform
Intex	negative	1	1	100.00
Syva Rapid Test	negative	4	4	100.00
Others	negative	24	23	96.00
Total		29	28	97.00

Ecstasy pos/neg

	Target Value	Participants	Conform	% Conform
Others	negative	15	15	100.00

Barbiturates

	Target Value	Participants	Conform	% Conform
Architect	positive	1	1	100.00
Cobas	positive	3	3	100.00
Other quantitativ	positive	10	10	100.00
Others	positive	23	22	96.00
Total		37	36	97.00

Benzodiazepines

	Target Value	Participants	Conform	% Conform
Architect	negative	2	2	100.00
Cobas	negative	4	4	100.00
Syva Rapid Test	negative	1	1	100.00
Intex	negative	9	9	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	17	17	100.00
Others	negative	61	59	97.00
Total		95	93	98.00

Cannabinoides

	Target Value	Participants	Conform	% Conform
Architect	positive	2	2	100.00
Cobas	positive	5	5	100.00
Syva Rapid Test	positive	5	5	100.00
Intex	positive	11	10	91.00
Syntron Quik	positive	1	1	100.00
Other quantitativ	positive	19	18	95.00
Others	positive	68	58	85.00
Total		111	99	89.00

Cocaine

	Target Value	Participants	Conform	% Conform
Architect	negative	2	2	100.00
Cobas	negative	4	4	100.00
Syva Rapid Test	negative	4	4	100.00
Intex	negative	7	7	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	18	18	100.00
Others	negative	65	61	94.00
Total		101	97	96.00

Methadone

	Target Value	Participants	Conform	% Conform
Architect	negative	2	2	100.00
Cobas	negative	4	4	100.00
Intex	negative	8	8	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	17	17	100.00
Others	negative	51	49	96.00
Total		83	81	98.00

EDDP

	Target Value	Participants	Conform	% Conform
all Participants	negative	2	2	100.00

Opiate

	Target Value	Participants	Conform	% Conform
Architect	positive	2	2	100.00
Cobas	positive	4	4	100.00
Syva Rapid Test	positive	4	4	100.00
Intex	positive	9	9	100.00
Syntron Quik	positive	1	1	100.00
Other quantitativ	positive	18	18	100.00
Others	positive	64	64	100.00
Total		102	102	100.00

Buprenorphin

	Target Value	Participants	Conform	% Conform
all Participants	negative	3	3	100.00

6-Acetylmorphin

	Target Value	Participants	Conform	% Conform
all Participants	positive	2	2	100.00

Phencyclidin

	Target Value	Participants	Conform	% Conform
Others	negative	5	5	100.00

TCA

	Target Value	Participants	Conform	% Conform
all Participants	negative	13	13	100.00

Commentary U3