

Report U03 Drugs of Abuse**Amphetamine**

	Target Value	Participants	Conform	% Conform
Architect	negative	2	2	100.00
Cobas	negative	3	3	100.00
Syva Rapid Test	negative	1	1	100.00
Intex	negative	2	2	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	16	16	100.00
Others	negative	54	53	98.00
Total		79	78	99.00

Methamphetamine

	Target Value	Participants	Conform	% Conform
Intex	negative	1	1	100.00
Syva Rapid Test	negative	4	4	100.00
Others	negative	28	27	96.00
Total		33	32	97.00

Ecstasy pos/neg

	Target Value	Participants	Conform	% Conform
Others	negative	17	17	100.00

Barbiturates

	Target Value	Participants	Conform	% Conform
Architect	negative	1	1	100.00
Cobas	negative	3	3	100.00
Other quantitativ	negative	10	10	100.00
Others	negative	24	24	100.00
Total		38	38	100.00

Benzodiazepines

	Target Value	Participants	Conform	% Conform
Architect	positive	2	2	100.00
Cobas	positive	4	4	100.00
Syva Rapid Test	positive	1	1	100.00
Intex	positive	7	7	100.00
Syntron Quik	positive	1	1	100.00
Other quantitativ	positive	17	17	100.00
Others	positive	61	61	100.00
Total		93	93	100.00

Cannabinoïdes

	Target Value	Participants	Conform	% Conform
Architect	negative	2	2	100.00
Cobas	negative	4	4	100.00
Syva Rapid Test	negative	4	4	100.00
Intex	negative	10	10	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	19	19	100.00
Others	negative	68	68	100.00
Total		108	108	100.00

Cocaine

	Target Value	Participants	Conform	% Conform
Architect	negative	2	2	100.00
Cobas	negative	4	4	100.00
Syva Rapid Test	negative	4	4	100.00
Intex	negative	5	5	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	18	18	100.00
Others	negative	65	65	100.00
Total		99	99	100.00

Methadone

	Target Value	Participants	Conform	% Conform
Architect	negative	2	2	100.00
Cobas	negative	4	4	100.00
Intex	negative	6	6	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	17	17	100.00
Others	negative	51	51	100.00
Total		81	81	100.00

Methaqualon

	Target Value	Participants	Conform	% Conform
all Participants	positive	1	1	100.00

EDDP

	Target Value	Participants	Conform	% Conform
all Participants	negative	2	2	100.00

Opiate

	Target Value	Participants	Conform	% Conform
Architect	negative	2	2	100.00
Cobas	negative	4	4	100.00
Syva Rapid Test	negative	4	4	100.00
Intex	negative	7	7	100.00
Syntron Quik	negative	1	1	100.00
Other quantitativ	negative	18	18	100.00
Others	negative	63	62	98.00
Total		99	98	99.00

Buprenorphin

	Target Value	Participants	Conform	% Conform
all Participants	negative	3	3	100.00

6-Acetylmorphin

	Target Value	Participants	Conform	% Conform
all Participants	negative	2	2	100.00

Phencyclidin

	Target Value	Participants	Conform	% Conform
Others	positive	7	6	86.00

TCA

	Target Value	Participants	Conform	% Conform
all Participants	positive	15	13	87.00

LSD

	Target Value	Participants	Conform	% Conform
Other quantitativ	negative	1	1	100.00

Commentary U3