

Verein für
Association pour le
Associazione per il



medizinische Qualitätskontrolle
contrôle de qualité médical
controllo di qualità medico

Survey Report

2018 - 3

Survey Specimens

The homogeneity and stability of all specimens were checked before and/or during shipment and no irregularities were noted. The suitability tests were performed by the laboratories of the Universitätsspital Zürich (University Hospital Zurich) (<http://www.uzl.usz.ch/>).

The following survey specimens were produced specifically for MQ by a sub-contractor:
B1 Strep A Test, B2 Uricult, H4 Parasitic Hematology, K14 tumor marker

Determination of target values

For each target value, the type of determination per ISO17043: 2010 B2.1 is indicated (column "type"):

- a Value known due to production.
- b Certified reference value for use with special specimens
- c Reference value determined by analysis
- d Consensus values of expert laboratories
- e Consensus values of the participants

For methods groups with more than 9 participants, consensus values of the participants ("e") are generally determined.

In order to calculate the target values, we use the mean value of the method group. Values that differ more than 1.5 times the QUALAB-tolerance are outliers and are not used to calculate the target value. Starting point for the elimination of outliers are the values of our suitability tests.

In order to provide all participants with target values that are as meaningful as possible, other methods may also be applied for smaller method groups.

Uncertainty of the determined target values

The standard uncertainty (u_x) is calculated using the following formula (ISO13528):

$$u_x = (\text{target value}/100) * (1.25/\text{square root of "number of participants"}) * \%CV$$

- u_x has the same unit as the target value
- u_x can be compared with the standard deviation of the participants' collective ($SD = \text{target value} * \%CV / 100$)
- For participant numbers >18, the standard uncertainty (u_x) is significantly lower than the scatter of the collective participants and can be neglected.


QUALAB and MQ tolerances

For all mandatory analyzes, QUALAB tolerances are used (www.qualab.ch, external quality control). For non-mandatory analyzes, the tolerances are specified by MQ's survey specimen leader.

If the determined uncertainty, u_x , of the target value is greater than 15% of the QUALAB or MQ tolerance, the letter indicating the type of target detection is marked with an additional star (example "e*"). Thereby, we are alerting the participants to the fact that the uncertainty of the target value can have an impact on the evaluation.

Graphics

The results are shown graphically as follows:



Comparison of Devices

The data in this report allows you to compare the performance of different devices. However, remember to consider the following:


- The chemical control K1 is a ready-to-use commercial control serum. Even if the specimen is of human origin, it is possible that matrix effects occur. These are device-specific and result in different target values.
- Only one specimen was measured. Since the scatter of the results is dependent on the nature of the specimen (matrix effects) and on the signal strength, the determined coefficient of variations (CV in %) cannot be applied generally.
- A large number of runaways is due to administrative errors (wrong unit, results mixed up) or to operator errors (wrong sample, not correctly taken up in solution, not mixed well) and has nothing to do with the type of device.

Zürich, 28.9.2018

Dr. R. Fried
Survey Director

Publication of this report or any portion thereof without our prior written consent is not permitted. The original is archived at www.mqzh.ch

INR




QUALAB tolerance : 15 %

INR ()


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Neoplastin Plus	6	100.0	0.0	0.0	2.09	4.6	e*
2	Innovin	11	81.8	18.2	0.0	1.68	8.4	e*
3	Recombiplastin 2G	18	100.0	0.0	0.0	1.72	4.7	e
4	Eurolyser	5	40.0	40.0	20.0	1.40	14.2	e*
5	Other methods	4	100.0	0.0	0.0	1.60	8.7	e*
6	Neoplastin R	9	100.0	0.0	0.0	1.84	4.2	e

Fibrinogen OA



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Other methods	7	100.0	0.0	0.0	1.11	6.4	e*
2	Siemens Thrombin	4	75.0	0.0	25.0	0.97	4.5	e*
3	Stago/STA	9	88.9	11.1	0.0	1.04	6.2	e*
4	Fibrinogen Q.F.A.	6	83.3	16.7	0.0	1.13	8.1	e*

Activated Prothrombin Time




QUALAB tolerance : 25 %

Activated Prothrombin Time (Sek)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Other methods	5	100.0	0.0	0.0	47.0	12.4	e*
2	Actin FS	7	100.0	0.0	0.0	42.0	6.9	e
3	Pathromtin SL	4	100.0	0.0	0.0	73.2	7.0	e*
4	Stago/STA	10	100.0	0.0	0.0	51.3	6.0	e
5	aPTT-SP	11	100.0	0.0	0.0	44.7	9.3	e

INR CoaguChek




QUALAB tolerance : 15 %

INR CoaguChek ()

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	CoaguChek Pro II	323	99.1	0.0	0.9	1.1	3.8	e

Prothrombin time NT




QUALAB tolerance : 15 %

Prothrombin time NT (%)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Neoplastin R	9	100.0	0.0	0.0	93	2.8	e
2	Neoplastin Plus	5	100.0	0.0	0.0	92	5.3	e*
3	Innovin	9	88.9	0.0	11.1	95	3.2	e
4	all Participants	5	100.0	0.0	0.0	100	4.0	e
5	Recombiplastin 2G	15	100.0	0.0	0.0	98	4.5	e

Fibrinogen N




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Siemens Thrombin	8	100.0	0.0	0.0	2.43	4.7	e
2	Stago/STA	12	100.0	0.0	0.0	2.59	6.3	e*
3	Fibrinogen Q.F.A.	11	100.0	0.0	0.0	2.52	8.3	e*
4	Fib Clauss (IL)	5	80.0	20.0	0.0	2.46	8.7	e*

aPTT N



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Actin FS	5	80.0	0.0	20.0	26.0	6.0	e
2	Other methods	4	100.0	0.0	0.0	34.5	9.1	e*
3	Stago/STA	12	100.0	0.0	0.0	31.2	6.9	e
4	aPTT-SP	17	100.0	0.0	0.0	25.3	3.5	e

Prothrombin time HT




QUALAB tolerance : 15 %

Prothrombin time HT (%)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Neoplastin R	7	100.0	0.0	0.0	91	3.2	e
2	Innovin	5	100.0	0.0	0.0	87	2.2	e
3	all Participants	8	100.0	0.0	0.0	94	7.6	e*
4	Recombiplastin 2G	9	88.9	11.1	0.0	97	8.6	e*

Fibrinogen H




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Siemens Thrombin	4	100.0	0.0	0.0	2.98	1.0	e
2	Stago/STA	9	100.0	0.0	0.0	2.92	3.9	e
3	Fibrinogen Q.F.A.	4	100.0	0.0	0.0	3.00	5.2	e*
4	Fib Clauss (IL)	7	85.7	14.3	0.0	2.86	9.4	e*

aPTT H



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Actin FS	7	100.0	0.0	0.0	61.3	11.5	e*
2	Stago/STA	6	100.0	0.0	0.0	61.9	3.6	e
3	aPTT-SP	7	85.7	14.3	0.0	77.0	12.1	e*

D-Dimer NC




QUALAB tolerance : 21 %

D-Dimer NC (mg/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	NycoCard	17	41.2	41.2	17.6	1.15	20.1	e*

D-dimer




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	STA Liatest	9	100.0	0.0	0.0	2.44	5.5	e
2	Siemens Innovance	4	100.0	0.0	0.0	4.02	8.5	e*
3	Eurolyser	13	76.9	0.0	23.1	5.11	7.5	e
4	ACL	4	100.0	0.0	0.0	3.62	3.6	e
5	AQT 90 FLEX	7	100.0	0.0	0.0	1.67	4.8	e
6	VIDAS	17	100.0	0.0	0.0	2.40	5.1	e

CoaguChek APTT



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	CoaguChek Pro II	9	88.9	11.1	0.0	93.7	13.8	e*

INR CCXS




QUALAB tolerance : 15 %

INR CCXS ()


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	CoaguChek XS	2061	99.3	0.5	0.2	2.4	4.4	e

INR HC



No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Hemochron j.	10	70.0	10.0	20.0	5.1	9.6	e*

INR MI




QUALAB tolerance : 15 %

INR MI ()

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	MicroINR	107	73.9	16.8	9.3	2.0	8.6	e

INR Xprecia




QUALAB tolerance : 15 %

INR Xprecia ()

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Xprecia	57	100.0	0.0	0.0	2.9	4.3	e

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Automat	29	89.7	3.4	6.9	123.0	3.8	e
2	Cyanmethemoglobin	36	97.2	2.8	0.0	120.1	3.3	e
3	System X	38	100.0	0.0	0.0	121.9	1.8	e
4	Advia 120	10	100.0	0.0	0.0	122.2	1.8	e
5	ABX Pentra	9	100.0	0.0	0.0	122.0	2.5	e
6	Reflotron	51	80.4	11.8	7.8	126.2	5.7	e
7	Hemocue	366	96.4	1.4	2.2	120.5	2.6	e
8	Dr. Lange	14	85.8	7.1	7.1	123.5	5.4	e*
9	Hemocontrol	14	100.0	0.0	0.0	121.9	3.0	e
10	Eurolyser	6	83.3	16.7	0.0	120.0	6.7	e*
11	DiaSpect	12	100.0	0.0	0.0	125.0	3.6	e
12	MS4	4	50.0	0.0	50.0	107.0	5.3	e*

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Micros 60	242	95.4	2.1	2.5	119.2	2.9	e
2	Sysmex KX21	309	95.8	0.6	3.6	121.8	2.0	e
3	Sysmex PochH - 100i	198	92.9	1.5	5.6	119.4	2.5	e
4	Sysmex XP 300	439	98.2	0.7	1.1	121.4	2.1	e
5	Mythic	268	93.7	0.7	5.6	117.7	3.1	e
6	Swelab	48	93.7	2.1	4.2	123.0	2.9	e
7	Abacus Junior	10	80.0	0.0	20.0	123.4	2.2	e
8	Medonic	10	100.0	0.0	0.0	121.9	2.8	e
9	Celltac Alpha (Nihon	75	97.3	0.0	2.7	122.1	2.1	e
10	Samsung HC10	42	90.5	2.4	7.1	120.3	3.6	e

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Automat	25	88.0	4.0	8.0	0.34	5.3	e
2	Centrifuge	10	80.0	20.0	0.0	0.33	7.1	e*
3	Sysmex X	38	100.0	0.0	0.0	0.35	2.9	e
4	Advia 120	10	90.0	10.0	0.0	0.31	6.5	e*
5	ABX Pentra	9	100.0	0.0	0.0	0.32	3.0	e
6	MS4	4	75.0	0.0	25.0	0.33	2.5	e*

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Micros 60	241	92.5	4.6	2.9	0.32	4.0	e
2	Sysmex KX21	309	97.7	0.0	2.3	0.32	2.5	e
3	Sysmex PochH - 100i	198	91.4	1.5	7.1	0.35	3.7	e
4	Sysmex XP 300	441	98.9	0.2	0.9	0.33	2.2	e
5	Mythic	268	91.8	2.6	5.6	0.34	4.0	e
6	Swelab	48	91.6	4.2	4.2	0.34	3.9	e
7	Abacus Junior	10	80.0	0.0	20.0	0.35	2.2	e
8	Medonic	10	100.0	0.0	0.0	0.34	3.7	e*
9	Celltac Alpha (Nihon	75	97.3	0.0	2.7	0.35	2.4	e
10	Samsung HC10	43	88.3	7.0	4.7	0.36	4.7	e

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Automat	24	95.8	4.2	0.0	3.93	8.2	e
2	Microscopic	5	100.0	0.0	0.0	3.60	11.8	e*
3	Sysmex X	38	100.0	0.0	0.0	4.02	2.2	e
4	Advia 120	10	100.0	0.0	0.0	3.73	4.5	e
5	ABX Pentra	9	100.0	0.0	0.0	3.93	2.4	e
6	MS4	4	75.0	0.0	25.0	3.95	3.3	e

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Micros 60	241	98.8	1.2	0.0	3.88	6.3	e
2	Sysmex KX21	309	98.1	0.3	1.6	3.99	3.0	e
3	Sysmex PochH - 100i	198	95.5	1.5	3.0	4.11	4.8	e
4	Sysmex XP 300	441	99.3	0.5	0.2	3.99	2.9	e
5	Mythic	268	95.9	1.1	3.0	3.87	5.5	e
6	Swelab	48	95.8	2.1	2.1	4.02	5.0	e
7	Abacus Junior	10	80.0	10.0	10.0	3.90	11.0	e*
8	Medonic	10	100.0	0.0	0.0	4.03	3.0	e
9	Celltac Alpha (Nihon	75	97.3	0.0	2.7	4.02	2.4	e
10	Samsung HC10	43	97.7	0.0	2.3	3.96	5.9	e

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Automat	21	85.7	0.0	14.3	5.10	12.7	e
2	Microscopic	31	87.1	9.7	3.2	5.05	13.8	e
3	Sysmex X	38	100.0	0.0	0.0	5.67	7.4	e
4	Advia 120 (Perox)	9	88.9	11.1	0.0	5.07	15.0	e*
5	ABX Pentra	9	100.0	0.0	0.0	5.80	10.5	e*
6	MS4	4	100.0	0.0	0.0	5.50	16.0	e*

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Micros 60	241	95.1	3.7	1.2	4.99	9.9	e
2	Sysmex KX21	309	98.7	1.3	0.0	5.38	8.4	e
3	Sysmex PochH - 100i	198	95.5	1.5	3.0	5.30	7.9	e
4	Sysmex XP 300	441	98.0	1.1	0.9	5.51	8.3	e
5	Mythic	267	95.9	2.6	1.5	5.11	9.7	e
6	Swelab	48	97.9	0.0	2.1	5.53	8.1	e
7	Abacus Junior	10	70.0	20.0	10.0	6.09	18.3	e*
8	Medonic	10	100.0	0.0	0.0	5.42	7.8	e
9	Celltac Alpha (Nihon	75	100.0	0.0	0.0	5.66	5.5	e
10	Samsung HC10	43	95.3	0.0	4.7	5.38	7.1	e

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Automat	19	89.5	0.0	10.5	171.4	9.2	e
2	Microscopic	22	86.4	9.1	4.5	186.3	16.0	e*
3	Sysmex X	38	97.4	0.0	2.6	171.4	4.3	e
4	Advia 120	10	100.0	0.0	0.0	173.9	5.9	e
5	ABX Pentra	9	100.0	0.0	0.0	167.0	5.1	e
6	MS4	4	75.0	0.0	25.0	175.0	11.0	e*

Thrombocytes



QUALAB tolerance : 25 %

Thrombocytes (G/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Micros 60	241	91.7	5.0	3.3	167.5	11.9	e
2	Sysmex KX21	309	99.4	0.6	0.0	174.7	7.2	e
3	Sysmex PochH - 100i	198	96.5	1.0	2.5	176.6	7.2	e
4	Sysmex XP 300	441	98.8	0.7	0.5	181.4	6.7	e
5	Mythic	268	92.5	3.4	4.1	161.9	11.4	e
6	Swelab	48	91.6	6.3	2.1	155.5	10.7	e
7	Abacus Junior	10	80.0	10.0	10.0	160.4	11.6	e*
8	Medonic	10	100.0	0.0	0.0	152.3	6.4	e
9	Celltac Alpha (Nihon	75	98.7	0.0	1.3	171.3	7.2	e
10	Samsung HC10	43	90.6	4.7	4.7	170.7	10.5	e

Hemoglobin H2



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Micros	221	95.9	1.4	2.7	101.0	3.2	e
2	Microsemi	634	98.6	0.3	1.1	104.1	1.9	e

Hematocrit H2




QUALAB tolerance : 9 %

Hematocrit H2 (l/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Micros	220	95.9	1.4	2.7	0.30	3.6	e
2	Microsemi	633	98.4	0.3	1.3	0.29	2.6	e

Leucocytes H2




QUALAB tolerance : 25 %

Leucocytes H2 (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Micros	221	99.5	0.0	0.5	5.31	5.0	e
2	Microsemi	634	99.7	0.0	0.3	5.52	3.2	e

Thrombocytes H2




QUALAB tolerance : 25 %

Thrombocytes H2 (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Micros	221	97.7	0.9	1.4	183.4	8.2	e
2	Microsemi	634	99.6	0.2	0.2	187.7	6.1	e

Erythrocytes H2




QUALAB tolerance : 25 %

Erythrocytes H2 (T/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Abx Micros	221	98.6	0.5	0.9	3.36	4.3	e
2	Microsemi	634	98.6	0.3	1.1	3.33	2.7	e

CRP H2




QUALAB tolerance : 21 %

CRP H2 (mg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Microsemi	628	97.2	1.8	1.0	37.6	8.6	e
2	Abx Micros	20	85.0	10.0	5.0	36.7	13.3	e*
3	ABX Micros CRP200	192	95.3	3.1	1.6	33.9	9.6	e

Hemoglobin BG




QUALAB tolerance : 9 %

Hemoglobin BG (g/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	iStat	5	100.0	0.0	0.0	187.0	2.3	e

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas	4	100.0	0.0	0.0	0.49	0.8	e
2	iStat	6	100.0	0.0	0.0	0.55	3.0	e*
3	EPOC	5	60.0	20.0	20.0	0.54	6.5	e*

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	44	100.0	0.0	0.0	92.8	1.4	e
2	Advia	12	100.0	0.0	0.0	94.5	1.3	e
3	ABX Pentra	10	100.0	0.0	0.0	95.2	2.4	e

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Sysmex	44	100.0	0.0	0.0	0.36	2.7	e
2 Advia	12	100.0	0.0	0.0	0.35	2.0	e
3 ABX Pentra	10	100.0	0.0	0.0	0.31	2.2	e

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	44	100.0	0.0	0.0	4.55	1.7	e
2	Advia	12	100.0	0.0	0.0	4.69	5.9	e
3	ABX Pentra	10	100.0	0.0	0.0	4.50	0.6	e

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	44	100.0	0.0	0.0	12.89	3.0	e
2	Advia	12	100.0	0.0	0.0	11.78	4.1	e
3	ABX Pentra	10	100.0	0.0	0.0	12.12	5.1	e

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	44	100.0	0.0	0.0	216.7	5.7	e
2	Advia	12	91.7	8.3	0.0	191.0	16.5	e*
3	ABX Pentra	10	100.0	0.0	0.0	226.3	5.8	e

Neutrophils




MQ tolerance : 25 %

Neutrophils (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	44	100.0	0.0	0.0	11.40	3.0	e
2	Advia	12	100.0	0.0	0.0	10.37	5.4	e
3	ABX Pentra	10	90.0	10.0	0.0	9.65	10.0	e*

Lymphocytes




MQ tolerance : 25 %

Lymphocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	44	97.7	2.3	0.0	0.71	11.4	a
2	Advia	12	100.0	0.0	0.0	0.59	14.9	a
3	ABX Pentra	10	70.0	10.0	20.0	0.92	24.1	a

Monocytes




MQ tolerance : 25 %

Monocytes (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	44	100.0	0.0	0.0	0.27	16.0	a
2	Advia	12	91.7	0.0	8.3	0.30	37.5	a
3	ABX Pentra	10	70.0	20.0	10.0	0.62	45.5	a

Eosinophils




MQ tolerance : 50 %

Eosinophils (G/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	44	100.0	0.0	0.0	0.30	12.2	a
2	Advia	12	100.0	0.0	0.0	0.25	25.7	a
3	ABX Pentra	10	100.0	0.0	0.0	0.25	25.3	a

Basophiles



MQ tolerance : 80 %
 (< 0.10: +/- 0.08 G/l)

Basophiles (G/l)


No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Sysmex	44	97.7	0.0	2.3	0.18	26.9	e
2 Advia	12	100.0	0.0	0.0	0.10	27.5	e
3 ABX Pentra	10	90.0	0.0	10.0	0.07	52.4	e*

Reticulocytes



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sysmex	24	95.8	4.2	0.0	150.2	8.7	e
2	Advia	9	88.9	11.1	0.0	92.4	18.6	a

Hämolyseindex Probe A




MQ tolerance : 15 %

Hämolyseindex Probe A ()

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas	15	93.3	0.0	6.7	89.1	7.0	e*

Hämolyseindex Probe B




MQ tolerance : 15 %

Hämolyseindex Probe B ()

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Cobas	15	86.7	0.0	13.3	22.0	6.9	e

Erythrocyte sedimentation rate 1h




MQ tolerance : 30 %

Erythrocyte sedimentation rate 1h (mm/h)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Sarstedt Sedivette	11	100.0	0.0	0.0	6	20.0	a
2	BD Seditainer	16	87.4	6.3	6.3	6	24.2	a
3	Other methods	4	75.0	25.0	0.0	7	26.1	a

CRP




QUALAB tolerance : 21 %

CRP (mg/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	IChroma	4	100.0	0.0	0.0	65.7	7.2	e*
2	Celltac chemi	25	100.0	0.0	0.0	57.6	4.4	e
3	Cobas b101	58	100.0	0.0	0.0	51.7	3.6	e
4	Cobas	17	100.0	0.0	0.0	55.3	12.4	e*
5	Turbidimetry	38	94.7	0.0	5.3	61.0	4.8	e
6	Afinion	1375	98.7	0.9	0.4	52.9	6.3	e
7	NycoCard SingleTest-	211	84.4	5.2	10.4	54.6	11.1	e
8	Quick Read go	144	97.2	0.7	2.1	58.9	6.5	e
9	Eurolyser	114	75.4	5.3	19.3	70.8	10.4	e
10	Fuji Dri-Chem	19	73.6	21.1	5.3	50.4	15.0	e*
11	Autolyser/DiaSys	10	90.0	0.0	10.0	52.3	9.3	e*
12	Piccolo	6	83.3	16.7	0.0	77.0	11.2	e*

CRP




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	QuikRead (Vollblut)	85	97.6	0.0	2.4	97.9	5.4	e

CRP



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	AQT 90 FLEX	7	100.0	0.0	0.0	64.0	4.8	e
2	Spotchem D-Concept	6	100.0	0.0	0.0	62.1	6.2	e*
3	Spotchem SI-3510	4	100.0	0.0	0.0	54.1	2.7	e
4	Other methods	7	85.7	0.0	14.3	60.0	7.6	e*

IgG




QUALAB tolerance : 25 %

IgG (g/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Turbidimetry	13	100.0	0.0	0.0	11.4	4.4	e
2	Nephelometry	5	100.0	0.0	0.0	11.9	7.9	e*

IgA



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Turbidimetry	14	100.0	0.0	0.0	2.1	3.2	e
2	Nephelometry	5	100.0	0.0	0.0	2.3	5.9	e

IgM



QUALAB tolerance : 25 %

IgM (g/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Turbidimetry	13	100.0	0.0	0.0	1.0	6.9	e
2	Nephelometry	5	100.0	0.0	0.0	1.1	4.5	e

IgE




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	5	100.0	0.0	0.0	124	10.3	e*
2	Cobas	5	100.0	0.0	0.0	149	4.7	e

Alpha-1-Antitrypsin




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	5	100.0	0.0	0.0	1.20	2.3	e

Anti-Streptolysin-Antibodies




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	12	91.7	8.3	0.0	178	11.0	e*

Complement C3



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	12	100.0	0.0	0.0	1.69	7.4	e

Complement C4



MQ tolerance : 25 %

Complement C4 (g/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	11	90.9	9.1	0.0	0.31	11.3	e*

Haptoglobin




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	15	100.0	0.0	0.0	1.43	4.3	e

Transferrin




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	21	100.0	0.0	0.0	2.43	2.4	e

Beta-2-Mikroglobulin




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	6	100.0	0.0	0.0	1.97	6.9	e

Rheumatoid factor



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	5	100.0	0.0	0.0	31.0	6.1	e

Ceruloplasmin




MQ tolerance : 25 %

Ceruloplasmin (mg/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	4	100.0	0.0	0.0	255.0	10.0	e*

Präalbumin




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	10	100.0	0.0	0.0	234.1	3.7	e

Soluble transferrin receptor




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	6	83.3	16.7	0.0	3.9	19.0	e*

CRP



No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 AFIAS	54	90.7	1.9	7.4	39.8	8.9	e

Albumine




QUALAB tolerance : 12 %

Albumine (g/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	6	100.0	0.0	0.0	40	2.0	e
2	Cobas	15	100.0	0.0	0.0	42	2.9	e
3	Fuji Dri-Chem	214	96.2	1.9	1.9	53	4.7	e
4	Spotchem/Ready	31	90.3	6.5	3.2	39	6.5	e
5	Spotchem D-Concept	127	94.5	3.9	1.6	47	5.5	e
6	Piccolo	45	95.6	0.0	4.4	44	3.7	e
7	Beckmann	14	92.9	7.1	0.0	39	6.0	e*
8	Skyla	6	100.0	0.0	0.0	39	3.9	e*
9	Abx Mira	6	100.0	0.0	0.0	41	3.0	e
10	Hitachi S40/M40	9	100.0	0.0	0.0	40	2.7	e
11	Autolyser/DiaSys	7	100.0	0.0	0.0	40	3.8	e

Alkaline phosphatase



No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC	4	100.0	0.0	0.0	186	6.3	e*
2 Cobas	17	100.0	0.0	0.0	166	4.1	e
3 Reflotron	566	96.4	2.7	0.9	272	8.1	e
4 Fuji Dri-Chem	760	99.2	0.1	0.7	216	5.4	e
5 Spotchem/Ready	71	90.1	8.5	1.4	329	10.1	e
6 Spotchem D-Concept	232	95.7	3.4	0.9	272	8.3	e
7 Hitachi S40/M40	16	93.7	0.0	6.3	149	5.1	e
8 Beckman	19	94.7	5.3	0.0	226	7.8	e
9 Piccolo	38	97.4	0.0	2.6	237	4.4	e
10 Abx Mira	9	100.0	0.0	0.0	208	9.1	e*
11 Autolyser/DiaSys	17	100.0	0.0	0.0	161	5.7	e

Amylase




QUALAB tolerance : 18 %

Amylase (U/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC	15	100.0	0.0	0.0	86	3.8	e
2 Cobas	5	100.0	0.0	0.0	85	3.2	e
3 Reflotron	154	99.4	0.6	0.0	76	5.4	e
4 Fuji Dri-Chem	551	99.4	0.2	0.4	78	3.7	e
5 Spotchem/Ready	48	95.8	2.1	2.1	95	7.6	e
6 Spotchem D-Concept	178	99.4	0.6	0.0	92	5.8	e
7 Piccolo	37	97.3	0.0	2.7	73	3.8	e
8 Abx Mira	4	100.0	0.0	0.0	78	1.3	e
9 Hitachi S40/M40	11	100.0	0.0	0.0	92	4.5	e
10 Autolyser/DiaSys	5	100.0	0.0	0.0	74	3.1	e

Pancreatic amylase




QUALAB tolerance : 18 %

Pancreatic amylase (U/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC	21	100.0	0.0	0.0	61	4.1	e
2 Cobas	10	100.0	0.0	0.0	64	3.0	e
3 Reflotron	376	98.2	0.5	1.3	59	5.5	e
4 Autolyser/DiaSys	10	100.0	0.0	0.0	61	4.9	e

Bilirubin




QUALAB tolerance : 18 %

Bilirubin (µmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	7	85.7	14.3	0.0	27.0	9.9	e*
2	Cobas	16	100.0	0.0	0.0	23.8	4.0	e
3	Reflotron	408	96.0	1.5	2.5	23.5	7.4	e
4	Fuji Dri-Chem	601	98.0	0.7	1.3	25.6	6.2	e
5	Spotchem/Ready	63	95.2	4.8	0.0	29.9	7.8	e
6	Spotchem D-Concept	188	97.8	1.1	1.1	24.5	5.3	e
7	Beckman	14	100.0	0.0	0.0	28.0	2.3	e
8	Piccolo	43	97.7	0.0	2.3	26.7	8.3	e
9	Skyla	4	75.0	25.0	0.0	31.7	10.3	e*
10	Abx Mira	10	100.0	0.0	0.0	24.0	9.0	e*
11	Hitachi S40/M40	12	100.0	0.0	0.0	26.0	6.8	e
12	Autolyser/DiaSys	15	100.0	0.0	0.0	23.0	6.6	e

Bilirubin direct




MQ tolerance : 18 %

Bilirubin direct (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Fuji Dri-Chem	27	74.1	7.4	18.5	11.3	11.3	e*

Calcium




QUALAB tolerance : 12 %
(< 2.00: +/- 0.24 mmol/l)

Calcium (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	29	100.0	0.0	0.0	2.15	2.2	e
2	Cobas	16	100.0	0.0	0.0	2.13	1.7	e
3	Fuji Dri-Chem	370	97.8	0.8	1.4	2.11	4.5	e
4	Spotchem/Ready	20	95.0	5.0	0.0	1.74	5.5	e
5	Spotchem D-Concept	89	97.8	2.2	0.0	1.79	5.4	e
6	Piccolo	41	97.6	2.4	0.0	2.14	3.8	e
7	Abx Mira	6	100.0	0.0	0.0	2.22	5.2	e*
8	Hitachi S40/M40	10	100.0	0.0	0.0	2.06	2.7	e
9	Autolyser/DiaSys	9	100.0	0.0	0.0	2.20	2.3	e

Calcium ISE




MQ tolerance : 12 %

Calcium ISE (mmol/l)


No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 ISE	4	100.0	0.0	0.0	0.88	3.6	e*
2 iStat Chem8	7	100.0	0.0	0.0	0.82	3.9	e*

Chloride



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ISE	27	100.0	0.0	0.0	94	1.1	e
2	Cobas	7	100.0	0.0	0.0	91	0.7	e
3	Fuji Dri-Chem	703	97.7	1.3	1.0	95	2.0	e
4	Spotchem D-Concept	208	97.2	1.4	1.4	106	2.2	e
5	Spotchem EL-SE 1520	88	92.0	5.7	2.3	105	2.9	e
6	Piccolo	19	94.7	0.0	5.3	93	2.3	e
7	iStat Chem8	6	100.0	0.0	0.0	93	1.1	e

Cholesterol total




QUALAB tolerance : 10 %

Cholesterol total (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	26	100.0	0.0	0.0	4.24	3.3	e
2	Cobas	17	100.0	0.0	0.0	4.04	2.3	e
3	Reflotron	551	98.4	1.1	0.5	4.07	3.4	e
4	Fuji Dri-Chem	761	97.3	1.4	1.3	4.18	4.3	e
5	Spotchem/Ready	98	95.9	1.0	3.1	4.15	4.6	e
6	Spotchem D-Concept	236	98.8	0.8	0.4	4.21	3.8	e
7	Piccolo	20	90.0	5.0	5.0	4.18	3.2	e
8	Cholestech LDX	145	94.4	2.8	2.8	4.33	4.2	e
9	Abx Mira	9	88.9	11.1	0.0	4.33	6.6	e*
10	Hitachi S40/M40	15	100.0	0.0	0.0	4.14	1.9	e
11	Autolyser/DiaSys	16	93.7	6.3	0.0	4.18	4.7	e
12	Other methods	4	100.0	0.0	0.0	3.76	4.4	e*

Cholesterin HDL




QUALAB tolerance : 21 %

Cholesterin HDL (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Pentra/Selectra	13	100.0	0.0	0.0	1.52	6.4	e
2	Wet chemistry, direc	16	100.0	0.0	0.0	1.58	3.7	e
3	Cobas	16	100.0	0.0	0.0	1.76	3.7	e
4	Reflotron	406	83.0	11.1	5.9	1.38	12.8	e
5	Fuji Dri-Chem	729	99.0	0.3	0.7	1.87	3.5	e
6	Spotchem/Ready	86	87.2	9.3	3.5	1.43	11.7	e
7	Spotchem D-Concept	231	98.3	0.4	1.3	1.80	7.6	e
8	Piccolo	19	78.9	5.3	15.8	1.58	9.4	e
9	Cholestech LDX	144	97.9	2.1	0.0	1.75	7.0	e
10	Hitachi S40/M40	14	100.0	0.0	0.0	1.88	5.2	e
11	Autolyser/DiaSys	16	100.0	0.0	0.0	1.75	6.3	e

Creatine kinase




QUALAB tolerance : 18 %

Creatine kinase (U/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	IFCC	28	96.4	0.0	3.6	251	4.7	e
2	Cobas	14	100.0	0.0	0.0	239	2.4	e
3	Reflotron	363	87.3	7.2	5.5	206	9.8	e
4	Fuji Dri-Chem	490	99.0	0.0	1.0	255	4.0	e
5	Spotchem/Ready	40	100.0	0.0	0.0	197	7.4	e
6	Spotchem D-Concept	150	98.6	0.7	0.7	190	5.7	e
7	Piccolo	17	100.0	0.0	0.0	251	2.2	e
8	Abx Mira	6	100.0	0.0	0.0	253	3.3	e
9	Hitachi S40/M40	8	75.0	12.5	12.5	196	10.2	e*
10	Autolyser/DiaSys	13	92.3	0.0	7.7	251	3.1	e

LDL Cholesterin




MQ tolerance : 25 %

LDL Cholesterin (mmol/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	5	100.0	0.0	0.0	2.2	5.0	e
2	Roche, Cobas	6	100.0	0.0	0.0	2.2	3.2	e
3	Hitachi S40/M40	8	100.0	0.0	0.0	1.7	3.6	e
4	Autolyser/DiaSys	13	92.3	7.7	0.0	1.8	10.8	e
5	Beckman	11	100.0	0.0	0.0	2.4	3.0	e

Iron



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	16	100.0	0.0	0.0	19	1.8	e
2	Cobas	8	100.0	0.0	0.0	19	4.5	e
3	Abx Mira	5	100.0	0.0	0.0	18	4.4	e

Gamma-glutamyltransferase




QUALAB tolerance : 18 %

Gamma-glutamyltransferase (U/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	IFCC	4	100.0	0.0	0.0	58	9.4	e*
2	Cobas	17	100.0	0.0	0.0	48	6.3	e
3	Reflotron	727	97.1	1.4	1.5	43	6.4	e
4	Fuji Dri-Chem	838	98.0	0.7	1.3	44	6.3	e
5	Spotchem/Ready	105	97.1	1.0	1.9	51	7.2	e
6	Spotchem D-Concept	258	99.2	0.0	0.8	44	5.7	e
7	Selectra/Biolis	6	100.0	0.0	0.0	55	6.6	e*
8	Architect	4	100.0	0.0	0.0	45	1.1	e
9	Dimension	13	100.0	0.0	0.0	52	5.5	e
10	IFCC Beckmann	7	100.0	0.0	0.0	51	1.9	e
11	Piccolo	35	94.3	0.0	5.7	47	3.7	e
12	Abx Mira	5	100.0	0.0	0.0	50	10.9	e*
13	Hitachi S40/M40	16	93.7	0.0	6.3	59	4.2	e
14	Autolysers/DiaSys	17	94.1	0.0	5.9	53	3.6	e

Glucose




QUALAB tolerance : 10 %

Glucose (mmol/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	36	97.2	0.0	2.8	6.2	3.2	e
2	Cobas	15	100.0	0.0	0.0	6.1	3.1	e
3	Reflotron	726	96.1	2.5	1.4	6.4	4.4	e
4	Fuji Dri-Chem	791	97.9	1.1	1.0	6.2	3.1	e
5	Spotchem/Ready	91	98.9	1.1	0.0	6.4	4.4	e
6	Spotchem D-Concept	242	98.0	1.2	0.8	6.1	4.0	e
7	Piccolo	49	95.9	0.0	4.1	6.2	1.8	e
8	Cholestech LDX	121	95.0	3.3	1.7	6.1	4.9	e
9	Abx Mira	9	88.9	11.1	0.0	6.2	7.0	e*
10	Hitachi S40/M40	17	94.1	5.9	0.0	6.2	5.1	e*
11	Autolyser/DiaSys	17	100.0	0.0	0.0	6.0	3.5	e
12	iStat Chem8	7	100.0	0.0	0.0	5.4	0.9	e

Glucose




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Accu-Chek Aviva	336	93.4	3.9	2.7	6.7	3.7	e
2	Accu-Chek Inform 2	380	97.9	2.1	0.0	6.8	3.6	e
3	Accu-Check Guide	110	98.2	0.9	0.9	5.4	3.3	e
4	Contour XT	1100	95.2	3.5	1.3	6.4	4.8	e
5	Skylla	4	100.0	0.0	0.0	6.6	1.5	e
6	Glucocard	17	100.0	0.0	0.0	7.8	4.8	e
7	Hemocue 201+ P-equiv	97	92.8	3.1	4.1	7.6	3.7	e
8	Hemocue 201RT P-equiv	74	94.6	2.7	2.7	7.6	3.6	e
9	FreeStyle Precision	7	100.0	0.0	0.0	6.5	5.7	e*
10	Freestyle Freedom li	6	100.0	0.0	0.0	6.4	2.4	e
11	Glucomen Lx	4	0.0	0.0	100.0	8.1	0.0	e
12	Sanofi BG Star	4	75.0	25.0	0.0	7.8	6.3	e*
13	Contour NEXT ONE	5	80.0	0.0	20.0	6.0	3.4	e*

Glucose B




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Hemocue 201+ (alt)	48	91.6	4.2	4.2	7.4	5.2	e
2	OneTouch Verio	31	100.0	0.0	0.0	5.9	2.4	e
3	Contour 2 (5s)	31	80.6	12.9	6.5	5.2	6.5	e
4	Contour (15s)	5	100.0	0.0	0.0	6.8	3.8	e*
5	Healthpro	39	89.7	0.0	10.3	9.8	5.2	e
6	Mylife UNIO	260	97.3	2.7	0.0	6.5	4.4	e
7	mylife Pura	69	97.1	2.9	0.0	6.2	3.2	e
8	Omnitest	18	94.4	0.0	5.6	8.3	3.9	e
9	Alpha Check	7	100.0	0.0	0.0	8.9	5.5	e*

Uric Acid



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	28	96.4	3.6	0.0	341	3.7	e
2	Cobas	13	100.0	0.0	0.0	329	2.0	e
3	Reflotron	639	97.5	1.7	0.8	360	4.4	e
4	Fuji Dri-Chem	789	98.6	0.6	0.8	349	3.4	e
5	Spotchem/Ready	83	97.6	1.2	1.2	335	4.4	e
6	Spotchem D-Concept	241	99.6	0.0	0.4	352	3.6	e
7	Piccolo	27	96.3	0.0	3.7	279	3.9	e
8	Skyla	5	100.0	0.0	0.0	324	3.2	e*
9	Abx Mira	8	87.5	12.5	0.0	326	6.2	e*
10	Hitachi S40/M40	15	100.0	0.0	0.0	334	2.6	e
11	Autolyser/DiaSys	15	93.3	6.7	0.0	339	5.5	e

Urea




QUALAB tolerance : 15 %

Urea (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	26	100.0	0.0	0.0	7.3	3.5	e
2	Cobas	15	100.0	0.0	0.0	7.1	3.6	e
3	Reflotron	290	95.1	2.8	2.1	7.6	5.9	e
4	Fuji Dri-Chem	470	98.9	0.0	1.1	7.6	2.3	e
5	Spotchem/Ready	54	85.2	11.1	3.7	6.0	9.0	e
6	Spotchem D-Concept	151	77.5	10.6	11.9	5.3	9.2	e
7	Piccolo	45	95.6	0.0	4.4	6.6	3.0	e
8	Skyla	5	100.0	0.0	0.0	6.6	4.4	e*
9	Abx Mira	7	100.0	0.0	0.0	7.3	6.9	e*
10	Hitachi S40/M40	12	100.0	0.0	0.0	7.2	2.1	e
11	Autolyser/DiaSys	9	100.0	0.0	0.0	7.7	5.0	e
12	iStat Chem8	7	100.0	0.0	0.0	8.4	3.5	e

Potassium




QUALAB tolerance : 6 %

Potassium (mmol/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ISE	45	100.0	0.0	0.0	4.03	2.2	e
2	Cobas	17	100.0	0.0	0.0	4.07	1.3	e
3	Reflotron	654	85.4	8.3	6.3	4.11	3.5	e
4	Fuji Dri-Chem	830	97.5	1.3	1.2	4.10	1.7	e
5	Spotchem D-Concept	242	96.7	2.5	0.8	4.06	2.0	e
6	Spotchem EL-SE 1520	91	96.7	2.2	1.1	4.06	1.8	e
7	Piccolo	36	69.4	13.9	16.7	4.08	4.5	e*
8	iStat Chem8	9	100.0	0.0	0.0	4.00	1.3	e

Creatinine



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	21	95.2	0.0	4.8	120	5.7	e
2	Cobas	17	100.0	0.0	0.0	122	4.3	e
3	Reflotron	837	97.8	1.1	1.1	129	6.0	e
4	Fuji Dri-Chem	863	97.6	1.0	1.4	104	6.6	e
5	Spotchem/Ready	114	98.2	0.0	1.8	101	4.0	e
6	Spotchem D-Concept	257	99.2	0.4	0.4	103	3.9	e
7	Enzymatic	6	100.0	0.0	0.0	119	4.0	e
8	Piccolo	47	95.7	4.3	0.0	122	8.5	e
9	Abx Mira	11	90.9	9.1	0.0	120	9.9	e*
10	Skyla	5	100.0	0.0	0.0	115	11.3	e*
11	Hitachi S40/M40	16	93.7	0.0	6.3	108	8.6	e
12	Autolyser/DiaSys	17	100.0	0.0	0.0	118	4.9	e

Creatinine E




QUALAB tolerance : 18 %

Creatinine E (µmol/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	iStat Chem8	11	90.9	0.0	9.1	114	3.5	e
2	EPOC	5	100.0	0.0	0.0	99	12.6	e*
3	ABL700/800	8	100.0	0.0	0.0	118	5.5	e

eGFR CKD-EPI




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	64	96.8	1.6	1.6	56	11.1	e
2	Spotchem/Ready	135	97.8	0.0	2.2	68	5.7	e
3	Reflotron	290	98.3	0.7	1.0	51	8.5	e
4	Fuji Dri-Chem	347	93.6	2.9	3.5	68	10.7	e

eGFR Cockcroft-Gault




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Spotchem/Ready	16	93.7	0.0	6.3	63	7.8	e
2	Reflotron	20	95.0	0.0	5.0	49	6.7	e
3	Fuji Dri-Chem	30	93.4	3.3	3.3	61	9.8	e

eGFR MDRD




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Reflotron	8	100.0	0.0	0.0	52	11.1	e*
2	Fuji Dri-Chem	5	80.0	0.0	20.0	69	10.7	e*

LDH



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	IFCC	30	100.0	0.0	0.0	197	4.3	e
2	Cobas	10	100.0	0.0	0.0	377	2.2	e
3	Fuji Dri-Chem	151	96.7	0.0	3.3	168	4.4	e
4	Spotchem/Ready	15	93.3	0.0	6.7	173	5.9	e
5	Spotchem D-Concept	45	93.4	2.2	4.4	170	8.0	e
6	Abx Mira	7	100.0	0.0	0.0	216	5.0	e
7	Hitachi S40/M40	6	100.0	0.0	0.0	197	4.9	e
8	Autolyser/DiaSys	8	100.0	0.0	0.0	215	7.6	e*

Magnesium




QUALAB tolerance : 12 %
 (< 0.70: +/- 0.09 mmol/l)

Magnesium (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	13	100.0	0.0	0.0	0.82	3.9	e
2	Cobas	11	100.0	0.0	0.0	0.86	2.4	e
3	Fuji Dri-Chem	114	94.7	4.4	0.9	0.75	4.9	e
4	Spotchem D-Concept	47	97.9	2.1	0.0	0.60	6.0	e
5	Spotchem/Ready	6	83.3	16.7	0.0	0.67	8.9	e*
6	Beckman	8	100.0	0.0	0.0	0.87	4.9	e*
7	Piccolo	5	100.0	0.0	0.0	0.80	5.6	e*
8	Abx Mira	4	75.0	0.0	25.0	0.92	5.0	e*

Sodium




QUALAB tolerance : 6 %

Sodium (mmol/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ISE	43	100.0	0.0	0.0	140	1.4	e
2	Cobas	17	100.0	0.0	0.0	139	1.2	e
3	Fuji Dri-Chem	770	98.3	1.2	0.5	141	1.4	e
4	Spotchem D-Concept	231	99.2	0.4	0.4	140	1.4	e
5	Spotchem EL-SE 1520	89	96.7	2.2	1.1	137	1.9	e
6	Piccolo	36	100.0	0.0	0.0	134	1.9	e
7	iStat Chem8	7	100.0	0.0	0.0	139	0.3	e

Phosphate




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	20	100.0	0.0	0.0	1.4	3.3	e
2	Cobas	12	100.0	0.0	0.0	1.4	3.7	e
3	Fuji Dri-Chem	83	100.0	0.0	0.0	1.5	3.7	e
4	Spotchem D-Concept	21	95.2	0.0	4.8	1.4	3.4	e
5	Spotchem/Ready	6	100.0	0.0	0.0	1.3	6.1	e*

Protein total



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	23	100.0	0.0	0.0	56.7	2.4	e
2	Cobas	13	100.0	0.0	0.0	54.4	3.0	e
3	Fuji Dri-Chem	183	98.4	0.5	1.1	57.3	2.9	e
4	Spotchem/Ready	27	100.0	0.0	0.0	58.6	6.1	e
5	Spotchem D-Concept	100	92.0	6.0	2.0	60.8	7.2	e
6	Piccolo	32	100.0	0.0	0.0	57.0	1.9	e
7	Skyla	5	100.0	0.0	0.0	58.0	1.9	e
8	Abx Mira	6	100.0	0.0	0.0	55.5	5.1	e*
9	Hitachi S40/M40	6	100.0	0.0	0.0	58.5	5.0	e*

Aspartate aminotransferase




QUALAB tolerance : 18 %

Aspartate aminotransferase (U/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC with PP	30	100.0	0.0	0.0	46	7.4	e
2 Cobas	11	100.0	0.0	0.0	46	8.2	e*
3 Reflotron	733	98.0	1.0	1.0	47	7.0	e
4 Fuji Dri-Chem	841	98.7	0.2	1.1	40	3.6	e
5 Spotchem/Ready	108	99.1	0.0	0.9	36	5.5	e
6 Spotchem D-Concept	254	97.6	1.2	1.2	34	6.9	e
7 IFCC without PP	6	100.0	0.0	0.0	40	2.8	e
8 Piccolo	48	97.9	0.0	2.1	45	4.6	e
9 Skyla	5	80.0	20.0	0.0	46	12.6	e*
10 Abx Mira	10	100.0	0.0	0.0	42	6.5	e
11 Hitachi S40/M40	18	100.0	0.0	0.0	39	3.2	e
12 Autolyser/DiaSys	17	100.0	0.0	0.0	40	6.2	e

Alanine aminotransferase




QUALAB tolerance : 18 %

Alanine aminotransferase (U/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	IFCC with PP	28	100.0	0.0	0.0	41	5.2	e
2	Cobas	18	100.0	0.0	0.0	39	4.0	e
3	Reflotron	762	98.3	1.2	0.5	35	6.0	e
4	Fuji Dri-Chem	855	98.3	0.4	1.3	40	5.1	e
5	Spotchem/Ready	112	93.7	4.5	1.8	36	7.4	e
6	Spotchem D-Concept	260	97.3	1.5	1.2	31	7.1	e
7	Piccolo	49	95.9	0.0	4.1	40	5.8	e
8	Skyla	6	66.6	16.7	16.7	40	11.8	e*
9	Abx Mira	9	88.9	0.0	11.1	40	4.9	e
10	Hitachi S40/M40	18	100.0	0.0	0.0	38	5.4	e
11	Autolyser/DiaSys	17	100.0	0.0	0.0	39	4.1	e

Triglycerides




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	26	100.0	0.0	0.0	1.18	4.6	e
2	Cobas	19	94.7	0.0	5.3	1.22	5.5	e
3	Reflotron	474	96.2	1.1	2.7	1.66	6.1	e
4	Fuji Dri-Chem	744	98.1	0.8	1.1	1.15	5.2	e
5	Spotchem/Ready	90	95.6	2.2	2.2	1.11	8.1	e
6	Spotchem D-Concept	230	99.2	0.4	0.4	1.24	5.0	e
7	Hitachi S40/M40	15	100.0	0.0	0.0	0.74	5.2	e
8	Piccolo	18	88.8	5.6	5.6	1.34	6.2	e
9	Cholestech LDX	144	98.6	0.0	1.4	1.25	5.6	e
10	Abx Mira	9	100.0	0.0	0.0	1.17	8.0	e*
11	Autolyser/DiaSys	16	93.7	6.3	0.0	1.15	10.4	e*

Lithium



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	16	100.0	0.0	0.0	1.03	4.5	e

Lactate




MQ tolerance : 18 %

Lactate (mmol/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	12	91.7	0.0	8.3	1.44	5.3	e

HbA1c sample A



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Roche, Cobas	13	100.0	0.0	0.0	5.8	2.0	a
2	HPLC	8	100.0	0.0	0.0	5.8	3.0	a
3	Afinion	715	99.2	0.4	0.4	5.7	2.4	e
4	Cobas b101	60	100.0	0.0	0.0	5.7	2.5	e
5	DCA2000/Vantage	195	97.9	2.1	0.0	5.6	3.3	e
6	Celltac chemi	8	75.0	25.0	0.0	5.8	3.9	a
7	NycoCard	55	90.9	1.8	7.3	5.8	4.6	a
8	Eurolyser	9	88.9	0.0	11.1	5.8	2.2	a
9	Hemocue HbA1c 501	9	66.7	11.1	22.2	6.3	6.4	e*
10	AFIAS	18	100.0	0.0	0.0	5.8	4.6	a
11	Others	18	100.0	0.0	0.0	5.8	4.5	a


HbA1c sample B




QUALAB tolerance : 9 %

HbA1c sample B (%)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Roche, Cobas	14	100.0	0.0	0.0	11.8	2.1	a
2	HPLC	8	100.0	0.0	0.0	11.8	2.6	a
3	Afinion	665	98.9	0.2	0.9	11.7	2.1	e
4	Cobas b101	44	100.0	0.0	0.0	10.8	2.2	e
5	DCA2000/Vantage	210	89.0	8.6	2.4	11.4	5.2	e
6	Celltac chemi	5	80.0	0.0	20.0	11.8	3.2	a
7	NycoCard	32	87.4	6.3	6.3	10.9	5.1	e
8	Eurolyser	15	86.7	0.0	13.3	11.8	6.5	a
9	Hemocue HbA1c 501	4	50.0	25.0	25.0	11.8	7.4	a
10	AFIAS	29	93.1	6.9	0.0	10.7	4.2	e
11	Others	16	93.7	0.0	6.3	11.8	4.3	a


pCO₂

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL80 FLEX	9	100.0	0.0	0.0	5.41	2.8	e
2	GEM	5	100.0	0.0	0.0	5.80	4.0	e
3	Cobas	21	95.2	4.8	0.0	5.68	5.4	e
4	iStat	39	94.9	0.0	5.1	4.97	3.8	e
5	EPOC	42	92.8	4.8	2.4	5.18	6.2	e
6	ABL700/800	75	96.0	0.0	4.0	5.35	3.0	e
7	ABL90 FLEX / PLUS	49	100.0	0.0	0.0	5.56	2.3	e
8	ABL80 FLEX CO-OX / O	15	93.3	0.0	6.7	5.77	5.2	e

pO₂


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL80 FLEX	9	100.0	0.0	0.0	14.30	6.9	e*
2	GEM	5	100.0	0.0	0.0	14.10	2.4	e
3	Cobas b221	5	100.0	0.0	0.0	15.69	8.2	e*
4	Cobas b121/123	13	92.3	0.0	7.7	13.60	8.3	e*
5	iStat	39	87.2	7.7	5.1	14.89	7.4	e
6	EPOC	42	88.1	7.1	4.8	13.30	8.2	e
7	ABL700/800	74	93.2	1.4	5.4	14.59	5.1	e
8	ABL90 FLEX / PLUS	51	92.2	0.0	7.8	13.13	4.0	e
9	ABL80 FLEX CO-OX / O	15	80.0	6.7	13.3	13.97	6.6	e

pH




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL80 FLEX	9	100.0	0.0	0.0	7.41	0.1	e
2	GEM	5	100.0	0.0	0.0	7.41	0.1	e
3	Cobas	20	100.0	0.0	0.0	7.41	0.2	e
4	iStat	40	100.0	0.0	0.0	7.44	0.1	e
5	EPOC	41	97.6	0.0	2.4	7.36	0.1	e
6	ABL700/800	75	98.7	0.0	1.3	7.40	0.1	e
7	ABL90 FLEX / PLUS	51	100.0	0.0	0.0	7.40	0.1	e
8	ABL80 FLEX CO-OX / O	15	100.0	0.0	0.0	7.39	0.2	e

Glucose BG




No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Cobas	10	100.0	0.0	0.0	5.6	2.3	e
2 iStat	9	100.0	0.0	0.0	5.2	1.2	e
3 EPOC	29	100.0	0.0	0.0	5.5	2.8	e
4 ABL700/800	64	98.4	0.0	1.6	5.7	2.5	e
5 ABL90 FLEX / PLUS	49	100.0	0.0	0.0	5.6	1.9	e

Hemoglobin BG




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	66	90.9	0.0	9.1	128.5	1.7	e
2	ABL90 FLEX / PLUS	47	100.0	0.0	0.0	128.9	0.5	e
3	ABL80 FLEX CO-OX / O	12	100.0	0.0	0.0	126.7	1.7	e

Potassium BG




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas	22	100.0	0.0	0.0	3.7	2.2	e
2	iStat	17	100.0	0.0	0.0	3.7	1.3	e
3	EPOC	35	100.0	0.0	0.0	3.7	1.6	e
4	ABL700/800	66	98.5	0.0	1.5	3.8	1.3	e
5	ABL90 FLEX / PLUS	50	100.0	0.0	0.0	3.7	1.3	e
6	ABL80 FLEX CO-OX / O	6	100.0	0.0	0.0	3.9	1.3	e

Sodium BG




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas	22	100.0	0.0	0.0	138.7	1.7	e
2	iStat	17	94.1	5.9	0.0	142.9	2.0	e
3	EPOC	33	100.0	0.0	0.0	139.9	0.9	e
4	ABL700/800	64	98.4	0.0	1.6	140.1	1.0	e
5	ABL90 FLEX / PLUS	51	100.0	0.0	0.0	141.1	0.3	e
6	ABL80 FLEX CO-OX / O	6	100.0	0.0	0.0	142.0	0.7	e

Chlorid-BG



No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Cobas	9	100.0	0.0	0.0	99.9	2.6	e*
2 ABL700/800	56	100.0	0.0	0.0	97.4	1.4	e
3 ABL90 FLEX / PLUS	50	100.0	0.0	0.0	97.2	0.4	e
4 ABL80 FLEX CO-OX / O	4	100.0	0.0	0.0	96.0	1.3	e

Calcium-BG




MQ tolerance : 12 %
(< 2.00: +/- 0.24 mmol/l)

Calcium-BG (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b123	8	100.0	0.0	0.0	0.39	12.0	e*
2	ABL80 FLEX	4	100.0	0.0	0.0	0.49	7.2	e*
3	GEM	4	100.0	0.0	0.0	0.45	1.8	e
4	Cobas	13	100.0	0.0	0.0	0.47	11.7	e*
5	iStat	9	100.0	0.0	0.0	0.47	0.7	e
6	EPOC	32	100.0	0.0	0.0	0.48	3.5	e
7	ABL700/800	65	100.0	0.0	0.0	0.53	5.4	e
8	ABL90 FLEX / PLUS	51	100.0	0.0	0.0	0.52	0.9	e
9	ABL80 FLEX CO-OX / O	5	100.0	0.0	0.0	0.53	10.0	e*

FHHb




MQ tolerance : 20 %

FHHb (%)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL80 FLEX CO-OX / O	6	100.0	0.0	0.0	2.650	7.9	e*

Lactate-BG




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b123	7	100.0	0.0	0.0	2.00	4.8	e
2	Cobas	5	100.0	0.0	0.0	2.00	0.0	e
3	IL	4	100.0	0.0	0.0	1.65	6.3	e*
4	EPOC	36	100.0	0.0	0.0	1.57	3.3	e
5	iStat	11	100.0	0.0	0.0	1.57	1.3	e
6	ABL700/800	68	100.0	0.0	0.0	1.72	3.3	e
7	ABL90 FLEX / PLUS	51	100.0	0.0	0.0	1.61	2.6	e

sO2 OR



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	iStat	9	100.0	0.0	0.0	99.000	0.5	e
2	ABL700/800	51	98.0	2.0	0.0	96.529	3.9	e
3	ABL90 FLEX / PLUS	44	100.0	0.0	0.0	96.936	0.1	e
4	ABL80 FLEX CO-OX / O	11	100.0	0.0	0.0	97.009	0.3	e

FO2Hb OR




MQ tolerance : 20 %

FO2Hb OR (%)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	48	97.9	0.0	2.1	92.130	0.1	e
2	ABL90 FLEX / PLUS	47	100.0	0.0	0.0	92.145	0.1	e
3	ABL80 FLEX CO-OX / O	14	100.0	0.0	0.0	92.136	0.1	e

FCOHb OR




MQ tolerance : 20 %

FCOHb OR (%)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 ABL700/800	50	96.0	0.0	4.0	3.150	7.1	e
2 ABL90 FLEX / PLUS	45	100.0	0.0	0.0	2.942	4.8	e
3 ABL80 FLEX CO-OX / O	14	100.0	0.0	0.0	2.900	9.2	e*

FMetHb OR




MQ tolerance : 20 %

FMetHb OR (%)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 ABL700/800	55	96.4	1.8	1.8	1.963	5.4	e
2 ABL90 FLEX / PLUS	43	100.0	0.0	0.0	2.005	2.4	e
3 ABL80 FLEX CO-OX / O	14	100.0	0.0	0.0	2.007	5.0	e

FHbF OR




MQ tolerance : 20 %

FHbF OR (%)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL90 FLEX / PLUS	9	100.0	0.0	0.0	80.000	2.2	e

Bilirubin OR




QUALAB tolerance : 18 %

Bilirubin OR (µmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ABL700/800	4	100.0	0.0	0.0	298.5	0.4	e
2	ABL90 FLEX / PLUS	15	100.0	0.0	0.0	297.9	0.4	e

Troponin I




QUALAB tolerance : 24 %

Troponin I (ng/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Vidas	14	100.0	0.0	0.0	2388.8	8.2	e
2	Architect High Sensi	10	100.0	0.0	0.0	1640.7	2.3	e
3	AQT 90 FLEX	5	100.0	0.0	0.0	270.0	4.0	e

Troponin T




QUALAB tolerance : 24 %

Troponin T (ng/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas hs	4	100.0	0.0	0.0	474.15	4.8	e
2	Cobas hs STAT	8	100.0	0.0	0.0	754.50	1.5	e

Myoglobin




QUALAB tolerance : 30 %

Myoglobin (µg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Cobas E / Elecsys	5	100.0	0.0	0.0	77.8	5.9	e

CK-MB mass




MQ tolerance : 40 %

CK-MB mass (µg/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Architect	4	100.0	0.0	0.0	9.8	4.4	e
2	VIDAS	5	100.0	0.0	0.0	15.5	1.5	e

NT-proBNP




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	AQT 90 FLEX	4	100.0	0.0	0.0	404.5	4.0	e
2	VIDAS	10	100.0	0.0	0.0	156.1	6.9	e
3	Other methods	6	83.3	16.7	0.0	308.0	19.0	e*
4	Cobas E / Elecsys	12	100.0	0.0	0.0	191.7	13.0	e*

TSH




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	12	100.0	0.0	0.0	4.39	2.7	e
2	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	3.86	1.5	e
3	Architect	12	100.0	0.0	0.0	3.54	3.3	e
4	VIDAS	14	100.0	0.0	0.0	4.54	4.3	e
5	AFIAS	28	96.4	0.0	3.6	4.87	3.8	e

T3




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	AFIAS	9	100.0	0.0	0.0	1.4	10.3	e*

T4




No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 AFIAS	12	83.3	16.7	0.0	236	14.1	e*

FT3



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	12	100.0	0.0	0.0	8.1	3.6	e
2	Architect	10	100.0	0.0	0.0	6.5	5.7	e
3	VIDAS	7	85.7	14.3	0.0	6.4	7.2	e*


FT4



QUALAB tolerance : 20 %

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Cobas E / Elecsys	13	100.0	0.0	0.0	27.4	2.5	e
2 Architect	12	100.0	0.0	0.0	21.1	4.8	e
3 VIDAS	7	100.0	0.0	0.0	26.5	6.5	e*
4 Other methods	4	100.0	0.0	0.0	23.5	9.5	e*

Testosterone




QUALAB tolerance : 30 %

Testosterone (nmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas	5	100.0	0.0	0.0	14	2.5	e
2	Architect	5	100.0	0.0	0.0	14	5.0	e

Estradiol




MQ tolerance : 30 %

Estradiol (pmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas	6	100.0	0.0	0.0	673	7.0	e
2	Architect	6	100.0	0.0	0.0	631	2.0	e

SHBG




MQ tolerance : 30 %

SHBG (nmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Architect	4	100.0	0.0	0.0	31.7	9.7	e*

Cortisol




QUALAB tolerance : 20 %

Cortisol (nmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	6	100.0	0.0	0.0	456	8.2	e*
2	ADVIA Centaur XP/CP	4	75.0	25.0	0.0	530	14.0	e*
3	Architect	6	100.0	0.0	0.0	398	3.0	e

Progesteron




MQ tolerance : 30 %

Progesteron (nmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Architect	5	100.0	0.0	0.0	19.8	5.7	e

DHEAS




MQ tolerance : 30 %

DHEAS (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Architect	4	100.0	0.0	0.0	4.35	6.8	e

Luteinizing hormone




QUALAB tolerance : 24 %

Luteinizing hormone (U/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Roche, Cobas	5	100.0	0.0	0.0	17.0	5.1	e
2	Architect	7	100.0	0.0	0.0	12.2	3.2	e

Follicle-stimulating hormone




QUALAB tolerance : 24 %

Follicle-stimulating hormone (U/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Roche, Cobas	5	100.0	0.0	0.0	14.0	3.3	e
2	Architect	7	100.0	0.0	0.0	13.9	3.5	e

Prolactine




QUALAB tolerance : 24 %

Prolactine (µg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas/Roche	5	100.0	0.0	0.0	18.8	4.7	e
2	Architect	6	100.0	0.0	0.0	15.4	6.1	e

HGH




MQ tolerance : 25 %

HGH (µg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	6	100.0	0.0	0.0	9.21	5.4	e

Freies Testosteron




MQ tolerance : 25 %

Freies Testosteron (pmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	6	66.7	0.0	33.3	25.8	6.1	a

IGF-1




MQ tolerance : 25 %

IGF-1 (µg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Liaison	5	100.0	0.0	0.0	121	16.4	e*

Troponin T CR




QUALAB tolerance : 24 %

Troponin T CR (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas h 232	1085	97.5	2.1	0.4	60.29	9.6	e
2	Cardiac Reader	14	92.9	7.1	0.0	56.29	10.2	e

D-dimer CR




QUALAB tolerance : 21 %

D-dimer CR (mg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas h 232	1095	89.5	8.6	1.9	0.80	11.9	e
2	Cardiac Reader	13	100.0	0.0	0.0	0.81	9.1	e

CKMB- K8




MQ tolerance : 40 %

CKMB- K8 (µg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Cobas h 232	12	100.0	0.0	0.0	8.2	8.7	e

NT-proBNP CR




QUALAB tolerance : 27 %

NT-proBNP CR (ng/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas h 232	689	91.4	7.7	0.9	231	16.0	e
2	Cardiac Reader	5	100.0	0.0	0.0	261	10.9	e*

PCO2 CCA




QUALAB tolerance : 12 %

PCO2 CCA (kPa)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	OPTI CCA	13	92.3	7.7	0.0	9.12	6.1	e*

PO2 CCA




QUALAB tolerance : 15 %

PO2 CCA (kPa)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	OPTI CCA	13	92.3	0.0	7.7	9.71	5.6	e

pH CCA



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	OPTI CCA	13	100.0	0.0	0.0	7.16	0.2	e

Potassium CCA




QUALAB tolerance : 6 %
(< 3.3: +/- 0.2 mmol/l)

Potassium CCA (mmol/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	OPTI CCA	4	75.0	25.0	0.0	2.6	6.3	e*

Ferritin



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Beckman	11	100.0	0.0	0.0	42.02	7.7	e
2	all Participants	6	83.3	0.0	16.7	29.95	5.8	e
3	Cobas E / Elecsys	11	100.0	0.0	0.0	43.70	7.6	e
4	Architect	5	100.0	0.0	0.0	38.00	3.6	e
5	Mira/DiaSys	4	100.0	0.0	0.0	36.00	17.5	e*
6	Mini Vidas	7	100.0	0.0	0.0	34.96	6.2	e
7	AFIAS	33	75.8	0.0	24.2	24.11	8.2	e
8	Eurolyser	14	92.9	0.0	7.1	46.00	6.3	e

Vitamin B12




QUALAB tolerance : 21 %
 (< 200.00: +/- 42.00 pmol/l)

Vitamin B12 (pmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	150.00	9.5	e*
2	Cobas E / Elecsys	8	100.0	0.0	0.0	152.29	11.6	e*
3	Architect	11	100.0	0.0	0.0	139.71	9.1	e*

Folate




QUALAB tolerance : 24 %
 (< 10.00: +/- 2.40 nmol/l)

Folate (nmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	8	100.0	0.0	0.0	4.10	17.9	e*
2	Architect	11	100.0	0.0	0.0	3.82	13.8	e*

Holotranscobalamine




MQ tolerance : 30 %

Holotranscobalamine (pmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Architect	14	92.9	7.1	0.0	39.2	13.1	e

Bilirubin total Neo




QUALAB tolerance : 18 %

Bilirubin total Neo (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	14	100.0	0.0	0.0	187	4.7	e

Bilirubin direct




QUALAB tolerance : 18 %

Bilirubin direct (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	17	100.0	0.0	0.0	89	7.6	e

Bilirubin neonatal




QUALAB tolerance : 18 %

Bilirubin neonatal (µmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	6	100.0	0.0	0.0	246	1.1	e
2	ABL700/800	4	100.0	0.0	0.0	207	4.4	e*

PSA




QUALAB tolerance : 25 %

PSA (µg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	11	100.0	0.0	0.0	4.34	5.0	e
2	VIDAS	4	100.0	0.0	0.0	4.59	10.8	a
3	Architect	11	100.0	0.0	0.0	3.85	3.8	e
4	Qualigen	5	100.0	0.0	0.0	5.10	12.4	e*
5	AFIAS	25	96.0	0.0	4.0	4.59	8.4	e

free PSA




QUALAB tolerance : 25 %

free PSA (µg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	6	100.0	0.0	0.0	1.57	4.2	e
2	Architect	9	100.0	0.0	0.0	1.68	4.3	e


CEA



QUALAB tolerance : 25 %

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	7	100.0	0.0	0.0	24.3	2.0	e
2	Architect	9	100.0	0.0	0.0	39.1	3.8	e

CA 125




MQ tolerance : 25 %

CA 125 (kIU/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Architect	6	100.0	0.0	0.0	231.0	2.9	a

CA 19-9




MQ tolerance : 25 %

CA 19-9 (kIU/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Architect	4	100.0	0.0	0.0	466.0	4.8	a

CA 15-3




MQ tolerance : 25 %

CA 15-3 (kIU/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	4	100.0	0.0	0.0	46.3	2.9	a
2	Architect	6	100.0	0.0	0.0	49.1	6.1	e

AFP




QUALAB tolerance : 25 %

AFP (µg/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	5	100.0	0.0	0.0	11.2	4.7	e
2	Architect	5	100.0	0.0	0.0	11.6	1.9	e

HCG qn




No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas E / Elecsys	6	100.0	0.0	0.0	44.0	3.8	e
2	VIDAS	8	100.0	0.0	0.0	8.6	8.7	e*
3	Architect	6	100.0	0.0	0.0	40.1	4.9	e
4	AFIAS	7	100.0	0.0	0.0	45.6	10.6	a

CK-MB



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Fuji Dri-Chem	35	100.0	0.0	0.0	93.7	7.8	e

BNP




QUALAB tolerance : 27 %

BNP (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Architect	6	100.0	0.0	0.0	456.0	12.2	e*

Cholesterin PTS




QUALAB tolerance : 10 %

Cholesterin PTS (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	CardioChek	8	87.5	12.5	0.0	4.70	7.3	e*

Cholesterin HDL PTS




QUALAB tolerance : 21 %

Cholesterin HDL PTS (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	CardioChek	8	87.5	12.5	0.0	1.86	10.5	e*

Triglyceride PTS




QUALAB tolerance : 20 %

Triglyceride PTS (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	CardioChek	8	100.0	0.0	0.0	2.25	7.0	e*

C-Peptid




MQ tolerance : 25 %

C-Peptid (nmol/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Liaison	5	100.0	0.0	0.0	1.0	2.2	e

Procalcitonin



No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Cobas	6	100.0	0.0	0.0	2.46	2.8	e
2 VIDAS	20	100.0	0.0	0.0	2.02	6.9	e
3 Liaison	7	71.4	14.3	14.3	5.40	17.2	e*

EPO




MQ tolerance : 25 %

EPO (U/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	4	100.0	0.0	0.0	152.5	9.8	e*

Parathyroid hormone



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Architect	4	100.0	0.0	0.0	63.2	12.0	e*
2	Cobas PTH STAT	4	100.0	0.0	0.0	37.1	5.7	e
3	Cobas	6	100.0	0.0	0.0	31.5	5.0	e
4	all Participants	4	100.0	0.0	0.0	55.3	8.1	e*
5	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	74.2	8.7	e*

Osmolality




QUALAB tolerance : 6 %

Osmolality (mosm/kg)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Cryoscopy	14	100.0	0.0	0.0	349	1.8	e

Potassium-K22




QUALAB tolerance : 6 %

Potassium-K22 (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	ISE	9	100.0	0.0	0.0	3.9	1.6	e

Sodium-K22




QUALAB tolerance : 6 %

Sodium-K22 (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	ISE	9	100.0	0.0	0.0	142	1.3	e

Glucose-K22




QUALAB tolerance : 10 %

Glucose-K22 (mmol/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	9	100.0	0.0	0.0	7.2	2.5	e

Urea-K22



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	9	100.0	0.0	0.0	4.9	4.6	e

Osmotic Gap




MQ tolerance : 20 %

Osmotic Gap (mmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Formel 1 (2Na+K+Glu+	7	100.0	0.0	0.0	50.5	10.9	a

Digoxin




QUALAB tolerance : 24 %

Digoxin (nmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Other methods	12	100.0	0.0	0.0	2.17	8.1	e

Vancomycin




MQ tolerance : 25 %

Vancomycin (mg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	4	100.0	0.0	0.0	20	22.2	a

Valproat




MQ tolerance : 24 %

Valproat (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	6	100.0	0.0	0.0	530.0	8.7	e*

Cystatin C




MQ tolerance : 24 %

Cystatin C (mg/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	9	100.0	0.0	0.0	3.6	4.1	e

Ethanol




QUALAB tolerance : 18 %

Ethanol (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	20	95.0	0.0	5.0	51.1	4.4	e

Ammonia




MQ tolerance : 25 %

Ammonia (µmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	4	100.0	0.0	0.0	296.5	8.2	e*

Calprotectin




MQ tolerance : 40 %

Calprotectin (µg/g)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Liaison	21	95.2	0.0	4.8	92	15.9	e
2	Ridas Screen DS2	4	100.0	0.0	0.0	77	21.8	e*
3	Bühlmann fCALturbo	5	100.0	0.0	0.0	135	14.3	e*
4	Other methods	4	100.0	0.0	0.0	135	29.3	e*
5	Bühlmann	14	92.9	0.0	7.1	142	15.5	e

Cholesterol total Af/b101




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b101	81	100.0	0.0	0.0	4.74	2.1	e
2	Afinion	398	99.2	0.8	0.0	4.89	3.5	e

Cholesterol HDL Af/b101



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b101	81	93.9	1.2	4.9	0.90	5.6	e
2	Afinion	396	91.7	0.0	8.3	1.01	4.2	e

Tryglycerides Af/b101




QUALAB tolerance : 20 %

Tryglycerides Af/b101 (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas b101	79	97.5	0.0	2.5	1.81	3.3	e
2	Afinion	399	98.9	0.3	0.8	1.75	6.2	e

Troponin I S




QUALAB tolerance : 24 %

Troponin I S (ng/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Samsung LABGEO IB10	28	92.9	7.1	0.0	4456.43	13.3	e
2	AFIAS	67	59.7	22.4	17.9	6982.00	20.8	e

D-dimer qn S




QUALAB tolerance : 21 %

D-dimer qn S (mg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Samsung LABGEO IB10	36	86.1	8.3	5.6	1.16	11.2	e
2	AFIAS	69	94.2	0.0	5.8	1.53	7.5	e

NT-proBNP S




QUALAB tolerance : 27 %

NT-proBNP S (ng/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	AFIAS (Gen. 1)	22	54.6	22.7	22.7	956.1	21.3	e*
2	Samsung LABGEO IB10	26	88.5	11.5	0.0	996.0	17.0	e*
3	AFIAS	32	93.7	6.3	0.0	5552.5	13.5	e

Homocystein




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	5	100.0	0.0	0.0	33.4	5.8	e

Lipase




No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Architect	4	100.0	0.0	0.0	57.5	7.5	e*
2 Beckman	11	100.0	0.0	0.0	59.5	3.6	e
3 Cobas	10	100.0	0.0	0.0	58.0	8.0	e*
4 Fuji Dri-Chem	118	96.7	0.8	2.5	62.0	4.8	e

Cholinesterase



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	5	80.0	20.0	0.0	5.6	14.7	e*

Glucose CSF




MQ tolerance : 20 %

Glucose CSF (mmol/l)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Other methods	5	100.0	0.0	0.0	2.20	4.5	e

Lactate CSF




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Other methods	4	100.0	0.0	0.0	3.22	5.1	e*

Protein CSF




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Other methods	5	100.0	0.0	0.0	0.79	5.0	e

Tacrolimus



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	5	100.0	0.0	0.0	14.1	8.7	e*

Totalprotein E




MQ tolerance : 12 %

Totalprotein E (g/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	5	100.0	0.0	0.0	48.0	3.0	e

Albumin E




MQ tolerance : 12 %

Albumin E (%)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Elektrophorese	7	100.0	0.0	0.0	43.8	3.7	e

alpha-1-Globuline




MQ tolerance : 30 %

alpha-1-Globuline (%)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Elektrophorese	5	60.0	0.0	40.0	7.5	7.1	e

alpha-2-Globuline




MQ tolerance : 30 %

alpha-2-Globuline (%)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Elektrophorese	8	100.0	0.0	0.0	21.1	11.2	e*

beta-Globuline




MQ tolerance : 30 %

beta-Globuline (%)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Elektrophorese	7	71.4	28.6	0.0	10.8	21.7	e*

gamma-Globuline




MQ tolerance : 30 %

gamma-Globuline (%)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Elektrophorese	7	100.0	0.0	0.0	12.6	6.6	e

Folate in Erythrocytes




MQ tolerance : 25 %

Folate in Erythrocytes (nmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Beckman	4	100.0	0.0	0.0	1216	9.4	e*
2	Architect	6	100.0	0.0	0.0	798	15.3	a

Gallensäure




MQ tolerance : 30 %

Gallensäure (µmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	9	100.0	0.0	0.0	48.2	7.8	e

BNP




QUALAB tolerance : 27 %

BNP (ng/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Triage	33	90.9	3.0	6.1	596.3	14.0	e

Troponin Triage




QUALAB tolerance : 24 %

Troponin Triage (ng/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Triage Next Gen	32	40.6	18.8	40.6	750.53	19.9	e*
2	Triage SOB/Cardiac	14	92.9	0.0	7.1	17546.15	9.6	e

NT-proBNP




QUALAB tolerance : 27 %

NT-proBNP (ng/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Triage	14	100.0	0.0	0.0	1247	10.1	e

D-dimer Triage




QUALAB tolerance : 21 %

D-dimer Triage (ng/ml)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Triage	46	95.7	4.3	0.0	1070.00	11.0	e

CK-MB Triage




MQ tolerance : 40 %

CK-MB Triage (µg/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Triage	13	84.6	7.7	7.7	12.5	22.3	e*

Myoglobin Triage




QUALAB tolerance : 30 %

Myoglobin Triage (µg/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 Triage	10	100.0	0.0	0.0	120.0	14.6	e*

25-OH Vitamin D




QUALAB tolerance : 27 %

25-OH Vitamin D (nmol/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas	6	100.0	0.0	0.0	55.5	7.8	e
2	VIDAS	6	100.0	0.0	0.0	67.3	12.0	e*
3	Architect	11	100.0	0.0	0.0	52.7	2.9	e

AMH




MQ tolerance : 25 %

AMH (pmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	8	100.0	0.0	0.0	3.4	9.4	e*

Calcitonin




MQ tolerance : 25 %

Calcitonin (pmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Liaison	4	100.0	0.0	0.0	4.2	7.0	e*

TRAK



MQ tolerance : 25 %

TRAK (IE/ml)


No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	all Participants	6	100.0	0.0	0.0	1.76	13.2	e*

Creatinine WB



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Statsensor i / Nova	35	82.8	8.6	8.6	118	10.0	e

Amylase-Urine




MQ tolerance : 25 %

Amylase-Urine (U/l)

No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC	4	100.0	0.0	0.0	199	10.4	e*

Pancreatic Amylase-Urine




QUALAB tolerance : 18 %
(< 25.0: +/- 5.0 U/l)

Pancreatic Amylase-Urine (U/l)


No. Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1 IFCC	4	100.0	0.0	0.0	8.5	10.9	e*

Calcium-Urine




No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	18	94.4	0.0	5.6	2.39	4.3	e

Chloride-Urine



No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas	7	100.0	0.0	0.0	180	3.3	e
2	Standard chemistry	7	100.0	0.0	0.0	187	1.8	e

Glucose-Urine




MQ tolerance : 10 %

Glucose-Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	16	93.7	6.3	0.0	16.2	4.2	e

Magnesium-Urine




MQ tolerance : 12 %

Magnesium-Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	9	100.0	0.0	0.0	3.84	4.0	e

Osmolality-Urine




MQ tolerance : 10 %

Osmolality-Urine (mosm/kg)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Cryoskopy	9	100.0	0.0	0.0	795	0.3	e

Phosphate-Urine




MQ tolerance : 15 %

Phosphate-Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	16	100.0	0.0	0.0	17.2	2.7	e

Potassium-Urine




MQ tolerance : 15 %

Potassium-Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	24	95.8	0.0	4.2	69	3.1	e

total Protein-Urine




MQ tolerance : 20 %

total Protein-Urine (mg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Cobas/Roche	10	100.0	0.0	0.0	493.2	2.6	e
2	Standard chemistry	11	100.0	0.0	0.0	597.6	4.3	e

Sodium-Urine




MQ tolerance : 15 %

Sodium-Urine (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	all Participants	23	100.0	0.0	0.0	174	1.4	e

Urea-Urine




MQ tolerance : 15 %

Urea-Urine (mmol/l)


No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Standard chemistry	19	100.0	0.0	0.0	249	4.4	e

Uric Acid-Urine



No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	Standard chemistry	16	100.0	0.0	0.0	0.98	5.9	e

Specific Gravity-Urine




MQ tolerance : 5 %

Specific Gravity-Urine ()

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Type
1	Refractometer	7	100.0	0.0	0.0	1.023	0.3	e

Creatinine U




QUALAB tolerance : 24 %

Creatinine U (mg/l)

No.	Methode	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	AFIAS	9	88.9	0.0	11.1	48.2	13.9	e*
2	Afinion	420	93.8	3.3	2.9	46.0	10.4	e
3	Other methods	4	25.0	0.0	75.0	30.0	0.0	a
4	NycoCard	6	83.3	0.0	16.7	41.4	10.0	a
5	Turbidimetry	21	95.2	4.8	0.0	41.5	8.2	e
6	DCA2000/Vantage	134	97.0	0.0	3.0	43.3	6.0	e
7	Siemens Clinitek	12	91.7	0.0	8.3	30.0	0.0	a

Creatinin Urin



QUALAB tolerance : 21 %

Creatinin Urin (mmol/l)

No.	Method	Total	% good	% insuff.	% outlier	Target value	CV%	Typ
1	DCA2000/Vantage	134	97.0	0.0	3.0	10.0	4.9	e
2	Afinion	420	98.6	0.0	1.4	8.9	5.7	e
3	Standard chemistry	36	100.0	0.0	0.0	9.6	5.0	e
4	Siemens Clinitek	12	41.7	0.0	58.3	8.8	0.0	a