

Report U03 Drugs of Abuse**Amphetamine**

| | Target Value | Participants | Conform | % Conform |
|-------------------|--------------|--------------|-----------|--------------|
| Architect | positive | 3 | 3 | 100.00 |
| Cobas | positive | 3 | 3 | 100.00 |
| Syva Rapid Test | positive | 1 | 1 | 100.00 |
| Intex | positive | 2 | 2 | 100.00 |
| Other quantitativ | positive | 15 | 13 | 87.00 |
| Others | positive | 54 | 46 | 85.00 |
| Total | | 78 | 68 | 87.00 |

Methamphetamine

| | Target Value | Participants | Conform | % Conform |
|-----------------|--------------|--------------|-----------|--------------|
| Intex | positive | 1 | 1 | 100.00 |
| Syva Rapid Test | positive | 3 | 2 | 67.00 |
| Others | positive | 27 | 25 | 93.00 |
| Total | | 31 | 28 | 90.00 |

Ecstasy pos/neg

| | Target Value | Participants | Conform | % Conform |
|--------|--------------|--------------|---------|-----------|
| Others | negative | 13 | 12 | 92.00 |

Barbiturates

| | Target Value | Participants | Conform | % Conform |
|-------------------|--------------|--------------|-----------|--------------|
| Architect | negative | 2 | 2 | 100.00 |
| Cobas | negative | 3 | 3 | 100.00 |
| Other quantitativ | negative | 8 | 8 | 100.00 |
| Others | negative | 23 | 22 | 96.00 |
| Total | | 36 | 35 | 97.00 |

Benzodiazepines

| | Target Value | Participants | Conform | % Conform |
|-------------------|--------------|--------------|-----------|--------------|
| Architect | negative | 3 | 3 | 100.00 |
| Cobas | negative | 4 | 4 | 100.00 |
| Intex | negative | 5 | 5 | 100.00 |
| Other quantitativ | negative | 16 | 16 | 100.00 |
| Others | negative | 61 | 60 | 98.00 |
| Total | | 89 | 88 | 99.00 |

Cannabinoides

| | Target Value | Participants | Conform | % Conform |
|-------------------|--------------|--------------|------------|--------------|
| Architect | negative | 3 | 3 | 100.00 |
| Cobas | negative | 4 | 4 | 100.00 |
| Syva Rapid Test | negative | 3 | 3 | 100.00 |
| Intex | negative | 8 | 8 | 100.00 |
| Other quantitativ | negative | 17 | 17 | 100.00 |
| Others | negative | 69 | 68 | 99.00 |
| Total | | 104 | 103 | 99.00 |

Cocaine

| | Target Value | Participants | Conform | % Conform |
|-------------------|--------------|--------------|-----------|--------------|
| Architect | negative | 3 | 3 | 100.00 |
| Cobas | negative | 4 | 4 | 100.00 |
| Syva Rapid Test | negative | 3 | 3 | 100.00 |
| Intex | negative | 3 | 3 | 100.00 |
| Other quantitativ | negative | 16 | 16 | 100.00 |
| Others | negative | 67 | 66 | 99.00 |
| Total | | 96 | 95 | 99.00 |

Methadone

| | Target Value | Participants | Conform | % Conform |
|-------------------|--------------|--------------|-----------|--------------|
| Architect | positive | 3 | 3 | 100.00 |
| Cobas | positive | 4 | 4 | 100.00 |
| Intex | positive | 5 | 5 | 100.00 |
| Rapitest Morwell | positive | 1 | 1 | 100.00 |
| Other quantitativ | positive | 14 | 13 | 93.00 |
| Others | positive | 50 | 50 | 100.00 |
| Total | | 77 | 76 | 99.00 |

EDDP

| | Target Value | Participants | Conform | % Conform |
|------------------|--------------|--------------|---------|-----------|
| all Participants | positive | 3 | 3 | 100.00 |

Opiate

| | Target Value | Participants | Conform | % Conform |
|-------------------|--------------|--------------|-----------|--------------|
| Architect | negative | 3 | 3 | 100.00 |
| Cobas | negative | 4 | 4 | 100.00 |
| Syva Rapid Test | negative | 3 | 3 | 100.00 |
| Intex | negative | 4 | 4 | 100.00 |
| Rapitest Morwell | negative | 1 | 1 | 100.00 |
| Other quantitativ | negative | 16 | 16 | 100.00 |
| Others | negative | 63 | 62 | 98.00 |
| Total | | 94 | 93 | 99.00 |

Buprenorphin

| | Target Value | Participants | Conform | % Conform |
|------------------|--------------|--------------|---------|-----------|
| all Participants | negative | 3 | 3 | 100.00 |

6-Acetylmorphin

| | Target Value | Participants | Conform | % Conform |
|------------------|--------------|--------------|---------|-----------|
| all Participants | negative | 3 | 3 | 100.00 |

Phencyclidin

| | Target Value | Participants | Conform | % Conform |
|--------|--------------|--------------|---------|-----------|
| Others | negative | 5 | 5 | 100.00 |

TCA

| | Target Value | Participants | Conform | % Conform |
|------------------|--------------|--------------|---------|-----------|
| all Participants | negative | 14 | 14 | 100.00 |

LSD

| | Target Value | Participants | Conform | % Conform |
|-------------------|--------------|--------------|---------|-----------|
| Other quantitativ | negative | 2 | 2 | 100.00 |

Commentary U3

To one liter of urine 650 µg methadon, 900 µg EDDP, 1600 µg d,l-amphetamine and 1600 µg d,l-methamphetamine were added. Accordingly methadon, amphetamine and methamphetamine are all positive and all the other analyses are negative. With amphetamine and methamphetamine some rapid tests resulted in "negative", because a higher decision limit is used as usual in Switzerland.