

**Report U03 Drugs of Abuse**

**Amphetamine**

	Target Value	Participants	Conform	% Conform
Architect	negative	3	3	100.00
Cobas	negative	3	3	100.00
Syva Rapid Test	negative	1	1	100.00
Intex	negative	2	2	100.00
Other quantitativ	negative	16	16	100.00
Others	negative	53	53	100.00
<b>Total</b>		<b>78</b>	<b>78</b>	<b>100.00</b>

**Methamphetamine**

	Target Value	Participants	Conform	% Conform
Intex	negative	1	1	100.00
Syva Rapid Test	negative	3	3	100.00
Others	negative	25	25	100.00
<b>Total</b>		<b>29</b>	<b>29</b>	<b>100.00</b>

**Ecstasy pos/neg**

	Target Value	Participants	Conform	% Conform
Others	negative	14	13	93.00

**Barbiturates**

	Target Value	Participants	Conform	% Conform
Architect	negative	2	2	100.00
Cobas	negative	3	3	100.00
Other quantitativ	negative	9	9	100.00
Others	negative	21	21	100.00
<b>Total</b>		<b>35</b>	<b>35</b>	<b>100.00</b>

**Benzodiazepines**

	Target Value	Participants	Conform	% Conform
Architect	negative	3	3	100.00
Cobas	negative	4	4	100.00
Intex	negative	5	5	100.00
Other quantitativ	negative	17	17	100.00
Others	negative	61	61	100.00
<b>Total</b>		<b>90</b>	<b>90</b>	<b>100.00</b>

**Cannabinoides**

	Target Value	Participants	Conform	% Conform
Architect	negative	3	3	100.00
Cobas	negative	4	4	100.00
Syva Rapid Test	negative	3	3	100.00
Intex	negative	8	8	100.00
Other quantitativ	negative	18	18	100.00
Others	negative	69	69	100.00
<b>Total</b>		<b>105</b>	<b>105</b>	<b>100.00</b>

**Cocaine**

	Target Value	Participants	Conform	% Conform
Architect	positive	3	3	100.00
Cobas	positive	4	4	100.00
Syva Rapid Test	positive	3	3	100.00
Intex	positive	3	3	100.00
Other quantitativ	positive	17	17	100.00
Others	positive	66	65	98.00
<b>Total</b>		<b>96</b>	<b>95</b>	<b>99.00</b>

**Methadone**

	Target Value	Participants	Conform	% Conform
Architect	negative	3	3	100.00
Cobas	negative	4	4	100.00
Syva Rapid Test		1	1	100.00
Intex	negative	5	5	100.00
Rapitest Morwell	negative	1	1	100.00
Syntron Quik		1	1	100.00
Other quantitativ	negative	15	15	100.00
Others	negative	49	49	100.00
<b>Total</b>		<b>79</b>	<b>79</b>	<b>100.00</b>

**Methaqualon**

	Target Value	Participants	Conform	% Conform
all Participants	positive	1	1	100.00

**EDDP**

	Target Value	Participants	Conform	% Conform
all Participants	negative	4	4	100.00

**Opiate**

	Target Value	Participants	Conform	% Conform
Architect	negative	3	3	100.00
Cobas	negative	5	4	80.00
Syva Rapid Test	negative	3	3	100.00
Intex	negative	4	4	100.00
Rapitest Morwell	negative	2	1	50.00
Other quantitativ	negative	17	17	100.00
Others	negative	63	63	100.00
<b>Total</b>		<b>97</b>	<b>95</b>	<b>98.00</b>

**Buprenorphin**

	Target Value	Participants	Conform	% Conform
all Participants	negative	3	3	100.00

**6-Acetylmorphin**

	Target Value	Participants	Conform	% Conform
all Participants	negative	3	3	100.00

**Phencyclidin**

	Target Value	Participants	Conform	% Conform
Others	negative	5	5	100.00

**TCA**

	Target Value	Participants	Conform	% Conform
all Participants	positive	13	12	92.00

**LSD**

	Target Value	Participants	Conform	% Conform
Other quantitativ	negative	2	2	100.00

**Commentary U3**