

**Report U03 Drugs of Abuse**

**Amphetamine**

	Target Value	Participants	Conform	% Conform
Architect	negative	3	3	100.00
Cobas	negative	3	3	100.00
Syva Rapid Test	negative	1	1	100.00
Intex	negative	2	2	100.00
Other quantitativ	negative	15	15	100.00
Others	negative	53	52	98.00
<b>Total</b>		<b>77</b>	<b>76</b>	<b>99.00</b>

**Methamphetamine**

	Target Value	Participants	Conform	% Conform
Intex	negative	1	1	100.00
Syva Rapid Test	negative	3	3	100.00
Others	negative	24	23	96.00
<b>Total</b>		<b>28</b>	<b>27</b>	<b>96.00</b>

**Ecstasy pos/neg**

	Target Value	Participants	Conform	% Conform
Others	negative	13	13	100.00

**Barbiturates**

	Target Value	Participants	Conform	% Conform
Architect	positive	2	2	100.00
Cobas	positive	3	3	100.00
Other quantitativ	positive	8	8	100.00
Others	positive	23	21	91.00
<b>Total</b>		<b>36</b>	<b>34</b>	<b>94.00</b>

**Benzodiazepines**

	Target Value	Participants	Conform	% Conform
Architect	negative	3	3	100.00
Cobas	negative	4	4	100.00
Intex	negative	5	5	100.00
Other quantitativ	negative	16	16	100.00
Others	negative	61	60	98.00
<b>Total</b>		<b>89</b>	<b>88</b>	<b>99.00</b>

**Cannabinoides**

	Target Value	Participants	Conform	% Conform
Architect	positive	3	3	100.00
Cobas	positive	4	4	100.00
Syva Rapid Test	positive	3	3	100.00
Intex	positive	8	8	100.00
Other quantitativ	positive	17	17	100.00
Others	positive	69	68	99.00
<b>Total</b>		<b>104</b>	<b>103</b>	<b>99.00</b>

**Cocaine**

	Target Value	Participants	Conform	% Conform
Architect	negative	3	3	100.00
Cobas	negative	4	4	100.00
Syva Rapid Test	negative	3	3	100.00
Intex	negative	3	3	100.00
Other quantitativ	negative	16	16	100.00
Others	negative	65	63	97.00
<b>Total</b>		<b>94</b>	<b>92</b>	<b>98.00</b>

**Methadone**

	Target Value	Participants	Conform	% Conform
Architect	negative	3	3	100.00
Cobas	negative	4	4	100.00
Intex	negative	4	3	75.00
Rapitest Morwell	negative	1	1	100.00
Other quantitativ	negative	14	14	100.00
Others	negative	49	48	98.00
<b>Total</b>		<b>75</b>	<b>73</b>	<b>97.00</b>

**Methaqualon**

	Target Value	Participants	Conform	% Conform
all Participants	negative	1	1	100.00

**EDDP**

	Target Value	Participants	Conform	% Conform
all Participants	negative	4	4	100.00

**Opiate**

	Target Value	Participants	Conform	% Conform
Architect	positive	3	3	100.00
Cobas	positive	4	4	100.00
Syva Rapid Test	positive	3	3	100.00
Intex	positive	4	3	75.00
Rapitest Morwell	positive	1	1	100.00
Other quantitativ	positive	17	17	100.00
Others	positive	62	58	94.00
<b>Total</b>		<b>94</b>	<b>89</b>	<b>95.00</b>

**Buprenorphin**

	Target Value	Participants	Conform	% Conform
all Participants	negative	4	3	75.00

**6-Acetylmorphin**

	Target Value	Participants	Conform	% Conform
all Participants	positive	3	3	100.00

**Phencyclidin**

	Target Value	Participants	Conform	% Conform
Others	positive	4	3	75.00

**TCA**

	Target Value	Participants	Conform	% Conform
all Participants	negative	13	12	92.00

**LSD**

	Target Value	Participants	Conform	% Conform
Other quantitativ	negative	3	3	100.00

**Commentary U3**

*To one liter of urine 600 µg Secobarbital, 400 µg 11-Nor-D9-THC, 400 µg D9-THC, 30 µg PCP and 800 µg 6-Acetylmorphin were added. Accordingly barbiturates, cannabis, PCP and opiates are all positive and all the other analyses are negative.*