

Address for correspondence

Dr. C. W. Weykamp

Queen Beatrix Hospital

Beatrix Park 1

7101 BN Winterswijk

The Netherlands

Telephone: + 31 543 54 47 74

Fax: + 31 543 52 42 65

e-mail: c.w.weykamp@skbwinterswijk.nl

Chamber of Commerce: 09101380

Winterswijk, 13 March 2020

Dear Dr. Fried,

Please find below the IFCC and NGSP certified values of your HbA1c samples. Values have been assigned in duplicate with three IFCC calibrated Secondary Reference Measurement Procedures. NGSP values are derived from the IFCC values using the Master Equation.

Sample ID	Assigned IFCC value mmol/mol	U (k=2)	Assigned NGSP value %	U (k=2)
RV 2020-01, HbA1c K3A	38,0	1,1	5,63	0,10
RV 2020-01, HbA1c K3B	40,8	1,0	5,88	0,09

Samples arrived on 11 March and were analysed the next day.

Yours sincerely,

Dr. Cas Weykamp
DirectorCarla Siebelder BSc.
Person in charge