



Address for correspondence
European Reference Laboratory
C. Siebelder
Beatrixpark 1
7101 BN Winterswijk
The Netherlands
Telephone: + 31 543 54 47 74
e-mail: c.siebelder@skbwinterswijk.nl
Chamber of Commerce: 09101380

Winterswijk, 15 November 2021

Dear Dr. Fried,

Please find below the IFCC and NGSP certified values of your HbA1c samples. Values have been assigned in duplicate with three IFCC calibrated Secondary Reference Measurement Procedures. NGSP values are derived from the IFCC values using the Master Equation.

Sample ID	Assigned IFCC value mmol/mol	U (k=2)	Assigned NGSP value %	U (k=2)
RV 2021-04, HbA1c K4A	37,9	1,0	5,62	0,09
RV 2021-04, HbA1c K4B	35,3	1,0	5,38	0,09

Samples arrived on 11 November and were analysed the next day.

Yours sincerely,

Carla Siebelder BSc.
Coordinator ERL

dr. Eline van der Hagen
Director MCA Laboratory



Dr. Ir. R.J. Slingerland
chair

Dr. C. W. Weykamp
secretary

Dr. H. van der Vuurst
treasurer