



G3/G4 Coagulation advanced 2023-1

Instructions

- The numbers show how many participants had a result in a class.
- The fields with a box show in which class the results of your lab were.
- Target values are calculated as consensus values after elimination of outliers.
- For thrombintime, we evaluate only the interpretation

Interpretation of the results

| No therapy | | Normal | Limit | Abnormal |
|------------|--------------|--------|-------|----------|
| G3 | PTT | 26 | 0 | 0 |
| G3 | Thrombintime | 23 | 2 | 1 |

| Heparin therapy | | No Heparin | Subtherap. | Therap. | To much |
|-----------------|--------------|------------|------------|---------|---------|
| G4 | PTT | 0 | 0 | 2 | 21 |
| G4 | Thrombintime | 1 | 0 | 4 | 22 |

28.03.2023nka

Distribution of aPTT results

Due to the large scatter of the quantitative results, it is difficult to calculate target values. As an additional help to interpret the results, we provide the following histograms that show the distribution of all values.

