



Address for correspondence  
 European Reference Laboratory  
 C. Siebelder  
 Beatrixpark 1  
 7101 BN Winterswijk  
 The Netherlands  
 Telephone: + 31 543 54 47 74  
 e-mail: c.siebelder@skbwinterswijk.nl  
 Chamber of Commerce: 09101380

Winterswijk, 7 June 2023

Dear Dr. Fried,

Please find below the IFCC and NGSP certified values of your HbA1c samples. Values have been assigned in duplicate with three IFCC calibrated Secondary Reference Measurement Procedures. NGSP values are derived from the IFCC values using the Master Equation.

Sample ID	Assigned IFCC value mmol/mol	U (k=2)	Assigned NGSP value %	U (k=2)
RV 2023-02, HbA1c K3A	35,2	1,0	5,37	0,09
RV 2023-02, HbA1c K3B	36,1	1,0	5,45	0,09

Samples arrived on 6 June and were analysed the next day.

Yours sincerely,

Carla Siebelder BSc.  
Coordinator ERL

dr. Eline van der Hagen  
Director MCA Laboratory



Dr. Ir. R. J. Slingerland  
chair

C. Siebelder  
secretary

J. Sigter  
treasurer