

**Report U03 Drugs of Abuse**

**Ethylglucuronid ql**

	Target Value	Participants	Conform	% Conform
Rapid Test	negative	2	2	100.00

**Amphetamine**

	Target Value	Participants	Conform	% Conform
Abbott	positive	4	2	50.00
Cobas	positive	5	5	100.00
Protzek	positive	5	2	40.00
Nal von Minden	positive	8	7	88.00
SureStep Abbott	positive	13	11	85.00
Intex	positive	3	2	67.00
Willi Fox	positive	3	2	67.00
MD Doctors Direct	positive	7	6	86.00
Clear Test	positive	4	4	100.00
Other quantitativ	positive	12	8	67.00
Others	positive	17	11	65.00
<b>Total</b>		<b>81</b>	<b>60</b>	<b>74.00</b>

**Methamphetamine**

	Target Value	Participants	Conform	% Conform
Syva Rapid Test	negative	2	2	100.00
Others	negative	26	24	92.00
<b>Total</b>		<b>28</b>	<b>26</b>	<b>93.00</b>

**Ecstasy pos/neg**

	Target Value	Participants	Conform	% Conform
Others	negative	21	21	100.00

**Barbiturates**

	Target Value	Participants	Conform	% Conform
Abbott	negative	3	3	100.00
Cobas	negative	5	5	100.00
Protzek	negative	2	2	100.00
Nal von Minden	negative	3	2	67.00
SureStep Abbott	negative	8	8	100.00
Willi Fox	negative	1	1	100.00
MD Doctors Direct	negative	3	3	100.00
Clear Test	negative	1	1	100.00
Other quantitativ	negative	6	6	100.00

Others	negative	11	11	100.00
<b>Total</b>		<b>43</b>	<b>42</b>	<b>98.00</b>

**Benzodiazepines**

	Target Value	Participants	Conform	% Conform
Abbott	negative	4	4	100.00
Cobas	negative	5	5	100.00
Protzek	negative	4	4	100.00
Nal von Minden	negative	8	8	100.00
SureStep Abbott	negative	12	11	92.00
Intex	negative	4	4	100.00
Willi Fox	negative	2	2	100.00
MD Doctors Direct	negative	8	8	100.00
Clear Test	negative	4	4	100.00
Other quantitativ	negative	15	15	100.00
Others	negative	16	16	100.00
<b>Total</b>		<b>82</b>	<b>81</b>	<b>99.00</b>

**Cannabinoïdes**

	Target Value	Participants	Conform	% Conform
Abbott	negative	5	5	100.00
Cobas	negative	5	5	100.00
Protzek	negative	5	5	100.00
Nal von Minden	negative	11	11	100.00
SureStep Abbott	negative	15	13	87.00
Syva Rapid Test	negative	2	2	100.00
Intex	negative	7	7	100.00
Willi Fox	negative	4	4	100.00
MD Doctors Direct	negative	9	9	100.00
Clear Test	negative	5	5	100.00
Other quantitativ	negative	14	14	100.00
Others	negative	25	24	96.00
<b>Total</b>		<b>107</b>	<b>104</b>	<b>97.00</b>

**Cocaine**

	Target Value	Participants	Conform	% Conform
Abbott	negative	4	4	100.00
Cobas	negative	5	5	100.00
Protzek	negative	5	5	100.00
Nal von Minden	negative	8	8	100.00
SureStep Abbott	negative	14	12	86.00
Syva Rapid Test	negative	1	1	100.00
Intex	negative	3	3	100.00
Willi Fox	negative	3	3	100.00
MD Doctors Direct	negative	8	8	100.00
Clear Test	negative	5	5	100.00
Other quantitativ	negative	14	14	100.00
Others	negative	19	18	95.00
<b>Total</b>		<b>89</b>	<b>86</b>	<b>97.00</b>

**Methadone**

	Target Value	Participants	Conform	% Conform
Abbott	positive	4	4	100.00
Cobas	positive	5	5	100.00
Protzek	positive	4	4	100.00
Nal von Minden	positive	6	6	100.00
SureStep Abbott	positive	12	9	75.00
Intex	positive	3	3	100.00
Willi Fox	positive	2	2	100.00
MD Doctors Direct	positive	9	8	89.00
Clear Test	positive	4	3	75.00
Other quantitativ	positive	11	11	100.00
Others	positive	18	14	78.00
<b>Total</b>		<b>78</b>	<b>69</b>	<b>88.00</b>

**Methaqualon**

	Target Value	Participants	Conform	% Conform
all Participants	negative	1	1	100.00

**EDDP**

	Target Value	Participants	Conform	% Conform
all Participants	positive	8	7	88.00

**Opiate**

	Target Value	Participants	Conform	% Conform
Abbott	negative	4	4	100.00
Cobas	negative	5	5	100.00
Protzek	negative	5	5	100.00
Nal von Minden	negative	8	8	100.00
SureStep Abbott	negative	11	9	82.00
Syva Rapid Test	negative	1	1	100.00
Intex	negative	3	3	100.00
Willi Fox	negative	3	3	100.00
MD Doctors Direct	negative	9	9	100.00
Clear Test	negative	5	4	80.00
Other quantitativ	negative	14	14	100.00
Others	negative	18	17	94.00
<b>Total</b>		<b>86</b>	<b>82</b>	<b>95.00</b>

**Buprenorphin**

	Target Value	Participants	Conform	% Conform
all Participants	negative	7	6	86.00

**6-Acetylmorphin**

	Target Value	Participants	Conform	% Conform
all Participants	negative	1	1	100.00

**Phencyclidin**

	Target Value	Participants	Conform	% Conform
Others	negative	4	4	100.00

**TCA**

	Target Value	Participants	Conform	% Conform
all Participants	negative	16	16	100.00

**LSD**

	Target Value	Participants	Conform	% Conform
Other quantitativ	positive	2	2	100.00

**Methamphetamine**

	Target Value	Participants	Conform	% Conform
Nal von Minden	negative	6	6	100.00

**Morphin**

	Target Value	Participants	Conform	% Conform
all Participants	negative	2	2	100.00

**Commentary U3**

*To one liter of urine 1650 µg Amphetamine, 3 ml Ethanol, 11.7 mg Paracetamol, 480 µg Methadone, 480 µg EDDP and 10 µg LSD attached.*