

Address for correspondence
European Reference Laboratory
C. Siebelder
Beatrixpark 1
7101 BN Winterswijk
The Netherlands
Telephone: + 31 543 54 47 74
e-mail: c.siebelder@skbwinterswijk.nl
Chamber of Commerce: 09101380

Winterswijk, 9 November 2023

Dear Dr. Fried,

Please find below the IFCC and NGSP certified values of your HbA1c samples. Values have been assigned in duplicate with three IFCC calibrated Secondary Reference Measurement Procedures. NGSP values are derived from the IFCC values using the Master Equation.

Sample ID	Assigned IFCC value mmol/mol	U (k=2)	Assigned NGSP value %	U (k=2)
RV 2023-04, HbA1c K3A	63,5	1,0	7,96	0,09
RV 2023-04, HbA1c K3B	42,7	1,0	6,06	0,09

Samples arrived on 8 November and were analysed the next day.

Yours sincerely.



Carla Siebelder BSc.
Coordinator ERL