

## Valori ideali MQ 2023 - 4

		Giusto	Tolleranza		Numero	Adempite
<b>H01 Ematologia</b>						
<b>Emoglobina</b>						
229	Hemocue Hb 801	132.0 g/l	120.1 - 143.9	9 %	5	5 (100.0 %)
286	MEK-1303/5	120.0 g/l	109.2 - 130.8	9 %	14	14 (100.0 %)
201	Automatico	118.0 g/l	107.4 - 128.6	9 %	9	8 (88.9 %)
204	Cianometemoglobina	118.5 g/l	107.8 - 129.1	9 %	16	15 (93.8 %)
274	Sysmex X	118.0 g/l	107.3 - 128.6	9 %	60	60 (100.0 %)
228	Hemocue	117.6 g/l	107.0 - 128.2	9 %	399	380 (95.2 %)
276	Hemocontrol	118.7 g/l	108.0 - 129.4	9 %	16	16 (100.0 %)
206	DiaSpect	124.6 g/l	113.4 - 135.8	9 %	13	13 (100.0 %)
<b>Emoglobina</b>						
261	Sysmex KX21	117.8 g/l	107.2 - 128.4	9 %	105	100 (95.2 %)
268	Sysmex Poch - 100i	116.1 g/l	105.6 - 126.5	9 %	184	172 (93.5 %)
280	Sysmex XP 300	117.6 g/l	107.0 - 128.2	9 %	599	582 (97.2 %)
270	Mythic	117.1 g/l	106.6 - 127.6	9 %	234	225 (96.2 %)
278	Sysmex XQ-320	118.4 g/l	107.8 - 129.1	9 %	103	100 (97.1 %)
264	Swelab	121.8 g/l	110.9 - 132.8	9 %	28	27 (96.4 %)
272	Medonic	120.0 g/l	109.2 - 130.8	9 %	4	4 (100.0 %)
273	Celltac Alpha (Nihon	120.9 g/l	110.0 - 131.8	9 %	70	65 (92.9 %)
281	Samsung HC10	119.0 g/l	108.3 - 129.7	9 %	9	7 (77.8 %)
284	Micros 60	116.2 g/l	105.7 - 126.6	9 %	61	60 (98.4 %)
<b>Ematocrito</b>						
230	MEK-1303/5	0.39 l/l	0.36 - 0.43	9 %	14	14 (100.0 %)
101	Automatico	0.31 l/l	0.28 - 0.34	9 %	6	5 (83.3 %)
174	Sysmex X	0.34 l/l	0.31 - 0.37	9 %	60	60 (100.0 %)
<b>Ematocrito</b>						
178	Sysmex XQ-320	0.32 l/l	0.30 - 0.35	9 %	104	101 (97.1 %)
161	Sysmex KX21	0.32 l/l	0.29 - 0.35	9 %	105	100 (95.2 %)
168	Sysmex Poch - 100i	0.34 l/l	0.31 - 0.37	9 %	184	172 (93.5 %)
180	Sysmex XP 300	0.32 l/l	0.29 - 0.35	9 %	600	586 (97.7 %)
170	Mythic	0.33 l/l	0.30 - 0.36	9 %	234	223 (95.3 %)
164	Swelab	0.33 l/l	0.30 - 0.36	9 %	28	27 (96.4 %)
172	Medonic	0.32 l/l	0.29 - 0.35	9 %	4	4 (100.0 %)
173	Celltac Alpha (Nihon	0.34 l/l	0.31 - 0.37	9 %	70	62 (88.6 %)
181	Samsung HC10	0.33 l/l	0.30 - 0.36	9 %	9	8 (88.9 %)
184	Micros 60	0.31 l/l	0.28 - 0.33	9 %	61	60 (98.4 %)
<b>Eritrociti</b>						
386	MEK-1303/5	3.85 T/l	2.89 - 4.81	25 %	14	14 (100.0 %)
301	Automatico	3.67 T/l	2.75 - 4.59	25 %	6	6 (100.0 %)
374	Sysmex X	3.69 T/l	2.77 - 4.61	25 %	60	60 (100.0 %)

## Valori ideali MQ 2023 - 4

		Giusto	Tolleranza		Numero	Adempite
<b>Eritrociti</b>						
378	Sysmex XQ-320	3.80 T/l	2.85 - 4.75	25 %	104	99 (95.2 %)
361	Sysmex KX21	3.71 T/l	2.79 - 4.64	25 %	105	102 (97.1 %)
368	Sysmex Poch - 100i	3.84 T/l	2.88 - 4.80	25 %	184	175 (95.1 %)
380	Sysmex XP 300	3.72 T/l	2.79 - 4.65	25 %	600	589 (98.2 %)
370	Mythic	3.67 T/l	2.75 - 4.59	25 %	234	229 (97.9 %)
364	Swelab	3.77 T/l	2.83 - 4.72	25 %	28	28 (100.0 %)
372	Medonic	3.71 T/l	2.78 - 4.63	25 %	4	4 (100.0 %)
473	Celltac Alpha (Nihon)	3.81 T/l	2.86 - 4.76	25 %	70	67 (95.7 %)
381	Samsung HC10	3.82 T/l	2.87 - 4.78	25 %	9	9 (100.0 %)
384	Micros 60	3.60 T/l	2.70 - 4.50	25 %	61	60 (98.4 %)
<b>Leucociti</b>						
486	MEK-1303/5	5.94 G/l	4.45 - 7.42	25 %	14	14 (100.0 %)
401	Automatico	5.90 G/l	4.43 - 7.38	25 %	5	5 (100.0 %)
403	Microscopio	5.70 G/l	4.28 - 7.13	25 %	5	3 (60.0 %)
474	Sysmex X	5.94 G/l	4.45 - 7.42	25 %	60	60 (100.0 %)
<b>Leucociti</b>						
478	Sysmex XQ-320	5.98 G/l	4.48 - 7.47	25 %	103	101 (98.1 %)
461	Sysmex KX21	5.88 G/l	4.41 - 7.34	25 %	105	104 (99.0 %)
468	Sysmex Poch - 100i	5.75 G/l	4.31 - 7.19	25 %	184	177 (96.2 %)
480	Sysmex XP 300	5.89 G/l	4.41 - 7.36	25 %	600	595 (99.2 %)
470	Mythic	5.59 G/l	4.20 - 6.99	25 %	233	229 (98.3 %)
464	Swelab	5.77 G/l	4.33 - 7.21	25 %	28	28 (100.0 %)
472	Medonic	5.85 G/l	4.39 - 7.31	25 %	4	4 (100.0 %)
373	Celltac Alpha (Nihon)	5.82 G/l	4.37 - 7.28	25 %	69	66 (95.7 %)
481	Samsung HC10	6.22 G/l	4.67 - 7.78	25 %	9	7 (77.8 %)
484	Micros 60	5.32 G/l	3.99 - 6.65	25 %	61	58 (95.1 %)
<b>Trombociti</b>						
586	MEK-1303/5	237.1 G/l	177.8 - 296.3	25 %	14	14 (100.0 %)
501	Automatico	212.0 G/l	159.0 - 265.0	25 %	5	5 (100.0 %)
503	Microscopio	234.5 G/l	175.9 - 293.1	25 %	4	3 (75.0 %)
574	Sysmex X	206.3 G/l	154.7 - 257.9	25 %	60	60 (100.0 %)
<b>Trombociti</b>						
578	Sysmex XQ-320	211.1 G/l	158.3 - 263.9	25 %	102	101 (99.0 %)
561	Sysmex KX21	213.6 G/l	160.2 - 267.1	25 %	105	103 (98.1 %)
568	Sysmex Poch - 100i	217.4 G/l	163.1 - 271.8	25 %	184	179 (97.3 %)
580	Sysmex XP 300	215.6 G/l	161.7 - 269.5	25 %	600	594 (99.0 %)
570	Mythic	192.8 G/l	144.6 - 241.0	25 %	235	229 (97.4 %)
564	Swelab	206.3 G/l	154.7 - 257.9	25 %	28	28 (100.0 %)
572	Medonic	200.0 G/l	150.0 - 250.0	25 %	4	4 (100.0 %)
573	Celltac Alpha (Nihon)	227.1 G/l	170.3 - 283.8	25 %	70	68 (97.1 %)
581	Samsung HC10	244.0 G/l	183.0 - 305.0	25 %	9	8 (88.9 %)
584	Micros 60	195.9 G/l	146.9 - 244.9	25 %	60	59 (98.3 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>H02 Ematologia Plus</b>					
<b>Emoglobina H2</b>					
285 Z3	103.1 g/l	93.8 - 112.4	9 %	219	208 (95.0 %)
263 Abx Micros	102.9 g/l	93.7 - 112.2	9 %	50	48 (96.0 %)
279 Microsemi	104.1 g/l	94.8 - 113.5	9 %	895	873 (97.5 %)
<b>Ematocrito H2</b>					
685 Z3	0.30 l/l	0.27 - 0.33	9 %	219	202 (92.2 %)
163 Abx Micros	0.27 l/l	0.25 - 0.30	9 %	50	46 (92.0 %)
179 Microsemi	0.28 l/l	0.25 - 0.30	9 %	895	869 (97.1 %)
<b>Leucociti H2</b>					
485 Z3	5.56 G/l	4.17 - 6.94	25 %	219	217 (99.1 %)
463 Abx Micros	5.58 G/l	4.19 - 6.98	25 %	50	49 (98.0 %)
479 Microsemi	5.74 G/l	4.31 - 7.18	25 %	895	888 (99.2 %)
<b>Trombociti H2</b>					
585 Z3	192.8 G/l	144.6 - 241.0	25 %	219	215 (98.2 %)
563 Abx Micros	193.4 G/l	145.1 - 241.8	25 %	50	45 (90.0 %)
579 Microsemi	195.0 G/l	146.3 - 243.8	25 %	894	880 (98.4 %)
<b>Eritrociti H2</b>					
385 Z3	3.23 T/l	2.43 - 4.04	25 %	219	211 (96.3 %)
363 Abx Micros	3.20 T/l	2.40 - 4.00	25 %	50	48 (96.0 %)
379 Microsemi	3.22 T/l	2.41 - 4.02	25 %	894	883 (98.8 %)
<b>CRP H2</b>					
1685 Z3	47.0 mg/l	37.1 - 56.8	21 %	203	183 (90.1 %)
1679 Microsemi	67.1 mg/l	53.0 - 81.2	21 %	877	821 (93.6 %)
1663 Abx Micros	67.7 mg/l	53.5 - 81.9	21 %	12	10 (83.3 %)
1664 ABX Micros CRP200	61.9 mg/l	48.9 - 74.9	21 %	38	36 (94.7 %)
<b>I01 CRP</b>					
<b>CRP</b>					
1604 Autolyser	54.1 mg/l	42.7 - 65.5	21 %	10	8 (80.0 %)
1602 Cobas b101	47.0 mg/l	37.2 - 56.9	21 %	400	397 (99.3 %)
1617 Cobas	52.9 mg/l	41.8 - 64.0	21 %	37	36 (97.3 %)
1643 Turbidimetrie	60.2 mg/l	47.6 - 72.8	21 %	5	5 (100.0 %)
1601 Afinion	52.7 mg/l	41.6 - 63.8	21 %	1103	1094 (99.2 %)
1630 NycoCard SingleTest-	50.0 mg/l	39.5 - 60.5	21 %	61	49 (80.3 %)
1616 Quick Read go	56.7 mg/l	44.8 - 68.6	21 %	83	80 (96.4 %)
1610 Eurolyser	66.9 mg/l	52.8 - 80.9	21 %	70	42 (60.0 %)
1632 Fuji Dri-Chem	55.6 mg/l	43.9 - 67.3	21 %	14	11 (78.6 %)
1613 Piccolo	77.0 mg/l	60.8 - 93.2	21 %	4	4 (100.0 %)
1673 Celltac chemi	54.0 mg/l	42.6 - 65.3	21 %	43	43 (100.0 %)
<b>CRP</b>					
1625 QuickRead (sangue)	90.9 mg/l	71.8 - 110.0	21 %	21	20 (95.2 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>CRP</b>					
1608 Spinit	46.2 mg/l	36.5 - 55.9	21 %	15	15 (100.0 %)
1609 Abbott	55.4 mg/l	43.8 - 67.0	21 %	14	14 (100.0 %)
1615 AQT 90 FLEX	57.0 mg/l	45.0 - 69.0	21 %	7	7 (100.0 %)
1635 Spotchem D-Concept	56.3 mg/l	44.5 - 68.2	21 %	6	6 (100.0 %)
1699 altro	54.0 mg/l	42.7 - 65.3	21 %	6	6 (100.0 %)
<b>I02 Proteine plasmatiche</b>					
<b>IgG</b>					
2343 Turbidimetrie	10.76 g/l	9.15 - 12.38	15 %	21	21 (100.0 %)
2344 Nephelometrie	10.75 g/l	9.14 - 12.36	15 %	4	4 (100.0 %)
2399 altro	10.90 g/l	9.27 - 12.54	15 %	5	5 (100.0 %)
<b>IgA</b>					
2443 Turbidimetrie	1.73 g/l	1.47 - 1.99	15 %	19	19 (100.0 %)
2444 Nephelometrie	1.91 g/l	1.62 - 2.20	15 %	5	5 (100.0 %)
2499 altro	1.68 g/l	1.42 - 1.93	15 %	4	4 (100.0 %)
<b>IgM</b>					
2543 Turbidimetrie	0.83 g/l	0.71 - 0.96	15 %	20	19 (95.0 %)
2544 Nephelometrie	0.84 g/l	0.71 - 0.96	15 %	5	5 (100.0 %)
<b>IgE</b>					
7007 tutti	76 kU/L	53 - 99	30 %	6	6 (100.0 %)
7009 Cobas	95 kU/L	66 - 123	30 %	6	6 (100.0 %)
<b>Alpha-1-Antitripsina</b>					
7002 tutti	1.21 g/l	0.91 - 1.51	25 %	9	9 (100.0 %)
<b>Anticorpi anti-streptolisina</b>					
7003 tutti	121 kIU/l	91 - 151	25 %	13	12 (92.3 %)
7020 altro	84 kIU/l	63 - 106	25 %	6	6 (100.0 %)
<b>Complemento C3</b>					
7004 tutti	1.38 g/l	1.17 - 1.59	15 %	19	19 (100.0 %)
<b>Complemento C4</b>					
6999 Alinity	0.25 g/l	0.21 - 0.29	15 %	5	5 (100.0 %)
7005 altro	0.24 g/l	0.21 - 0.28	15 %	15	15 (100.0 %)
<b>Aptoglobina</b>					
7006 tutti	1.17 g/l	0.88 - 1.46	25 %	29	29 (100.0 %)
<b>Transferrina</b>					
7008 tutti	2.23 g/l	1.67 - 2.78	25 %	34	34 (100.0 %)
<b>Beta-2 microglobulina</b>					
7011 tutti	1.53 mg/l	1.15 - 1.91	25 %	11	11 (100.0 %)
<b>Fattore reumatoide</b>					
7024 Architect	29.7 U/ml	22.3 - 37.1	25 %	6	6 (100.0 %)
7025 altro	30.0 U/ml	22.5 - 37.5	25 %	6	6 (100.0 %)
<b>Ceruloplasmina</b>					
7021 Siemens	302.5 mg/l	226.88 - 378.13	25 %	8	8 (100.0 %)
7012 altro	220.0 mg/l	165.00 - 275.00	25 %	4	4 (100.0 %)
<b>Prealbumina</b>					
7013 tutti	199.9 mg/l	149.96 - 249.93	25 %	18	18 (100.0 %)

## Valori ideali MQ 2023 - 4

		Giusto	Tolleranza		Numero	Adempite
<b>Recettore solubile della transferrina</b>						
7026	tutti	2.2 mg/l	1.7 - 2.8	25 %	11	11 (100.0 %)
<b>catene leggere libere kappa</b>						
7034	N Latex	7.84 mg/l	6.27 - 9.40	20 %	6	6 (100.0 %)
7027	Freelite	14.46 mg/l	11.57 - 17.35	20 %	12	12 (100.0 %)
<b>catene leggere libere lambda</b>						
7035	N Latex	7.34 mg/l	5.87 - 8.80	20 %	6	6 (100.0 %)
7028	Freelite	12.48 mg/l	9.99 - 14.98	20 %	12	11 (91.7 %)
<b>K01 Chimica</b>						
<b>Albumina</b>						
627	Hitachi S40/M40	30.5 g/l	26.8 - 34.2	12 %	4	4 (100.0 %)
601	Abbott	36.4 g/l	32.0 - 40.8	12 %	10	10 (100.0 %)
604	Autolyser	37.5 g/l	33.0 - 41.9	12 %	8	8 (100.0 %)
610	Beckman	37.0 g/l	32.5 - 41.4	12 %	4	4 (100.0 %)
632	Fuji Dri-Chem	46.2 g/l	40.6 - 51.7	12 %	236	233 (98.7 %)
603	Piccolo	39.4 g/l	34.7 - 44.1	12 %	60	58 (96.7 %)
623	Roche	38.1 g/l	33.6 - 42.7	12 %	29	29 (100.0 %)
624	Selectra Pro	36.1 g/l	31.7 - 40.4	12 %	11	11 (100.0 %)
616	Siemens	37.0 g/l	32.6 - 41.4	12 %	5	5 (100.0 %)
614	Skyla	37.0 g/l	32.6 - 41.4	12 %	5	5 (100.0 %)
635	Spotchem D-Concept	45.6 g/l	40.1 - 51.1	12 %	233	229 (98.3 %)
608	Spotchem SP-4430	41.2 g/l	36.2 - 46.1	12 %	31	31 (100.0 %)
<b>Fosfatasi alcalina</b>						
712	Abbott	92 U/l	75 - 108	18 %	14	13 (92.9 %)
704	Autolyser	83 U/l	68 - 98	18 %	21	21 (100.0 %)
714	Beckman	95 U/l	78 - 112	18 %	6	6 (100.0 %)
732	Fuji Dri-Chem	94 U/l	77 - 111	18 %	1020	1000 (98.0 %)
719	Piccolo	100 U/l	82 - 118	18 %	52	51 (98.1 %)
723	Roche	82 U/l	68 - 97	18 %	29	29 (100.0 %)
724	Selectra Pro	96 U/l	79 - 114	18 %	15	15 (100.0 %)
717	Siemens	81 U/l	66 - 96	18 %	6	6 (100.0 %)
736	Skyla	97 U/l	80 - 114	18 %	5	5 (100.0 %)
735	Spotchem D-Concept	81 U/l	66 - 95	18 %	540	534 (98.9 %)
708	Spotchem SP-4430	67 U/l	55 - 78	18 %	93	87 (93.5 %)
<b>Amilasi</b>						
817	Abbott	115 U/l	94 - 136	18 %	7	7 (100.0 %)
804	Autolyser	115 U/l	94 - 136	18 %	8	7 (87.5 %)
818	Beckman	143 U/l	117 - 168	18 %	4	4 (100.0 %)
832	Fuji Dri-Chem	90 U/l	74 - 106	18 %	742	732 (98.7 %)
819	Piccolo	122 U/l	100 - 144	18 %	52	51 (98.1 %)
823	Roche	133 U/l	109 - 157	18 %	8	8 (100.0 %)
824	Selectra Pro	175 U/l	144 - 207	18 %	9	8 (88.9 %)
835	Spotchem D-Concept	63 U/l	51 - 74	18 %	397	377 (95.0 %)
808	Spotchem SP-4430	22 U/l	13 - 31	18 %	61	57 (93.4 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>Amilasi pancreatica</b>					
921 Abbott	93 U/l	76 - 109	18 %	12	12 (100.0 %)
904 Autolyser	100 U/l	82 - 118	18 %	11	11 (100.0 %)
923 Roche	109 U/l	89 - 128	18 %	14	14 (100.0 %)
<b>Bilirubina totale</b>					
1009 Abbott	25.4 µmol/l	20.9 - 30.0	18 %	14	14 (100.0 %)
1004 Autolyser	22.1 µmol/l	18.1 - 26.1	18 %	18	18 (100.0 %)
1010 Beckman	31.3 µmol/l	25.7 - 36.9	18 %	6	5 (83.3 %)
1032 Fuji Dri-Chem	28.0 µmol/l	23.0 - 33.0	18 %	827	800 (96.7 %)
1013 Piccolo	25.0 µmol/l	20.5 - 29.5	18 %	58	52 (89.7 %)
1005 Reflotron	24.8 µmol/l	20.3 - 29.3	18 %	4	4 (100.0 %)
1023 Roche	24.1 µmol/l	19.8 - 28.5	18 %	28	28 (100.0 %)
1024 Selectra Pro	18.5 µmol/l	15.2 - 21.8	18 %	15	10 (66.7 %)
1007 Siemens	30.2 µmol/l	24.7 - 35.6	18 %	6	6 (100.0 %)
1014 Skyla	31.8 µmol/l	26.1 - 37.5	18 %	5	5 (100.0 %)
1035 Spotchem D-Concept	23.3 µmol/l	19.1 - 27.5	18 %	426	418 (98.1 %)
1008 Spotchem SP-4430	29.0 µmol/l	23.8 - 34.2	18 %	82	76 (92.7 %)
<b>Bilirubina diretto</b>					
1031 Autolyser	18.5 µmol/l	15.2 - 21.8	18 %	8	8 (100.0 %)
1033 Fuji Dri-Chem	14.1 µmol/l	11.6 - 16.6	18 %	21	18 (85.7 %)
<b>Calcio</b>					
1100 Beckman	2.02 mmol/l	1.84 - 2.20	9 %	5	5 (100.0 %)
1129 Siemens	1.98 mmol/l	1.80 - 2.16	9 %	6	6 (100.0 %)
1109 Abbott	1.99 mmol/l	1.81 - 2.17	9 %	14	14 (100.0 %)
1104 Autolyser	2.07 mmol/l	1.88 - 2.26	9 %	9	9 (100.0 %)
1132 Fuji Dri-Chem	2.04 mmol/l	1.86 - 2.22	9 %	296	285 (96.3 %)
1113 Piccolo	1.99 mmol/l	1.81 - 2.17	9 %	51	49 (96.1 %)
1123 Roche	1.99 mmol/l	1.81 - 2.17	9 %	29	29 (100.0 %)
1124 Selectra Pro	2.00 mmol/l	1.82 - 2.18	9 %	6	6 (100.0 %)
1135 Spotchem D-Concept	1.42 mmol/l	1.24 - 1.60	9 %	78	76 (97.4 %)
1108 Spotchem SP-4430	1.79 mmol/l	1.61 - 1.97	9 %	15	11 (73.3 %)
<b>Calcium ISE</b>					
4694 iStat Chem8	0.77 mmol/l	0.68 - 0.86	12 %	7	6 (85.7 %)
<b>Cloruri</b>					
1230 Abbott	105 mmol/l	98 - 111	6 %	16	16 (100.0 %)
1200 Beckman	102 mmol/l	96 - 108	6 %	5	5 (100.0 %)
1232 Fuji Dri-Chem	110 mmol/l	103 - 117	6 %	933	908 (97.3 %)
1213 Piccolo	105 mmol/l	98 - 111	6 %	27	27 (100.0 %)
1223 Roche	103 mmol/l	97 - 109	6 %	20	18 (90.0 %)
1209 Siemens	106 mmol/l	100 - 112	6 %	6	6 (100.0 %)
1235 Spotchem D-Concept	113 mmol/l	106 - 120	6 %	427	417 (97.7 %)
1208 Spotchem EL-SE 1520	118 mmol/l	111 - 125	6 %	76	68 (89.5 %)
4693 iStat Chem8	112 mmol/l	105 - 119	6 %	7	7 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>Colesterolo</b>					
1310 Beckman	4.43 mmol/l	3.98 - 4.87	10 %	4	4 (100.0 %)
1311 Siemens	4.20 mmol/l	3.78 - 4.62	10 %	5	5 (100.0 %)
1309 Abbott	4.27 mmol/l	3.84 - 4.69	10 %	15	15 (100.0 %)
1304 Autolyser	4.15 mmol/l	3.73 - 4.56	10 %	21	20 (95.2 %)
1332 Fuji Dri-Chem	3.93 mmol/l	3.54 - 4.32	10 %	975	937 (96.1 %)
1313 Piccolo	4.14 mmol/l	3.72 - 4.55	10 %	29	28 (96.6 %)
1305 Reflotron	3.90 mmol/l	3.51 - 4.28	10 %	8	7 (87.5 %)
1323 Roche	4.19 mmol/l	3.77 - 4.61	10 %	27	27 (100.0 %)
1324 Selectra Pro	4.26 mmol/l	3.84 - 4.69	10 %	14	13 (92.9 %)
1335 Spotchem D-Concept	4.17 mmol/l	3.75 - 4.59	10 %	452	437 (96.7 %)
1308 Spotchem SP-4430	4.00 mmol/l	3.60 - 4.40	10 %	89	86 (96.6 %)
1320 Cholestech LDX	4.27 mmol/l	3.84 - 4.69	10 %	280	270 (96.4 %)
1399 altro	3.55 mmol/l	3.20 - 3.91	10 %	4	4 (100.0 %)
<b>Colesterolo HDL</b>					
1403 Siemens	1.72 mmol/l	1.36 - 2.08	21 %	4	4 (100.0 %)
1428 Abbott	1.01 mmol/l	0.80 - 1.22	21 %	14	14 (100.0 %)
1404 Autolyser	1.66 mmol/l	1.32 - 2.01	21 %	21	21 (100.0 %)
1410 Beckman	1.18 mmol/l	0.93 - 1.42	21 %	8	8 (100.0 %)
1432 Fuji Dri-Chem	1.48 mmol/l	1.17 - 1.79	21 %	940	926 (98.5 %)
1413 Piccolo	1.10 mmol/l	0.87 - 1.33	21 %	29	29 (100.0 %)
1423 Roche	0.97 mmol/l	0.77 - 1.18	21 %	24	23 (95.8 %)
1415 Selectra Pro	1.76 mmol/l	1.39 - 2.12	21 %	14	12 (85.7 %)
1435 Spotchem D-Concept	0.79 mmol/l	0.62 - 0.96	21 %	435	403 (92.6 %)
1408 Spotchem SP-4430	0.72 mmol/l	0.57 - 0.87	21 %	80	76 (95.0 %)
1420 Cholestech LDX	1.06 mmol/l	0.84 - 1.28	21 %	280	256 (91.4 %)
<b>Creatina chinasi</b>					
1507 Beckman	173 U/l	142 - 204	18 %	5	5 (100.0 %)
1511 Abbott	173 U/l	142 - 204	18 %	14	14 (100.0 %)
1504 Autolyser	185 U/l	151 - 218	18 %	18	18 (100.0 %)
1532 Fuji Dri-Chem	181 U/l	149 - 214	18 %	658	645 (98.0 %)
1513 Piccolo	174 U/l	142 - 205	18 %	18	18 (100.0 %)
1523 Roche	166 U/l	136 - 196	18 %	26	26 (100.0 %)
1524 Selectra Pro	174 U/l	142 - 205	18 %	10	10 (100.0 %)
1528 Siemens	160 U/l	131 - 189	18 %	6	6 (100.0 %)
1535 Spotchem D-Concept	150 U/l	123 - 177	18 %	319	315 (98.7 %)
1508 Spotchem SP-4430	155 U/l	127 - 183	18 %	45	45 (100.0 %)
<b>Colesterolo LDL</b>					
1438 Autolyser	2.7 mmol/l	2.3 - 3.2	18 %	10	8 (80.0 %)
1442 Selectra	2.7 mmol/l	2.2 - 3.1	18 %	4	3 (75.0 %)
1430 Chimica umida	2.1 mmol/l	1.7 - 2.4	18 %	14	14 (100.0 %)
1431 Roche, Cobas	2.8 mmol/l	2.3 - 3.3	18 %	15	15 (100.0 %)
1439 Beckman	2.8 mmol/l	2.3 - 3.3	18 %	4	4 (100.0 %)
<b>Ferro</b>					
1709 Abbott	37 µmol/l	30 - 45	20 %	11	11 (100.0 %)
1723 Roche	37 µmol/l	29 - 44	20 %	19	19 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto		Tolleranza		Numero	Adempite
<b>Gamma-GT</b>						
1810 Abbott	42	U/l	34 - 49	18 %	15	15 (100.0 %)
1823 Cobas	38	U/l	30 - 46	18 %	30	29 (96.7 %)
1804 Autolyser	39	U/l	31 - 47	18 %	21	21 (100.0 %)
1812 Beckman	40	U/l	33 - 47	18 %	6	6 (100.0 %)
1832 Fuji Dri-Chem	47	U/l	38 - 55	18 %	1128	1110 (98.4 %)
1813 Piccolo	36	U/l	28 - 44	18 %	56	56 (100.0 %)
1805 Reflotron	47	U/l	39 - 56	18 %	13	11 (84.6 %)
1801 Selectra Pro	37	U/l	29 - 45	18 %	15	15 (100.0 %)
1811 Siemens	43	U/l	35 - 51	18 %	5	5 (100.0 %)
1814 Skyla	43	U/l	35 - 51	18 %	5	5 (100.0 %)
1835 Spotchem D-Concept	44	U/l	36 - 52	18 %	613	600 (97.9 %)
1808 Spotchem SP-4430	47	U/l	39 - 56	18 %	135	133 (98.5 %)
<b>Glucosio</b>						
1909 Abbott	6.1	mmol/l	5.5 - 6.6	9 %	15	15 (100.0 %)
1904 Autolyser	6.1	mmol/l	5.5 - 6.6	9 %	19	15 (78.9 %)
1932 Fuji Dri-Chem	5.8	mmol/l	5.3 - 6.3	9 %	1066	1053 (98.8 %)
1913 Piccolo	6.2	mmol/l	5.6 - 6.7	9 %	68	68 (100.0 %)
1905 Reflotron	5.9	mmol/l	5.4 - 6.5	9 %	12	11 (91.7 %)
1923 Roche	6.1	mmol/l	5.6 - 6.7	9 %	30	30 (100.0 %)
1924 Selectra Pro	6.3	mmol/l	5.8 - 6.9	9 %	16	15 (93.8 %)
1907 Siemens	6.4	mmol/l	5.8 - 7.0	9 %	6	6 (100.0 %)
1935 Spotchem D-Concept	6.1	mmol/l	5.6 - 6.7	9 %	575	555 (96.5 %)
1908 Spotchem SP-4430	6.3	mmol/l	5.7 - 6.8	9 %	118	113 (95.8 %)
4695 iStat Chem8	5.7	mmol/l	5.2 - 6.2	9 %	8	8 (100.0 %)
1920 Cholestech LDX	5.8	mmol/l	5.3 - 6.4	9 %	277	271 (97.8 %)
2087 Cobas Pulse	5.6	mmol/l	5.1 - 6.1	9 %	11	11 (100.0 %)
1925 Beckman	6.2	mmol/l	5.6 - 6.8	9 %	6	6 (100.0 %)
<b>Glucosio</b>						
2086 Accu-Chek Instant	5.7	mmol/l	5.2 - 6.3	9 %	117	115 (98.3 %)
2065 Accu-Chek Aviva	6.4	mmol/l	5.9 - 7.0	9 %	125	108 (86.4 %)
2070 Accu-Chek Inform 2	6.9	mmol/l	6.3 - 7.5	9 %	917	903 (98.5 %)
2085 Accu-Check Guide	5.6	mmol/l	5.1 - 6.1	9 %	331	316 (95.5 %)
2074 Contour XT	6.1	mmol/l	5.6 - 6.6	9 %	1409	1340 (95.1 %)
1914 Skyla	6.3	mmol/l	5.7 - 6.9	9 %	5	5 (100.0 %)
1928 Statstrip/Xpress	5.9	mmol/l	5.3 - 6.4	9 %	92	91 (98.9 %)
2021 Glucocard	8.3	mmol/l	7.6 - 9.0	9 %	9	8 (88.9 %)
2030 Hemocue 201+ P-equiv	7.7	mmol/l	7.0 - 8.4	9 %	118	113 (95.8 %)
2032 Hemocue 201RT P-equiv	7.7	mmol/l	7.0 - 8.4	9 %	131	129 (98.5 %)
2064 CardioChek	8.0	mmol/l	7.3 - 8.7	9 %	4	2 (50.0 %)
2069 Freestyle Freedom li	6.4	mmol/l	5.8 - 7.0	9 %	4	4 (100.0 %)
2084 Contour NEXT	5.8	mmol/l	5.3 - 6.3	9 %	42	38 (90.5 %)



## Valori ideali MQ 2023 - 4

		Giusto	Tolleranza		Numero	Adempite	
<b>Glucosio</b>							
2028	Hemocue 201+ (alt)	7.6	mmol/l	6.9 - 8.3	9 %	31	30 (96.8 %)
2057	OneTouch Verio	6.0	mmol/l	5.5 - 6.5	9 %	24	23 (95.8 %)
2066	Contour 2 (5s)	4.9	mmol/l	4.4 - 5.3	9 %	8	7 (87.5 %)
2072	Healthpro	10.2	mmol/l	9.2 - 11.1	9 %	22	18 (81.8 %)
2078	Mylife UNIO	7.3	mmol/l	6.6 - 8.0	9 %	441	433 (98.2 %)
2031	mylife Pura	6.7	mmol/l	6.1 - 7.3	9 %	104	99 (95.2 %)
2076	Alpha Check	9.2	mmol/l	8.3 - 10.0	9 %	10	9 (90.0 %)
<b>Acido urico</b>							
2109	Abbott	272	µmol/l	239 - 305	12 %	15	15 (100.0 %)
2104	Autolyser	278	µmol/l	245 - 311	12 %	19	18 (94.7 %)
2107	Beckman	280	µmol/l	246 - 314	12 %	6	6 (100.0 %)
2132	Fuji Dri-Chem	309	µmol/l	272 - 347	12 %	1048	1035 (98.8 %)
2113	Piccolo	203	µmol/l	179 - 228	12 %	36	31 (86.1 %)
2123	Roche	269	µmol/l	237 - 301	12 %	27	27 (100.0 %)
2124	Selectra Pro	264	µmol/l	232 - 295	12 %	16	16 (100.0 %)
2199	Siemens	272	µmol/l	239 - 305	12 %	6	6 (100.0 %)
2105	Reflotron	298	µmol/l	262 - 334	12 %	7	7 (100.0 %)
2114	Skyla	276	µmol/l	243 - 309	12 %	5	5 (100.0 %)
2135	Spotchem D-Concept	257	µmol/l	226 - 288	12 %	571	559 (97.9 %)
2108	Spotchem SP-4430	234	µmol/l	206 - 262	12 %	115	111 (96.5 %)
<b>Urea</b>							
2211	Abbott	7.7	mmol/l	6.5 - 8.8	15 %	14	14 (100.0 %)
2204	Autolyser	7.8	mmol/l	6.6 - 8.9	15 %	15	15 (100.0 %)
2299	Beckman	7.4	mmol/l	6.2 - 8.5	15 %	8	8 (100.0 %)
2232	Fuji Dri-Chem	7.5	mmol/l	6.4 - 8.7	15 %	629	624 (99.2 %)
2213	Piccolo	6.5	mmol/l	5.6 - 7.5	15 %	61	61 (100.0 %)
2205	Reflotron	7.5	mmol/l	6.3 - 8.6	15 %	5	5 (100.0 %)
2223	Roche	7.2	mmol/l	6.1 - 8.3	15 %	27	27 (100.0 %)
2224	Selectra Pro	7.5	mmol/l	6.4 - 8.6	15 %	12	12 (100.0 %)
2209	Siemens	7.6	mmol/l	6.5 - 8.7	15 %	7	7 (100.0 %)
2214	Skyla	7.5	mmol/l	6.4 - 8.6	15 %	5	5 (100.0 %)
2235	Spotchem D-Concept	6.8	mmol/l	5.8 - 7.8	15 %	328	315 (96.0 %)
2208	Spotchem SP-4430	6.7	mmol/l	5.7 - 7.7	15 %	60	56 (93.3 %)
4696	iStat Chem8	9.8	mmol/l	8.3 - 11.3	15 %	9	8 (88.9 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>Potassio</b>					
2630 Abbott	3.01 mmol/l	2.81 - 3.21	6 %	14	14 (100.0 %)
2636 Autolyser	2.82 mmol/l	2.62 - 3.02	6 %	20	17 (85.0 %)
2607 Beckman	3.00 mmol/l	2.80 - 3.20	6 %	6	6 (100.0 %)
2632 Fuji Dri-Chem	3.00 mmol/l	2.80 - 3.20	6 %	1102	1043 (94.6 %)
2613 Piccolo	3.20 mmol/l	3.00 - 3.40	6 %	39	23 (59.0 %)
2623 Roche	3.03 mmol/l	2.83 - 3.23	6 %	29	29 (100.0 %)
2699 Siemens	3.00 mmol/l	2.80 - 3.20	6 %	6	6 (100.0 %)
2605 Reflotron	3.30 mmol/l	3.10 - 3.50	6 %	9	4 (44.4 %)
2635 Spotchem D-Concept	2.73 mmol/l	2.53 - 2.93	6 %	546	536 (98.2 %)
2608 Spotchem EL-SE 1520	2.72 mmol/l	2.52 - 2.92	6 %	100	96 (96.0 %)
4692 iStat Chem8	2.97 mmol/l	2.77 - 3.17	6 %	11	11 (100.0 %)
<b>Creatinina</b>					
2701 Abbott	177 µmol/l	145 - 209	18 %	14	14 (100.0 %)
2704 Autolyser	169 µmol/l	138 - 199	18 %	21	21 (100.0 %)
2707 Beckman	179 µmol/l	147 - 211	18 %	6	6 (100.0 %)
2732 Fuji Dri-Chem	161 µmol/l	132 - 190	18 %	1159	1147 (99.0 %)
2719 Piccolo	178 µmol/l	146 - 210	18 %	65	65 (100.0 %)
2705 Reflotron	188 µmol/l	154 - 221	18 %	23	22 (95.7 %)
2723 Roche	175 µmol/l	144 - 207	18 %	30	30 (100.0 %)
2724 Selectra Pro	166 µmol/l	136 - 196	18 %	16	16 (100.0 %)
2799 Siemens	179 µmol/l	146 - 211	18 %	6	6 (100.0 %)
2726 Skyla	161 µmol/l	132 - 190	18 %	5	5 (100.0 %)
2735 Spotchem D-Concept	81 µmol/l	66 - 96	18 %	621	611 (98.4 %)
2708 Spotchem SP-4430	79 µmol/l	65 - 93	18 %	158	156 (98.7 %)
4860 EPOC	174 µmol/l	142 - 205	18 %	11	7 (63.6 %)
<b>Creatinina E</b>					
4697 iStat Chem8	173 µmol/l	142 - 204	18 %	37	37 (100.0 %)
6916 ABL700/800	178 µmol/l	146 - 210	18 %	13	12 (92.3 %)
<b>eGFR CKD-EPI</b>					
2740 Chimica umida	40	28 - 52	30 %	55	53 (96.4 %)
2741 Reflotron	40	28 - 52	30 %	8	7 (87.5 %)
2742 Fuji Dri-Chem	45	32 - 59	30 %	407	391 (96.1 %)
2743 Spotchem	104	73 - 135	30 %	281	271 (96.4 %)
<b>eGFR Cockcroft-Gault</b>					
2752 Fuji Dri-Chem	49	34 - 64	30 %	39	38 (97.4 %)
2753 Spotchem	108	75 - 140	30 %	14	14 (100.0 %)
<b>eGFR MDRD</b>					
2762 Fuji Dri-Chem	48	33 - 62	30 %	4	3 (75.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>LDH</b>					
2809 Abbott	129 U/l	105 - 152	18 %	14	14 (100.0 %)
2804 Autolyser	135 U/l	111 - 159	18 %	7	6 (85.7 %)
2807 Beckman	132 U/l	108 - 155	18 %	4	4 (100.0 %)
2832 Fuji Dri-Chem	114 U/l	93 - 134	18 %	117	115 (98.3 %)
2813 Piccolo	110 U/l	90 - 130	18 %	4	4 (100.0 %)
2823 Roche	137 U/l	112 - 162	18 %	27	27 (100.0 %)
2811 Siemens	138 U/l	113 - 163	18 %	5	5 (100.0 %)
2835 Spotchem D-Concept	71 U/l	58 - 84	18 %	42	30 (71.4 %)
2808 Spotchem SP-4430	69 U/l	56 - 81	18 %	14	12 (85.7 %)
<b>Magnesio</b>					
2908 Spotchem SP-4430	0.55 mmol/l	0.46 - 0.64	12 %	5	5 (100.0 %)
2909 Abbott	0.63 mmol/l	0.54 - 0.72	12 %	13	13 (100.0 %)
2910 Beckman	0.66 mmol/l	0.57 - 0.75	12 %	5	5 (100.0 %)
2932 Fuji Dri-Chem	0.70 mmol/l	0.62 - 0.79	12 %	76	72 (94.7 %)
2913 Piccolo	0.65 mmol/l	0.56 - 0.74	12 %	4	4 (100.0 %)
2923 Roche	0.64 mmol/l	0.55 - 0.73	12 %	22	22 (100.0 %)
2900 Siemens	0.63 mmol/l	0.54 - 0.72	12 %	5	5 (100.0 %)
2935 Spotchem D-Concept	0.48 mmol/l	0.39 - 0.57	12 %	36	36 (100.0 %)
<b>Sodio</b>					
3025 Autolyser	147 mmol/l	138 - 156	6 %	9	9 (100.0 %)
3030 Abbott	144 mmol/l	135 - 153	6 %	15	15 (100.0 %)
3009 Beckman	142 mmol/l	133 - 151	6 %	6	6 (100.0 %)
3032 Fuji Dri-Chem	146 mmol/l	137 - 155	6 %	1012	991 (97.9 %)
3023 Roche	143 mmol/l	134 - 151	6 %	29	29 (100.0 %)
3000 Siemens	144 mmol/l	135 - 153	6 %	6	6 (100.0 %)
3035 Spotchem D-Concept	139 mmol/l	130 - 147	6 %	475	474 (99.8 %)
3008 Spotchem EL-SE 1520	139 mmol/l	131 - 147	6 %	84	84 (100.0 %)
3013 Piccolo	140 mmol/l	132 - 149	6 %	37	36 (97.3 %)
4691 iStat Chem8	143 mmol/l	134 - 151	6 %	10	10 (100.0 %)
<b>Fosfati</b>					
3109 Abbott	1.1 mmol/l	0.9 - 1.2	15 %	13	13 (100.0 %)
3100 Beckman	1.1 mmol/l	0.9 - 1.3	15 %	4	4 (100.0 %)
3132 Fuji Dri-Chem	1.1 mmol/l	1.0 - 1.3	15 %	75	72 (96.0 %)
3113 Piccolo	1.4 mmol/l	1.2 - 1.6	15 %	6	5 (83.3 %)
3123 Roche	1.1 mmol/l	0.9 - 1.3	15 %	25	24 (96.0 %)
3108 Siemens	1.1 mmol/l	1.0 - 1.3	15 %	9	9 (100.0 %)
3135 Spotchem D-Concept	1.2 mmol/l	1.0 - 1.4	15 %	14	14 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>Proteine totali</b>					
3209 Abbott	57.6 g/l	50.7 - 64.5	12 %	13	13 (100.0 %)
3225 Beckman	58.0 g/l	51.0 - 65.0	12 %	4	4 (100.0 %)
3232 Fuji Dri-Chem	55.8 g/l	49.1 - 62.5	12 %	189	184 (97.4 %)
3213 Piccolo	58.8 g/l	51.8 - 65.9	12 %	49	47 (95.9 %)
3223 Roche	58.6 g/l	51.5 - 65.6	12 %	25	25 (100.0 %)
3224 Selectra Pro	57.3 g/l	50.4 - 64.2	12 %	8	7 (87.5 %)
3227 Siemens	61.0 g/l	53.7 - 68.3	12 %	6	6 (100.0 %)
3214 Skyla	62.0 g/l	54.6 - 69.4	12 %	5	5 (100.0 %)
3235 Spotchem D-Concept	55.0 g/l	48.4 - 61.6	12 %	187	178 (95.2 %)
3208 Spotchem SP-4430	60.3 g/l	53.0 - 67.5	12 %	27	25 (92.6 %)
<b>Transaminasi GOT/AST</b>					
3313 Abbott	72 U/l	59 - 85	18 %	15	15 (100.0 %)
3304 Autolyser	66 U/l	54 - 77	18 %	21	21 (100.0 %)
3314 Beckman	73 U/l	60 - 86	18 %	6	6 (100.0 %)
3332 Fuji Dri-Chem	68 U/l	56 - 80	18 %	1138	1118 (98.2 %)
3319 Piccolo	65 U/l	53 - 77	18 %	71	69 (97.2 %)
3305 Reflotron	85 U/l	70 - 101	18 %	17	13 (76.5 %)
3323 Roche	71 U/l	58 - 84	18 %	30	30 (100.0 %)
3324 Selectra Pro	73 U/l	60 - 86	18 %	16	16 (100.0 %)
3399 Siemens	70 U/l	57 - 83	18 %	6	6 (100.0 %)
3320 Skyla	70 U/l	57 - 83	18 %	5	5 (100.0 %)
3435 Spotchem D-Concept	43 U/l	35 - 50	18 %	613	606 (98.9 %)
3308 Spotchem SP-4430	40 U/l	33 - 48	18 %	147	145 (98.6 %)
<b>Transaminasi GPT/ALT</b>					
3413 Abbott	50 U/l	41 - 60	18 %	15	15 (100.0 %)
3404 Autolyser	55 U/l	45 - 65	18 %	21	21 (100.0 %)
3414 Beckman	58 U/l	48 - 69	18 %	6	6 (100.0 %)
3432 Fuji Dri-Chem	60 U/l	49 - 71	18 %	1151	1126 (97.8 %)
3419 Piccolo	54 U/l	44 - 64	18 %	69	67 (97.1 %)
3405 Reflotron	50 U/l	41 - 59	18 %	12	10 (83.3 %)
3423 Roche	56 U/l	46 - 66	18 %	30	30 (100.0 %)
3424 Selectra Pro	62 U/l	51 - 74	18 %	16	14 (87.5 %)
3499 Siemens	60 U/l	49 - 71	18 %	6	6 (100.0 %)
3420 Skyla	60 U/l	49 - 71	18 %	5	5 (100.0 %)
3335 Spotchem D-Concept	36 U/l	30 - 42	18 %	620	588 (94.8 %)
3408 Spotchem SP-4430	35 U/l	29 - 42	18 %	150	133 (88.7 %)

## Valori ideali MQ 2023 - 4

		Giusto	Tolleranza		Numero	Adempite
<b>Trigliceridi</b>						
3525	Beckman	2.36 mmol/l	1.93 - 2.78	18 %	4	4 (100.0 %)
3511	Abbott	2.21 mmol/l	1.81 - 2.61	18 %	13	13 (100.0 %)
3504	Autolyser	2.13 mmol/l	1.75 - 2.51	18 %	21	21 (100.0 %)
3532	Fuji Dri-Chem	2.75 mmol/l	2.25 - 3.24	18 %	957	945 (98.7 %)
3513	Piccolo	2.56 mmol/l	2.10 - 3.02	18 %	27	26 (96.3 %)
3505	Reflotron	3.32 mmol/l	2.72 - 3.92	18 %	5	4 (80.0 %)
3523	Roche	2.32 mmol/l	1.91 - 2.74	18 %	26	26 (100.0 %)
3524	Selectra Pro	2.35 mmol/l	1.92 - 2.77	18 %	13	12 (92.3 %)
3509	Siemens	2.30 mmol/l	1.89 - 2.71	18 %	7	7 (100.0 %)
3535	Spotchem D-Concept	1.45 mmol/l	1.19 - 1.71	18 %	440	431 (98.0 %)
3508	Spotchem SP-4430	1.43 mmol/l	1.17 - 1.68	18 %	83	81 (97.6 %)
3520	Cholestech LDX	2.26 mmol/l	1.85 - 2.67	18 %	280	277 (98.9 %)
<b>Litio</b>						
6521	Cobas Integra 800/40	1.11 mmol/l	0.94 - 1.28	15 %	4	4 (100.0 %)
6520	altro	1.12 mmol/l	0.95 - 1.28	15 %	19	19 (100.0 %)
<b>Laktat</b>						
4685	tutti	2.34 mmol/l	1.92 - 2.76	18 %	20	20 (100.0 %)
<b>K03 HbA1c</b>						
<b>HbA1c campione A</b>						
4703	Abbott	7.9 %	7.1 - 8.6	9 %	6	6 (100.0 %)
4756	Roche, Cobas	7.9 %	7.2 - 8.6	9 %	17	17 (100.0 %)
4754	HPLC	7.9 %	7.2 - 8.6	9 %	9	9 (100.0 %)
4701	Afinion	7.8 %	7.1 - 8.5	9 %	508	501 (98.6 %)
4710	Cobas b101	7.4 %	6.7 - 8.1	9 %	197	188 (95.4 %)
4752	DCA2000/Vantage	7.8 %	7.1 - 8.5	9 %	129	122 (94.6 %)
4771	Celltac chemi	7.4 %	6.8 - 8.1	9 %	19	19 (100.0 %)
4726	NycoCard	8.0 %	7.2 - 8.7	9 %	8	7 (87.5 %)
4708	Eurolyser	8.0 %	7.3 - 8.7	9 %	9	9 (100.0 %)
4767	A1c Now	7.2 %	6.6 - 7.9	9 %	217	185 (85.3 %)
4769	AFIAS	7.2 %	6.6 - 7.8	9 %	97	80 (82.5 %)
4753	Andere	7.8 %	7.1 - 8.5	9 %	21	19 (90.5 %)
4772	Spinit	8.0 %	7.3 - 8.7	9 %	14	11 (78.6 %)
<b>K04 Gas sanguini</b>						
<b>pCO2</b>						
6901	ABL700/800	5.36 kPa	4.72 - 6.00	12 %	107	105 (98.1 %)
6971	ABL80 FLEX CO-OX / O	6.01 kPa	5.29 - 6.73	12 %	5	4 (80.0 %)
6951	ABL90 FLEX / PLUS	5.57 kPa	4.90 - 6.24	12 %	115	114 (99.1 %)
4046	Cobas b 123	5.60 kPa	4.93 - 6.28	12 %	15	15 (100.0 %)
4045	Cobas b 221	5.72 kPa	5.03 - 6.40	12 %	7	7 (100.0 %)
4002	GEM	5.70 kPa	5.02 - 6.38	12 %	8	8 (100.0 %)
4051	iStat	4.81 kPa	4.23 - 5.39	12 %	44	43 (97.7 %)
4851	EPOC	5.28 kPa	4.65 - 5.92	12 %	54	50 (92.6 %)
4048	IL	5.70 kPa	5.02 - 6.38	12 %	4	4 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>pO2</b>					
6902 ABL700/800	14.60 kPa	12.41 - 16.79	15 %	106	101 (95.3 %)
6972 ABL80 FLEX CO-OX / O	14.26 kPa	12.12 - 16.40	15 %	5	5 (100.0 %)
6952 ABL90 FLEX / PLUS	13.56 kPa	11.53 - 15.59	15 %	116	111 (95.7 %)
4146 Cobas b 123	14.37 kPa	12.21 - 16.53	15 %	17	17 (100.0 %)
4145 Cobas b 221	16.30 kPa	13.85 - 18.74	15 %	4	4 (100.0 %)
4003 GEM	14.50 kPa	12.33 - 16.68	15 %	8	8 (100.0 %)
4151 iStat	15.92 kPa	13.53 - 18.30	15 %	41	39 (95.1 %)
4852 EPOC	14.79 kPa	12.57 - 17.01	15 %	54	50 (92.6 %)
4148 IL	14.60 kPa	12.41 - 16.79	15 %	4	4 (100.0 %)
<b>pH</b>					
6900 ABL700/800	7.39	7.33 - 7.46	1 %	107	105 (98.1 %)
6970 ABL80 FLEX CO-OX / O	7.37	7.30 - 7.44	1 %	5	5 (100.0 %)
6950 ABL90 FLEX / PLUS	7.40	7.33 - 7.46	1 %	116	115 (99.1 %)
4246 Cobas b 123	7.41	7.34 - 7.48	1 %	16	16 (100.0 %)
4245 Cobas b 221	7.40	7.33 - 7.47	1 %	7	7 (100.0 %)
4001 GEM	7.40	7.33 - 7.47	1 %	8	8 (100.0 %)
4251 iStat	7.43	7.36 - 7.49	1 %	45	44 (97.8 %)
4850 EPOC	7.39	7.32 - 7.46	1 %	53	52 (98.1 %)
4248 IL	7.42	7.35 - 7.49	1 %	5	5 (100.0 %)
<b>Glucosio GS</b>					
4007 GEM	5.8 mmol/l	5.3 - 6.3	9 %	4	4 (100.0 %)
4346 Cobas b 123	5.3 mmol/l	4.9 - 5.8	9 %	10	9 (90.0 %)
4351 iStat	5.4 mmol/l	4.9 - 5.9	9 %	10	10 (100.0 %)
4856 EPOC	5.6 mmol/l	5.1 - 6.1	9 %	42	40 (95.2 %)
6914 ABL700/800	5.7 mmol/l	5.2 - 6.2	9 %	98	98 (100.0 %)
6964 ABL90 FLEX / PLUS	5.6 mmol/l	5.1 - 6.1	9 %	102	102 (100.0 %)
<b>Emoglobina BG</b>					
6903 ABL700/800	128.0 g/l	116.5 - 139.5	9 %	98	98 (100.0 %)
6953 ABL90 FLEX / PLUS	128.1 g/l	116.6 - 139.7	9 %	103	101 (98.1 %)
6973 ABL80 FLEX CO-OX / O	126.5 g/l	115.1 - 137.9	9 %	6	6 (100.0 %)
<b>Potassio BG</b>					
4005 GEM	3.7 mmol/l	3.5 - 3.9	6 %	5	5 (100.0 %)
4546 Cobas b 123	3.7 mmol/l	3.5 - 4.0	6 %	20	20 (100.0 %)
4551 iStat	3.8 mmol/l	3.6 - 4.0	6 %	17	17 (100.0 %)
4854 EPOC	3.6 mmol/l	3.4 - 3.9	6 %	46	45 (97.8 %)
6910 ABL700/800	3.8 mmol/l	3.6 - 4.0	6 %	99	99 (100.0 %)
6960 ABL90 FLEX / PLUS	3.7 mmol/l	3.5 - 3.9	6 %	110	110 (100.0 %)
<b>Sodio BG</b>					
4004 GEM	138.0 mmol/l	129.7 - 146.3	6 %	6	6 (100.0 %)
4646 Cobas b 123	138.5 mmol/l	130.1 - 146.8	6 %	20	20 (100.0 %)
4651 iStat	142.5 mmol/l	133.9 - 151.0	6 %	17	17 (100.0 %)
4853 EPOC	141.0 mmol/l	132.5 - 149.4	6 %	44	44 (100.0 %)
6911 ABL700/800	140.8 mmol/l	132.3 - 149.2	6 %	98	98 (100.0 %)
6961 ABL90 FLEX / PLUS	141.2 mmol/l	132.7 - 149.7	6 %	109	109 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>Cloruro-BG</b>					
4661 Cobas b 123	100.5 mmol/l	94.5 - 106.5	6 %	9	9 (100.0 %)
4849 EPOC	100.8 mmol/l	94.7 - 106.8	6 %	15	14 (93.3 %)
6913 ABL700/800	98.0 mmol/l	92.1 - 103.9	6 %	93	90 (96.8 %)
6963 ABL90 FLEX / PLUS	97.0 mmol/l	91.1 - 102.8	6 %	103	103 (100.0 %)
<b>Calcio-BG</b>					
4006 GEM	0.43 mmol/l	0.19 - 0.67	12 %	5	5 (100.0 %)
4670 Cobas b123	0.38 mmol/l	0.14 - 0.62	12 %	12	12 (100.0 %)
4671 Roche, Cobas	0.45 mmol/l	0.21 - 0.69	12 %	7	7 (100.0 %)
4673 iStat	0.48 mmol/l	0.24 - 0.72	12 %	13	13 (100.0 %)
4855 EPOC	0.50 mmol/l	0.26 - 0.74	12 %	40	39 (97.5 %)
6912 ABL700/800	0.53 mmol/l	0.29 - 0.77	12 %	98	98 (100.0 %)
6962 ABL90 FLEX / PLUS	0.52 mmol/l	0.28 - 0.76	12 %	107	107 (100.0 %)
<b>FHHb</b>					
6966 ABL90 FLEX / PLUS	3.000 %	2.400 - 3.600	20 %	9	9 (100.0 %)
<b>Lattato-BG</b>					
4008 GEM	1.90 mmol/l	1.56 - 2.24	18 %	5	5 (100.0 %)
4680 Cobas b123	1.95 mmol/l	1.60 - 2.30	18 %	8	8 (100.0 %)
4681 Roche, Cobas	1.90 mmol/l	1.56 - 2.24	18 %	6	6 (100.0 %)
4683 IL	1.70 mmol/l	1.39 - 2.01	18 %	4	4 (100.0 %)
4857 EPOC	1.83 mmol/l	1.50 - 2.16	18 %	40	38 (95.0 %)
4859 iStat	1.62 mmol/l	1.33 - 1.91	18 %	15	15 (100.0 %)
6915 ABL700/800	1.69 mmol/l	1.39 - 2.00	18 %	100	100 (100.0 %)
6965 ABL90 FLEX / PLUS	1.59 mmol/l	1.30 - 1.88	18 %	110	110 (100.0 %)
<b>sO2 OR</b>					
4751 iStat	98.92 %	79.143 - 118.714	20 %	15	14 (93.3 %)
6904 ABL700/800	96.84 %	77.475 - 116.212	20 %	88	88 (100.0 %)
6954 ABL90 FLEX / PLUS	96.91 %	77.534 - 116.300	20 %	94	94 (100.0 %)
6974 ABL80 FLEX CO-OX / O	96.85 %	77.480 - 116.220	20 %	4	4 (100.0 %)
<b>FO2Hb OR</b>					
6905 ABL700/800	92.36 %	73.891 - 110.837	20 %	86	86 (100.0 %)
6955 ABL90 FLEX / PLUS	92.16 %	73.732 - 110.598	20 %	95	95 (100.0 %)
6975 ABL80 FLEX CO-OX / O	92.10 %	73.680 - 110.520	20 %	4	4 (100.0 %)
<b>FCOHb OR</b>					
6906 ABL700/800	2.743 %	2.194 - 3.291	20 %	88	81 (92.0 %)
6956 ABL90 FLEX / PLUS	2.901 %	2.321 - 3.481	20 %	95	94 (98.9 %)
6976 ABL80 FLEX CO-OX / O	2.800 %	2.240 - 3.360	20 %	4	4 (100.0 %)
<b>FMetHb OR</b>					
6907 ABL700/800	1.992 %	1.594 - 2.390	20 %	88	87 (98.9 %)
6957 ABL90 FLEX / PLUS	2.020 %	1.616 - 2.424	20 %	95	94 (98.9 %)
6977 ABL80 FLEX CO-OX / O	2.100 %	1.680 - 2.520	20 %	4	4 (100.0 %)
<b>FHbF OR</b>					
6958 ABL90 FLEX / PLUS	77.72 %	62.180 - 93.270	20 %	40	40 (100.0 %)
<b>Bilirubin OR</b>					
6909 ABL700/800	294.5 µmol/l	241.5 - 347.5	18 %	6	6 (100.0 %)
6959 ABL90 FLEX / PLUS	295.8 µmol/l	242.5 - 349.0	18 %	32	32 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>U01 Urina quantitativa</b>					
<b>Amilasi-urine</b>					
4821 IFCC	107 U/l	80 - 133	25 %	4	4 (100.0 %)
<b>Amilasi pancr. urina</b>					
4921 IFCC	4.0 U/l	0.0 - 9.0	18 %	4	4 (100.0 %)
<b>Calcio-urine</b>					
5001 Roche, Cobas	2.15 mmol/l	1.96 - 2.35	9 %	12	12 (100.0 %)
5002 Abbott	2.02 mmol/l	1.84 - 2.20	9 %	7	7 (100.0 %)
5009 altro	2.11 mmol/l	1.92 - 2.30	9 %	7	7 (100.0 %)
<b>Cloro-urine</b>					
5109 Abbott	138 mmol/l	122 - 155	12 %	10	10 (100.0 %)
5110 Roche, Cobas	130 mmol/l	115 - 146	12 %	11	11 (100.0 %)
<b>Glucosio-urine</b>					
5309 Chimica umida	8.1 mmol/l	7.4 - 8.9	9 %	21	21 (100.0 %)
<b>Magnesio-urine</b>					
5709 Chimica umida	2.55 mmol/l	2.24 - 2.86	12 %	15	15 (100.0 %)
<b>Osmolalità-urine</b>					
6059 Cryoscopie	604 mosm/kg	544 - 665	10 %	19	19 (100.0 %)
<b>Fosforo-urine</b>					
6209 Chimica umida	11.6 mmol/l	9.9 - 13.4	15 %	23	23 (100.0 %)
<b>Potassio-urine</b>					
5630 tutti	49 mmol/l	42 - 56	15 %	32	32 (100.0 %)
<b>Proteina-urina</b>					
6301 Cobas/Roche	323.2 mg/l	274.8 - 371.7	15 %	18	18 (100.0 %)
6309 Abbott	394.1 mg/l	335.0 - 453.2	15 %	14	14 (100.0 %)
<b>Sodio-urine</b>					
5930 tutti	117 mmol/l	100 - 135	15 %	32	32 (100.0 %)
<b>Urea-urine</b>					
5509 Chimica umida	202 mmol/l	172 - 233	15 %	30	30 (100.0 %)
<b>Acido urico-urine</b>					
5409 Chimica umida	0.78 mmol/l	0.66 - 0.89	15 %	21	21 (100.0 %)
<b>Peso Specifico-urine</b>					
6460 Refraktometer	1.017	0.966 - 1.068	5 %	7	7 (100.0 %)
<b>G02 CoaguChek</b>					
<b>INR CoaguChek</b>					
3670 CoaguChek Pro II	1.4	1.2 - 1.6	15 %	803	795 (99.0 %)



## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>G01 Quick</b>					
<b>Quick OA</b>					
3638 Innovin	1.53	1.30 - 1.76	15 %	16	16 (100.0 %)
3668 Neoplastin R	1.57	1.33 - 1.80	15 %	16	16 (100.0 %)
3634 Neoplastin Plus	1.91	1.62 - 2.20	15 %	7	7 (100.0 %)
3632 STA-NeoPTimal	1.67	1.42 - 1.91	15 %	4	4 (100.0 %)
3643 Recombiplastin 2G	1.60	1.36 - 1.84	15 %	11	11 (100.0 %)
3699 altro	1.57	1.34 - 1.81	15 %	15	14 (93.3 %)
<b>Fibrinogeno OA</b>					
3964 Siemens Thrombin	1.10 g/l	0.94 - 1.27	15 %	9	9 (100.0 %)
3966 Stago/STA	1.16 g/l	0.99 - 1.33	15 %	19	19 (100.0 %)
3967 Fibrinogen Q.F.A.	1.20 g/l	1.02 - 1.38	15 %	7	7 (100.0 %)
3901 altro	1.17 g/l	0.99 - 1.35	15 %	5	5 (100.0 %)
<b>aPTT OA</b>					
3762 Actin FS	37.0 Sek	27.8 - 46.3	25 %	7	7 (100.0 %)
3763 Pathromtin SL	62.0 Sek	46.5 - 77.5	25 %	7	7 (100.0 %)
3764 Stago/STA	49.7 Sek	37.3 - 62.1	25 %	22	22 (100.0 %)
3765 aPTT-SP	39.6 Sek	29.7 - 49.5	25 %	7	7 (100.0 %)
3701 altro	39.5 Sek	29.7 - 49.4	25 %	11	10 (90.9 %)
<b>G03 Coagulazione</b>					
<b>Quick N</b>					
8138 Innovin	100 %	85 - 115	15 %	12	12 (100.0 %)
8132 Neoplastin R	95 %	81 - 109	15 %	13	13 (100.0 %)
8134 Neoplastin Plus	90 %	77 - 104	15 %	4	4 (100.0 %)
8131 STA-NeoPTimal	87 %	74 - 100	15 %	8	7 (87.5 %)
8146 Recombiplastin 2G	100 %	85 - 115	15 %	10	10 (100.0 %)
8142 altro	98 %	83 - 113	15 %	15	15 (100.0 %)
<b>Fibrinogeno N</b>					
8000 Siemens Thrombin	2.85 g/l	2.42 - 3.28	15 %	8	8 (100.0 %)
8003 Stago/STA	3.00 g/l	2.55 - 3.45	15 %	19	19 (100.0 %)
8004 Fibrinogen Q.F.A.	2.90 g/l	2.47 - 3.34	15 %	14	14 (100.0 %)
8006 Fib Clauss (IL)	2.91 g/l	2.47 - 3.35	15 %	5	5 (100.0 %)
<b>Faktor V</b>					
8151 tutti	94.5 %	70.9 - 118.1	25 %	8	8 (100.0 %)
<b>aPTT N</b>					
8024 Actin FS	22.8 Sek	17.1 - 28.5	25 %	10	10 (100.0 %)
8025 Pathromtin SL	37.1 Sek	27.8 - 46.4	25 %	5	5 (100.0 %)
8027 Stago/STA	31.5 Sek	23.7 - 39.4	25 %	21	21 (100.0 %)
8028 aPTT-SP	25.5 Sek	19.1 - 31.9	25 %	12	12 (100.0 %)
8026 altro	32.6 Sek	24.4 - 40.7	25 %	8	8 (100.0 %)
<b>Faktor VII</b>					
8152 tutti	97.0 %	72.8 - 121.3	25 %	5	5 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>G04 eparina non frazionata</b>					
<b>Quick H</b>					
8238 Innovin	100 %	85 - 115	15 %	13	13 (100.0 %)
8232 Neoplastin R	96 %	81 - 110	15 %	13	13 (100.0 %)
8243 Recombiplastin 2G	98 %	83 - 113	15 %	12	12 (100.0 %)
8242 altro	96 %	82 - 111	15 %	11	11 (100.0 %)
<b>Anti-FXa (unfrakt-Heparin)</b>					
8159 Stago/STA	0.27 IU/ml	0.22 - 0.32	20 %	11	10 (90.9 %)
8160 ACL	0.45 IU/ml	0.36 - 0.54	20 %	22	21 (95.5 %)
8153 altro	0.48 IU/ml	0.38 - 0.58	20 %	9	9 (100.0 %)
<b>Fibrinogeno H</b>					
8010 Siemens Thrombin	3.34 g/l	2.84 - 3.84	15 %	4	4 (100.0 %)
8013 Stago/STA	3.33 g/l	2.83 - 3.83	15 %	15	15 (100.0 %)
8014 Fibrinogen Q.F.A.	3.30 g/l	2.81 - 3.80	15 %	14	14 (100.0 %)
8011 altro	3.20 g/l	2.72 - 3.68	15 %	9	9 (100.0 %)
<b>aPTT H</b>					
8034 Actin FS	46.7 Sek	35.0 - 58.3	25 %	4	4 (100.0 %)
8040 Actin FSL	48.1 Sek	36.1 - 60.1	25 %	5	5 (100.0 %)
8037 Stago/STA	77.4 Sek	58.0 - 96.7	25 %	14	14 (100.0 %)
8038 aPTT-SP	72.4 Sek	54.3 - 90.5	25 %	16	16 (100.0 %)
8036 altro	46.5 Sek	34.9 - 58.1	25 %	4	4 (100.0 %)
<b>K05 Marker cardiaco</b>					
<b>Troponina I</b>					
8214 Pathfast	6118. ng/l	4650.3 - 7587.3	24 %	15	15 (100.0 %)
8101 Vidas	13629ng/l	10358.7 - 16901.1	24 %	12	12 (100.0 %)
8102 Architect High Sensi	1363. ng/l	1036.5 - 1691.1	24 %	10	10 (100.0 %)
<b>Troponina T</b>					
8114 Cobas hs	892.0 ng/l	677.92 - 1106.08	24 %	9	9 (100.0 %)
8116 Cobas hs STAT	871.1 ng/l	662.09 - 1080.26	24 %	12	12 (100.0 %)
<b>Mioglobina</b>					
8125 Cobas E / Elecsys	123.5 µg/l	86.5 - 160.6	30 %	8	8 (100.0 %)
8144 Abbott	168.0 µg/l	117.6 - 218.4	30 %	5	5 (100.0 %)
<b>CK-MB massa</b>					
8095 Cobas E / Elecsys	25.6 µg/l	15.4 - 35.8	40 %	9	9 (100.0 %)
<b>BNP</b>					
4799 altro	860.3 ng/l	628.0 - 1092.6	27 %	4	4 (100.0 %)
<b>NT-proBNP</b>					
7478 Pathfast	4234. ng/l	3091.2 - 5377.9	27 %	11	10 (90.9 %)
7415 AQT 90 FLEX	2825. ng/l	2062.3 - 3587.8	27 %	4	4 (100.0 %)
7416 VIDAS	1179. ng/l	860.7 - 1497.3	27 %	9	9 (100.0 %)
7467 Cobas E / Elecsys	1160. ng/l	847.3 - 1474.1	27 %	19	19 (100.0 %)
7477 Abbott	1531. ng/l	1117.6 - 1944.4	27 %	9	9 (100.0 %)

## Valori ideali MQ 2023 - 4

		Giusto	Tolleranza		Numero	Adempite
<b>G06 D-Dimeri</b>						
<b>D-Dimeri</b>						
7128	Cobas (Zitratplasma)	3.07 mg/l	2.42 - 3.71	21 %	6	6 (100.0 %)
7101	STA Liatest	2.94 mg/l	2.33 - 3.56	21 %	17	17 (100.0 %)
7102	Siemens Innovance	4.82 mg/l	3.81 - 5.83	21 %	13	13 (100.0 %)
7109	Pathfast	5.00 mg/l	3.95 - 6.05	21 %	14	14 (100.0 %)
7112	ACL	4.54 mg/l	3.59 - 5.49	21 %	12	12 (100.0 %)
7115	AQT 90 FLEX	2.16 mg/l	1.71 - 2.61	21 %	7	7 (100.0 %)
7127	VIDAS	3.41 mg/l	2.69 - 4.13	21 %	17	17 (100.0 %)
<b>K06 Tiroidea</b>						
<b>TSH</b>						
7201	Cobas E / Elecsys	6.60 mU/l	5.41 - 7.78	18 %	24	24 (100.0 %)
7204	Abbott	5.37 mU/l	4.40 - 6.33	18 %	11	11 (100.0 %)
7205	VIDAS	7.27 mU/l	5.96 - 8.58	18 %	14	14 (100.0 %)
7208	Dimension	5.76 mU/l	4.72 - 6.80	18 %	5	5 (100.0 %)
7257	AFIAS	9.54 mU/l	7.82 - 11.25	18 %	16	12 (75.0 %)
7209	altro	6.41 mU/l	5.26 - 7.57	18 %	5	5 (100.0 %)
<b>T3</b>						
7214	Abbott	2.9 nmol/l	2.3 - 3.5	20 %	4	4 (100.0 %)
<b>T4</b>						
7224	Abbott	152 nmol/l	122 - 182	20 %	4	4 (100.0 %)
<b>FT3</b>						
7231	Cobas E / Elecsys	12.0 pmol/l	9.8 - 14.1	18 %	21	21 (100.0 %)
7233	ADVIA Centaur XP/CP	12.3 pmol/l	10.0 - 14.5	18 %	4	4 (100.0 %)
7234	Abbott	9.9 pmol/l	8.2 - 11.7	18 %	10	10 (100.0 %)
7235	VIDAS	11.7 pmol/l	9.6 - 13.8	18 %	7	7 (100.0 %)
<b>FT4</b>						
7241	Cobas E / Elecsys	41.0 pmol/l	32.8 - 49.3	20 %	22	22 (100.0 %)
7244	Abbott	27.1 pmol/l	21.7 - 32.5	20 %	11	11 (100.0 %)
7246	VIDAS	43.1 pmol/l	34.4 - 51.7	20 %	9	9 (100.0 %)
7249	altro	35.9 pmol/l	28.7 - 43.0	20 %	8	8 (100.0 %)
<b>Testosterone</b>						
7390	Cobas	14.3 nmol/l	10.0 - 18.6	30 %	11	11 (100.0 %)
7391	ADVIA Centaur XP/CP	13.7 nmol/l	9.6 - 17.8	30 %	5	5 (100.0 %)
<b>Estradiolo</b>						
7370	Cobas	1190 pmol/l	833 - 1547	30 %	9	9 (100.0 %)
7371	ADVIA Centaur XP/CP	1877 pmol/l	1314 - 2440	30 %	5	5 (100.0 %)
7375	tutti	1157 pmol/l	810 - 1504	30 %	4	4 (100.0 %)
7372	Abbott	1100 pmol/l	770 - 1429	30 %	6	6 (100.0 %)
<b>SHBG</b>						
7360	Cobas	32.3 nmol/l	22.6 - 42.0	30 %	15	15 (100.0 %)
7362	Abbott	33.4 nmol/l	23.4 - 43.4	30 %	5	5 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>Cortisolo</b>					
7261 Cobas E / Elecsys	512 nmol/l	410 - 614	20 %	19	19 (100.0 %)
7264 Abbott	480 nmol/l	384 - 576	20 %	5	5 (100.0 %)
<b>Progesteron</b>					
7350 Cobas	38.0 nmol/l	26.6 - 49.4	30 %	6	6 (100.0 %)
7352 Abbott	28.7 nmol/l	20.1 - 37.3	30 %	5	5 (100.0 %)
7355 altro	36.9 nmol/l	25.8 - 47.9	30 %	4	4 (100.0 %)
<b>DHEAS</b>					
7340 Cobas	5.84 µmol/l	4.09 - 7.59	30 %	11	11 (100.0 %)
7342 Abbott	5.07 µmol/l	3.55 - 6.59	30 %	4	4 (100.0 %)
<b>Luteinisiertes Hormon</b>					
8181 Roche, Cobas	23.0 U/l	17.5 - 28.5	24 %	9	9 (100.0 %)
8182 ADVIA Centaur XP/CP	18.3 U/l	13.9 - 22.7	24 %	5	5 (100.0 %)
8183 Abbott	14.5 U/l	11.0 - 17.9	24 %	6	6 (100.0 %)
<b>Follikelstimulierendes Hormon</b>					
8171 Roche, Cobas	17.1 U/l	13.0 - 21.2	24 %	10	10 (100.0 %)
8172 ADVIA Centaur XP/CP	17.5 U/l	13.3 - 21.7	24 %	5	5 (100.0 %)
8173 Architect	17.2 U/l	13.1 - 21.4	24 %	7	7 (100.0 %)
<b>Prolaktin (PRL)</b>					
7270 ADVIA Centaur XP/CP	13.0 µg/l	9.9 - 16.2	24 %	4	4 (100.0 %)
7271 Cobas/Roche	22.0 µg/l	16.7 - 27.3	24 %	11	11 (100.0 %)
7272 Abbott	18.3 µg/l	13.9 - 22.6	24 %	5	5 (100.0 %)
<b>Insulin</b>					
7380 Cobas	502 pmol/l	376 - 627	25 %	14	14 (100.0 %)
7382 Abbott	408 pmol/l	306 - 510	25 %	4	4 (100.0 %)
<b>HGH</b>					
6830 tutti	11.97 µg/l	8.98 - 14.96	25 %	6	6 (100.0 %)
6831 Liaison	10.30 µg/l	7.73 - 12.88	25 %	4	4 (100.0 %)
<b>IGF-1</b>					
6846 Liaison	111 µg/l	83 - 139	25 %	7	7 (100.0 %)
6845 altro	74 µg/l	55 - 92	25 %	5	5 (100.0 %)
<b>K08 Marker infarto cardiaco h232</b>					
<b>Troponina T CR</b>					
7445 Cobas h 232	314.2 ng/l	238.85 - 389.70	24 %	1182	1107 (93.7 %)
<b>D-Dimeri CR</b>					
7442 Cobas h 232	0.83 mg/l	0.65 - 1.00	21 %	1196	968 (80.9 %)
7122 Lumira Dx	0.72 mg/l	0.57 - 0.87	21 %	8	8 (100.0 %)
<b>CKMB- K8</b>					
7448 Cobas h 232	25.1 µg/l	15.1 - 35.2	40 %	8	8 (100.0 %)
<b>NT-proBNP CR</b>					
7446 Cobas h 232	346 ng/l	253 - 439	27 %	751	689 (91.7 %)
<b>K09 Gas sanguini Opti CCA</b>					
<b>PCO2 CCA</b>					
4066 OPTI CCA	9.54 kPa	8.39 - 10.68	12 %	11	11 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>PO2 CCA</b>					
4166 OPTI CCA	9.31 kPa	7.91 - 10.70	15 %	11	10 (90.9 %)
<b>pH CCA</b>					
4266 OPTI CCA	7.14	7.07 - 7.20	1 %	10	10 (100.0 %)
<b>Potassio CCA</b>					
4549 OPTI CCA	2.9 mmol/l	2.7 - 3.1	6 %	5	4 (80.0 %)
<b>Sodio CCA</b>					
4649 OPTI CCA	127.4 mmol/l	119.8 - 135.0	6 %	4	3 (75.0 %)
<b>K10 Anemia</b>					
<b>Ferritina</b>					
7047 Dimension	99.40 µg/l	75.54 - 123.26	24 %	5	5 (100.0 %)
7048 Beckman	74.90 µg/l	56.92 - 92.88	24 %	5	5 (100.0 %)
7050 tutti	91.20 µg/l	69.32 - 113.09	24 %	4	3 (75.0 %)
7052 Cobas E / Elecsys	95.25 µg/l	72.39 - 118.11	24 %	26	26 (100.0 %)
7053 Abbott	122.4 µg/l	93.02 - 151.77	24 %	13	13 (100.0 %)
7057 Mini Vidas	70.00 µg/l	53.20 - 86.80	24 %	7	5 (71.4 %)
7046 AFIAS	78.14 µg/l	59.38 - 96.89	24 %	26	26 (100.0 %)
<b>Vitamina B12</b>					
7060 tutti	291.0 pmol/l	229.89 - 352.11	21 %	5	5 (100.0 %)
7062 Cobas E / Elecsys	309.7 pmol/l	244.73 - 374.84	21 %	19	18 (94.7 %)
7063 Abbott	277.0 pmol/l	218.86 - 335.22	21 %	11	11 (100.0 %)
<b>Acido folico</b>					
7072 Cobas E / Elecsys	9.26 nmol/l	6.86 - 11.66	24 %	21	21 (100.0 %)
7073 Abbott	13.32 nmol/l	10.12 - 16.52	24 %	10	10 (100.0 %)
<b>Holotranscobalamine</b>					
7081 Abbott	111.0 pmol/l	77.7 - 144.3	30 %	16	15 (93.8 %)
7082 Cobas	83.1 pmol/l	58.1 - 108.0	30 %	30	30 (100.0 %)
<b>I03 Allergologia</b>					
<b>IgE Arachidi qn</b>					
7602 tutti	0.37 kU/L	0.01 - 0.82	30 %	15	15 (100.0 %)
<b>IgE Betulla, qn</b>					
7604 tutti	0.19 kU/L	0.01 - 0.64	30 %	15	15 (100.0 %)
<b>IgE Epitelio di gatto, qn</b>					
7606 tutti	1.04 kU/L	0.59 - 1.49	30 %	14	14 (100.0 %)
<b>IgE totali</b>					
7620 tutti	258 kU/L	181 - 336	30 %	14	14 (100.0 %)
<b>IgE sx1 qn</b>					
7622 tutti	13.67 kU/L	9.57 - 17.78	30 %	10	10 (100.0 %)
<b>IgE fx5 qn</b>					
7624 tutti	2.01 kU/L	1.41 - 2.61	30 %	11	10 (90.9 %)
<b>IgE rx1qn</b>					
7626 tutti	0.44 kU/L	0.01 - 0.89	30 %	8	8 (100.0 %)
<b>IgE rx2 qn</b>					
7628 tutti	9.26 kU/L	6.48 - 12.04	30 %	8	8 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>IgE D. pteronyssinus qn</b>					
7645 tutti	15.94 kU/L	11.16 - 20.72	30 %	12	12 (100.0 %)
<b>G09 CoaguChek APTT</b>					
<b>CoaguChek APTT</b>					
3770 CoaguChek Pro II	86.2 Sek	64.7 - 107.8	25 %	8	7 (87.5 %)
<b>K12 Bilirubina neonatale</b>					
<b>Bilirubina totale Neo</b>					
1050 tutti	227 µmol/l	187 - 268	18 %	14	14 (100.0 %)
1056 Dimension	269 µmol/l	221 - 318	18 %	5	5 (100.0 %)
<b>Bilirubina diretta</b>					
1051 tutti	133 µmol/l	109 - 157	18 %	18	18 (100.0 %)
<b>Bilirubin neonatale</b>					
1054 ABL700/800	268 µmol/l	220 - 316	18 %	9	9 (100.0 %)
1053 altro	288 µmol/l	236 - 340	18 %	13	13 (100.0 %)
<b>K15 Creatinkinase Aktivität</b>					
<b>CK-MB</b>					
6504 Fuji Dri-Chem	62.5 U/l	43.7 - 81.2	30 %	24	24 (100.0 %)
6507 Cobas/Roche	20.0 U/l	14.0 - 26.0	30 %	7	7 (100.0 %)
<b>K14 Tumormarker</b>					
<b>PSA</b>					
6591 Cobas E / Elecsys	8.57 µg/l	6.77 - 10.37	21 %	18	18 (100.0 %)
6598 Abbott	8.52 µg/l	6.73 - 10.31	21 %	7	7 (100.0 %)
6696 AFIAS	8.81 µg/l	6.96 - 10.66	21 %	12	11 (91.7 %)
<b>PSA frei</b>					
6631 Cobas E / Elecsys	2.34 µg/l	1.85 - 2.84	21 %	10	10 (100.0 %)
6633 ADVIA Centaur XP/CP	2.59 µg/l	2.05 - 3.13	21 %	4	4 (100.0 %)
6639 Abbott	2.86 µg/l	2.26 - 3.47	21 %	5	5 (100.0 %)
<b>CEA</b>					
6601 Cobas E / Elecsys	7.5 µg/l	5.9 - 9.1	21 %	12	12 (100.0 %)
6603 ADVIA Centaur XP/CP	8.8 µg/l	7.0 - 10.7	21 %	4	4 (100.0 %)
6608 Abbott	10.8 µg/l	8.5 - 13.1	21 %	6	6 (100.0 %)
<b>CA 125</b>					
6611 Cobas E / Elecsys	70.0 kIU/l	52.5 - 87.5	25 %	8	8 (100.0 %)
6618 Abbott	117.0 kIU/l	87.8 - 146.3	25 %	5	5 (100.0 %)
<b>CA 19-9</b>					
6661 Cobas E / Elecsys	17.4 kIU/l	13.1 - 21.8	25 %	7	7 (100.0 %)
6663 ADVIA Centaur XP	38.6 kIU/l	29.0 - 48.3	25 %	4	4 (100.0 %)
<b>CA 15-3</b>					
6621 Cobas E / Elecsys	33.5 kIU/l	25.1 - 41.9	25 %	10	10 (100.0 %)
6623 ADVIA Centaur XP/CP	45.9 kIU/l	34.4 - 57.4	25 %	4	4 (100.0 %)
6628 Abbott	41.5 kIU/l	31.1 - 51.9	25 %	5	5 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>AFP</b>					
6641 Cobas E / Elecsys	86.5 µg/l	64.9 - 108.1	25 %	8	8 (100.0 %)
6648 Abbott	80.6 µg/l	60.4 - 100.7	25 %	4	4 (100.0 %)
<b>HCG qn</b>					
6651 Cobas E / Elecsys	125.0 U/l	93.8 - 156.3	25 %	13	13 (100.0 %)
6656 VIDAS	48.1 U/l	36.1 - 60.2	25 %	10	10 (100.0 %)
6658 Architect	130.6 U/l	98.0 - 163.3	25 %	9	9 (100.0 %)
6659 AFIAS	180.0 U/l	135.0 - 225.0	25 %	7	7 (100.0 %)
<b>HCG intakt</b>					
6670 Cobas	43.4 U/l	32.5 - 54.2	25 %	4	4 (100.0 %)
<b>S100</b>					
6675 Cobas	1.49 µg/l	1.12 - 1.86	25 %	6	6 (100.0 %)
<b>NSE</b>					
6677 Cobas	2.4 ng/ml	1.8 - 3.0	25 %	5	5 (100.0 %)
<b>Tirreoglobulina</b>					
6683 Cobas	84.1 µg/l	63.1 - 105.1	25 %	4	4 (100.0 %)
6684 altro	95.0 µg/l	71.3 - 118.8	25 %	4	4 (100.0 %)
<b>K19 CardioChek Lipidpanel</b>					
<b>Colesterolo PTS</b>					
1321 CardioChek	3.96 mmol/l	3.56 - 4.36	10 %	14	7 (50.0 %)
<b>Colesterolo HDL PTS</b>					
1421 CardioChek	1.32 mmol/l	1.04 - 1.59	21 %	14	10 (71.4 %)
<b>Trigliceridi PTS</b>					
3521 CardioChek	0.67 mmol/l	0.49 - 0.85	18 %	14	13 (92.9 %)
<b>U05 Urina albumina/creatinina</b>					
<b>Microalbumina</b>					
5804 Abbott	26.0 mg/l	19.7 - 32.2	24 %	12	12 (100.0 %)
5805 Roche, Cobas	24.2 mg/l	18.4 - 30.0	24 %	14	14 (100.0 %)
5218 Aution	10.0 mg/l	4.0 - 16.0	24 %	8	8 (100.0 %)
5800 AFIAS	22.7 mg/l	17.3 - 28.2	24 %	20	20 (100.0 %)
5803 Afinion	22.5 mg/l	17.1 - 27.9	24 %	487	455 (93.4 %)
5810 Sysmex U	24.0 mg/l	9.6 - 38.4	24 %	16	14 (87.5 %)
5843 Turbidimetrie	25.3 mg/l	19.2 - 31.3	24 %	4	4 (100.0 %)
5852 DCA2000/Vantage	23.7 mg/l	18.0 - 29.4	24 %	146	138 (94.5 %)
5220 Siemens Clinitek	24.0 mg/l	9.6 - 38.4	24 %	26	25 (96.2 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>Creatinina urina</b>					
5211 Abbott	3.2 mmol/l	2.5 - 3.8	21 %	12	12 (100.0 %)
5212 Roche	3.4 mmol/l	2.7 - 4.1	21 %	21	21 (100.0 %)
5213 Beckman	3.5 mmol/l	2.7 - 4.2	21 %	4	4 (100.0 %)
5214 Siemens	3.0 mmol/l	2.4 - 3.6	21 %	5	5 (100.0 %)
5201 DCA2000/Vantage	3.5 mmol/l	2.8 - 4.3	21 %	144	134 (93.1 %)
5203 Afinion	3.0 mmol/l	2.4 - 3.7	21 %	484	481 (99.4 %)
5210 Sysmex U	4.4 mmol/l	3.5 - 5.3	21 %	13	9 (69.2 %)
5219 Aution	4.4 mmol/l	3.5 - 5.3	21 %	8	4 (50.0 %)
5221 Siemens Clinitek	4.4 mmol/l	3.5 - 5.3	21 %	24	19 (79.2 %)
5222 altro	0.9 mmol/l	0.5 - 1.3	21 %	4	4 (100.0 %)
<b>G11 CoaguChek XS INR</b>					
<b>INR CCXS</b>					
3685 CoaguChek XS	2.0	1.7 - 2.3	15 %	1421	1405 (98.9 %)
<b>G12 Hemochron</b>					
<b>INR HC</b>					
3681 Hemochron j.	1.5	1.3 - 1.7	15 %	6	4 (66.7 %)
<b>K22 Osmo</b>					
<b>Osmolalità</b>					
6080 Cryoscopie	370 mosm/kg	348 - 392	6 %	21	21 (100.0 %)
<b>Kalium-K22</b>					
6081 ISE	3.7 mmol/l	3.5 - 3.9	6 %	12	12 (100.0 %)
<b>Natrium-K22</b>					
6082 ISE	146 mmol/l	137 - 155	6 %	12	12 (100.0 %)
<b>Glukose-K22</b>					
6083 Chimica umida	4.5 mmol/l	4.1 - 4.9	9 %	12	12 (100.0 %)
<b>Harnstoff-K22</b>					
6084 Chimica umida	4.3 mmol/l	3.7 - 5.0	15 %	12	12 (100.0 %)
<b>Osmotische Lücke</b>					
6085 Formel 1 (2Na+K+Glu+	67.9 mmol/l	54.4 - 81.5	20 %	10	10 (100.0 %)
<b>K20 PCT/C-Peptid</b>					
<b>C-Peptid</b>					
6826 Cobas	5.43 nmol/l	4.07 - 6.78	25 %	8	8 (100.0 %)
6825 altro	4.96 nmol/l	3.72 - 6.20	25 %	5	5 (100.0 %)
<b>ACTH</b>					
6841 Cobas	19.28 ng/l	14.46 - 24.11	25 %	8	8 (100.0 %)
<b>Procalcitonina</b>					
7320 Abbott	14.80 µg/l	10.80 - 18.80	27 %	9	9 (100.0 %)
7280 Cobas	9.99 µg/l	7.29 - 12.68	27 %	12	12 (100.0 %)
7281 VIDAS	7.97 µg/l	5.82 - 10.12	27 %	13	13 (100.0 %)
7285 Liaison	19.20 µg/l	14.02 - 24.38	27 %	5	5 (100.0 %)



## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>K21 PTH/EPO</b>					
<b>EPO</b>					
6820 tutti	116.4 U/l	87.3 - 145.4	25 %	4	4 (100.0 %)
<b>Parathormon</b>					
7293 Cobas PTH STAT	17.0 pmol/l	12.9 - 21.0	24 %	10	10 (100.0 %)
7295 Cobas	14.0 pmol/l	10.6 - 17.3	24 %	10	10 (100.0 %)
7287 Abbott	31.5 pmol/l	24.0 - 39.1	24 %	5	5 (100.0 %)
<b>K24 Farmaci</b>					
<b>Digoxin</b>					
9020 altro	1.50 nmol/l	1.14 - 1.86	24 %	9	9 (100.0 %)
<b>Phenytoin</b>					
9018 tutti	35 µmol/l	26 - 44	25 %	4	4 (100.0 %)
<b>Phenobarbital</b>					
9017 tutti	91 µmol/l	68 - 114	25 %	4	4 (100.0 %)
<b>Vancomycin</b>					
9033 Roche, Cobas	8.3 µmol/l	6.3 - 10.4	25 %	5	5 (100.0 %)
<b>Valproat</b>					
9021 tutti	359.0 µmol/l	272.8 - 445.2	24 %	9	9 (100.0 %)
<b>Carbamazepin</b>					
9022 tutti	26.8 µmol/l	20.3 - 33.2	24 %	4	4 (100.0 %)
<b>K25 Cystatina C</b>					
<b>Cystatin C</b>					
7014 tutti	1.78 mg/l	1.35 - 2.20	24 %	17	16 (94.1 %)
<b>H05 Ematologia BG</b>					
<b>Emoglobina BG</b>					
4502 iStat	117.7 g/l	107.1 - 128.3	9 %	12	12 (100.0 %)
<b>Ematocrito</b>					
4503 iStat	0.35 l/l	0.32 - 0.38	9 %	18	18 (100.0 %)
4858 EPOC	0.36 l/l	0.32 - 0.39	9 %	13	12 (92.3 %)
<b>I05 CRP/Lp (a)</b>					
<b>CRP HS</b>					
1680 Turbidimetrie	5.52 mg/l	3.52 - 7.52	21 %	10	10 (100.0 %)
<b>Lipoprotein (a)</b>					
7289 tutti	56 nmol/l	42 - 70	25 %	7	7 (100.0 %)
8222 Andere	69 nmol/l	52 - 87	25 %	4	4 (100.0 %)
<b>K28 Alcol/Ammoniaca</b>					
<b>Etanolo</b>					
7193 Andere	14.3 mmol/l	11.7 - 16.9	18 %	5	4 (80.0 %)
7179 Abbott	15.3 mmol/l	12.5 - 18.0	18 %	7	7 (100.0 %)
7191 Roche, Cobas	15.0 mmol/l	12.3 - 17.6	18 %	19	19 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>Ammoniaca</b>					
7198 Abbott	93.5 µmol/l	73.9 - 113.1	21 %	4	4 (100.0 %)
7195 tutti	117.0 µmol/l	92.4 - 141.6	21 %	6	6 (100.0 %)
<b>K29 Calprotectina</b>					
<b>Calprotectina</b>					
7185 Bühlmann fCALturbo	497 µg/g	298 - 695	40 %	21	19 (90.5 %)
7183 Bühlmann Quantum Blu	542 µg/g	325 - 759	40 %	5	5 (100.0 %)
7187 Liaison	216 µg/g	130 - 302	40 %	20	19 (95.0 %)
<b>K30 Lipidi Af/b101</b>					
<b>Colesterolo Af/b101</b>					
1302 Cobas b101	4.99 mmol/l	4.49 - 5.49	10 %	304	302 (99.3 %)
1301 Afinion	4.93 mmol/l	4.44 - 5.43	10 %	417	411 (98.6 %)
<b>Colesterolo HDL Af/b101</b>					
1402 Cobas b101	0.91 mmol/l	0.72 - 1.10	21 %	301	275 (91.4 %)
1401 Afinion	0.88 mmol/l	0.70 - 1.07	21 %	414	383 (92.5 %)
<b>Trigliceridi Af/b101</b>					
3502 Cobas b101	1.67 mmol/l	1.37 - 1.97	18 %	300	297 (99.0 %)
3501 Afinion	1.62 mmol/l	1.33 - 1.91	18 %	417	412 (98.8 %)
<b>S01 Sangue occulto</b>					
<b>Sangue occulto qn</b>					
5018 OC-Sensor	50 ng/ml	35 - 65	30 %	7	7 (100.0 %)
<b>K31 Marker per infarto cardiaco Af/Nx</b>					
<b>Troponina I S</b>					
7431 AFIAS	1087. ng/l	826.73 - 1348.87	24 %	298	252 (84.6 %)
<b>D-Dimeri qn S</b>					
7428 AFIAS	0.91 mg/l	0.72 - 1.11	21 %	297	243 (81.8 %)
<b>NT-proBNP S</b>					
7427 AFIAS	2069. ng/l	1510.5 - 2627.9	27 %	223	108 (48.4 %)
<b>G14 MicroINR</b>					
<b>INR MI</b>					
3677 MicroINR	2.1	1.8 - 2.4	15 %	139	108 (77.7 %)
<b>K32 Omocisteina</b>					
<b>Homocystein</b>					
8209 Beckman	31.0 µmol/l	21.7 - 40.3	30 %	6	6 (100.0 %)
8210 tutti	26.6 µmol/l	18.6 - 34.6	30 %	4	4 (100.0 %)

## Valori ideali MQ 2023 - 4

		Giusto	Tolleranza		Numero	Adempite
<b>K34 Chimica clinica 2</b>						
<b>Lipasi</b>						
6499	Abbott	102.5 U/l	84.1 - 121.0	18 %	13	13 (100.0 %)
6500	Beckman	96.0 U/l	78.7 - 113.3	18 %	5	5 (100.0 %)
6501	Roche	101.8 U/l	83.5 - 120.1	18 %	23	23 (100.0 %)
6503	Fuji Dri-Chem	82.6 U/l	67.7 - 97.4	18 %	170	167 (98.2 %)
<b>Bicarbonat</b>						
4090	Cobas	26.7 mmol/l	22.7 - 30.7	15 %	5	5 (100.0 %)
4095	Abbott	25.9 mmol/l	22.0 - 29.7	15 %	4	4 (100.0 %)
<b>Colinesterasi</b>						
6516	ADVIA Centaur XP/CP	7.9 kU/L	5.5 - 10.3	30 %	4	4 (100.0 %)
<b>G16 INR Xprecia</b>						
<b>INR Xprecia</b>						
3688	Xprecia	1.2	1.0 - 1.4	15 %	48	48 (100.0 %)
<b>H06 Emogramma, automatico, 5-part</b>						
<b>Emoglobina</b>						
105	Sysmex	118.2 g/l	107.6 - 128.9	9 %	110	110 (100.0 %)
120	Advia	121.0 g/l	110.1 - 131.9	9 %	4	4 (100.0 %)
150	Yumizen/Pentra	119.9 g/l	109.1 - 130.7	9 %	13	13 (100.0 %)
<b>Ematocrito</b>						
106	Sysmex	0.36 l/l	0.33 - 0.39	9 %	112	112 (100.0 %)
121	Advia	0.37 l/l	0.33 - 0.40	9 %	4	4 (100.0 %)
151	Yumizen/Pentra	0.34 l/l	0.31 - 0.38	9 %	13	13 (100.0 %)
<b>Eritrociti</b>						
107	Sysmex	3.87 T/l	2.91 - 4.84	25 %	111	111 (100.0 %)
122	Advia	3.96 T/l	2.97 - 4.94	25 %	4	4 (100.0 %)
152	Yumizen/Pentra	3.86 T/l	2.90 - 4.83	25 %	13	13 (100.0 %)
<b>Leucociti</b>						
108	Sysmex	4.75 G/l	3.56 - 5.94	25 %	111	110 (99.1 %)
123	Advia	4.70 G/l	3.52 - 5.87	25 %	4	4 (100.0 %)
153	Yumizen/Pentra	3.73 G/l	2.79 - 4.66	25 %	13	10 (76.9 %)
<b>Trombociti</b>						
109	Sysmex	163.7 G/l	122.7 - 204.6	25 %	110	110 (100.0 %)
124	Advia	147.5 G/l	110.6 - 184.4	25 %	4	4 (100.0 %)
154	Yumizen/Pentra	162.4 G/l	121.8 - 203.0	25 %	13	13 (100.0 %)
<b>Neutrofili</b>						
110	Sysmex	2.69 G/l	1.35 - 4.04	25 %	112	112 (100.0 %)
125	Advia	2.77 G/l	1.39 - 4.16	25 %	4	4 (100.0 %)
155	Yumizen/Pentra	1.28 G/l	0.64 - 1.91	25 %	12	12 (100.0 %)
<b>Linfociti</b>						
111	Sysmex	1.61 G/l	0.81 - 2.42	25 %	112	112 (100.0 %)
126	Advia	1.45 G/l	0.73 - 2.18	25 %	4	4 (100.0 %)
156	Yumizen/Pentra	1.64 G/l	0.82 - 2.46	25 %	12	12 (100.0 %)

## Valori ideali MQ 2023 - 4

		Giusto	Tolleranza		Numero	Adempite
<b>Monociti</b>						
112	Sysmex	0.29 G/l	0.06 - 0.52	40 %	111	111 (100.0 %)
127	Advia	0.28 G/l	0.06 - 0.50	40 %	4	4 (100.0 %)
157	Yumizen/Pentra	0.18 G/l	0.04 - 0.32	40 %	12	11 (91.7 %)
<b>Eosinofili</b>						
113	Sysmex	0.04 G/l	0.01 - 0.06	80 %	111	111 (100.0 %)
128	Advia	0.07 G/l	0.01 - 0.13	80 %	4	4 (100.0 %)
158	Yumizen/Pentra	0.02 G/l	0.00 - 0.03	80 %	12	12 (100.0 %)
<b>Basofili</b>						
114	Sysmex	0.13 G/l	0.03 - 0.23	80 %	112	112 (100.0 %)
129	Advia	0.02 G/l	0.01 - 0.10	80 %	4	4 (100.0 %)
159	Yumizen/Pentra	0.05 G/l	0.01 - 0.13	80 %	12	11 (91.7 %)
<b>MCV</b>						
404	Sysmex	91.4 fl	68.5 - 114.2	25 %	101	101 (100.0 %)
420	Yumizen/Pentra	90.0 fl	67.5 - 112.5	25 %	9	9 (100.0 %)
<b>MCH</b>						
405	Sysmex	30.5 pg	22.9 - 38.1	25 %	101	101 (100.0 %)
421	Yumizen/Pentra	31.1 pg	23.3 - 38.9	25 %	10	10 (100.0 %)
<b>MCHC</b>						
406	Sysmex	334 g/l	251 - 418	25 %	102	101 (99.0 %)
422	Yumizen/Pentra	349 g/l	262 - 436	25 %	10	10 (100.0 %)
<b>RDW</b>						
407	Sysmex	12.4 %	9.3 - 15.5	25 %	98	98 (100.0 %)
423	Yumizen/Pentra	13.7 %	10.3 - 17.1	25 %	9	9 (100.0 %)
<b>Immature Granulocytes</b>						
408	Sysmex	0.07 G/l	0.01 - 0.12	25 %	94	86 (91.5 %)
<b>H07 Reticolociti, Automatico</b>						
<b>Reticolociti</b>						
115	Sysmex	61.4 G/l	43.0 - 79.9	30 %	62	60 (96.8 %)
<b>H08 Indici del siero HIL</b>						
<b>Hämolyseindex Probe A</b>						
2940	Cobas	24.52	17.17 - 31.88	30 %	24	23 (95.8 %)
2948	Abbott	0.18	0.13 - 0.23	30 %	6	6 (100.0 %)
<b>Hämolyseindex Probe B</b>						
2941	Cobas	12.50	8.75 - 16.25	30 %	24	22 (91.7 %)
2949	Abbott	0.09	0.06 - 0.11	30 %	6	6 (100.0 %)
<b>Indice di lipemia A</b>						
2950	Cobas	34.50	24.15 - 44.85	30 %	6	6 (100.0 %)
<b>Indice di lipemia B</b>						
2951	Cobas	15.00	10.50 - 19.50	30 %	6	5 (83.3 %)
<b>Indice Icteria A</b>						
2960	Cobas	7.00	4.90 - 9.10	30 %	6	6 (100.0 %)
<b>Indice Icteria B</b>						
2961	Cobas	3.00	1.50 - 4.50	30 %	6	5 (83.3 %)

## Valori ideali MQ 2023 - 4

		Giusto	Tolleranza		Numero	Adempite	
<b>H10 Velocità di eritrosedimentazione</b>							
<b>Velocità di eritrosedimentazione 1h</b>							
399	MINI-CUBE	13	mm/h	5 - 20	40 %	24	24 (100.0 %)
389	miniiSed	1	mm/h	0 - 2	40 %	8	8 (100.0 %)
390	Sarstedt Sedivette	9	mm/h	4 - 14	40 %	19	19 (100.0 %)
391	Sarstedt Microvette	11	mm/h	4 - 18	40 %	4	4 (100.0 %)
392	BD Seditainer	9	mm/h	3 - 14	40 %	42	40 (95.2 %)
393	altro	10	mm/h	4 - 16	40 %	19	17 (89.5 %)
<b>Velocità di eritrosedimentazione 2h</b>							
395	Sarstedt Sedivette	19	mm/2h	8 - 30	40 %	7	7 (100.0 %)
397	BD Seditainer	19	mm/2h	7 - 30	40 %	4	3 (75.0 %)
398	altro	20	mm/2h	12 - 28	40 %	4	4 (100.0 %)
<b>K35 Liquido cerebrospinale</b>							
<b>Glucosio CSF</b>							
8930	Cobas	3.45	mmol/l	3.14 - 3.76	9 %	10	10 (100.0 %)
8931	altro	3.38	mmol/l	3.07 - 3.68	9 %	13	13 (100.0 %)
<b>Lattato CSF</b>							
8932	Cobas	1.82	mmol/l	1.49 - 2.15	18 %	8	8 (100.0 %)
8933	altro	1.79	mmol/l	1.47 - 2.12	18 %	11	10 (90.9 %)
<b>Proteine CSF</b>							
8934	Cobas	0.35	g/l	0.30 - 0.40	15 %	15	15 (100.0 %)
8935	altro	0.42	g/l	0.35 - 0.48	15 %	4	4 (100.0 %)
<b>Albumina CSF</b>							
8944	Cobas	212.0	mg/l	169.60 - 254.40	20 %	5	5 (100.0 %)
<b>K36 CDT (transferrina carboidrato carente)</b>							
<b>CDT</b>							
7192	tutti	1.00	%	0.80 - 1.20	20 %	7	6 (85.7 %)
<b>K37 Immunosoppressore</b>							
<b>Tacrolimus</b>							
9127	tutti	8.4	µg/l	6.3 - 10.5	25 %	9	9 (100.0 %)
<b>K38 Elettroforesi delle sieroproteine</b>							
<b>Totalproteina E</b>							
7900	tutti	76.6	g/l	67.4 - 85.8	12 %	19	19 (100.0 %)
<b>albumina E</b>							
7901	elettroforesi	65.1	%	57.3 - 73.0	12 %	30	30 (100.0 %)
<b>alfa-1-globuline</b>							
7902	elettroforesi	1.4	%	1.0 - 1.8	30 %	11	11 (100.0 %)
7912	elettroforesi capill	2.7	%	1.9 - 3.5	30 %	19	19 (100.0 %)
<b>alfa-2-globuline</b>							
7903	elettroforesi	8.0	%	5.6 - 10.4	30 %	30	30 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>beta-globuline</b>					
7904 elettroforesi	10.8 %	7.6 - 14.1	30 %	16	16 (100.0 %)
<b>Beta-1-Globulin</b>					
7913 elettroforesi	6.2 %	4.3 - 8.1	30 %	8	7 (87.5 %)
<b>Beta-2-Globulin</b>					
7914 elettroforesi	5.9 %	4.1 - 7.7	30 %	7	7 (100.0 %)
<b>gamma-Globuline</b>					
7905 elettroforesi	12.7 %	8.9 - 16.5	30 %	7	7 (100.0 %)
<b>Gamma-Globuline+P</b>					
7916 elettroforesi	11.6 %	8.1 - 15.0	30 %	23	23 (100.0 %)
<b>Beta-Globuline+P</b>					
7917 elettroforesi	12.1 %	8.5 - 15.7	30 %	6	6 (100.0 %)
<b>Immunfixation</b>					
7915 Interpretation	8 Code	8 - 8	1 %	25	24 (96.0 %)
<b>K39 Folati eritrocitari</b>					
<b>Folati eritrocitari</b>					
7093 Abbott	143 nmol/l	86 - 200	40 %	10	10 (100.0 %)
7094 Roche, Cobas	972 nmol/l	583 - 1360	40 %	21	19 (90.5 %)
7096 ADVIA Centaur XP/CP	473 nmol/l	284 - 662	40 %	4	4 (100.0 %)
<b>K40 Acidi biliari nel siero</b>					
<b>Gallensäure</b>					
3541 altro	11.5 µmol/l	8.1 - 15.0	30 %	11	9 (81.8 %)
3540 tutti	8.8 µmol/l	6.2 - 11.5	30 %	16	16 (100.0 %)
<b>K41 Marker per infarto cardiaco, Triage</b>					
<b>BNP</b>					
7460 Triage	20.5 ng/l	0.5 - 40.5	27 %	9	9 (100.0 %)
<b>Troponin Triage</b>					
8199 Triage high sensitiv	376.1 ng/l	285.84 - 466.36	24 %	95	85 (89.5 %)
8197 Triage Next Gen	10.00 ng/l	7.60 - 12.40	24 %	11	8 (72.7 %)
<b>NT-pro BNP</b>					
7414 Triage	403 ng/l	294 - 512	27 %	53	50 (94.3 %)
<b>D-Dimere Triage</b>					
8191 Triage	536.5 ng/ml	423.84 - 649.17	21 %	90	46 (51.1 %)
<b>K42 Vitamina D</b>					
<b>Vitamina D 25 (OH)</b>					
7312 LCMS	111.8 nmol/l	81.6 - 142.0	27 %	5	5 (100.0 %)
7294 Cobas	97.8 nmol/l	71.4 - 124.2	27 %	15	15 (100.0 %)
7279 VIDAS	113.3 nmol/l	82.7 - 143.9	27 %	6	6 (100.0 %)
7288 altro	106.8 nmol/l	78.0 - 135.6	27 %	10	9 (90.0 %)
7296 Architect	119.1 nmol/l	87.0 - 151.3	27 %	10	10 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>K43 AMH</b>					
<b>AMH</b>					
6800 tutti	17.4 pmol/l	13.0 - 21.7	25 %	18	18 (100.0 %)
<b>K44 Inhibina B</b>					
<b>Inhibin B</b>					
6805 tutti	160.0 ng/l	112.0 - 208.0	25 %	6	5 (83.3 %)
<b>K45 Calcitonina</b>					
<b>Calcitonin</b>					
6811 Liaison	6.4 pmol/l	4.8 - 8.0	25 %	5	5 (100.0 %)
6810 altro	22.0 pmol/l	16.5 - 27.5	25 %	7	7 (100.0 %)
<b>K46 IGF-BP3 / Renin / Aldosteron</b>					
<b>IGF-BP3</b>					
6816 Cobas	3.27 mg/l	2.45 - 4.09	25 %	4	4 (100.0 %)
6815 tutti	3.79 mg/l	2.84 - 4.74	25 %	4	3 (75.0 %)
<b>Renin</b>					
6817 Liaison	67.0 mU/l	50.2 - 83.7	25 %	7	7 (100.0 %)
<b>K47 Schilddrüsenautoantikörper</b>					
<b>Anti Thyreoglobulin</b>					
6866 Alinity	73 IU/ml	55 - 92	25 %	7	7 (100.0 %)
6852 Cobas	360 IU/ml	270 - 450	25 %	9	9 (100.0 %)
<b>Anti TPO</b>					
6857 Cobas	90 IU/ml	68 - 113	25 %	10	10 (100.0 %)
6858 Abbott	56 IU/ml	42 - 71	25 %	7	7 (100.0 %)
<b>TRAK</b>					
6867 Abbott	4.17 IU/l	3.13 - 5.21	25 %	6	6 (100.0 %)
6861 Roche, Cobas	2.00 IU/l	1.50 - 2.50	25 %	7	7 (100.0 %)
<b>I07 CRP WB</b>					
<b>CRP</b>					
1614 AFIAS	69.2 mg/l	54.7 - 83.8	21 %	158	144 (91.1 %)
<b>K48 Creatinina sang intero</b>					
<b>Creatinina WB</b>					
2720 Statsensor i / Nova	205 µmol/l	168 - 242	18 %	69	58 (84.1 %)
<b>G18 LMW-Heparin</b>					
<b>Anti-FXa (LMW-Heparin)</b>					
8163 Stago/STA	1.68 IU/ml	1.34 - 2.02	20 %	9	9 (100.0 %)
8164 ACL	1.49 IU/ml	1.19 - 1.79	20 %	15	15 (100.0 %)
8154 altro	1.74 IU/ml	1.39 - 2.08	20 %	15	15 (100.0 %)

## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>G19 Rivaroxaban</b>					
<b>Anti-FXa (Rivaroxaban)</b>					
8161 Stago/STA	298.0 µg/l	238.40 - 357.60	20 %	9	9 (100.0 %)
8162 ACL	296.2 µg/l	237.00 - 355.50	20 %	8	8 (100.0 %)
8155 altro	282.3 µg/l	225.84 - 338.76	20 %	6	5 (83.3 %)
<b>G20 Apixaban</b>					
<b>Anti-FXa (Apixaban)</b>					
8166 ACL	274.7 µg/l	219.76 - 329.64	20 %	4	4 (100.0 %)
<b>K03 Emoglobina glicata (HbA1c)</b>					
<b>HbA1c campione B</b>					
4704 Abbott	6.0 %	5.5 - 6.5	9 %	5	5 (100.0 %)
4766 Roche, Cobas	6.0 %	5.5 - 6.6	9 %	18	18 (100.0 %)
4764 HPLC	6.0 %	5.5 - 6.5	9 %	9	9 (100.0 %)
4702 Afinion	5.9 %	5.4 - 6.4	9 %	790	785 (99.4 %)
4760 Cobas b101	5.7 %	5.2 - 6.3	9 %	199	192 (96.5 %)
4762 DCA2000/Vantage	6.1 %	5.6 - 6.7	9 %	192	189 (98.4 %)
4722 Celltac chemi	5.5 %	5.0 - 6.0	9 %	15	15 (100.0 %)
4761 NycoCard	6.1 %	5.5 - 6.6	9 %	6	3 (50.0 %)
4709 Eurolyser	6.1 %	5.5 - 6.6	9 %	5	4 (80.0 %)
4757 A1c Now	5.6 %	5.1 - 6.1	9 %	8	8 (100.0 %)
4770 AFIAS	5.8 %	5.3 - 6.3	9 %	146	133 (91.1 %)
4723 Spinit	6.0 %	5.5 - 6.6	9 %	10	10 (100.0 %)
4763 Andere	5.9 %	5.4 - 6.5	9 %	15	14 (93.3 %)
4777 Quick Read go	6.1 %	5.5 - 6.6	9 %	5	5 (100.0 %)
<b>G17 INR LumiraDX</b>					
<b>INR Lumira Dx</b>					
3693 Lumira Dx	1.2	1.0 - 1.4	15 %	11	11 (100.0 %)
<b>K49 IL6</b>					
<b>IL6</b>					
7335 Roche, Cobas	4504. ng/l	3152.8 - 5855.2	30 %	5	5 (100.0 %)
<b>I10 celiachia</b>					
<b>Anti deam. Gliadin IgA</b>					
7614 altro	3.30 U/ml	2.31 - 4.29	30 %	4	4 (100.0 %)
<b>Anti tTG IgG</b>					
7615 altro	37.00 U/ml	25.90 - 48.10	30 %	4	4 (100.0 %)
<b>Anti tTG IgA</b>					
7616 altro	109.5 U/ml	76.65 - 142.35	30 %	6	6 (100.0 %)



## Valori ideali MQ 2023 - 4

	Giusto	Tolleranza		Numero	Adempite
<b>H12 Hemoscreen</b>					
<b>Hemoglobina HS</b>					
190 PixCell HemoScreen	155.4 g/l	141.4 - 169.3	9 %	17	16 (94.1 %)
<b>Hematocrito HS</b>					
191 PixCell HemoScreen	0.4 l/l	0.4 - 0.4	9 %	17	16 (94.1 %)
<b>Eritrociti HS</b>					
192 PixCell HemoScreen	4.49 T/l	3.37 - 5.61	25 %	17	17 (100.0 %)
<b>Leucociti HS</b>					
193 PixCell HemoScreen	7.97 G/l	5.98 - 9.96	25 %	17	17 (100.0 %)
<b>Thrombociti HS</b>					
194 PixCell HemoScreen	240.2 G/l	180.1 - 300.2	25 %	17	17 (100.0 %)
<b>K51 Pankreas Elastase</b>					
<b>Pankreas Elastase</b>					
7180 Liaison	164 ug/g	99 - 230	40 %	10	10 (100.0 %)
<b>I11 CRP Lumira</b>					
<b>CRP Lumira</b>					
1603 Lumira Dx	180.4 mg/l	142.6 - 218.3	21 %	13	10 (76.9 %)
<b>K52 Copeptin</b>					
<b>Copeptin</b>					
7386 Kryptor	2.9 pmol/l	2.0 - 3.8	30 %	7	7 (100.0 %)