

Verein für
Association pour le
Associazione per il



medizinische Qualitätskontrolle
contrôle de qualité médical
controllo di qualità medico

Survey Report

2024 - 3

Survey Specimens

The homogeneity and stability of all specimens were checked before and/or during shipment and no irregularities were noted. The suitability tests were performed by the laboratories of the Universitätsspital Zürich (University Hospital Zurich) (<http://www.uzl.usz.ch/>).

The following survey specimens were produced specifically for MQ by a sub-contractor:
B2 Uricult, H4 Parasitic Hematology, K14 tumor marker

Determination of target values

For each target value, the type of determination per ISO17043: 2010 B2.1 is indicated (column "type"):

- a Value known due to production.
- b Certified reference value for use with special specimens
- c Reference value determined by analysis
- d Consensus values of expert laboratories
- e Consensus values of the participants

For methods groups with more than 9 participants, consensus values of the participants ("e") are generally determined.

In order to calculate the target values, we use the mean value of the method group. Values that differ more than 1.5 times the QUALAB-tolerance are outliers and are not used to calculate the target value. Starting point for the elimination of outliers are the values of our suitability tests.

In order to provide all participants with target values that are as meaningful as possible, other methods may also be applied for smaller method groups.

Uncertainty of the determined target values

The standard uncertainty (u_x) is calculated using the following formula (ISO13528):

$$u_x = (\text{target value}/100) * (1.25/\text{square root of "number of participants"}) * \%CV$$

- u_x has the same unit as the target value
- u_x can be compared with the standard deviation of the participants' collective ($SD = \text{target value} * \%CV / 100$)
- For participant numbers >18, the standard uncertainty (u_x) is significantly lower than the scatter of the collective participants and can be neglected.


QUALAB and MQ tolerances

For all mandatory analyzes, QUALAB tolerances are used (www.qualab.ch, external quality control). For non-mandatory analyzes, the tolerances are specified by MQ's survey specimen leader.

If the determined uncertainty, u_x , of the target value is greater than 15% of the QUALAB or MQ tolerance, the letter indicating the type of target detection is marked with an additional star (example "e*"). Thereby, we are alerting the participants to the fact that the uncertainty of the target value can have an impact on the evaluation.

Graphics

The results are shown graphically as follows:



Comparison of Devices

The data in this report allows you to compare the performance of different devices. However, remember to consider the following:


- The chemical control K1 is a ready-to-use commercial control serum. Even if the specimen is of human origin, it is possible that matrix effects occur. These are device-specific and result in different target values.
- Only one specimen was measured. Since the scatter of the results is dependent on the nature of the specimen (matrix effects) and on the signal strength, the determined coefficient of variations (CV in %) cannot be applied generally.
- A large number of runaways is due to administrative errors (wrong unit, results mixed up) or to operator errors (wrong sample, not correctly taken up in solution, not mixed well) and has nothing to do with the type of device.

Zürich, 1.10.2024

Dr. R. Fried
Survey Director

Publication of this report or any portion thereof without our prior written consent is not permitted. The original is archived at www.mqzh.ch


INR



No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Innovin	16	100.0	0.0	0.0	1.53	4.0	e
2 Neoplastin R	15	100.0	0.0	0.0	1.67	5.0	e
3 Neoplastin Plus	8	100.0	0.0	0.0	2.08	6.0	e*
4 Recombiplastin 2G	14	100.0	0.0	0.0	1.68	4.4	e
5 Other methods	13	100.0	0.0	0.0	1.64	6.1	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)


Fibrinogen OA



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens Thrombin	8	75.0	25.0	0.0	1.01	11.9	e*
2	Stago/STA	20	100.0	0.0	0.0	1.14	5.0	e
3	Fibrinogen Q.F.A.	8	100.0	0.0	0.0	1.06	8.1	e*
4	Other methods	6	100.0	0.0	0.0	1.13	5.9	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Activated Prothrombin Time




QUALAB tolerance : 25 %

Activated Prothrombin Time (Sek)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Actin FS	6	100.0	0.0	0.0	38.1	5.7	e
2	Pathromtin SL	7	100.0	0.0	0.0	58.6	4.6	e
3	Stago/STA	22	100.0	0.0	0.0	50.3	5.4	e
4	aPTT-SP	9	100.0	0.0	0.0	40.0	5.6	e
5	Other methods	11	100.0	0.0	0.0	39.4	8.7	e

INR CoaguChek




QUALAB tolerance : 15 %
(< 1.3: +/- 0.2)

INR CoaguChek ()

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CoaguChek Pro II	886	98.3	0.7	1.0	1.2	2.8	e

Prothrombin time NT




QUALAB tolerance : 15 %

Prothrombin time NT (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Innovin	12	100.0	0.0	0.0	100	0.0	e
2	Neoplastin R	12	100.0	0.0	0.0	97	3.5	e
3	Neoplastin Plus	4	100.0	0.0	0.0	99	3.7	e*
4	STA-NeoPTimal	7	100.0	0.0	0.0	92	5.0	e*
5	Recombiplastin 2G	10	100.0	0.0	0.0	100	0.0	e
6	Other methods	15	100.0	0.0	0.0	99	1.4	e

Fibrinogen N




QUALAB tolerance : 15 %

Fibrinogen N (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens Thrombin	10	100.0	0.0	0.0	2.66	8.2	e*
2	Stago/STA	17	100.0	0.0	0.0	3.00	5.0	e
3	Fibrinogen Q.F.A.	14	100.0	0.0	0.0	2.67	4.7	e
4	Fib Clauss (IL)	5	100.0	0.0	0.0	2.86	7.3	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Faktor V




MQ tolerance : 25 %

Faktor V (%)


No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	8	87.5	12.5	0.0	95.5	15.8	e*

aPTT N



No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Actin FS	12	100.0	0.0	0.0	22.8	6.1	e
2 Pathromtin SL	6	100.0	0.0	0.0	37.7	7.1	e
3 Stago/STA	20	100.0	0.0	0.0	31.5	5.8	e
4 aPTT-SP	11	90.9	9.1	0.0	27.7	9.6	e
5 Other methods	5	100.0	0.0	0.0	36.0	1.9	e

Faktor VII




MQ tolerance : 25 %

Faktor VII (%)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	5	100.0	0.0	0.0	109.4	5.8	e

Prothrombin time HT




QUALAB tolerance : 15 %

Prothrombin time HT (%)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Innovin	13	100.0	0.0	0.0	104	4.5	e
2 Neoplastin R	13	100.0	0.0	0.0	97	5.6	e
3 Recombiplastin 2G	11	100.0	0.0	0.0	101	3.0	e
4 Other methods	11	90.9	9.1	0.0	103	7.3	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti-FXa (unfrakt-Heparin)




MQ tolerance : 20 %

Anti-FXa (unfrakt-Heparin) (IU/ml)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Stago/STA	12	100.0	0.0	0.0	0.08	21.5	a
2	ACL	25	100.0	0.0	0.0	0.24	11.4	a
3	Other methods	12	100.0	0.0	0.0	0.26	13.4	a

Fibrinogen H




QUALAB tolerance : 15 %

Fibrinogen H (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens Thrombin	4	100.0	0.0	0.0	3.19	6.6	e*
2	Stago/STA	16	100.0	0.0	0.0	3.22	4.6	e
3	Fibrinogen Q.F.A.	12	100.0	0.0	0.0	2.97	4.5	e
4	Other methods	10	100.0	0.0	0.0	2.99	5.2	e


4 additional results were submitted but not published because the method groups were too small. (< results per group)

aPTT H



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Actin FS	4	100.0	0.0	0.0	31.0	1.6	e
2	Actin FSL	5	100.0	0.0	0.0	32.6	5.2	e
3	Pathromtin SL	7	100.0	0.0	0.0	55.0	6.6	e
4	Stago/STA	14	100.0	0.0	0.0	48.8	7.3	e
5	aPTT-SP	11	100.0	0.0	0.0	45.8	4.5	e
6	Other methods	4	100.0	0.0	0.0	33.0	3.1	e


D-dimer



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche (Zitratplasma)	9	88.9	11.1	0.0	0.58	12.3	e*
2	STA Liatest	17	94.1	0.0	5.9	0.52	11.8	e*
3	Siemens Innovance	12	100.0	0.0	0.0	1.02	8.2	a
4	Pathfast	23	82.7	4.3	13.0	2.78	11.3	e
5	ACL	16	93.7	0.0	6.3	1.06	9.7	e
6	AQT 90 FLEX	5	100.0	0.0	0.0	0.39	6.3	e*
7	VIDAS	16	93.7	6.3	0.0	0.77	8.4	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

D-Dimer Triage




QUALAB tolerance : 21 %

D-Dimer Triage (ng/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Triage	195	93.3	4.1	2.6	385.17	10.7	e

D-dimer qn AFIAS




QUALAB tolerance : 21 %

D-dimer qn AFIAS (mg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 AFIAS	441	78.6	7.3	14.1	0.55	11.9 e

CoaguChek APTT




QUALAB tolerance : 25 %

CoaguChek APTT (Sek)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CoaguChek Pro II	7	100.0	0.0	0.0	40.4	7.1	e

INR CCXS




QUALAB tolerance : 15 %

INR CCXS ()

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CoaguChek XS	1322	99.3	0.2	0.5	2.8	3.6	e

INR HC




QUALAB tolerance : 15 %

INR HC ()

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Hemochron j.	6	83.3	0.0	16.7	5.2	7.1	e*

INR MI




QUALAB tolerance : 15 %

INR MI ()

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	microINR Expert	14	78.6	21.4	0.0	2.0	9.6	e*
2	MicroINR	112	74.1	8.9	17.0	2.0	7.5	e

INR Xprecia




QUALAB tolerance : 15 %

INR Xprecia ()

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Xprecia	44	95.5	0.0	4.5	3.0	3.2	e

INR Lumira Dx




QUALAB tolerance : 15 %

INR Lumira Dx ()

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 Lumira Dx	7	85.7	14.3	0.0	4.7	13.3 e*

Anti-FXa (LMW-Heparin)




MQ tolerance : 20 %

Anti-FXa (LMW-Heparin) (IU/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Stago/STA	14	100.0	0.0	0.0	1.80	6.4	e
2 ACL	23	100.0	0.0	0.0	1.75	8.9	e
3 Other methods	12	91.7	0.0	8.3	1.67	12.4	e*

Anti-FXa (Rivaroxaban)




MQ tolerance : 20 %

Anti-FXa (Rivaroxaban) (µg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Stago/STA	14	100.0	0.0	0.0	320.26	4.5	e
2	ACL	10	100.0	0.0	0.0	312.16	8.5	e*

Anti-FXa (Apixaban)




MQ tolerance : 20 %

Anti-FXa (Apixaban) (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Stago/STA	8	100.0	0.0	0.0	61.34	9.6	a
2	ACL	5	100.0	0.0	0.0	72.80	3.2	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti-FXa (Edoxaban)




MQ tolerance : 20 %

Anti-FXa (Edoxaban) (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	5	100.0	0.0	0.0	44.00	8.2	e*

Anti-FIIa (Dabigatran)




MQ tolerance : 20 %

Anti-FIIa (Dabigatran) (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 all Participants	4	100.0	0.0	0.0	56.00	10.3 e*


Hemoglobin



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Hemocue Hb 801	6	83.3	0.0	16.7	137.0	3.1	e*
2	Automat	9	100.0	0.0	0.0	117.0	3.9	e*
3	Cyanmethemoglobin	13	92.3	7.7	0.0	112.2	4.6	e*
4	Sysmex X	58	98.3	1.7	0.0	116.3	2.3	e
5	Hemocue	386	94.0	1.3	4.7	114.0	3.0	e
6	Hemocontrol	14	100.0	0.0	0.0	114.4	2.7	e
7	DiaSpect	13	76.9	7.7	15.4	110.0	5.8	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex KX21	92	100.0	0.0	0.0	115.8	1.4	e
2	Sysmex PochH - 100i	179	97.8	1.1	1.1	114.6	2.7	e
3	Sysmex XP 300	565	97.7	0.4	1.9	115.7	1.7	e
4	Mythic	217	96.7	0.5	2.8	112.7	2.7	e
5	Sysmex XQ-320	152	94.7	0.0	5.3	116.9	1.5	e
6	Swelab	27	92.6	3.7	3.7	119.3	3.3	e
7	Celltac Alpha (Nihon	52	96.2	0.0	3.8	118.9	2.6	e
8	Samsung HC10	6	100.0	0.0	0.0	116.0	1.7	e
9	Micros 60	53	100.0	0.0	0.0	112.7	2.8	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Automat	5	80.0	0.0	20.0	0.32	1.6	e
2	Sysmex X	58	100.0	0.0	0.0	0.33	2.9	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex XQ-320	151	95.4	1.3	3.3	0.32	2.8	e
2	Sysmex KX21	92	98.9	1.1	0.0	0.32	2.5	e
3	Sysmex PochH - 100i	179	97.8	1.1	1.1	0.33	3.2	e
4	Sysmex XP 300	567	97.7	0.7	1.6	0.31	2.3	e
5	Mythic	217	96.3	0.0	3.7	0.32	3.3	e
6	Swelab	27	92.6	0.0	7.4	0.33	3.6	e
7	Celltac Alpha (Nihon	52	82.7	9.6	7.7	0.32	4.2	e
8	Samsung HC10	6	100.0	0.0	0.0	0.34	4.1	e*
9	Micros 60	53	96.2	3.8	0.0	0.30	4.0	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Automat	6	83.3	16.7	0.0	3.73	15.2	e*
2	Sysmex X	58	100.0	0.0	0.0	3.69	2.8	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex XQ-320	151	97.4	1.3	1.3	3.85	5.4	e
2	Sysmex KX21	92	100.0	0.0	0.0	3.73	2.3	e
3	Sysmex Poch - 100i	179	98.8	0.6	0.6	3.81	3.9	e
4	Sysmex XP 300	567	98.0	0.4	1.6	3.71	3.2	e
5	Mythic	217	97.7	0.5	1.8	3.71	4.3	e
6	Swelab	27	96.3	0.0	3.7	3.69	3.9	e
7	Celltac Alpha (Nihon	52	98.1	0.0	1.9	3.82	2.8	e
8	Samsung HC10	6	100.0	0.0	0.0	3.81	3.7	e
9	Micros 60	53	96.2	1.9	1.9	3.63	5.0	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Automat	5	100.0	0.0	0.0	4.60	8.0	e*
2	Microscopic	4	100.0	0.0	0.0	4.50	10.9	e*
3	Sysmex X	58	98.3	1.7	0.0	4.97	10.5	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex XQ-320	151	98.6	0.7	0.7	4.76	8.2	e
2	Sysmex KX21	92	100.0	0.0	0.0	4.66	6.9	e
3	Sysmex PochH - 100i	179	98.9	1.1	0.0	4.46	8.2	e
4	Sysmex XP 300	567	98.9	0.9	0.2	4.61	7.9	e
5	Mythic	217	96.4	1.8	1.8	4.46	8.7	e
6	Swelab	27	96.3	0.0	3.7	4.80	8.3	e
7	Celltac Alpha (Nihon	52	96.2	1.9	1.9	4.98	9.0	e
8	Samsung HC10	6	83.3	0.0	16.7	4.61	6.8	e
9	Micros 60	53	98.1	0.0	1.9	4.46	8.5	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Automat	5	100.0	0.0	0.0	168.0	10.1	e*
2	Sysmex X	58	100.0	0.0	0.0	160.3	6.9	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex XQ-320	151	98.6	0.7	0.7	156.2	6.9	e
2	Sysmex KX21	92	98.9	1.1	0.0	173.7	8.4	e
3	Sysmex PochH - 100i	179	99.4	0.6	0.0	169.1	6.6	e
4	Sysmex XP 300	567	97.8	1.1	1.1	176.7	6.8	e
5	Mythic	218	94.5	1.8	3.7	178.5	9.2	e
6	Swelab	27	96.3	0.0	3.7	158.4	11.4	e
7	Celltac Alpha (Nihon	52	96.2	3.8	0.0	154.2	9.3	e
8	Samsung HC10	6	100.0	0.0	0.0	172.5	8.3	e*
9	Micros 60	53	98.1	0.0	1.9	166.1	9.5	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Hemoglobin H2




QUALAB tolerance : 9 %

Hemoglobin H2 (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	5	80.0	20.0	0.0	104.0	5.5	e*
2	Microsemi	902	95.7	1.1	3.2	104.4	2.4	e
3	Abx Micros	36	97.2	2.8	0.0	103.5	3.7	e
4	Z3	243	91.8	2.9	5.3	105.1	3.2	e
5	MEK-1303/5	42	95.2	0.0	4.8	107.0	2.0	e

Hematocrit H2




QUALAB tolerance : 9 %

Hematocrit H2 (l/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	5	80.0	0.0	20.0	0.28	3.1	e*
2	Microsemi	902	95.2	1.4	3.4	0.27	3.1	e
3	Abx Micros	36	97.2	0.0	2.8	0.27	3.2	e
4	Z3	244	91.4	2.9	5.7	0.29	4.3	e
5	MEK-1303/5	42	90.5	7.1	2.4	0.33	3.8	e

Leucocytes H2




QUALAB tolerance : 25 %

Leucocytes H2 (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	5	100.0	0.0	0.0	4.35	7.0	e*
2	Microsemi	902	98.2	0.7	1.1	4.37	7.3	e
3	Abx Micros	36	100.0	0.0	0.0	4.38	6.7	e
4	Z3	244	99.6	0.0	0.4	4.38	6.0	e
5	MEK-1303/5	42	92.8	4.8	2.4	4.58	8.4	e

Thrombocytes H2




QUALAB tolerance : 25 %

Thrombocytes H2 (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	5	100.0	0.0	0.0	180.0	13.2	e*
2	Microsemi	902	97.0	1.7	1.3	154.5	8.5	e
3	Abx Micros	36	97.2	2.8	0.0	155.5	8.0	e
4	Z3	243	97.6	0.8	1.6	170.2	7.8	e
5	MEK-1303/5	42	97.6	0.0	2.4	159.9	6.1	e

Erythrocytes H2




QUALAB tolerance : 25 %

Erythrocytes H2 (T/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	5	100.0	0.0	0.0	3.42	6.8	e*
2	Microsemi	902	97.6	0.8	1.6	3.33	4.7	e
3	Abx Micros	36	100.0	0.0	0.0	3.31	4.1	e
4	Z3	244	94.7	0.0	5.3	3.30	3.6	e
5	MEK-1303/5	42	95.2	0.0	4.8	3.47	2.6	e

CRP H2




QUALAB tolerance : 21 %

CRP H2 (mg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	4	100.0	0.0	0.0	24.4	9.7	e*
2	Microsemi	883	93.0	3.6	3.4	27.7	10.4	e
3	Abx Micros	11	90.9	0.0	9.1	28.7	5.2	e
4	ABX Micros CRP200	22	95.5	4.5	0.0	27.5	9.8	e
5	Z3	224	91.5	5.4	3.1	26.1	10.9	e
6	MEK-1303/5	37	94.6	0.0	5.4	22.8	7.9	e

Hemoglobin BG




QUALAB tolerance : 9 %

Hemoglobin BG (g/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 iStat	14	100.0	0.0	0.0	121.8	3.1	e

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 iStat	19	100.0	0.0	0.0	0.35	3.6	e
2 EPOC	16	100.0	0.0	0.0	0.34	2.3	e

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	122	100.0	0.0	0.0	118.5	1.3	e
2	Advia	4	100.0	0.0	0.0	123.0	1.4	e
3	Beckman	5	100.0	0.0	0.0	119.0	0.6	e
4	Yumizen/Pentra	13	100.0	0.0	0.0	120.5	1.3	e

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	122	97.5	2.5	0.0	0.38	3.4	e
2 Advia	4	100.0	0.0	0.0	0.40	0.9	e
3 Beckman	5	100.0	0.0	0.0	0.36	2.9	e*
4 Yumizen/Pentra	13	100.0	0.0	0.0	0.35	2.9	e

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	124	100.0	0.0	0.0	3.92	1.6	e
2	Advia	4	100.0	0.0	0.0	4.06	1.8	e
3	Beckman	5	100.0	0.0	0.0	3.95	0.9	e
4	Yumizen/Pentra	13	100.0	0.0	0.0	3.91	2.7	e

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	123	100.0	0.0	0.0	5.15	6.4	e
2	Advia	4	100.0	0.0	0.0	5.20	3.6	e
3	Beckman	5	100.0	0.0	0.0	5.10	11.6	e*
4	Yumizen/Pentra	13	100.0	0.0	0.0	4.37	10.7	e

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	122	97.6	1.6	0.8	154.1	8.5	e
2	Advia	4	100.0	0.0	0.0	156.5	2.1	e
3	Beckman	5	100.0	0.0	0.0	140.4	17.0	a
4	Yumizen/Pentra	13	100.0	0.0	0.0	154.2	6.2	e

Neutrophils




MQ tolerance : 25 %

Neutrophils (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	123	98.4	1.6	0.0	2.96	7.0	e
2	Advia	4	100.0	0.0	0.0	3.01	2.1	e
3	Beckman	5	100.0	0.0	0.0	2.96	15.8	a
4	Yumizen/Pentra	12	66.7	25.0	8.3	2.05	22.4	e*

Lymphocytes




MQ tolerance : 25 %

Lymphocytes (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	123	100.0	0.0	0.0	1.76	7.1	e
2	Advia	4	100.0	0.0	0.0	1.66	1.2	e
3	Beckman	5	100.0	0.0	0.0	1.55	18.2	a
4	Yumizen/Pentra	12	100.0	0.0	0.0	2.03	12.5	e*

Monocytes




MQ tolerance : 80 %

Monocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	123	99.2	0.8	0.0	0.30	28.3	e
2	Advia	4	100.0	0.0	0.0	0.36	27.2	e*
3	Beckman	5	100.0	0.0	0.0	0.19	56.6	a
4	Yumizen/Pentra	12	83.3	16.7	0.0	0.21	50.6	e*

Eosinophils




MQ tolerance : 80 %
 (< 0.13: +/- 0.10 G/l)

Eosinophils (G/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	122	96.7	3.3	0.0	0.05	30.8	a
2 Advia	4	100.0	0.0	0.0	0.05	36.1	e*
3 Beckman	5	100.0	0.0	0.0	0.10	46.7	e*
4 Yumizen/Pentra	12	100.0	0.0	0.0	0.03	65.9	e*

Basophiles




MQ tolerance : 80 %
 (< 0.13: +/- 0.10 G/l)

Basophiles (G/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	122	93.5	4.9	1.6	0.10	62.2	a
2 Advia	4	100.0	0.0	0.0	0.03	0.0	e
3 Beckman	5	100.0	0.0	0.0	0.05	29.3	e*
4 Yumizen/Pentra	12	100.0	0.0	0.0	0.04	60.2	e*

MCV




MQ tolerance : 25 %

MCV (fl)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	112	100.0	0.0	0.0	95.8	3.5	e
2	Beckman	4	100.0	0.0	0.0	93.2	3.2	e
3	Yumizen/Pentra	9	100.0	0.0	0.0	88.4	3.3	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

MCH




MQ tolerance : 25 %

MCH (pg)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	113	100.0	0.0	0.0	30.2	1.4	e
2	Beckman	4	100.0	0.0	0.0	30.1	0.5	e
3	Yumizen/Pentra	10	100.0	0.0	0.0	31.0	2.8	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

MCHC




MQ tolerance : 25 %

MCHC (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	114	99.1	0.0	0.9	316	3.6	e
2	Beckman	4	100.0	0.0	0.0	322	2.6	e
3	Yumizen/Pentra	10	100.0	0.0	0.0	348	3.3	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

RDW




MQ tolerance : 25 %

RDW (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	108	100.0	0.0	0.0	13.0	4.4	e
2	Beckman	4	100.0	0.0	0.0	14.3	4.7	e
3	Yumizen/Pentra	10	100.0	0.0	0.0	13.5	9.7	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Immature Granulocytes




MQ tolerance : 25 %
 (< 1.30: +/- 0.10 G/l)

Immature Granulocytes (G/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	105	97.1	0.0	2.9	0.03	62.2	e*

Reticulocytes




MQ tolerance : 30 %

Reticulocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	66	98.5	0.0	1.5	40.7	15.4	a
2	Advia	4	100.0	0.0	0.0	27.5	41.0	a
3	Beckman	5	100.0	0.0	0.0	27.5	18.8	a

Hämolyseindex Probe A




MQ tolerance : 30 % Hämolyseindex Probe A (Index (mg/))
 (< 20.000: +/- 8.000 Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	23	100.0	0.0	0.0	17.609	11.3	e
2	Abbott	7	100.0	0.0	0.0	2.000	49.5	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Hämolyseindex Probe B




MQ tolerance : 30 % Hämolyseindex Probe B (Index (mg/))
 (< 20.000: +/- 8.000 Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	23	100.0	0.0	0.0	0.010	0.0	e
2	Abbott	7	100.0	0.0	0.0	0.010	0.0	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Lipemia index A




MQ tolerance : 30 %

Lipemia index A (Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	9	88.9	11.1	0.0	9.00	16.9	e*
2	Abbott	6	100.0	0.0	0.0	5.16	11.4	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Lipemia index B




MQ tolerance : 30 %

Lipemia index B (Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	9	77.8	0.0	22.2	20.00	7.4	e
2	Abbott	6	33.4	33.3	33.3	3.00	29.7	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Icteria Index A




MQ tolerance : 30 %
 (< 5.00: +/- 1.50 Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	9	100.0	0.0	0.0	0.01	0.0	e
2	Abbott	6	100.0	0.0	0.0	0.59	9.2	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Icteria Index B




MQ tolerance : 30 %

Icteria Index B (Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	9	88.9	11.1	0.0	12.00	14.5	e*
2	Abbott	6	100.0	0.0	0.0	12.11	1.8	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

Erythrocyte sedimentation rate 1h




MQ tolerance : 40 %

Erythrocyte sedimentation rate 1h (mm/h)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	MINI-CUBE	28	85.7	10.7	3.6	13	35.1	a
2	miniiSed	11	90.9	0.0	9.1	1	0.0	a
3	Sarstedt Sedivette	17	94.1	0.0	5.9	12	10.3	a
4	Sarstedt Microvette	4	75.0	25.0	0.0	12	34.5	a
5	BD Seditainer	43	88.4	11.6	0.0	10	25.6	a
6	Other methods	20	95.0	5.0	0.0	12	24.9	a

Erythrocyte sedimentation rate 2h



MQ tolerance : 40 %

Erythrocyte sedimentation rate 2h (mm/2h)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sarstedt Sedivette	6	100.0	0.0	0.0	27	6.6	a
2	BD Seditainer	4	75.0	25.0	0.0	31	45.2	a
3	Other methods	4	100.0	0.0	0.0	24	10.7	a

Hemoglobin HS




MQ tolerance : 9 %

Hemoglobin HS (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	PixCell HemoScreen	20	85.0	0.0	15.0	158.2	3.8	e

Hematocrit HS




MQ tolerance : 9 %

Hematocrit HS (l/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	PixCell HemoScreen	20	85.0	0.0	15.0	0.4	3.3	e

Erythrocytes HS




MQ tolerance : 25 %

Erythrocytes HS (T/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	PixCell HemoScreen	20	85.0	5.0	10.0	4.53	10.2	e

Leucocytes HS



MQ tolerance : 25 %

Leucocytes HS (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	PixCell HemoScreen	20	95.0	0.0	5.0	7.16	8.9	e

Trombocytes HS




MQ tolerance : 25 %

Trombocytes HS (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	PixCell HemoScreen	20	100.0	0.0	0.0	234.6	4.3	e

Leucocytes BF




MQ tolerance : 40 %

Leucocytes BF (G/l)

No.Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	14	100.0	0.0	0.0	0.060	18.2	a

Erythrocytes BF




MQ tolerance : 40 %

Erythrocytes BF (G/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	13	100.0	0.0	0.0	22.427	10.2	e

Total cells (TC)




MQ tolerance : 40 %

Total cells (TC) (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	11	100.0	0.0	0.0	0.060	18.0	a

Mononuclear cells (MN)




MQ tolerance : 40 %

Mononuclear cells (MN) (G/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	14	100.0	0.0	0.0	0.048	10.8	a

Polynuclear cells (PMN)




MQ tolerance : 40 %

Polynuclear cells (PMN) (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	14	92.9	0.0	7.1	0.004	61.8	a


CRP



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Autolyser	9	88.9	0.0	11.1	23.4	5.0	e
2	Cobas b101	415	99.8	0.0	0.2	22.1	5.0	e
3	Siemens	8	100.0	0.0	0.0	25.4	7.8	e*
4	Cobas	42	100.0	0.0	0.0	24.2	5.0	e
5	Turbidimetry	4	50.0	0.0	50.0	25.2	2.0	a
6	Afinion	1076	99.1	0.1	0.8	24.7	4.7	e
7	NycoCard SingleTest-	50	80.0	6.0	14.0	23.6	11.6	e
8	Quick Read go	86	98.8	0.0	1.2	24.0	8.0	e
9	Eurolyser	55	78.2	3.6	18.2	30.4	10.5	e
10	Fuji Dri-Chem	12	83.3	0.0	16.7	25.4	9.1	e*
11	Piccolo	6	66.6	16.7	16.7	33.6	12.7	e*
12	Celltac chemi	20	100.0	0.0	0.0	23.1	4.8	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

CRP




QUALAB tolerance : 21 %

CRP (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	QuikRead (Vollblut)	10	100.0	0.0	0.0	42.3	3.5	e

CRP




QUALAB tolerance : 21 %

CRP (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Spinit	8	100.0	0.0	0.0	21.2	6.6	e
2	Abbott	14	100.0	0.0	0.0	25.7	2.4	e
3	Beckman	5	80.0	0.0	20.0	23.0	8.9	e*
4	AQT 90 FLEX	5	100.0	0.0	0.0	26.0	4.2	e
5	Spotchem D-Concept	5	100.0	0.0	0.0	26.8	6.4	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

IgG




QUALAB tolerance : 15 %

IgG (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Turbidimetry	22	100.0	0.0	0.0	10.57	2.9	e
2	Nephelometry	5	100.0	0.0	0.0	11.20	8.9	a
3	Other methods	4	100.0	0.0	0.0	9.76	6.9	e*

IgA




QUALAB tolerance : 15 %

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Turbidimetry	20	100.0	0.0	0.0	1.76	2.8	e
2	Nephelometry	7	100.0	0.0	0.0	1.94	6.6	a

3 additional results were submitted but not published because the method groups were too small. (< results per group)

IgM




QUALAB tolerance : 15 %

IgM (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Turbidimetry	22	100.0	0.0	0.0	0.87	4.8	e
2	Nephelometry	7	100.0	0.0	0.0	0.90	6.1	e*


IgE



QUALAB tolerance : 30 %


No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	5	100.0	0.0	0.0	88	12.8	e*
2	Cobas	6	100.0	0.0	0.0	122	5.7	e

Alpha-1-Antitrypsin



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Nephelometry	4	100.0	0.0	0.0	1.03	4.4	e
2	all Participants	9	100.0	0.0	0.0	1.05	7.9	e

Anti-Streptolysin-Antibodies




MQ tolerance : 25 %

Anti-Streptolysin-Antibodies (kIU/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	11	100.0	0.0	0.0	112	8.4	e
2	Other methods	6	83.3	0.0	16.7	84	1.7	e

Complement C3




QUALAB tolerance : 15 %

Complement C3 (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	20	100.0	0.0	0.0	1.30	5.1	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Complement C4




QUALAB tolerance : 15 %

Complement C4 (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Alinity	6	100.0	0.0	0.0	0.25	4.6	e*
2	Other methods	15	100.0	0.0	0.0	0.25	8.9	a

Haptoglobin




MQ tolerance : 25 %

Haptoglobin (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	29	100.0	0.0	0.0	1.19	3.2	e

Transferrin




MQ tolerance : 25 %

Transferrin (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	37	100.0	0.0	0.0	1.90	2.3	e

Beta-2 microglobuline




MQ tolerance : 25 %

Beta-2 microglobuline (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	12	100.0	0.0	0.0	1.38	6.6	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)


Rheumatoid factor



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	7	100.0	0.0	0.0	29.0	2.3	e
2	Other methods	11	100.0	0.0	0.0	26.9	4.6	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Ceruloplasmin




MQ tolerance : 25 %

Ceruloplasmin (mg/l)


No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens	9	100.0	0.0	0.0	261.00	5.2	e
2	Other methods	4	100.0	0.0	0.0	191.50	2.4	e

Prealbumin



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	19	100.0	0.0	0.0	175.26	5.5	e

Soluble transferrin receptor




MQ tolerance : 25 %

Soluble transferrin receptor (mg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	14	100.0	0.0	0.0	2.2	2.7	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

free light chain kappa




QUALAB tolerance : 20 %

free light chain kappa (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	N Latex	6	100.0	0.0	0.0	9.39	3.4	e
2	Freelite	13	100.0	0.0	0.0	18.69	5.8	e

free light chain lambda




QUALAB tolerance : 20 %

free light chain lambda (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	N Latex	6	100.0	0.0	0.0	10.40	5.9	e*
2	Freelite	13	100.0	0.0	0.0	15.35	6.5	e

IgE peanut qn




MQ tolerance : 30 %

IgE peanut qn (kU/L)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	14	78.6	21.4	0.0	51.11	19.9	e*

IgE birch qn




QUALAB tolerance : 30 %

IgE birch qn (kU/L)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	15	86.7	13.3	0.0	11.51	16.9	e*

IgE cat qn




QUALAB tolerance : 30 %

IgE cat qn (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	14	92.9	7.1	0.0	6.96	14.8	e*

IgE total




QUALAB tolerance : 30 %

IgE total (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	14	100.0	0.0	0.0	521	12.0	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

IgE sx1 qn




MQ tolerance : 30 %

IgE sx1 qn (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	10	80.0	0.0	20.0	15.15	13.0	e*

IgE fx5 qn




MQ tolerance : 30 %

IgE fx5 qn (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	10	100.0	0.0	0.0	50.66	14.3	e*

IgE rx1qn




MQ tolerance : 30 %

IgE rx1qn (kU/L)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 all Participants	8	100.0	0.0	0.0	6.60	17.0 a

IgE rx2 qn




MQ tolerance : 30 %

IgE rx2 qn (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	8	100.0	0.0	0.0	7.77	18.5	e*

IgE D. pteronyssinus qn




QUALAB tolerance : 30 %
 (< 1.50: +/- 0.45 kU/L)

IgE D. pteronyssinus qn (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	12	100.0	0.0	0.0	1.39	11.9	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

CRP HS




QUALAB tolerance : 21 %
 (< 10.00: +/- 2.00 mg/l)

CRP HS (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Turbidimetry	10	100.0	0.0	0.0	5.37	10.5	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Lipoprotein (a)




MQ tolerance : 25 %

Lipoprotein (a) (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	6	100.0	0.0	0.0	50	2.8	e
2	Others	6	100.0	0.0	0.0	65	10.2	e*

CRP




QUALAB tolerance : 21 %

CRP (mg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 AFIAS	193	93.8	2.1	4.1	29.7	9.0	e

Anti deam. Gliadin IgG




MQ tolerance : 30 %

Anti deam. Gliadin IgG (U/ml)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Phadia	8	100.0	0.0	0.0	26.00	10.0	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti deam. Gliadin IgA




MQ tolerance : 30 %

Anti deam. Gliadin IgA (U/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Phadia	7	100.0	0.0	0.0	32.00	15.6	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti tTG IgG




MQ tolerance : 30 %

Anti tTG IgG (U/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Other methods	4	100.0	0.0	0.0	135.00	8.1	e*

Anti tTG IgA




MQ tolerance : 30 %

Anti tTG IgA (U/ml)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	8	75.0	0.0	25.0	128.00	0.0	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

CRP Lumira




QUALAB tolerance : 21 %

CRP Lumira (mg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 Lumira Dx	11	81.8	18.2	0.0	12.3	16.0 a

Albumine




QUALAB tolerance : 12 %

Albumine (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	11	100.0	0.0	0.0	35.6	1.4	e
2	Beckman	4	100.0	0.0	0.0	37.0	1.5	e
3	Roche	35	100.0	0.0	0.0	38.3	3.0	e
4	Siemens	5	100.0	0.0	0.0	36.0	7.3	e*
5	Autolyser	8	100.0	0.0	0.0	37.2	3.7	e
6	Selectra Pro	10	100.0	0.0	0.0	36.0	2.1	e
7	Fuji Dri-Chem	247	98.4	1.2	0.4	46.1	4.4	e
8	Spotchem D-Concept	236	95.4	3.8	0.8	44.7	5.9	e
9	Spotchem SP-4430	29	96.6	3.4	0.0	42.6	4.6	e
10	Piccolo	59	98.3	0.0	1.7	39.3	4.1	e
11	Skylla	4	100.0	0.0	0.0	39.5	2.4	e
12	Hitachi S40/M40	5	80.0	0.0	20.0	31.0	1.7	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Alkaline phosphatase




QUALAB tolerance : 18 %

Alkaline phosphatase (U/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	14	100.0	0.0	0.0	96	3.1	e
2	Beckman	6	100.0	0.0	0.0	96	5.4	e*
3	Roche	35	100.0	0.0	0.0	85	4.5	e
4	Siemens	7	100.0	0.0	0.0	87	5.2	e
5	Autolyser	21	100.0	0.0	0.0	91	6.8	e
6	Selectra Pro	13	92.3	7.7	0.0	99	7.5	e
7	Fuji Dri-Chem	1017	97.5	2.0	0.5	100	6.9	e
8	Spotchem D-Concept	555	98.4	0.5	1.1	85	6.0	e
9	Spotchem SP-4430	80	97.5	2.5	0.0	81	8.2	e
10	Piccolo	50	100.0	0.0	0.0	104	6.4	e
11	Skylla	5	100.0	0.0	0.0	101	4.9	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Amylase




QUALAB tolerance : 18 %
(< 50: +/- 9 U/l)

Amylase (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	10	90.0	0.0	10.0	121	2.8	e
2	Beckman	4	100.0	0.0	0.0	147	5.7	e*
3	Roche	14	100.0	0.0	0.0	136	2.5	e
4	Autolyser	8	100.0	0.0	0.0	118	5.6	e
5	Selectra Pro	9	100.0	0.0	0.0	180	4.3	e
6	Fuji Dri-Chem	734	99.5	0.4	0.1	95	4.4	e
7	Spotchem D-Concept	400	96.0	3.5	0.5	63	7.7	e
8	Spotchem SP-4430	63	85.7	1.6	12.7	21	18.7	e*
9	Piccolo	55	100.0	0.0	0.0	123	3.1	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Pancreatic amylase




QUALAB tolerance : 18 %

Pancreatic amylase (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	97	2.5	e
2	Roche	17	100.0	0.0	0.0	112	2.1	e
3	Siemens	4	100.0	0.0	0.0	117	15.0	e*
4	Autolyser	10	100.0	0.0	0.0	103	4.5	e

Bilirubin




QUALAB tolerance : 18 %

Bilirubin (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	14	92.9	7.1	0.0	22.3	8.5	e*
2	Beckman	6	100.0	0.0	0.0	29.4	8.9	a
3	Roche	34	97.1	2.9	0.0	21.5	6.7	e
4	Siemens	7	85.7	14.3	0.0	28.0	10.4	e*
5	Autolyser	18	88.8	5.6	5.6	22.0	9.7	e*
6	Selectra Pro	15	73.3	6.7	20.0	18.8	11.0	e*
7	Fuji Dri-Chem	829	92.3	4.9	2.8	26.2	8.9	e
8	Spotchem D-Concept	442	95.5	2.5	2.0	27.9	7.8	e
9	Spotchem SP-4430	75	90.7	5.3	4.0	26.4	8.8	e
10	Piccolo	57	96.5	0.0	3.5	22.3	9.0	e
11	Skyla	5	60.0	40.0	0.0	26.6	15.7	a

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Bilirubin direct




MQ tolerance : 18 %

Bilirubin direct (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Autolyser	12	91.7	0.0	8.3	15.9	8.3	e*
2	Fuji Dri-Chem	23	78.3	13.0	8.7	12.8	11.6	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Calcium




QUALAB tolerance : 9 %
(< 2.00: +/- 0.18 mmol/l)

Calcium (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	17	100.0	0.0	0.0	2.03	2.9	e
2	Beckman	5	100.0	0.0	0.0	2.02	2.2	e
3	Roche	37	100.0	0.0	0.0	2.00	1.7	e
4	Siemens	5	100.0	0.0	0.0	2.03	3.8	e*
5	Autolyser	9	100.0	0.0	0.0	2.10	4.6	a
6	Fuji Dri-Chem	265	97.8	1.1	1.1	2.03	4.1	e
7	Spotchem D-Concept	75	90.6	6.7	2.7	1.43	6.5	e
8	Spotchem SP-4430	12	91.7	0.0	8.3	1.91	2.0	e
9	Piccolo	50	98.0	0.0	2.0	2.00	3.2	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

Calcium ISE




MQ tolerance : 12 %

Calcium ISE (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	iStat Chem8	11	90.9	9.1	0.0	0.75	4.3	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Chloride




QUALAB tolerance : 6 %

Chloride (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	17	100.0	0.0	0.0	105	1.9	e
2	Beckman	5	80.0	20.0	0.0	102	3.4	e*
3	Roche	27	96.3	3.7	0.0	103	1.7	e
4	Siemens	5	100.0	0.0	0.0	105	0.9	e
5	Fuji Dri-Chem	937	98.4	0.9	0.7	112	1.7	e
6	Spotchem D-Concept	423	98.6	0.9	0.5	113	1.7	e
7	Spotchem EL-SE 1520	61	93.5	1.6	4.9	118	2.6	e
8	Piccolo	27	100.0	0.0	0.0	104	2.0	e
9	Exias	21	100.0	0.0	0.0	106	1.5	e
10	iStat Chem8	11	100.0	0.0	0.0	111	0.9	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Cholesterol total




QUALAB tolerance : 10 %

Cholesterol total (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	18	100.0	0.0	0.0	4.22	2.7	e
2	Beckman	4	100.0	0.0	0.0	4.18	6.9	e*
3	Roche	29	100.0	0.0	0.0	4.14	2.2	e
4	Siemens	4	100.0	0.0	0.0	4.29	1.9	e
5	Autolyser	21	100.0	0.0	0.0	4.21	4.2	e
6	Selectra Pro	13	100.0	0.0	0.0	4.31	3.2	e
7	Fuji Dri-Chem	970	96.3	1.8	1.9	3.91	3.9	e
8	Spotchem D-Concept	470	96.8	1.3	1.9	4.11	3.5	e
9	Spotchem SP-4430	80	96.2	1.3	2.5	3.93	3.7	e
10	Piccolo	24	100.0	0.0	0.0	4.19	2.2	e
11	Reflotron	4	75.0	25.0	0.0	3.94	8.3	e*
12	Cholestech LDX	265	94.7	1.9	3.4	3.88	4.1	e
13	Other methods	4	100.0	0.0	0.0	3.44	1.5	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

Cholesterin HDL




QUALAB tolerance : 21 %

Cholesterin HDL (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	0.97	4.1	e
2	Beckman	6	100.0	0.0	0.0	1.21	5.5	e
3	Roche	27	100.0	0.0	0.0	0.93	4.2	e
4	Dimension	4	100.0	0.0	0.0	1.71	5.9	e*
5	Autolyser	21	95.2	4.8	0.0	1.72	11.4	e
6	Selectra Pro	13	84.6	0.0	15.4	1.65	11.2	e*
7	Fuji Dri-Chem	938	98.6	0.2	1.2	1.50	4.3	e
8	Spotchem D-Concept	454	95.2	3.5	1.3	0.73	9.4	e
9	Spotchem SP-4430	73	90.5	6.8	2.7	0.69	9.8	e
10	Piccolo	23	95.7	0.0	4.3	1.17	10.3	e
11	Cholestech LDX	265	93.2	4.2	2.6	1.04	9.7	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

Creatine kinase




QUALAB tolerance : 18 %

Creatine kinase (U/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	176	1.8	e
2	Beckman	5	100.0	0.0	0.0	178	3.2	e
3	Roche	32	100.0	0.0	0.0	169	1.7	e
4	Siemens	7	85.7	0.0	14.3	167	2.1	e
5	Autolyser	17	94.1	0.0	5.9	192	3.5	e
6	Selectra Pro	8	100.0	0.0	0.0	168	3.9	e
7	Fuji Dri-Chem	663	98.0	1.1	0.9	182	4.9	e
8	Spotchem D-Concept	318	98.1	0.6	1.3	149	6.5	e
9	Spotchem SP-4430	39	87.2	7.7	5.1	159	9.5	e
10	Piccolo	19	94.7	0.0	5.3	172	6.9	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Cholesterol LDL




QUALAB tolerance : 18 %

Cholesterol LDL (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	2.1	4.7	e
2	Beckman	5	100.0	0.0	0.0	2.7	5.8	e*
3	Roche, Cobas	17	100.0	0.0	0.0	2.9	3.6	e
4	Autolyser	10	90.0	0.0	10.0	2.9	7.3	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Iron




QUALAB tolerance : 20 %

Iron (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	36	2.6	e
2	Roche	20	100.0	0.0	0.0	36	0.9	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Gamma-glutamyltransferase




QUALAB tolerance : 18 %
(< 40: +/- 8 U/l)

Gamma-glutamyltransferase (U/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Abbott	15	100.0	0.0	0.0	44	6.1	e
2 Beckman	6	83.3	16.7	0.0	41	8.4	e*
3 Cobas	36	97.2	2.8	0.0	38	7.2	e
4 Siemens	7	100.0	0.0	0.0	45	6.7	e*
5 Autolyser	21	95.2	0.0	4.8	39	2.7	e
6 Selectra Pro	14	92.9	7.1	0.0	34	13.5	e*
7 Fuji Dri-Chem	1130	99.6	0.0	0.4	47	4.9	e
8 Spotchem D-Concept	623	98.5	0.5	1.0	44	6.9	e
9 Spotchem SP-4430	129	95.3	3.9	0.8	50	8.6	e
10 Piccolo	60	100.0	0.0	0.0	36	5.5	e
11 Skyla	5	100.0	0.0	0.0	36	5.3	e*

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Glucose




QUALAB tolerance : 9 %

Glucose (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	16	100.0	0.0	0.0	6.3	1.2	e
2	Beckman	6	83.3	16.7	0.0	6.3	6.8	e*
3	Roche	38	100.0	0.0	0.0	6.3	2.0	e
4	Siemens	6	83.3	0.0	16.7	6.4	1.1	e
5	Autolyser	19	94.7	0.0	5.3	6.3	2.8	e
6	Selectra Pro	15	86.7	13.3	0.0	6.5	5.3	e*
7	Fuji Dri-Chem	1074	98.9	0.3	0.8	6.0	2.5	e
8	Spotchem D-Concept	584	96.7	2.4	0.9	6.2	4.1	e
9	Spotchem SP-4430	107	96.3	0.9	2.8	6.6	4.3	e
10	Piccolo	71	100.0	0.0	0.0	6.4	1.3	e
11	Cholestech LDX	260	95.0	2.3	2.7	6.0	3.7	e
12	iStat Chem8	12	100.0	0.0	0.0	5.8	2.0	e
13	Cobas Pulse	54	94.4	3.7	1.9	5.8	3.5	e

8 additional results were submitted but not published because the method groups were too small. (< results per group)


Glucose



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Accu-Chek Instant	127	96.8	0.8	2.4	6.0	3.4	e
2	Accu-Chek Aviva	102	90.2	2.9	6.9	7.0	3.7	e
3	Accu-Chek Inform 2	896	98.2	1.0	0.8	7.4	3.1	e
4	Accu-Check Guide	341	97.6	0.6	1.8	5.9	3.1	e
5	Contour XT	1393	96.5	2.6	0.9	6.3	3.6	e
6	Skylla	5	100.0	0.0	0.0	6.5	1.7	e
7	Statstrip/Xpress	87	100.0	0.0	0.0	6.0	3.0	e
8	Glucocard	6	66.7	33.3	0.0	8.7	6.9	e*
9	Hemocue 201+ P-equiv	148	96.6	0.0	3.4	7.9	2.8	e
10	Hemocue 201RT P-equiv	131	96.1	0.8	3.1	7.9	2.7	e
11	Contour NEXT	48	93.7	6.3	0.0	6.0	3.8	e

11 additional results were submitted but not published because the method groups were too small. (< results per group)

Glucose B




QUALAB tolerance : 9 %

Glucose B (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	OneTouch Verio	29	93.2	3.4	3.4	6.2	4.9	e
2	Contour 2 (5s)	8	100.0	0.0	0.0	5.0	3.8	e*
3	Healthpro	21	90.5	0.0	9.5	10.8	4.7	e
4	Mylife UNIO	345	96.9	1.7	1.4	7.2	3.7	e
5	mylife Pura	89	84.3	10.1	5.6	7.2	6.1	e
6	Alpha Check	25	88.0	4.0	8.0	9.8	4.2	e

Uric Acid




QUALAB tolerance : 12 %

Uric Acid (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	17	100.0	0.0	0.0	274	2.8	e
2	Beckman	6	100.0	0.0	0.0	288	4.3	e*
3	Roche	33	100.0	0.0	0.0	272	2.3	e
4	Siemens	5	100.0	0.0	0.0	287	2.6	e
5	Autolyser	18	100.0	0.0	0.0	292	3.5	e
6	Selectra Pro	15	100.0	0.0	0.0	268	5.1	e
7	Fuji Dri-Chem	1047	98.8	0.4	0.8	307	2.7	e
8	Spotchem D-Concept	587	99.3	0.2	0.5	270	4.1	e
9	Spotchem SP-4430	103	92.2	3.9	3.9	231	5.8	e
10	Piccolo	36	94.4	0.0	5.6	209	5.5	e
11	Skylla	5	100.0	0.0	0.0	281	3.3	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Urea




QUALAB tolerance : 15 %

Urea (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	7.9	3.8	e
2	Beckman	6	100.0	0.0	0.0	7.6	3.9	e
3	Roche	33	97.0	3.0	0.0	7.3	5.2	e
4	Siemens	7	100.0	0.0	0.0	7.8	6.2	e*
5	Autolyser	16	93.7	0.0	6.3	8.1	6.1	e
6	Selectra Pro	11	100.0	0.0	0.0	7.8	5.6	e
7	Fuji Dri-Chem	617	98.9	0.0	1.1	7.8	2.5	e
8	Spotchem D-Concept	329	93.6	4.9	1.5	7.1	7.4	e
9	Spotchem SP-4430	55	94.6	3.6	1.8	7.3	7.4	e
10	Piccolo	65	98.5	0.0	1.5	6.8	3.0	e
11	Skyla	5	100.0	0.0	0.0	7.8	2.7	e
12	iStat Chem8	9	100.0	0.0	0.0	10.1	2.8	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Potassium




QUALAB tolerance : 6 %
(< 3.30: +/- 0.20 mmol/l)

Potassium (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	17	100.0	0.0	0.0	3.08	1.1	e
2	Beckman	6	100.0	0.0	0.0	3.08	3.3	e*
3	Roche	36	100.0	0.0	0.0	3.08	1.5	e
4	Siemens	5	100.0	0.0	0.0	3.00	1.5	e
5	Autolyser	16	75.0	12.5	12.5	2.92	5.3	e*
6	Fuji Dri-Chem	1100	99.5	0.3	0.2	2.99	1.3	e
7	Spotchem D-Concept	543	96.7	2.2	1.1	2.80	2.0	e
8	Spotchem EL-SE 1520	79	93.7	6.3	0.0	2.80	2.4	e
9	Piccolo	41	48.8	31.7	19.5	3.20	5.6	e*
10	iStat Chem8	15	100.0	0.0	0.0	3.01	0.9	e
11	Exias	23	100.0	0.0	0.0	3.08	1.1	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

Creatinine




QUALAB tolerance : 18 %

Creatinine (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	183	3.3	e
2	Beckman	6	100.0	0.0	0.0	187	6.5	e*
3	Roche	35	100.0	0.0	0.0	184	2.9	e
4	Siemens	7	100.0	0.0	0.0	185	2.8	e
5	Enzymatic	4	100.0	0.0	0.0	183	6.3	e*
6	Autolyser	21	100.0	0.0	0.0	176	3.9	e
7	Selectra Pro	15	100.0	0.0	0.0	178	6.9	e
8	Fuji Dri-Chem	1162	99.0	0.3	0.7	166	4.4	e
9	Spotchem D-Concept	641	98.7	0.2	1.1	89	5.2	e
10	Spotchem SP-4430	148	98.0	0.0	2.0	86	5.1	e
11	Piccolo	67	98.5	0.0	1.5	185	5.3	e
12	Skyla	5	100.0	0.0	0.0	178	7.1	e*
13	Reflotron	6	66.6	16.7	16.7	201	10.8	e*
14	EPOC	11	72.7	0.0	27.3	178	9.2	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Creatinine E




QUALAB tolerance : 18 %

Creatinine E (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	iStat Chem8	46	97.8	0.0	2.2	181	2.7	e
2	ABL700/800	15	100.0	0.0	0.0	180	7.7	e

eGFR CKD-EPI




MQ tolerance : 30 %

eGFR CKD-EPI ()

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	56	98.2	0.0	1.8	30	6.3	e
2	Fuji Dri-Chem	393	95.9	1.0	3.1	34	7.0	e
3	Spotchem	264	91.7	3.4	4.9	75	10.3	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

eGFR Cockcroft-Gault




MQ tolerance : 30 %

eGFR Cockcroft-Gault ()

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Fuji Dri-Chem	33	84.9	3.0	12.1	38	10.6	e
2	Spotchem	16	93.7	0.0	6.3	72	4.5	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

LDH




QUALAB tolerance : 18 %

LDH (U/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	17	100.0	0.0	0.0	129	5.3	e
2	Beckman	5	100.0	0.0	0.0	139	8.4	e*
3	Roche	34	100.0	0.0	0.0	138	2.2	e
4	Siemens	4	100.0	0.0	0.0	139	8.9	e*
5	Autolyser	7	100.0	0.0	0.0	137	6.6	e*
6	Fuji Dri-Chem	113	96.4	0.9	2.7	119	5.0	e
7	Spotchem D-Concept	42	78.6	14.3	7.1	64	12.6	e
8	Spotchem SP-4430	12	91.7	8.3	0.0	75	10.5	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Magnesium




QUALAB tolerance : 12 %
 (< 0.70: +/- 0.09 mmol/l)

Magnesium (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	16	100.0	0.0	0.0	0.66	3.2	e
2	Beckman	4	100.0	0.0	0.0	0.67	1.9	e
3	Roche	28	89.3	10.7	0.0	0.69	6.4	e
4	Siemens	4	100.0	0.0	0.0	0.68	7.2	a
5	Fuji Dri-Chem	74	93.2	5.4	1.4	0.72	4.6	e
6	Spotchem D-Concept	39	100.0	0.0	0.0	0.51	4.6	e
7	Spotchem SP-4430	4	100.0	0.0	0.0	0.56	1.5	e

9 additional results were submitted but not published because the method groups were too small. (< results per group)


Sodium



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Autolyser	6	100.0	0.0	0.0	148	1.4	e
2	Abbott	18	100.0	0.0	0.0	143	1.1	e
3	Beckman	6	83.3	16.7	0.0	142	3.3	e*
4	Roche	36	100.0	0.0	0.0	143	1.3	e
5	Siemens	5	100.0	0.0	0.0	144	0.9	e
6	Fuji Dri-Chem	1017	99.2	0.3	0.5	145	1.1	e
7	Spotchem D-Concept	469	99.6	0.2	0.2	137	1.2	e
8	Spotchem EL-SE 1520	67	98.5	1.5	0.0	137	1.4	e
9	Piccolo	37	100.0	0.0	0.0	140	1.4	e
10	Exias	23	100.0	0.0	0.0	143	1.0	e
11	iStat Chem8	15	100.0	0.0	0.0	142	0.7	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Phosphate




QUALAB tolerance : 15 %

Phosphate (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	16	100.0	0.0	0.0	1.17	5.4	e
2	Beckman	4	100.0	0.0	0.0	1.17	5.0	e*
3	Roche	33	100.0	0.0	0.0	1.14	3.5	e
4	Siemens	4	100.0	0.0	0.0	1.23	4.8	e*
5	Fuji Dri-Chem	73	90.5	2.7	6.8	1.21	6.6	e
6	Spotchem D-Concept	14	85.8	7.1	7.1	1.18	7.7	e*
7	Piccolo	4	100.0	0.0	0.0	1.52	1.0	e

10 additional results were submitted but not published because the method groups were too small. (< results per group)

Protein total




QUALAB tolerance : 12 %

Protein total (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	93.3	6.7	0.0	57.7	4.1	e
2	Beckman	4	100.0	0.0	0.0	59.1	3.2	e*
3	Roche	30	93.3	0.0	6.7	57.7	2.3	e
4	Siemens	5	100.0	0.0	0.0	60.0	2.2	e
5	Selectra Pro	8	100.0	0.0	0.0	58.0	2.3	e
6	Fuji Dri-Chem	185	98.9	0.0	1.1	56.0	3.3	e
7	Spotchem D-Concept	183	96.2	2.7	1.1	59.0	5.6	e
8	Spotchem SP-4430	26	92.3	0.0	7.7	61.4	3.1	e
9	Piccolo	49	100.0	0.0	0.0	58.3	1.7	e
10	Skylla	5	100.0	0.0	0.0	62.0	2.7	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Aspartate aminotransferase




QUALAB tolerance : 18 %

Aspartate aminotransferase (U/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	17	100.0	0.0	0.0	75	4.1	e
2	Beckman	5	100.0	0.0	0.0	75	7.5	e*
3	Roche	37	100.0	0.0	0.0	73	4.4	e
4	Siemens	5	100.0	0.0	0.0	72	3.0	e
5	Autolyser	21	95.2	0.0	4.8	65	5.0	e
6	Selectra Pro	15	100.0	0.0	0.0	78	6.3	e
7	Fuji Dri-Chem	1139	98.9	0.4	0.7	71	4.3	e
8	Spotchem D-Concept	626	99.3	0.2	0.5	47	5.0	e
9	Spotchem SP-4430	139	98.6	0.7	0.7	48	5.6	e
10	Piccolo	73	100.0	0.0	0.0	63	4.7	e
11	Skylla	5	100.0	0.0	0.0	66	5.2	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Alanine aminotransferase




QUALAB tolerance : 18 %

Alanine aminotransferase (U/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	53	5.6	e
2	Beckman	6	100.0	0.0	0.0	60	6.4	e*
3	Roche	37	100.0	0.0	0.0	60	4.2	e
4	Siemens	7	85.7	14.3	0.0	64	8.9	e*
5	Autolyser	21	95.2	0.0	4.8	60	5.3	e
6	Selectra Pro	15	93.3	0.0	6.7	69	8.9	e*
7	Fuji Dri-Chem	1157	98.5	0.5	1.0	63	5.0	e
8	Spotchem D-Concept	631	98.9	0.5	0.6	40	5.8	e
9	Spotchem SP-4430	139	97.2	1.4	1.4	46	7.3	e
10	Piccolo	71	98.6	0.0	1.4	54	5.8	e
11	Skylla	5	100.0	0.0	0.0	63	4.3	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Triglycerides




QUALAB tolerance : 18 %

Triglycerides (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	14	100.0	0.0	0.0	2.24	2.6	e
2	Beckman	5	100.0	0.0	0.0	2.30	4.9	e*
3	Roche	29	100.0	0.0	0.0	2.33	1.6	e
4	Siemens	6	100.0	0.0	0.0	2.28	4.5	e
5	Autolyser	21	95.2	4.8	0.0	2.21	7.2	e
6	Selectra Pro	12	100.0	0.0	0.0	2.40	5.7	e
7	Fuji Dri-Chem	957	98.6	1.0	0.4	2.83	5.4	e
8	Spotchem D-Concept	454	95.8	0.9	3.3	1.39	5.3	e
9	Spotchem SP-4430	75	93.4	1.3	5.3	1.36	5.1	e
10	Piccolo	21	100.0	0.0	0.0	2.63	3.1	e
11	Cholestech LDX	264	98.5	1.1	0.4	2.20	4.9	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Lithium




QUALAB tolerance : 15 %

Lithium (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	6	100.0	0.0	0.0	1.08	4.7	e*
2	Roche	14	100.0	0.0	0.0	1.15	3.0	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Lactate




QUALAB tolerance : 18 %

Lactate (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	7	100.0	0.0	0.0	2.48	1.2	e
2	Roche	17	100.0	0.0	0.0	2.43	3.0	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

HbA1c sample A




QUALAB tolerance : 9 %

HbA1c sample A (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	5	100.0	0.0	0.0	5.9	2.4	e
2	Roche, Cobas	19	100.0	0.0	0.0	5.9	4.1	e
3	HPLC	11	100.0	0.0	0.0	5.8	2.5	e
4	Afinion	522	99.0	0.4	0.6	5.8	2.5	e
5	Cobas b101	210	98.6	1.4	0.0	5.5	3.5	e
6	DCA2000/Vantage	138	96.4	2.2	1.4	5.9	3.6	e
7	Celltac chemi	20	100.0	0.0	0.0	5.4	4.2	e
8	NycoCard	7	71.4	0.0	28.6	5.9	3.7	c
9	Eurolyser	6	83.3	16.7	0.0	5.8	5.7	e*
10	A1c Now	219	91.3	5.5	3.2	5.4	4.4	e
11	AFIAS	127	78.8	15.7	5.5	5.1	6.0	e
12	Others	19	100.0	0.0	0.0	5.8	4.1	e
13	Quick Read go	4	75.0	0.0	25.0	5.9	1.0	c

2 additional results were submitted but not published because the method groups were too small. (< results per group)

HbA1c sample B




QUALAB tolerance : 9 %

HbA1c sample B (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	5.3	2.4	e*
2	Roche, Cobas	18	100.0	0.0	0.0	5.4	3.8	e
3	HPLC	11	100.0	0.0	0.0	5.4	3.3	e
4	Afinion	731	99.5	0.1	0.4	5.3	2.5	e
5	Cobas b101	207	98.1	1.9	0.0	5.1	3.2	e
6	DCA2000/Vantage	189	95.3	4.2	0.5	5.5	3.9	e
7	Celltac chemi	22	100.0	0.0	0.0	5.0	4.3	e
8	NycoCard	5	80.0	20.0	0.0	5.5	5.7	c
9	Eurolyser	5	100.0	0.0	0.0	5.6	5.4	e*
10	A1c Now	8	100.0	0.0	0.0	5.5	4.8	c
11	AFIAS	184	81.6	9.2	9.2	5.0	5.5	e
12	Others	13	84.6	0.0	15.4	5.4	2.3	e
13	Quick Read go	5	100.0	0.0	0.0	5.5	3.8	c

3 additional results were submitted but not published because the method groups were too small. (< results per group)

pCO₂



QUALAB tolerance : 12 %


pCO₂ (kPa)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	109	99.1	0.0	0.9	5.36	2.7	e
2	ABL80 FLEX CO-OX / O	4	100.0	0.0	0.0	6.02	5.5	e*
3	ABL90 FLEX / PLUS	117	99.1	0.9	0.0	5.53	2.1	e
4	Cobas b 123	16	100.0	0.0	0.0	5.59	1.9	e
5	Cobas b 221	5	100.0	0.0	0.0	5.93	3.2	e
6	GEM	8	100.0	0.0	0.0	5.70	2.3	e
7	iStat	51	96.1	3.9	0.0	4.92	4.2	e
8	EPOC	53	96.2	0.0	3.8	5.21	2.8	e
9	IL	4	100.0	0.0	0.0	5.60	3.1	e*

8 additional results were submitted but not published because the method groups were too small. (< results per group)

K04 Blood gases

pO2




No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	108	97.2	0.0	2.8	14.63	3.9	e
2	ABL80 FLEX CO-OX / O	4	100.0	0.0	0.0	13.43	7.6	e*
3	ABL90 FLEX / PLUS	118	95.8	3.4	0.8	13.48	5.5	e
4	Cobas b 123	17	94.1	5.9	0.0	14.08	6.1	e
5	GEM	8	87.5	0.0	12.5	14.20	1.7	e
6	iStat	48	95.8	2.1	2.1	15.57	6.0	e
7	EPOC	53	92.4	3.8	3.8	14.48	7.2	e
8	IL	4	100.0	0.0	0.0	14.20	1.8	e

10 additional results were submitted but not published because the method groups were too small. (< results per group)

K04 Blood gases

pH




QUALAB tolerance : 1 %

pH ()

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	108	99.1	0.0	0.9	7.39	0.1	e
2	ABL80 FLEX CO-OX / O	4	100.0	0.0	0.0	7.38	0.1	e
3	ABL90 FLEX / PLUS	118	100.0	0.0	0.0	7.40	0.1	e
4	Cobas b 123	17	100.0	0.0	0.0	7.41	0.1	e
5	Cobas b 221	5	100.0	0.0	0.0	7.40	0.2	e
6	GEM	8	100.0	0.0	0.0	7.40	0.1	e
7	iStat	53	100.0	0.0	0.0	7.43	0.1	e
8	EPOC	53	98.1	1.9	0.0	7.38	0.2	e
9	IL	4	100.0	0.0	0.0	7.42	0.3	e*

8 additional results were submitted but not published because the method groups were too small. (< results per group)

Glucose BG




QUALAB tolerance : 9 %

Glucose BG (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	GEM	5	100.0	0.0	0.0	5.8	1.0	e
2	Cobas b 123	10	100.0	0.0	0.0	5.5	2.6	e
3	iStat	10	100.0	0.0	0.0	5.5	1.8	e
4	EPOC	42	97.6	0.0	2.4	5.5	3.0	e
5	ABL700/800	99	100.0	0.0	0.0	5.7	2.1	e
6	ABL90 FLEX / PLUS	104	100.0	0.0	0.0	5.6	1.6	e

8 additional results were submitted but not published because the method groups were too small. (< results per group)


Hemoglobin BG



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	101	96.0	0.0	4.0	128.9	1.7	e
2	ABL90 FLEX / PLUS	106	100.0	0.0	0.0	129.3	0.5	e
3	ABL80 FLEX CO-OX / O	5	100.0	0.0	0.0	125.0	2.5	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Potassium BG




QUALAB tolerance : 6 %

Potassium BG (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	GEM	6	100.0	0.0	0.0	3.8	1.3	e
2	Cobas b 123	18	100.0	0.0	0.0	3.8	0.7	e
3	iStat	18	94.4	0.0	5.6	3.8	1.5	e
4	EPOC	43	100.0	0.0	0.0	3.8	1.4	e
5	ABL700/800	100	99.0	0.0	1.0	3.8	1.6	e
6	ABL90 FLEX / PLUS	110	100.0	0.0	0.0	3.7	1.0	e

9 additional results were submitted but not published because the method groups were too small. (< results per group)

Sodium BG




QUALAB tolerance : 6 %

Sodium BG (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	GEM	7	85.7	0.0	14.3	137.0	0.6	e
2	Cobas b 123	18	100.0	0.0	0.0	138.0	0.6	e
3	iStat	18	94.4	0.0	5.6	142.4	0.6	e
4	EPOC	42	100.0	0.0	0.0	140.6	0.6	e
5	ABL700/800	99	100.0	0.0	0.0	140.7	1.0	e
6	ABL90 FLEX / PLUS	109	100.0	0.0	0.0	141.6	0.4	e

8 additional results were submitted but not published because the method groups were too small. (< results per group)

Chlorid-BG




QUALAB tolerance : 6 %

Chlorid-BG (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas b 123	11	100.0	0.0	0.0	100.4	0.9	e
2	EPOC	19	100.0	0.0	0.0	100.8	1.5	e
3	ABL700/800	95	100.0	0.0	0.0	98.8	1.7	e
4	ABL90 FLEX / PLUS	106	100.0	0.0	0.0	97.3	0.5	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Calcium-BG




MQ tolerance : 12 %
 (< 2.00: +/- 0.24 mmol/l)

Calcium-BG (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	GEM	6	100.0	0.0	0.0	0.44	2.0	e
2	Cobas b123	10	100.0	0.0	0.0	0.37	10.6	e*
3	Roche, Cobas	4	100.0	0.0	0.0	0.41	20.3	e*
4	iStat	15	100.0	0.0	0.0	0.48	1.3	e
5	EPOC	39	97.4	0.0	2.6	0.53	3.6	e
6	ABL700/800	99	99.0	0.0	1.0	0.53	3.7	e
7	ABL90 FLEX / PLUS	107	100.0	0.0	0.0	0.52	1.0	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

FHHb




MQ tolerance : 20 %

FHHb (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL90 FLEX / PLUS	18	100.0	0.0	0.0	2.911	2.6	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Lactate-BG




QUALAB tolerance : 18 %

Lactate-BG (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	GEM	6	100.0	0.0	0.0	1.80	2.2	e
2	Cobas b123	9	100.0	0.0	0.0	1.90	2.6	e
3	IL	4	100.0	0.0	0.0	1.70	5.7	e*
4	EPOC	41	97.6	0.0	2.4	1.74	2.6	e
5	iStat	19	100.0	0.0	0.0	1.60	1.7	e
6	ABL700/800	103	100.0	0.0	0.0	1.70	4.4	e
7	ABL90 FLEX / PLUS	112	98.2	0.9	0.9	1.61	3.2	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

sO2 OR




MQ tolerance : 20 %

sO2 OR (%)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 iStat	25	96.0	0.0	4.0	98.875	0.3	e
2 ABL700/800	90	100.0	0.0	0.0	96.728	2.0	e
3 ABL90 FLEX / PLUS	96	100.0	0.0	0.0	96.912	0.2	e
4 ABL80 FLEX CO-OX / O	4	100.0	0.0	0.0	96.950	0.2	e

FO2Hb OR




MQ tolerance : 20 %

FO2Hb OR (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	88	98.9	0.0	1.1	92.233	0.1	e
2	ABL90 FLEX / PLUS	97	100.0	0.0	0.0	92.256	1.2	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

FCOHb OR




MQ tolerance : 20 %

FCOHb OR (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	89	94.4	4.5	1.1	2.873	9.9	e
2	ABL90 FLEX / PLUS	97	98.0	1.0	1.0	2.894	5.8	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

FMetHb OR




MQ tolerance : 20 %

FMetHb OR (%)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 ABL700/800	89	100.0	0.0	0.0	1.969	4.2	e
2 ABL90 FLEX / PLUS	97	100.0	0.0	0.0	2.020	3.5	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

FHbF OR




MQ tolerance : 20 %

FHbF OR (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL90 FLEX / PLUS	36	100.0	0.0	0.0	79.139	2.2	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Bilirubin OR




QUALAB tolerance : 18 %

Bilirubin OR (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	5	100.0	0.0	0.0	295.0	2.0	e
2	ABL90 FLEX / PLUS	34	100.0	0.0	0.0	298.2	0.5	e

Troponin I




QUALAB tolerance : 24 %

Troponin I (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Pathfast	26	92.4	3.8	3.8	7945.0	11.2	e
2	Vidas	13	92.3	7.7	0.0	14760.9	13.3	e*
3	Abbott	12	91.7	8.3	0.0	1760.2	9.6	e

8 additional results were submitted but not published because the method groups were too small. (< results per group)

Troponin T




QUALAB tolerance : 24 %

Troponin T (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas hs	11	100.0	0.0	0.0	791.80	5.8	e
2	Cobas hs STAT	15	100.0	0.0	0.0	793.80	5.1	e

Myoglobin




QUALAB tolerance : 30 %

Myoglobin (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas E / Elecsys	6	100.0	0.0	0.0	115.0	5.5	e
2	Abbott	5	100.0	0.0	0.0	169.6	1.6	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

CK-MB mass




MQ tolerance : 40 %

CK-MB mass (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	21.8	7.0	e
2	Cobas E / Elecsys	10	100.0	0.0	0.0	24.0	2.0	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

BNP




QUALAB tolerance : 27 %

BNP (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	6	83.3	16.7	0.0	1027.3	14.2	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

NT-proBNP




QUALAB tolerance : 27 %

NT-proBNP (ng/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Pathfast	19	94.7	0.0	5.3	3483.8	6.4	e
2	AQT 90 FLEX	4	75.0	25.0	0.0	2585.0	19.6	e*
3	VIDAS	9	100.0	0.0	0.0	1111.0	4.9	e
4	Cobas E / Elecsys	23	100.0	0.0	0.0	1147.2	4.4	e
5	Abbott	9	100.0	0.0	0.0	1300.0	4.5	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

TSH




QUALAB tolerance : 18 %

TSH (mU/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	30	100.0	0.0	0.0	6.24	4.3	e
2	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	5.26	4.4	e*
3	Abbott	13	100.0	0.0	0.0	4.76	6.2	e
4	VIDAS	15	100.0	0.0	0.0	6.71	4.9	e
5	Dimension	5	100.0	0.0	0.0	5.86	6.0	e*
6	AFIAS	17	64.7	11.8	23.5	9.10	10.7	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

T3




MQ tolerance : 20 %

T3 (nmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	2.8	4.5	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

T4




MQ tolerance : 20 %

T4 (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	5	100.0	0.0	0.0	145	3.7	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

FT3




QUALAB tolerance : 18 %

FT3 (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	26	100.0	0.0	0.0	10.6	2.9	e
2	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	10.1	10.6	e*
3	Abbott	11	100.0	0.0	0.0	8.9	4.9	e
4	VIDAS	7	100.0	0.0	0.0	10.0	5.5	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

FT4




QUALAB tolerance : 20 %

FT4 (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	26	100.0	0.0	0.0	36.7	3.2	e
2	Abbott	11	100.0	0.0	0.0	27.1	6.7	e
3	VIDAS	8	100.0	0.0	0.0	34.6	4.1	e
4	Other methods	8	100.0	0.0	0.0	30.4	7.3	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Testosterone




QUALAB tolerance : 30 %

Testosterone (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	14	100.0	0.0	0.0	14.0	3.7	e
2	Siemens	7	85.7	14.3	0.0	13.6	14.5	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Estradiol




QUALAB tolerance : 30 %

Estradiol (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	12	100.0	0.0	0.0	1178	3.4	e
2	Siemens	5	100.0	0.0	0.0	1643	6.2	e
3	all Participants	5	100.0	0.0	0.0	951	11.3	e*
4	Abbott	5	100.0	0.0	0.0	1096	2.5	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

SHBG




MQ tolerance : 30 %

SHBG (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	16	100.0	0.0	0.0	40.9	4.3	e
2	Abbott	5	100.0	0.0	0.0	45.0	2.4	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Cortisol




QUALAB tolerance : 20 %

Cortisol (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	21	100.0	0.0	0.0	616	4.4	e
2	Abbott	5	100.0	0.0	0.0	566	6.9	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Progesteron




MQ tolerance : 30 %

Progesteron (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	9	100.0	0.0	0.0	39.6	4.4	e
2	Abbott	4	100.0	0.0	0.0	29.5	3.4	e
3	Other methods	6	100.0	0.0	0.0	38.3	9.3	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

DHEAS




MQ tolerance : 30 %

DHEAS (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	12	100.0	0.0	0.0	5.08	5.1	e
2	Abbott	4	100.0	0.0	0.0	4.33	9.5	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Luteinizing hormone




QUALAB tolerance : 24 %

Luteinizing hormone (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	14	100.0	0.0	0.0	23.7	6.5	e
2	Siemens	5	100.0	0.0	0.0	20.2	5.8	e
3	Abbott	5	100.0	0.0	0.0	15.7	4.6	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Follicle-stimulating hormone




QUALAB tolerance : 24 %

Follicle-stimulating hormone (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	13	100.0	0.0	0.0	17.4	4.1	e
2	Abbott	8	100.0	0.0	0.0	17.1	2.5	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Prolactine




QUALAB tolerance : 24 %

Prolactine (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	12.3	5.7	e
2	Cobas/Roche	14	100.0	0.0	0.0	21.0	4.6	e
3	Abbott	6	100.0	0.0	0.0	17.9	9.4	e*

Insulin




MQ tolerance : 25 %

Insulin (pmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	16	100.0	0.0	0.0	496	9.6	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

HGH




MQ tolerance : 25 %

HGH (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	7	100.0	0.0	0.0	11.63	1.7	e
2 Liaison	4	100.0	0.0	0.0	11.25	6.0	e

Freies Testosteron




MQ tolerance : 25 %

Freies Testosteron (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	RIA	4	100.0	0.0	0.0	27.0	11.0	e*

5 additional results were submitted but not published because the method groups were too small. (< results per group)

IGF-1




MQ tolerance : 25 %

IGF-1 (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Liaison	9	88.9	11.1	0.0	105	15.3	e*
2	Other methods	4	100.0	0.0	0.0	73	3.2	e

Troponin T CR




QUALAB tolerance : 24 %

Troponin T CR (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas h 232	822	88.8	7.4	3.8	242.00	13.0	e

D-dimer CR




QUALAB tolerance : 21 %

D-dimer CR (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas h 232	929	82.8	12.8	4.4	0.61	13.9	e
2	Lumira Dx	6	83.3	0.0	16.7	0.57	3.4	e

CKMB- K8




MQ tolerance : 40 %

CKMB- K8 (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 Cobas h 232	6	100.0	0.0	0.0	25.7	18.1 e*

NT-proBNP CR




QUALAB tolerance : 27 %

NT-proBNP CR (ng/l)


No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Lumira Dx	4	100.0	0.0	0.0	960	7.0	e*
2	Cobas h 232	567	81.4	12.3	6.3	255	17.5	e

PCO2 CCA



No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 OPTI CCA	11	100.0	0.0	0.0	5.63	5.3	e*

PO2 CCA




QUALAB tolerance : 15 %

PO2 CCA (kPa)


No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 OPTI CCA	11	90.9	0.0	9.1	12.41	5.8	e

pH CCA



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	OPTI CCA	10	100.0	0.0	0.0	7.40	0.2	e

Potassium CCA




QUALAB tolerance : 6 %

Potassium CCA (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	OPTI CCA	5	100.0	0.0	0.0	4.8	1.6	e

Sodium CCA




QUALAB tolerance : 6 %

Sodium CCA (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	OPTI CCA	4	100.0	0.0	0.0	147.7	1.6	e*

Ferritin




QUALAB tolerance : 24 %

Ferritin (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dimension	6	100.0	0.0	0.0	120.00	9.6	e*
2	Beckman	7	100.0	0.0	0.0	88.20	8.5	e*
3	all Participants	5	60.0	20.0	20.0	120.30	17.3	e*
4	Roche, Cobas	30	96.7	3.3	0.0	110.92	8.8	e
5	Abbott	12	100.0	0.0	0.0	153.20	4.7	e
6	Mini Vidas	7	85.7	14.3	0.0	88.00	10.6	e*
7	AFIAS	26	96.2	3.8	0.0	92.50	9.1	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Vitamin B12




QUALAB tolerance : 21 %

Vitamin B12 (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	6	100.0	0.0	0.0	286.00	4.6	e
2	Roche, Cobas	20	100.0	0.0	0.0	307.48	3.7	e
3	Abbott	11	100.0	0.0	0.0	275.76	7.0	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Folate




QUALAB tolerance : 24 %
 (< 10.00: +/- 2.40 nmol/l)

Folate (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens	5	100.0	0.0	0.0	7.00	10.3	e*
2	Roche, Cobas	23	100.0	0.0	0.0	7.78	13.2	e
3	Abbott	11	90.9	9.1	0.0	13.45	15.0	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Holotranscobalamine




MQ tolerance : 30 %

Holotranscobalamine (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	30	100.0	0.0	0.0	98.1	3.5	e
2	Abbott	17	100.0	0.0	0.0	120.2	5.5	e
3	Cobas Biotin suppress	7	100.0	0.0	0.0	67.6	7.4	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Bilirubin total Neo




QUALAB tolerance : 18 %

Bilirubin total Neo (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	16	93.7	0.0	6.3	88	2.8	e
2	Dimension	6	100.0	0.0	0.0	106	5.4	e*

Bilirubin direct



QUALAB tolerance : 18 %

Bilirubin direct (µmol/l)


No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	18	100.0	0.0	0.0	63	6.4	a
2	Dimension	4	100.0	0.0	0.0	52	1.8	e

Bilirubin neonatal



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	10	100.0	0.0	0.0	108	6.6	e
2	Other methods	12	100.0	0.0	0.0	123	3.5	e

PSA




QUALAB tolerance : 21 %

PSA (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	17	100.0	0.0	0.0	22.86	3.2	e
2	Abbott	8	100.0	0.0	0.0	21.06	4.7	e
3	AFIAS	11	90.9	9.1	0.0	22.73	8.5	e*

9 additional results were submitted but not published because the method groups were too small. (< results per group)

free PSA




QUALAB tolerance : 21 %

free PSA (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	11	100.0	0.0	0.0	4.56	4.6	e
2	Abbott	7	100.0	0.0	0.0	5.02	5.3	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

CEA




QUALAB tolerance : 21 %

CEA (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	12	100.0	0.0	0.0	25.9	3.9	e
2	Abbott	7	100.0	0.0	0.0	39.5	4.2	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

CA 125




MQ tolerance : 25 %

CA 125 (kIU/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	8	100.0	0.0	0.0	196.0	4.9	e
2	Abbott	7	100.0	0.0	0.0	300.3	3.4	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

CA 19-9




MQ tolerance : 25 %

CA 19-9 (kIU/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	7	100.0	0.0	0.0	126.0	3.8	e
2	Abbott	4	100.0	0.0	0.0	1028.5	4.9	e

8 additional results were submitted but not published because the method groups were too small. (< results per group)

CA 15-3



MQ tolerance : 25 %


CA 15-3 (kIU/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	10	100.0	0.0	0.0	64.3	5.2	e
2	Siemens	4	100.0	0.0	0.0	71.4	7.0	e*
3	Abbott	7	100.0	0.0	0.0	84.0	7.2	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

K14 Tumor Markers

AFP




QUALAB tolerance : 21 %

AFP (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	7	100.0	0.0	0.0	28.2	10.7	d
2	Abbott	5	100.0	0.0	0.0	27.0	3.7	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)


HCG qn



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	15	100.0	0.0	0.0	65.7	4.2	e
2	VIDAS	12	91.7	8.3	0.0	28.0	11.3	e*
3	Abbott	7	100.0	0.0	0.0	59.7	5.8	e
4	AFIAS	9	88.9	11.1	0.0	115.0	14.4	e*

9 additional results were submitted but not published because the method groups were too small. (< results per group)

HCG intact




QUALAB tolerance : 25 %

HCG intact (U/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	5	100.0	0.0	0.0	21.5	3.1	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

S100




MQ tolerance : 25 %

S100 (µg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	6	100.0	0.0	0.0	3.24	3.0	e

NSE




MQ tolerance : 25 %

NSE (ng/ml)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	5	100.0	0.0	0.0	26.4	6.6	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

Thyreoglobulin




MQ tolerance : 25 %

Thyreoglobulin (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	4	100.0	0.0	0.0	159.5	3.5	e
2	Other methods	4	100.0	0.0	0.0	195.4	7.9	e*

K15 Creatinkinase Activity

CK-MB




MQ tolerance : 30 %

CK-MB (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Fuji Dri-Chem	20	100.0	0.0	0.0	59.5	6.3	e
2	Cobas/Roche	7	100.0	0.0	0.0	17.8	4.2	e


2 additional results were submitted but not published because the method groups were too small. (< results per group)

C-Peptid



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	7	100.0	0.0	0.0	2.67	4.0	e
2	Other methods	5	100.0	0.0	0.0	2.26	10.6	e*

ACTH




MQ tolerance : 25 %

ACTH (ng/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	8	100.0	0.0	0.0	15.70	9.9	e*
2	Liaison	4	100.0	0.0	0.0	31.50	14.5	e*

Procalcitonin




QUALAB tolerance : 27 %

Procalcitonin (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	6	100.0	0.0	0.0	6.43	5.2	e
2	Cobas	17	100.0	0.0	0.0	4.95	3.6	e
3	VIDAS	15	100.0	0.0	0.0	4.38	4.4	e
4	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	5.76	16.0	e*
5	Liaison	4	100.0	0.0	0.0	8.91	8.3	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

EPO




MQ tolerance : 25 %

EPO (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	4	100.0	0.0	0.0	124.3	4.4	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)


Parathyroid hormone



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas PTH STAT	11	100.0	0.0	0.0	15.8	6.5	e
2	Cobas	11	100.0	0.0	0.0	12.7	3.2	e
3	IDS	5	80.0	0.0	20.0	23.3	6.8	e*
4	Abbott	7	100.0	0.0	0.0	28.8	12.2	a

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Osmolality




QUALAB tolerance : 6 %

Osmolality (mosm/kg)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Cryoskopy	22	100.0	0.0	0.0	350	1.3	e

Potassium-K22




QUALAB tolerance : 6 %

Potassium-K22 (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ISE	11	100.0	0.0	0.0	3.7	1.5	e

Sodium-K22




QUALAB tolerance : 6 %

Sodium-K22 (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ISE	11	100.0	0.0	0.0	145	1.0	e

Glucose-K22




QUALAB tolerance : 9 %

Glucose-K22 (mmol/l)


No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	11	100.0	0.0	0.0	4.4	1.7	e

Urea-K22



No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Standard chemistry	11	100.0	0.0	0.0	4.3	6.1	e*

Osmotic Gap




MQ tolerance : 40 %

Osmotic Gap (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Formel 1 (2Na+K+Glu+	9	100.0	0.0	0.0	48.0	12.9	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Digoxin




QUALAB tolerance : 24 %

Digoxin (nmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Alinity	4	100.0	0.0	0.0	3.54	5.0	e
2	Other methods	9	100.0	0.0	0.0	3.70	5.7	e

Phenytoin




MQ tolerance : 25 %

Phenytoin (µmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	4	100.0	0.0	0.0	84	7.9	e*

Phenobarbital




MQ tolerance : 25 %

Phenobarbital (µmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	4	100.0	0.0	0.0	212	5.5	e

Paracetamol




MQ tolerance : 25 %

Paracetamol (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	4	100.0	0.0	0.0	578.5	5.9	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Vancomycin




MQ tolerance : 25 %

Vancomycin (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	5	100.0	0.0	0.0	20.2	6.7	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Valproat




MQ tolerance : 24 %

Valproat (µmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	9	100.0	0.0	0.0	823.0	4.7	e

Carbamazepin




MQ tolerance : 24 %

Carbamazepin (µmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	5	100.0	0.0	0.0	51.0	6.2	e

Cystatin C




MQ tolerance : 24 %

Cystatin C (mg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	19	94.7	0.0	5.3	1.27	6.1	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Ethanol




QUALAB tolerance : 18 %

Ethanol (mmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Others	5	100.0	0.0	0.0	17.4	5.6	e*
2 Abbott	9	100.0	0.0	0.0	17.2	5.2	e
3 Roche, Cobas	21	100.0	0.0	0.0	16.3	4.4	e

Ammonia




QUALAB tolerance : 21 %

Ammonia (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	85.3	2.9	e
2	all Participants	6	100.0	0.0	0.0	107.0	4.3	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)


Calprotectin



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Bühlmann fCALturbo	21	95.2	0.0	4.8	271	12.6	e
2	Bühlmann Quantum Blu	5	60.0	0.0	40.0	265	19.2	a
3	Liaison	15	100.0	0.0	0.0	120	8.9	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Cholesterol total Af/b101




QUALAB tolerance : 10 %

Cholesterol total Af/b101 (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas b101	320	99.7	0.0	0.3	5.82	2.2	e
2	Afinion	450	98.2	0.9	0.9	5.72	3.5	e

Cholesterol HDL Af/b101




QUALAB tolerance : 21 %

Cholesterol HDL Af/b101 (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas b101	316	94.6	0.0	5.4	1.24	4.0	e
2	Afinion	449	92.9	0.4	6.7	1.15	5.4	e

Tryglicerides Af/b101




QUALAB tolerance : 18 %

Tryglicerides Af/b101 (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas b101	317	99.4	0.0	0.6	2.16	2.5	e
2	Afinion	451	99.4	0.4	0.2	2.16	4.6	e

Cholesterol PTS




QUALAB tolerance : 10 %

Cholesterol PTS (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CardioChek	4	50.0	50.0	0.0	5.82	12.1	a

Cholesterol HDL PTS




QUALAB tolerance : 21 %

Cholesterol HDL PTS (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CardioChek	4	50.0	0.0	50.0	2.33	24.3	e*

Triglycerides




QUALAB tolerance : 18 %

Triglycerides (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CardioChek	4	100.0	0.0	0.0	2.08	7.2	e*

Troponin I AFIAS




QUALAB tolerance : 24 %

Troponin I AFIAS (ng/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 AFIAS	455	84.0	7.0	9.0	788.13	13.3 e

NT-proBNP AFIAS




QUALAB tolerance : 27 %

NT-proBNP AFIAS (ng/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 AFIAS	345	92.4	3.8	3.8	462.4	11.1 e

Homocystein




MQ tolerance : 30 %

Homocystein (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Beckman	7	100.0	0.0	0.0	10.8	1.8	e
2	all Participants	6	100.0	0.0	0.0	8.3	11.6	e*

Lipase




QUALAB tolerance : 18 %

Lipase (U/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	14	100.0	0.0	0.0	31.7	6.5	e
2	Beckman	5	100.0	0.0	0.0	32.0	5.0	e*
3	Roche	27	100.0	0.0	0.0	34.8	5.1	e
4	Fuji Dri-Chem	173	97.1	0.6	2.3	40.9	4.4	e

13 additional results were submitted but not published because the method groups were too small. (< results per group)


Fructosamine



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	4	100.0	0.0	0.0	292	4.3	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Bicarbonat




MQ tolerance : 15 %

Bicarbonat (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	6	100.0	0.0	0.0	16.4	3.2	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

Glucose CSF




QUALAB tolerance : 9 %
(< 3.30: +/- 0.30 mmol/l)

Glucose CSF (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	15	100.0	0.0	0.0	2.98	1.2	e
2	Other methods	14	100.0	0.0	0.0	2.91	3.0	e

Lactate CSF




QUALAB tolerance : 18 %

Lactate CSF (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	13	92.3	0.0	7.7	2.53	4.3	e
2	Other methods	10	90.0	0.0	10.0	2.65	5.0	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Protein CSF




QUALAB tolerance : 15 %

Protein CSF (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	19	100.0	0.0	0.0	0.55	4.2	e
2	Other methods	7	100.0	0.0	0.0	0.55	5.4	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Albumine CSF




MQ tolerance : 20 %

Albumine CSF (mg/l)


No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	5	100.0	0.0	0.0	265.00	3.6	e
2	Other methods	4	100.0	0.0	0.0	285.85	1.7	e

CDT



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Immunoassay	5	100.0	0.0	0.0	1.10	4.8	e
2	all Participants	5	80.0	0.0	20.0	0.60	9.5	e*

Tacrolimus




MQ tolerance : 25 %

Tacrolimus (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	10	100.0	0.0	0.0	8.9	9.4	e

Totalprotein E




MQ tolerance : 12 %

Totalprotein E (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	20	100.0	0.0	0.0	71.8	2.9	e

Albumin E




MQ tolerance : 12 %

Albumin E (%)


No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	35	97.1	0.0	2.9	62.2	5.2	e

alpha-1-Globuline



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	12	100.0	0.0	0.0	2.4	6.9	e
2	capillary electropho	23	100.0	0.0	0.0	3.9	9.0	e

alpha-2-Globuline




MQ tolerance : 30 %

alpha-2-Globuline (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	35	97.1	2.9	0.0	10.0	9.4	e

beta-Globuline




MQ tolerance : 30 %

beta-Globuline (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	17	100.0	0.0	0.0	8.2	20.7	e*

Beta-1-Globulin




MQ tolerance : 30 %

Beta-1-Globulin (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	13	100.0	0.0	0.0	6.9	7.7	e

Beta-2-Globulin




MQ tolerance : 30 %

Beta-2-Globulin (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	12	100.0	0.0	0.0	3.5	12.4	e

gamma-Globuline




MQ tolerance : 30 %

gamma-Globuline (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	15	100.0	0.0	0.0	13.2	17.1	e*

Gamma-Globuline+P




MQ tolerance : 30 %

Gamma-Globuline+P (%)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Electrophoresis	21	100.0	0.0	0.0	15.7	6.0	e

Paraprotein




MQ tolerance : 30 %

Paraprotein (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	5	100.0	0.0	0.0	2.6	88.2	a

Beta-Globuline+P




MQ tolerance : 30 %

Beta-Globuline+P (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	5	100.0	0.0	0.0	10.0	5.0	e

Immunifixation




QUALAB tolerance : 1 %

Immunifixation (Code)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Interpretation	30	96.7	0.0	3.3	5	0.0	e

Folate in Erythrocytes




MQ tolerance : 40 %

Folate in Erythrocytes (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Alinity	6	100.0	0.0	0.0	394	18.0	e*
2	Siemens	4	100.0	0.0	0.0	563	12.5	e*
3	Abbott	5	100.0	0.0	0.0	553	7.0	e
4	Roche, Cobas	22	100.0	0.0	0.0	1553	16.9	e

Gallensäure




MQ tolerance : 30 %

Gallensäure (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	9	100.0	0.0	0.0	68.1	6.8	e
2	all Participants	17	100.0	0.0	0.0	54.5	6.2	e

BNP




QUALAB tolerance : 27 %
 (< 75.0: +/- 20.0 ng/l)

BNP (ng/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Triage	7	100.0	0.0	0.0	12.4	42.3	e*

Troponin I Triage




QUALAB tolerance : 24 %

Troponin I Triage (ng/l)


No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Triage high sensitiv	265	88.3	7.9	3.8	731.00	13.0	e
2	Triage Next Gen	10	60.0	10.0	30.0	70.00	16.5	e*

NT-pro BNP



No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 Triage	144	91.6	4.2	4.2	287	13.3 e

Vitamin D 25 (OH)




QUALAB tolerance : 27 %

Vitamin D 25 (OH) (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	LCMS	4	100.0	0.0	0.0	60.8	6.4	e
2	AFIAS	5	100.0	0.0	0.0	60.4	8.8	e*
3	Cobas	18	100.0	0.0	0.0	59.8	7.7	e
4	VIDAS	7	100.0	0.0	0.0	75.0	8.7	e*
5	Other methods	6	66.7	0.0	33.3	63.0	23.3	a
6	Abbott	10	100.0	0.0	0.0	65.7	3.4	e
7	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	66.4	2.4	e

Vitamin D 1,25-(OH)2




MQ tolerance : 27 %

Vitamin D 1,25-(OH)2 (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	4	100.0	0.0	0.0	48.3	3.4	e

AMH




MQ tolerance : 25 %

AMH (pmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	19	100.0	0.0	0.0	19.8	4.0	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Inhibin B




MQ tolerance : 25 %

Inhibin B (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	7	100.0	0.0	0.0	98.9	11.6	e*

Calcitonin




MQ tolerance : 25 %

Calcitonin (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Liaison	5	100.0	0.0	0.0	4.7	3.9	e
2	Other methods	7	100.0	0.0	0.0	15.2	4.9	e

IGF-BP3




MQ tolerance : 25 %

IGF-BP3 (mg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	6	100.0	0.0	0.0	3.19	4.4	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Renin




MQ tolerance : 25 %

Renin (mU/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Liaison	8	87.5	12.5	0.0	66.6	16.7	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Aldosteron




MQ tolerance : 30 %

Aldosteron (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Liaison	5	100.0	0.0	0.0	50.4	15.0	a

3 additional results were submitted but not published because the method groups were too small. (< results per group)


Anti Thyreoglobulin



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Alinity	6	83.3	16.7	0.0	16	15.8	e*
2	Roche, Cobas	11	100.0	0.0	0.0	63	5.8	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti TPO




QUALAB tolerance : 25 %

Anti TPO (IU/ml)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	12	100.0	0.0	0.0	120	8.4	e
2	Abbott	7	100.0	0.0	0.0	168	9.7	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

TRAK




MQ tolerance : 25 %

TRAK (IU/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Abbott	6	100.0	0.0	0.0	6.64	5.0	e
2 Roche, Cobas	8	100.0	0.0	0.0	4.66	6.8	e
3 Kryptor	4	100.0	0.0	0.0	4.05	10.2	e*

Creatinine WB




QUALAB tolerance : 18 %

Creatinine WB (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Statsensor i / Nova	81	75.4	16.0	8.6	160	13.1	e

eGFR CDK-EPI WB




MQ tolerance : 30 %

eGFR CDK-EPI WB ()

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Statsensor i / Nova	4	100.0	0.0	0.0	34	8.5	e*

IL6




MQ tolerance : 30 %

IL6 (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	5	100.0	0.0	0.0	8.6	5.4	e

Pankreas Elastase




MQ tolerance : 40 %

Pankreas Elastase (ug/g)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Liaison	15	93.3	0.0	6.7	247	19.9	a

Copeptin




MQ tolerance : 30 %

Copeptin (pmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 Kryptor	7	100.0	0.0	0.0	2.5	15.7 a

Occult blood qn




MQ tolerance : 30 %

Occult blood qn (ng/ml)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	OC-Sensor	4	100.0	0.0	0.0	1	0.0	e

Amylase-Urine




MQ tolerance : 25 %

Amylase-Urine (U/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 IFCC	7	100.0	0.0	0.0	61	5.7	e

Pancreatic Amylase-Urine




QUALAB tolerance : 18 %
(< 25.0: +/- 5.0 U/l)

Pancreatic Amylase-Urine (U/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 IFCC	4	100.0	0.0	0.0	2.1	20.4	e*

Calcium-Urine




MQ tolerance : 9 %
 (< 2.00: +/- 0.18 mmol/l)

Calcium-Urine (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	17	100.0	0.0	0.0	1.72	3.2	e
2	Abbott	12	100.0	0.0	0.0	1.68	1.6	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Chloride-Urine




MQ tolerance : 12 %

Chloride-Urine (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	12	100.0	0.0	0.0	97	0.9	e
2	Roche, Cobas	14	100.0	0.0	0.0	89	2.2	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Glucose-Urine




QUALAB tolerance : 9 %
(< 3.3: +/- 0.3 mmol/l)

Glucose-Urine (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	26	100.0	0.0	0.0	1.5	2.8	e

Magnesium-Urine




MQ tolerance : 12 %

Magnesium-Urine (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	16	100.0	0.0	0.0	1.66	2.4	e
2	AAS	5	100.0	0.0	0.0	1.74	4.6	e*

Osmolality-Urine




MQ tolerance : 10 %

Osmolality-Urine (mosm/kg)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Cryoskopy	21	100.0	0.0	0.0	438	0.9	e

Phosphate-Urine




MQ tolerance : 15 %

Phosphate-Urine (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	28	96.4	3.6	0.0	8.1	4.3	e

Potassium-Urine




MQ tolerance : 15 %

Potassium-Urine (mmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	38	100.0	0.0	0.0	31	2.0	e

Protein-Urine




QUALAB tolerance : 15 %

Protein-Urine (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	22	100.0	0.0	0.0	181.1	5.5	e
2	Abbott	18	100.0	0.0	0.0	212.9	3.0	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Sodium-Urine




MQ tolerance : 15 %

Sodium-Urine (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	38	100.0	0.0	0.0	80	2.4	e

Urea-Urine




MQ tolerance : 15 %

Urea-Urine (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	36	97.2	2.8	0.0	150	6.4	e

Uric Acid-Urine




MQ tolerance : 15 %

Uric Acid-Urine (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	25	100.0	0.0	0.0	0.63	4.3	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Specific Gravity-Urine




MQ tolerance : 5 %

Specific Gravity-Urine ()

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Refractometer	5	100.0	0.0	0.0	1.013	0.1	e

Ethylglucuronid




MQ tolerance : 25 %

Ethylglucuronid (mg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	6	83.3	0.0	16.7	0.42	11.1	e*

Creatinine U




QUALAB tolerance : 24 %

Creatinine U (mg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	14	100.0	0.0	0.0	121.1	3.9	e
2	Roche, Cobas	20	100.0	0.0	0.0	112.2	2.4	e
3	Aution	9	77.8	0.0	22.2	80.0	0.0	e
4	AFIAS	17	94.1	0.0	5.9	119.2	7.1	e
5	Afinion	502	95.8	1.6	2.6	113.1	9.1	e
6	Sysmex U	16	81.2	0.0	18.8	150.0	0.0	a
7	Other methods	4	100.0	0.0	0.0	80.0	0.0	e
8	DCA2000/Vantage	159	93.0	1.3	5.7	116.7	6.4	e
9	Siemens Clinitek	26	80.8	0.0	19.2	80.0	0.0	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Creatinin Urin




QUALAB tolerance : 21 %

Creatinin Urin (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	14	100.0	0.0	0.0	3.2	2.6	e
2	Roche	26	100.0	0.0	0.0	3.5	4.9	e
3	Beckman	4	100.0	0.0	0.0	3.7	9.5	e*
4	Siemens	5	100.0	0.0	0.0	3.2	8.6	e*
5	DCA2000/Vantage	157	93.0	2.5	4.5	3.8	6.9	e
6	Afinion	497	98.6	0.4	1.0	3.2	5.1	e
7	Standard chemistry	4	100.0	0.0	0.0	3.5	8.8	e*
8	Sysmex U	13	69.2	0.0	30.8	4.4	0.0	a
9	Aution	9	77.8	0.0	22.2	4.4	0.0	e
10	Siemens Clinitek	24	83.3	0.0	16.7	4.4	0.2	e
11	Other methods	5	80.0	0.0	20.0	4.4	0.0	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

HBV NAT qn




HBV NAT qn (Log10 IU/m)

QUALAB tolerance : +/- 0.50 Log10 IU/m

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	7	100.0	0.0	0.0	3.38	4.5	e*

HCV NAT qn




HCV NAT qn (Log10 IU/m)

QUALAB tolerance : +/- 0.50 Log10 IU/m

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	8	100.0	0.0	0.0	0.00	0.0	e

HIV1 NAT qn



HIV1 NAT qn (Log10 cp/m)

QUALAB tolerance : +/- 0.50 Log10 cp/m

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	6	100.0	0.0	0.0	4.17	4.4	e*