

Verein für
Association pour le
Associazione per il



medizinische Qualitätskontrolle
contrôle de qualité médical
controllo di qualità medico

Survey Report

2024 - 4

Survey Specimens

The homogeneity and stability of all specimens were checked before and/or during shipment and no irregularities were noted. The suitability tests were performed by the laboratories of the Universitätsspital Zürich (University Hospital Zurich) (<http://www.uzl.usz.ch/>).

The following survey specimens were produced specifically for MQ by a sub-contractor:
B2 Uricult, H4 Parasitic Hematology, K14 tumor marker

Determination of target values

For each target value, the type of determination per ISO17043: 2010 B2.1 is indicated (column "type"):

- a Value known due to production.
- b Certified reference value for use with special specimens
- c Reference value determined by analysis
- d Consensus values of expert laboratories
- e Consensus values of the participants

For methods groups with more than 9 participants, consensus values of the participants ("e") are generally determined.

In order to calculate the target values, we use the mean value of the method group. Values that differ more than 1.5 times the QUALAB-tolerance are outliers and are not used to calculate the target value. Starting point for the elimination of outliers are the values of our suitability tests.

In order to provide all participants with target values that are as meaningful as possible, other methods may also be applied for smaller method groups.

Uncertainty of the determined target values

The standard uncertainty (u_x) is calculated using the following formula (ISO13528):

$$u_x = (\text{target value}/100) * (1.25/\text{square root of "number of participants"}) * \%CV$$

- u_x has the same unit as the target value
- u_x can be compared with the standard deviation of the participants' collective ($SD = \text{target value} * \%CV / 100$)
- For participant numbers >18, the standard uncertainty (u_x) is significantly lower than the scatter of the collective participants and can be neglected.


QUALAB and MQ tolerances

For all mandatory analyzes, QUALAB tolerances are used (www.qualab.ch, external quality control). For non-mandatory analyzes, the tolerances are specified by MQ's survey specimen leader.

If the determined uncertainty, u_x , of the target value is greater than 15% of the QUALAB or MQ tolerance, the letter indicating the type of target detection is marked with an additional star (example "e*"). Thereby, we are alerting the participants to the fact that the uncertainty of the target value can have an impact on the evaluation.

Graphics

The results are shown graphically as follows:



Comparison of Devices

The data in this report allows you to compare the performance of different devices. However, remember to consider the following:


- The chemical control K1 is a ready-to-use commercial control serum. Even if the specimen is of human origin, it is possible that matrix effects occur. These are device-specific and result in different target values.
- Only one specimen was measured. Since the scatter of the results is dependent on the nature of the specimen (matrix effects) and on the signal strength, the determined coefficient of variations (CV in %) cannot be applied generally.
- A large number of runaways is due to administrative errors (wrong unit, results mixed up) or to operator errors (wrong sample, not correctly taken up in solution, not mixed well) and has nothing to do with the type of device.

Zürich, 8.12.2024

Dr. R. Fried
Survey Director

Publication of this report or any portion thereof without our prior written consent is not permitted. The original is archived at www.mqzh.ch

INR




QUALAB tolerance : 15 %

INR ()

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Innovin	16	100.0	0.0	0.0	1.56	3.9	e
2 Neoplastin R	15	100.0	0.0	0.0	1.66	4.7	e
3 Neoplastin Plus	9	100.0	0.0	0.0	2.00	6.2	e*
4 Recombiplastin 2G	15	100.0	0.0	0.0	1.68	5.9	e
5 Other methods	13	92.3	0.0	7.7	1.62	6.6	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Fibrinogen OA




QUALAB tolerance : 15 %

Fibrinogen OA (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens Thrombin	9	100.0	0.0	0.0	1.03	6.8	e*
2	Stago/STA	21	100.0	0.0	0.0	1.12	5.9	e
3	Fibrinogen Q.F.A.	16	93.7	0.0	6.3	1.08	7.3	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Activated Prothrombin Time




QUALAB tolerance : 25 %

Activated Prothrombin Time (Sek)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Actin FS	6	100.0	0.0	0.0	37.8	7.1	e
2	Pathromtin SL	7	100.0	0.0	0.0	59.1	3.5	e
3	Stago/STA	23	100.0	0.0	0.0	50.9	5.4	e
4	aPTT-SP	11	100.0	0.0	0.0	40.4	5.6	e
5	Other methods	10	100.0	0.0	0.0	39.7	7.8	e

INR CoaguChek




QUALAB tolerance : 15 %

INR CoaguChek ()

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CoaguChek Pro II	910	98.4	0.9	0.7	2.9	4.1	e

Prothrombin time NT




QUALAB tolerance : 15 %

Prothrombin time NT (%)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Innovin	13	100.0	0.0	0.0	100	0.8	e
2 Neoplastin R	12	100.0	0.0	0.0	96	4.4	e
3 Neoplastin Plus	4	100.0	0.0	0.0	93	4.6	e*
4 STA-NeoPTimal	7	100.0	0.0	0.0	91	3.3	e
5 Recombiplastin 2G	11	100.0	0.0	0.0	99	3.7	e
6 Other methods	15	100.0	0.0	0.0	98	4.2	e

Fibrinogen N




QUALAB tolerance : 15 %

Fibrinogen N (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens Thrombin	11	90.9	9.1	0.0	2.67	7.3	e*
2	Stago/STA	19	100.0	0.0	0.0	2.98	5.2	e
3	Fibrinogen Q.F.A.	18	100.0	0.0	0.0	2.69	4.7	e
4	Fib Clauss (IL)	6	100.0	0.0	0.0	2.62	7.1	e*

Faktor V




MQ tolerance : 25 %

Faktor V (%)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	8	100.0	0.0	0.0	94.0	7.3	e

aPTT N




QUALAB tolerance : 25 %

aPTT N (Sek)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Actin FS	10	90.0	0.0	10.0	22.9	3.0	e
2	Pathromtin SL	7	100.0	0.0	0.0	39.1	7.4	e
3	Stago/STA	20	100.0	0.0	0.0	32.0	6.9	e
4	aPTT-SP	13	100.0	0.0	0.0	25.9	7.2	e
5	Other methods	6	100.0	0.0	0.0	36.2	1.7	e

Faktor VII




MQ tolerance : 25 %

Faktor VII (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	5	100.0	0.0	0.0	100.0	0.4	e

Prothrombin time HT




QUALAB tolerance : 15 %

Prothrombin time HT (%)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Innovin	14	100.0	0.0	0.0	98	4.0	e
2 Neoplastin R	14	100.0	0.0	0.0	94	5.6	e
3 Recombiplastin 2G	14	100.0	0.0	0.0	98	2.1	e
4 Other methods	9	100.0	0.0	0.0	100	2.2	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti-FXa (unfrakt-Heparin)




MQ tolerance : 20 %

Anti-FXa (unfrakt-Heparin) (IU/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Stago/STA	12	83.4	8.3	8.3	0.10	14.5	a
2 ACL	21	100.0	0.0	0.0	0.24	13.6	a
3 Other methods	20	80.0	15.0	5.0	0.25	20.5	a

Fibrinogen H




QUALAB tolerance : 15 %

Fibrinogen H (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens Thrombin	4	100.0	0.0	0.0	3.12	6.2	e*
2	Stago/STA	16	100.0	0.0	0.0	3.17	5.2	e
3	Fibrinogen Q.F.A.	14	92.9	0.0	7.1	2.99	7.6	e*
4	Other methods	11	100.0	0.0	0.0	2.90	5.1	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

aPTT H




QUALAB tolerance : 25 %

aPTT H (Sek)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Actin FS	4	100.0	0.0	0.0	31.2	2.4	e
2	Actin FSL	6	83.3	0.0	16.7	33.3	4.8	e
3	Pathromtin SL	7	100.0	0.0	0.0	56.3	9.3	e*
4	Stago/STA	14	100.0	0.0	0.0	50.5	8.5	e
5	aPTT-SP	12	91.7	8.3	0.0	45.8	11.7	e*
6	Other methods	4	100.0	0.0	0.0	32.8	3.9	e

D-dimer



QUALAB tolerance : 21 %

D-dimer (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche (Zitratplasma)	9	88.9	11.1	0.0	0.39	11.9	e*
2	STA Liatest	17	88.2	11.8	0.0	0.38	13.7	e*
3	Siemens Innovance	13	92.3	7.7	0.0	0.65	8.8	e
4	Pathfast	22	95.5	4.5	0.0	1.83	10.0	e
5	ACL	16	100.0	0.0	0.0	0.79	7.3	e
6	AQT 90 FLEX	5	100.0	0.0	0.0	0.27	5.5	e
7	VIDAS	16	100.0	0.0	0.0	0.53	7.7	a

5 additional results were submitted but not published because the method groups were too small. (< results per group)

D-Dimer Triage




QUALAB tolerance : 21 %

D-Dimer Triage (ng/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Triage	209	84.2	10.5	5.3	202.42	13.0	e

D-dimer qn AFIAS




QUALAB tolerance : 21 %

D-dimer qn AFIAS (mg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 AFIAS	487	80.1	8.2	11.7	0.42	12.5	e

CoaguChek APTT




QUALAB tolerance : 25 %

CoaguChek APTT (Sek)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CoaguChek Pro II	7	85.7	0.0	14.3	85.0	7.2	e

INR CCXS




QUALAB tolerance : 15 %

INR CCXS ()

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CoaguChek XS	1318	99.2	0.3	0.5	2.8	3.7	e

INR HC




QUALAB tolerance : 15 %

INR HC ()

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Hemochron j.	5	100.0	0.0	0.0	5.2	6.2	e*

INR MI




QUALAB tolerance : 15 %

INR MI ()

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	microINR Expert	15	60.0	13.3	26.7	2.0	11.6	e*
2	MicroINR	114	81.6	4.4	14.0	2.1	8.4	e

INR Xprecia




QUALAB tolerance : 15 %
(< 1.3: +/- 0.2)

INR Xprecia ()

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Xprecia	41	97.6	2.4	0.0	1.2	6.3	e

INR Lumira Dx




QUALAB tolerance : 15 %

INR Lumira Dx ()


No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Lumira Dx	7	100.0	0.0	0.0	1.3	7.0	e*

Anti-FXa (LMW-Heparin)



No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Stago/STA	13	84.6	15.4	0.0	0.64	13.7	e*
2 ACL	23	87.0	13.0	0.0	0.67	13.0	e*
3 Other methods	12	100.0	0.0	0.0	0.71	9.5	e*

Anti-FXa (Rivaroxaban)




MQ tolerance : 20 %

Anti-FXa (Rivaroxaban) (µg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Stago/STA	12	91.7	8.3	0.0	71.12	11.6	e*
2	ACL	11	100.0	0.0	0.0	77.05	9.6	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Anti-FXa (Apixaban)




MQ tolerance : 20 %

Anti-FXa (Apixaban) (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Stago/STA	8	100.0	0.0	0.0	283.50	8.5	e*
2 ACL	5	100.0	0.0	0.0	261.70	7.7	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti-FXa (Edoxaban)




MQ tolerance : 20 %

Anti-FXa (Edoxaban) (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	5	100.0	0.0	0.0	105.00	11.8	e*

Anti-FIIa (Dabigatran)




MQ tolerance : 20 %

Anti-FIIa (Dabigatran) (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	4	100.0	0.0	0.0	218.40	15.4	e*

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Hemocue Hb 801	6	100.0	0.0	0.0	122.0	4.5	e*
2	Automat	7	85.7	0.0	14.3	106.0	2.4	e
3	Cyanmethemoglobin	14	92.9	0.0	7.1	103.4	3.2	e
4	Sysmex X	62	98.4	0.0	1.6	105.6	1.7	e
5	Hemocue	381	91.1	2.9	6.0	104.3	3.1	e
6	Hemocontrol	14	92.9	0.0	7.1	105.5	1.5	e
7	DiaSpect	14	64.3	0.0	35.7	108.0	3.1	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex KX21	7	100.0	0.0	0.0	106.0	0.7	e
2	Sysmex PochH - 100i	178	98.9	0.0	1.1	104.9	2.3	e
3	Sysmex XP 300	622	98.4	0.2	1.4	105.8	1.7	e
4	Mythic	218	98.6	0.0	1.4	103.0	2.6	e
5	Sysmex XQ-320	181	97.2	0.6	2.2	106.7	1.8	e
6	Swelab	27	96.3	3.7	0.0	111.1	3.2	e
7	Celltac Alpha (Nihon	50	94.0	0.0	6.0	108.1	2.2	e
8	Samsung HC10	5	100.0	0.0	0.0	106.0	3.2	e*
9	Micros 60	53	94.3	0.0	5.7	103.0	3.0	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Automat	4	100.0	0.0	0.0	0.29	2.9	e*
2	Sysmex X	62	98.4	0.0	1.6	0.30	2.8	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex XQ-320	178	96.6	0.0	3.4	0.29	2.5	e
2	Sysmex KX21	7	100.0	0.0	0.0	0.28	2.9	e*
3	Sysmex PochH - 100i	178	99.4	0.0	0.6	0.30	2.5	e
4	Sysmex XP 300	621	97.9	0.3	1.8	0.29	1.9	e
5	Mythic	217	96.8	0.9	2.3	0.29	3.2	e
6	Swelab	27	96.3	3.7	0.0	0.30	4.0	e
7	Celltac Alpha (Nihon	50	86.0	6.0	8.0	0.30	4.0	e
8	Samsung HC10	5	100.0	0.0	0.0	0.32	2.5	e*
9	Micros 60	53	86.8	5.7	7.5	0.27	4.2	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Automat	5	100.0	0.0	0.0	3.51	4.5	e
2	Sysmex X	62	98.4	0.0	1.6	3.47	1.9	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex XQ-320	179	97.7	0.6	1.7	3.62	3.6	e
2	Sysmex KX21	7	100.0	0.0	0.0	3.52	1.9	e
3	Sysmex PochH - 100i	178	99.4	0.6	0.0	3.63	2.9	e
4	Sysmex XP 300	623	98.3	0.3	1.4	3.50	2.6	e
5	Mythic	216	99.1	0.9	0.0	3.50	4.4	e
6	Swelab	27	100.0	0.0	0.0	3.53	3.2	e
7	Celltac Alpha (Nihon	50	96.0	0.0	4.0	3.66	3.9	e
8	Samsung HC10	5	100.0	0.0	0.0	3.59	2.1	e
9	Micros 60	53	94.3	0.0	5.7	3.40	4.1	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Automat	4	100.0	0.0	0.0	5.87	9.1	e*
2	Microscopic	4	100.0	0.0	0.0	5.05	14.0	e*
3	Sysmex X	62	100.0	0.0	0.0	6.24	6.7	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex XQ-320	179	98.3	1.1	0.6	6.23	6.1	e
2	Sysmex KX21	7	100.0	0.0	0.0	5.90	8.5	e*
3	Sysmex Poch - 100i	177	99.4	0.6	0.0	5.88	6.1	e
4	Sysmex XP 300	623	98.6	0.6	0.8	6.04	5.9	e
5	Mythic	217	98.2	0.9	0.9	5.99	6.9	e
6	Swelab	27	96.3	0.0	3.7	6.38	5.1	e
7	Celltac Alpha (Nihon	50	98.0	0.0	2.0	6.41	5.2	e
8	Samsung HC10	5	80.0	0.0	20.0	6.29	5.1	e
9	Micros 60	53	96.2	0.0	3.8	5.89	6.0	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Automat	4	100.0	0.0	0.0	194.8	3.7	e
2	Sysmex X	62	100.0	0.0	0.0	178.8	6.2	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex XQ-320	179	97.7	0.6	1.7	174.2	7.7	e
2	Sysmex KX21	7	100.0	0.0	0.0	186.0	8.7	e*
3	Sysmex PochH - 100i	177	99.4	0.6	0.0	193.0	6.1	e
4	Sysmex XP 300	623	99.2	0.5	0.3	195.6	6.0	e
5	Mythic	218	98.6	0.5	0.9	180.2	8.9	e
6	Swelab	27	96.3	3.7	0.0	186.1	11.3	e
7	Celltac Alpha (Nihon	50	96.0	2.0	2.0	192.6	7.2	e
8	Samsung HC10	5	100.0	0.0	0.0	215.0	6.6	e
9	Micros 60	53	90.5	5.7	3.8	183.3	11.9	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Hemoglobin H2




QUALAB tolerance : 9 %

Hemoglobin H2 (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	6	100.0	0.0	0.0	100.0	4.8	e*
2	Microsemi	939	97.0	0.7	2.3	97.6	2.4	e
3	Abx Micros	6	83.3	0.0	16.7	97.0	4.8	e*
4	Z3	252	96.0	0.0	4.0	98.4	3.2	e
5	MEK-1303/5	47	91.5	0.0	8.5	98.9	2.2	e

Hematocrit H2




QUALAB tolerance : 9 %

Hematocrit H2 (l/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	6	66.6	16.7	16.7	0.27	7.8	e*
2	Microsemi	939	94.9	1.9	3.2	0.26	3.0	e
3	Abx Micros	6	83.3	0.0	16.7	0.26	1.0	e
4	Z3	253	94.8	1.2	4.0	0.28	3.4	e
5	MEK-1303/5	47	89.4	0.0	10.6	0.32	2.5	e

Leucocytes H2




QUALAB tolerance : 25 %

Leucocytes H2 (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	6	100.0	0.0	0.0	5.84	6.3	e
2	Microsemi	939	98.4	0.3	1.3	5.71	5.2	e
3	Abx Micros	6	100.0	0.0	0.0	5.70	3.0	e
4	Z3	253	97.6	1.2	1.2	5.54	6.0	e
5	MEK-1303/5	46	97.8	0.0	2.2	5.91	4.7	e

Thrombocytes H2




QUALAB tolerance : 25 %

Thrombocytes H2 (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	6	100.0	0.0	0.0	195.0	8.2	e*
2	Microsemi	938	98.4	0.5	1.1	175.1	7.1	e
3	Abx Micros	6	100.0	0.0	0.0	175.0	9.3	e*
4	Z3	253	97.2	0.4	2.4	185.6	6.9	e
5	MEK-1303/5	46	95.7	0.0	4.3	185.8	6.1	e

Erythrocytes H2




QUALAB tolerance : 25 %

Erythrocytes H2 (T/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	6	100.0	0.0	0.0	3.49	7.8	e*
2	Microsemi	939	96.8	0.3	2.9	3.22	3.6	e
3	Abx Micros	6	83.3	16.7	0.0	3.29	13.3	e*
4	Z3	253	96.0	0.4	3.6	3.18	4.1	e
5	MEK-1303/5	47	91.5	0.0	8.5	3.37	2.8	e

CRP H2




QUALAB tolerance : 21 %

CRP H2 (mg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dymind DP-H10	5	100.0	0.0	0.0	24.9	13.1	e*
2	Microsemi	921	92.6	4.5	2.9	29.7	10.2	e
3	Z3	236	94.1	4.2	1.7	30.6	10.0	e
4	MEK-1303/5	42	90.5	0.0	9.5	24.5	5.7	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Hemoglobin BG




QUALAB tolerance : 9 %

Hemoglobin BG (g/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 iStat	13	100.0	0.0	0.0	194.2	2.1	e

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 iStat	17	100.0	0.0	0.0	0.57	2.7	e
2 EPOC	16	87.5	0.0	12.5	0.49	7.1	a

Hemoglobin




QUALAB tolerance : 9 %

Hemoglobin (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	122	98.4	0.8	0.8	138.0	1.5	e
2	Beckman	5	100.0	0.0	0.0	138.0	1.1	e
3	Yumizen/Pentra	13	100.0	0.0	0.0	140.8	1.7	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Hematocrit




QUALAB tolerance : 9 %

Hematocrit (l/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	121	97.5	1.7	0.8	0.42	3.3	e
2 Beckman	5	100.0	0.0	0.0	0.41	2.1	e
3 Yumizen/Pentra	13	100.0	0.0	0.0	0.41	3.4	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Erythrocytes




QUALAB tolerance : 25 %

Erythrocytes (T/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	122	100.0	0.0	0.0	4.81	2.6	e
2	Beckman	5	100.0	0.0	0.0	4.79	1.2	e
3	Yumizen/Pentra	13	100.0	0.0	0.0	4.81	2.6	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Leucocytes




QUALAB tolerance : 25 %

Leucocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	121	100.0	0.0	0.0	6.62	6.2	e
2	Beckman	5	80.0	0.0	20.0	6.70	4.8	e
3	Yumizen/Pentra	13	92.3	7.7	0.0	5.27	12.6	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Thrombocytes




QUALAB tolerance : 25 %

Thrombocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	122	98.4	0.8	0.8	240.1	7.8	e
2	Beckman	5	100.0	0.0	0.0	237.0	4.4	e
3	Yumizen/Pentra	13	100.0	0.0	0.0	232.8	5.8	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Neutrophils




MQ tolerance : 25 %

Neutrophils (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	122	97.5	2.5	0.0	3.76	8.4	e
2	Beckman	5	80.0	0.0	20.0	4.12	2.2	e
3	Yumizen/Pentra	12	83.4	8.3	8.3	2.32	14.2	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Lymphocytes




MQ tolerance : 25 %

Lymphocytes (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	122	99.2	0.0	0.8	2.12	5.8	e
2	Beckman	5	100.0	0.0	0.0	2.10	13.8	e*
3	Yumizen/Pentra	12	100.0	0.0	0.0	2.35	12.4	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Monocytes




MQ tolerance : 80 %

Monocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	122	100.0	0.0	0.0	0.50	14.1	e
2	Beckman	5	100.0	0.0	0.0	0.36	21.6	e*
3	Yumizen/Pentra	12	91.7	8.3	0.0	0.42	33.2	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Eosinophils




MQ tolerance : 80 %
 (< 0.13: +/- 0.10 G/l)

Eosinophils (G/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	121	100.0	0.0	0.0	0.14	23.9	a
2 Beckman	5	100.0	0.0	0.0	0.20	17.4	e
3 Yumizen/Pentra	12	100.0	0.0	0.0	0.09	48.3	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Basophiles




MQ tolerance : 80 %
 (< 0.13: +/- 0.10 G/l)

Basophiles (G/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	121	100.0	0.0	0.0	0.07	54.1	a
2 Beckman	5	100.0	0.0	0.0	0.01	8.6	e
3 Yumizen/Pentra	12	100.0	0.0	0.0	0.04	78.4	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

MCV




MQ tolerance : 25 %

MCV (fl)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	111	100.0	0.0	0.0	86.6	2.8	e
2	Beckman	4	100.0	0.0	0.0	85.3	1.0	e
3	Yumizen/Pentra	9	100.0	0.0	0.0	85.0	2.6	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

MCH




MQ tolerance : 25 %

MCH (pg)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	112	100.0	0.0	0.0	28.7	1.3	e
2	Beckman	4	100.0	0.0	0.0	28.6	1.1	e
3	Yumizen/Pentra	10	100.0	0.0	0.0	29.5	3.0	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

MCHC




MQ tolerance : 25 %

MCHC (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	113	100.0	0.0	0.0	332	3.0	e
2	Beckman	4	100.0	0.0	0.0	334	0.9	e
3	Yumizen/Pentra	10	100.0	0.0	0.0	347	4.1	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

RDW




MQ tolerance : 25 %

RDW (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	107	97.2	0.0	2.8	14.1	2.1	e
2	Beckman	4	100.0	0.0	0.0	14.2	1.1	e
3	Yumizen/Pentra	10	100.0	0.0	0.0	14.4	7.7	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Immature Granulocytes




MQ tolerance : 25 %
 (< 1.30: +/- 0.10 G/l)

Immature Granulocytes (G/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	105	95.2	1.0	3.8	0.06	46.0	e*

Reticulocytes




MQ tolerance : 30 %

Reticulocytes (G/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	67	98.5	0.0	1.5	161.1	9.6	e
2	Advia	4	100.0	0.0	0.0	174.9	7.7	e*
3	Beckman	5	100.0	0.0	0.0	158.1	12.3	e*

Hämolyseindex Probe A




MQ tolerance : 30 % Hämolyseindex Probe A (Index (mg/))
 (< 20.000: +/- 8.000 Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	24	95.8	0.0	4.2	112.913	5.0	e
2	Abbott	7	100.0	0.0	0.0	0.850	14.6	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Hämolyseindex Probe B




MQ tolerance : 30 % Hämolyseindex Probe B (Index (mg/))
 (< 20.000: +/- 8.000 Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	20	95.0	0.0	5.0	0.020	145.7	e*
2	Abbott	7	100.0	0.0	0.0	0.150	14.7	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Lipemia index A




MQ tolerance : 30 %

Lipemia index A (Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	9	100.0	0.0	0.0	25.00	12.3	e*
2	Abbott	6	83.3	0.0	16.7	27.29	8.9	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Lipemia index B




MQ tolerance : 30 %

Lipemia index B (Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	9	88.9	0.0	11.1	454.00	16.3	e*
2	Abbott	6	83.3	0.0	16.7	293.17	9.7	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Icteria Index A




MQ tolerance : 30 %
 (< 5.00: +/- 1.50 Index (mg/))

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	8	100.0	0.0	0.0	0.01	0.0	e
2	Abbott	6	100.0	0.0	0.0	1.40	13.0	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Icteria Index B




MQ tolerance : 30 %
 (< 5.00: +/- 1.50 Index (mg/))

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Cobas	8	100.0	0.0	0.0	0.01	0.0	e
2 Abbott	6	100.0	0.0	0.0	1.54	14.3	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Erythrocyte sedimentation rate 1h




MQ tolerance : 40 %

Erythrocyte sedimentation rate 1h (mm/h)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	MINI-CUBE	28	92.9	7.1	0.0	118	23.0	e
2	Sarstedt Sedivette	18	83.3	5.6	11.1	75	20.5	e*
3	Sarstedt Microvette	4	100.0	0.0	0.0	70	16.1	e*
4	BD Seditainer	42	90.4	4.8	4.8	79	18.8	e
5	Other methods	18	88.9	11.1	0.0	77	21.2	e*

Erythrocyte sedimentation rate 2h




MQ tolerance : 40 %

Erythrocyte sedimentation rate 2h (mm/2h)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sarstedt Sedivette	6	100.0	0.0	0.0	115	16.6	e*
2	BD Seditainer	4	100.0	0.0	0.0	115	16.2	e*
3	Other methods	4	100.0	0.0	0.0	123	22.4	e*

Hemoglobin HS




MQ tolerance : 9 %

Hemoglobin HS (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	PixCell HemoScreen	25	60.0	0.0	40.0	161.3	2.6	e

Hematocrit HS




MQ tolerance : 9 %

Hematocrit HS (l/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	PixCell HemoScreen	25	60.0	0.0	40.0	0.4	2.3	e

Erythrocytes HS




MQ tolerance : 25 %

Erythrocytes HS (T/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	PixCell HemoScreen	25	60.0	0.0	40.0	4.51	7.3	e

Leucocytes HS




MQ tolerance : 25 %

Leucocytes HS (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	PixCell HemoScreen	25	100.0	0.0	0.0	7.64	6.2	e

Trombocytes HS




MQ tolerance : 25 %

Trombocytes HS (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	PixCell HemoScreen	25	100.0	0.0	0.0	245.7	4.4	e

Leucocytes BF




MQ tolerance : 40 %

Leucocytes BF (G/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	14	100.0	0.0	0.0	0.294	5.6	e

Erythrocytes BF




MQ tolerance : 40 %

Erythrocytes BF (G/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Sysmex	13	100.0	0.0	0.0	72.245	6.4	e

Total cells (TC)




MQ tolerance : 40 %

Total cells (TC) (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	11	100.0	0.0	0.0	0.300	4.9	e

Mononuclear cells (MN)




MQ tolerance : 40 %

Mononuclear cells (MN) (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	13	84.6	15.4	0.0	0.134	38.6	a

Polynuclear cells (PMN)




MQ tolerance : 40 %

Polynuclear cells (PMN) (G/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Sysmex	13	84.6	15.4	0.0	0.179	46.9	a

Erythrocyte sedimentation rate




MQ tolerance : 40 %

Erythrocyte sedimentation rate (mm/h)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 miniiSed	9	100.0	0.0	0.0	19	27.9	a


CRP



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Autolyser	10	90.0	0.0	10.0	35.6	6.6	e
2	Cobas b101	431	97.2	0.9	1.9	32.6	6.3	e
3	Siemens	8	100.0	0.0	0.0	37.2	2.9	e
4	Cobas	44	100.0	0.0	0.0	36.4	5.6	e
5	Turbidimetry	4	100.0	0.0	0.0	39.3	8.0	e*
6	Afinion	1066	99.1	0.5	0.4	37.0	5.0	e
7	NycoCard SingleTest-	48	77.1	8.3	14.6	35.0	11.8	e
8	Quick Read go	85	94.1	2.4	3.5	36.9	7.9	e
9	Eurolyser	53	83.0	5.7	11.3	42.2	10.8	e
10	Fuji Dri-Chem	12	91.7	8.3	0.0	38.1	9.7	e*
11	Piccolo	5	100.0	0.0	0.0	49.1	8.3	e*
12	Celltac chemi	20	90.0	0.0	10.0	35.8	4.8	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)


CRP



QUALAB tolerance : 21 %

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	QuikRead (Vollblut)	10	90.0	0.0	10.0	63.4	6.9	e

CRP




QUALAB tolerance : 21 %

CRP (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Spinit	7	100.0	0.0	0.0	30.4	2.6	e
2	Abbott	12	100.0	0.0	0.0	37.9	3.3	e
3	Beckman	5	80.0	0.0	20.0	34.1	1.0	e
4	AQT 90 FLEX	5	100.0	0.0	0.0	37.0	6.7	e*
5	Spotchem D-Concept	6	100.0	0.0	0.0	39.9	3.2	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

IgG




QUALAB tolerance : 15 %

IgG (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	5	100.0	0.0	0.0	13.10	0.7	e
2	Roche, Cobas	10	100.0	0.0	0.0	13.13	2.6	e
3	Turbidimetry	7	100.0	0.0	0.0	13.10	2.0	e
4	Siemens	5	100.0	0.0	0.0	13.60	4.9	e*
5	Other methods	4	100.0	0.0	0.0	12.25	6.7	e*

IgA




QUALAB tolerance : 15 %

IgA (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	2.40	2.2	e
2	Roche, Cobas	10	100.0	0.0	0.0	2.44	2.3	e
3	Siemens	6	100.0	0.0	0.0	2.67	5.2	e*
4	Turbidimetry	6	100.0	0.0	0.0	2.43	1.9	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

IgM




QUALAB tolerance : 15 %

IgM (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Turbidimetry	10	100.0	0.0	0.0	1.23	5.4	e
2	Siemens	8	100.0	0.0	0.0	1.26	5.4	e*
3	Roche, Cobas	11	100.0	0.0	0.0	1.18	3.8	e

IgE




QUALAB tolerance : 30 %

IgE (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	4	100.0	0.0	0.0	177	4.5	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

Alpha-1-Antitrypsin




MQ tolerance : 25 %

Alpha-1-Antitrypsin (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	1.26	4.1	e
2	Roche, Cobas	5	100.0	0.0	0.0	1.26	4.7	e
3	Siemens	5	100.0	0.0	0.0	1.34	1.6	e

Anti-Streptolysin-Antibodies




MQ tolerance : 25 %

Anti-Streptolysin-Antibodies (kIU/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	6	100.0	0.0	0.0	105	3.4	e
2	Roche, Cobas	9	100.0	0.0	0.0	144	5.5	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Complement C3




QUALAB tolerance : 15 %

Complement C3 (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	1.66	1.0	e
2	Roche, Cobas	7	100.0	0.0	0.0	1.75	4.4	e
3	Siemens	5	80.0	20.0	0.0	1.74	11.1	e*
4	Other methods	4	100.0	0.0	0.0	1.80	5.3	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Complement C4




QUALAB tolerance : 15 %

Complement C4 (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Alinity	5	100.0	0.0	0.0	0.35	2.4	e
2	Roche, Cobas	6	100.0	0.0	0.0	0.32	2.6	e
3	Siemens	5	100.0	0.0	0.0	0.38	3.3	e
4	Other methods	5	100.0	0.0	0.0	0.35	6.6	e*


Haptoglobin



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	1.59	1.4	e
2	Roche, Cobas	21	100.0	0.0	0.0	1.62	2.1	e
3	Other methods	4	100.0	0.0	0.0	1.61	1.8	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Transferrin




MQ tolerance : 25 %

Transferrin (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	6	100.0	0.0	0.0	2.23	3.3	e
2	Roche, Cobas	22	100.0	0.0	0.0	2.24	2.5	e
3	Siemens	5	100.0	0.0	0.0	2.21	2.5	e
4	Other methods	5	100.0	0.0	0.0	2.29	3.3	e


Beta-2 microglobuline



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	1.95	3.3	e
2	Roche, Cobas	6	100.0	0.0	0.0	1.85	3.5	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)


Rheumatoid factor



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	10	100.0	0.0	0.0	30.1	3.9	e
2	Abbott	9	77.8	0.0	22.2	35.4	2.4	a


3 additional results were submitted but not published because the method groups were too small. (< results per group)

Ceruloplasmin



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens	5	100.0	0.0	0.0	344.00	5.2	e
2	Abbott	4	100.0	0.0	0.0	350.00	0.0	e
3	Roche, Cobas	4	100.0	0.0	0.0	250.00	2.3	e


Prealbumin



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	7	100.0	0.0	0.0	209.00	3.1	e
2	all Participants	7	100.0	0.0	0.0	220.00	5.8	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)


Soluble transferrin receptor



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	14	100.0	0.0	0.0	3.2	1.9	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

free light chain kappa




QUALAB tolerance : 20 %

free light chain kappa (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	N Latex	6	100.0	0.0	0.0	11.00	5.5	e
2	Freelite	12	100.0	0.0	0.0	20.41	8.3	e*

free light chain lambda




QUALAB tolerance : 20 %

free light chain lambda (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	N Latex	6	100.0	0.0	0.0	10.60	3.2	e
2	Freelite	12	100.0	0.0	0.0	17.88	7.7	e

IgE peanut qn




MQ tolerance : 30 %
 (< 1.50: +/- 0.45 kU/L)

IgE peanut qn (kU/L)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	14	100.0	0.0	0.0	0.14	45.5	e*

IgE birch qn




QUALAB tolerance : 30 %
(< 1.50: +/- 0.45 kU/L)

IgE birch qn (kU/L)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	15	100.0	0.0	0.0	0.12	57.6	e*

IgE cat qn




QUALAB tolerance : 30 %
 (< 1.50: +/- 0.45 kU/L)

IgE cat qn (kU/L)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	14	100.0	0.0	0.0	0.12	58.8	e*

IgE total




QUALAB tolerance : 30 %

IgE total (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	14	100.0	0.0	0.0	585	7.7	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

IgE sx1 qn




MQ tolerance : 30 %

IgE sx1 qn (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	10	100.0	0.0	0.0	11.54	14.7	a

IgE fx5 qn




MQ tolerance : 30 %
 (< 1.50: +/- 0.45 kU/L)

IgE fx5 qn (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	10	100.0	0.0	0.0	0.20	32.4	e*

IgE rx1qn




MQ tolerance : 30 %

IgE rx1qn (kU/L)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	8	100.0	0.0	0.0	1.77	17.4	a

IgE rx2 qn




MQ tolerance : 30 %

IgE rx2 qn (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	8	100.0	0.0	0.0	2.49	15.8	a

IgE D. pteronyssinus qn




QUALAB tolerance : 30 % IgE D. pteronyssinus qn (kU/L)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	12	100.0	0.0	0.0	4.64	14.5	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

CRP HS




QUALAB tolerance : 21 %
 (< 10.00: +/- 2.00 mg/l)

CRP HS (mg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Turbidimetry	10	100.0	0.0	0.0	1.71	7.8	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

Lipoprotein (a)




MQ tolerance : 25 %

Lipoprotein (a) (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	7	100.0	0.0	0.0	24	4.6	e
2	Others	5	100.0	0.0	0.0	37	7.9	e*

Apolipoprotein A1




MQ tolerance : 30 %

Apolipoprotein A1 (g/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Alinity	4	100.0	0.0	0.0	1.18	5.7	e

Apolipoprotein B




MQ tolerance : 30 %

Apolipoprotein B (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Alinity	4	100.0	0.0	0.0	1.0	5.4	e

CRP




QUALAB tolerance : 21 %

CRP (mg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 AFIAS	204	93.2	3.4	3.4	31.5	8.7 e

Anti deam. Gliadin IgG




MQ tolerance : 30 %
 (< 2.00: +/- 1.00 U/ml)

Anti deam. Gliadin IgG (U/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Phadia	8	100.0	0.0	0.0	0.50	69.6	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti deam. Gliadin IgA




MQ tolerance : 30 %

Anti deam. Gliadin IgA (U/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Phadia	7	100.0	0.0	0.0	4.00	12.9	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti tTG IgG




MQ tolerance : 30 %
 (< 2.00: +/- 1.00 U/ml)

Anti tTG IgG (U/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Other methods	5	100.0	0.0	0.0	0.80	29.4	e*

Anti tTG IgA




MQ tolerance : 30 %
 (< 2.00: +/- 1.00 U/ml)

Anti tTG IgA (U/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Other methods	9	100.0	0.0	0.0	0.60	36.7	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

CRP Lumira




QUALAB tolerance : 21 %

CRP Lumira (mg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Lumira Dx	11	100.0	0.0	0.0	98.2	7.3	e

Albumine




QUALAB tolerance : 12 %
(< 30.0: +/- 3.6 g/l)

Albumine (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Turbidimetry	5	100.0	0.0	0.0	31.0	3.4	e*
2	Abbott	9	100.0	0.0	0.0	34.0	2.3	e
3	Beckman	4	100.0	0.0	0.0	35.0	1.4	e
4	Roche	37	97.3	2.7	0.0	36.3	3.7	e
5	Autolyser	8	87.5	12.5	0.0	35.0	5.2	e*
6	Selectra Pro	10	100.0	0.0	0.0	34.1	3.8	e
7	Fuji Dri-Chem	245	98.8	0.0	1.2	44.2	4.2	e
8	Spotchem D-Concept	235	99.1	0.9	0.0	42.2	5.2	e
9	Spotchem SP-4430	28	96.4	3.6	0.0	41.1	4.7	e
10	Piccolo	63	98.4	0.0	1.6	36.4	4.1	e
11	Skyla	4	100.0	0.0	0.0	36.5	3.9	e*
12	Hitachi S40/M40	4	100.0	0.0	0.0	29.0	1.0	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Alkaline phosphatase




QUALAB tolerance : 18 %

Alkaline phosphatase (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	12	100.0	0.0	0.0	130	5.3	e
2	Beckman	6	83.3	16.7	0.0	129	10.7	e*
3	Roche	37	100.0	0.0	0.0	110	4.2	e
4	Siemens	7	100.0	0.0	0.0	114	2.9	e
5	Autolyser	22	100.0	0.0	0.0	113	6.2	e
6	Selectra Pro	15	100.0	0.0	0.0	133	6.6	e
7	Fuji Dri-Chem	1030	98.4	0.8	0.8	132	6.2	e
8	Spotchem D-Concept	559	99.3	0.2	0.5	114	5.3	e
9	Spotchem SP-4430	77	97.4	2.6	0.0	115	7.3	e
10	Piccolo	50	96.0	4.0	0.0	141	6.0	e
11	Skylla	5	100.0	0.0	0.0	151	3.3	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

Amylase




QUALAB tolerance : 18 %
(< 50: +/- 9 U/l)

Amylase (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	9	100.0	0.0	0.0	174	3.8	e
2	Beckman	4	100.0	0.0	0.0	193	0.7	e
3	Roche	15	100.0	0.0	0.0	178	2.1	e
4	Autolyser	9	100.0	0.0	0.0	160	5.3	e
5	Selectra Pro	10	90.0	10.0	0.0	229	7.7	e*
6	Fuji Dri-Chem	736	99.2	0.7	0.1	137	4.2	e
7	Spotchem D-Concept	396	99.0	1.0	0.0	92	7.0	e
8	Spotchem SP-4430	61	62.2	23.0	14.8	46	16.1	e
9	Piccolo	59	100.0	0.0	0.0	166	4.1	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Pancreatic amylase




QUALAB tolerance : 18 %

Pancreatic amylase (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	141	2.3	e
2	Roche	17	100.0	0.0	0.0	156	2.2	e
3	Siemens	4	100.0	0.0	0.0	181	8.0	e*
4	Autolyser	10	100.0	0.0	0.0	147	4.2	e

Bilirubin




QUALAB tolerance : 18 %

Bilirubin (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	12	100.0	0.0	0.0	33.9	6.5	e
2	Beckman	6	100.0	0.0	0.0	42.2	5.1	e
3	Roche	37	94.6	5.4	0.0	32.9	7.2	e
4	Siemens	7	100.0	0.0	0.0	43.3	4.2	e
5	Autolyser	19	100.0	0.0	0.0	33.2	5.5	e
6	Selectra Pro	15	86.6	6.7	6.7	30.1	9.3	e*
7	Fuji Dri-Chem	839	95.5	3.3	1.2	37.9	7.9	e
8	Spotchem D-Concept	450	98.0	1.3	0.7	39.7	6.4	e
9	Spotchem SP-4430	72	98.6	0.0	1.4	39.6	6.7	e
10	Piccolo	58	98.3	1.7	0.0	33.2	6.8	e
11	Skyla	7	100.0	0.0	0.0	35.9	7.9	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Bilirubin direct




MQ tolerance : 18 %

Bilirubin direct (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	12	100.0	0.0	0.0	23.1	8.9	a
2	Fuji Dri-Chem	22	100.0	0.0	0.0	22.3	6.4	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Calcium




QUALAB tolerance : 9 %
(< 2.00: +/- 0.18 mmol/l)

Calcium (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	2.22	1.5	e
2	Beckman	5	100.0	0.0	0.0	2.25	3.7	e*
3	Roche	39	100.0	0.0	0.0	2.21	1.9	e
4	Siemens	5	100.0	0.0	0.0	2.20	2.2	e
5	Autolyser	9	100.0	0.0	0.0	2.20	3.4	e*
6	Fuji Dri-Chem	261	95.8	2.3	1.9	2.32	3.9	e
7	Spotchem D-Concept	71	90.2	7.0	2.8	1.60	5.9	e
8	Spotchem SP-4430	12	100.0	0.0	0.0	2.15	3.9	e*
9	Piccolo	50	100.0	0.0	0.0	2.23	2.5	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

Calcium ISE




MQ tolerance : 12 %

Calcium ISE (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	iStat Chem8	11	100.0	0.0	0.0	0.87	1.1	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)


Chloride



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	101	1.9	e
2	Beckman	5	100.0	0.0	0.0	98	0.9	e
3	Roche	29	96.6	0.0	3.4	98	0.8	e
4	Siemens	5	100.0	0.0	0.0	102	1.8	e*
5	Fuji Dri-Chem	950	97.5	1.7	0.8	108	2.0	e
6	Spotchem D-Concept	434	98.3	0.5	1.2	111	1.6	e
7	Spotchem EL-SE 1520	61	80.3	13.1	6.6	117	3.5	e
8	Piccolo	24	100.0	0.0	0.0	101	2.3	e
9	Exias	24	95.8	0.0	4.2	101	1.8	e
10	iStat Chem8	11	100.0	0.0	0.0	107	0.6	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Cholesterol total




QUALAB tolerance : 10 %

Cholesterol total (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	16	100.0	0.0	0.0	3.80	2.2	e
2	Beckman	4	100.0	0.0	0.0	3.72	3.6	e*
3	Roche	31	100.0	0.0	0.0	3.69	2.7	e
4	Siemens	4	75.0	0.0	25.0	3.72	2.6	e*
5	Autolyser	22	95.5	4.5	0.0	3.72	3.9	e
6	Selectra Pro	13	92.3	0.0	7.7	3.78	3.1	e
7	Fuji Dri-Chem	981	97.1	1.2	1.7	3.47	3.7	e
8	Spotchem D-Concept	479	97.9	1.3	0.8	3.66	3.5	e
9	Spotchem SP-4430	76	97.4	1.3	1.3	3.50	3.8	e
10	Piccolo	26	100.0	0.0	0.0	3.78	3.1	e
11	Reflotron	4	100.0	0.0	0.0	3.41	7.8	e*
12	Cholestech LDX	256	96.1	3.1	0.8	3.55	4.3	e
13	Other methods	4	100.0	0.0	0.0	2.89	1.5	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Cholesterin HDL




QUALAB tolerance : 21 %

Cholesterin HDL (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	0.84	6.3	e
2	Beckman	6	100.0	0.0	0.0	1.10	4.2	e
3	Roche	29	100.0	0.0	0.0	0.81	4.1	e
4	Dimension	5	100.0	0.0	0.0	1.49	5.0	e
5	Autolyser	22	100.0	0.0	0.0	1.47	8.0	e
6	Selectra Pro	12	58.4	8.3	33.3	1.35	13.1	e*
7	Fuji Dri-Chem	948	98.4	0.3	1.3	1.31	4.0	e
8	Spotchem D-Concept	464	93.9	3.7	2.4	0.55	10.2	e
9	Spotchem SP-4430	69	91.4	4.3	4.3	0.51	10.6	e
10	Piccolo	25	96.0	4.0	0.0	0.93	8.8	e
11	Cholestech LDX	255	86.7	7.8	5.5	0.87	12.1	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Creatine kinase




QUALAB tolerance : 18 %

Creatine kinase (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	252	2.0	e
2	Beckman	5	100.0	0.0	0.0	249	2.2	e
3	Roche	34	100.0	0.0	0.0	242	2.2	e
4	Siemens	7	100.0	0.0	0.0	239	2.4	e
5	Autolyser	17	94.1	0.0	5.9	269	4.0	e
6	Selectra Pro	8	100.0	0.0	0.0	250	2.0	e
7	Fuji Dri-Chem	666	98.7	0.5	0.8	267	4.8	e
8	Spotchem D-Concept	319	95.3	3.1	1.6	205	8.4	e
9	Spotchem SP-4430	38	97.4	0.0	2.6	246	7.5	e
10	Piccolo	20	95.0	5.0	0.0	250	5.4	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

Cholesterol LDL




QUALAB tolerance : 18 %

Cholesterol LDL (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	2.0	3.4	e
2	Beckman	5	100.0	0.0	0.0	2.5	4.5	e
3	Roche, Cobas	17	100.0	0.0	0.0	2.6	3.6	e
4	Autolyser	11	90.9	0.0	9.1	2.5	4.8	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Iron




QUALAB tolerance : 20 %

Iron (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	12	100.0	0.0	0.0	32	3.3	e
2	Roche	23	100.0	0.0	0.0	32	1.9	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Gamma-glutamyltransferase




QUALAB tolerance : 18 %

Gamma-glutamyltransferase (U/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	59	4.2	e
2	Beckman	6	100.0	0.0	0.0	54	3.2	e
3	Cobas	38	100.0	0.0	0.0	51	6.9	e
4	Siemens	7	100.0	0.0	0.0	57	6.3	e*
5	Autolyser	22	100.0	0.0	0.0	53	4.3	e
6	Selectra Pro	14	100.0	0.0	0.0	51	5.4	e
7	Fuji Dri-Chem	1144	99.1	0.3	0.6	64	5.1	e
8	Spotchem D-Concept	630	99.0	0.5	0.5	61	6.0	e
9	Spotchem SP-4430	125	97.6	0.8	1.6	66	7.2	e
10	Piccolo	62	100.0	0.0	0.0	47	6.8	e
11	Skylla	5	100.0	0.0	0.0	49	3.9	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Glucose




QUALAB tolerance : 9 %

Glucose (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	14	100.0	0.0	0.0	8.6	1.7	e
2	Beckman	6	83.3	16.7	0.0	8.9	4.6	e*
3	Roche	40	100.0	0.0	0.0	8.7	2.0	e
4	Siemens	6	100.0	0.0	0.0	8.7	3.2	e*
5	Autolyser	20	95.0	0.0	5.0	8.5	3.6	e
6	Selectra Pro	14	100.0	0.0	0.0	8.9	3.5	e
7	Fuji Dri-Chem	1083	98.8	0.5	0.7	8.2	2.5	e
8	Spotchem D-Concept	592	94.3	5.4	0.3	8.5	4.4	e
9	Spotchem SP-4430	102	95.1	2.9	2.0	8.7	4.5	e
10	Piccolo	75	100.0	0.0	0.0	8.7	1.2	e
11	Cholestech LDX	253	92.5	2.8	4.7	8.4	4.3	e
12	iStat Chem8	12	100.0	0.0	0.0	7.9	1.9	e
13	Cobas Pulse	137	98.5	1.5	0.0	8.4	3.5	e
14	Other methods	4	100.0	0.0	0.0	8.9	3.5	e*

4 additional results were submitted but not published because the number of groups were too small. (< results per group)

Glucose




QUALAB tolerance : 9 %

Glucose (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Accu-Chek Instant	143	97.2	2.1	0.7	8.1	3.3	e
2	Accu-Chek Aviva	100	87.0	6.0	7.0	9.4	4.0	e
3	Accu-Chek Inform 2	949	97.3	1.5	1.2	10.2	3.3	e
4	Accu-Check Guide	343	97.9	0.9	1.2	8.2	3.0	e
5	Contour XT	1338	94.8	3.8	1.4	8.9	4.1	e
6	Skyla	5	100.0	0.0	0.0	8.8	1.4	e
7	Statstrip/Xpress	66	100.0	0.0	0.0	7.9	2.9	e
8	Glucocard	7	42.8	14.3	42.9	12.2	6.3	e*
9	Hemocue 201+ P-equiv	145	94.5	1.4	4.1	10.5	2.8	e
10	Hemocue 201RT P-equiv	132	97.7	1.5	0.8	10.4	2.6	e
11	Contour NEXT	60	98.3	1.7	0.0	8.4	4.2	e

11 additional results were submitted but not published because the method groups were too small. (< results per group)

Glucose B




QUALAB tolerance : 9 %

Glucose B (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	OneTouch Verio	29	100.0	0.0	0.0	8.6	2.5	e
2	Contour 2 (5s)	8	87.5	0.0	12.5	7.1	1.8	e
3	Healthpro	20	95.0	0.0	5.0	14.9	4.2	e
4	Mylife UNIO	237	95.4	2.5	2.1	10.1	4.2	e
5	mylife Pura	87	78.2	13.8	8.0	9.9	5.7	e
6	Alpha Check	34	88.2	5.9	5.9	14.1	5.1	e

Uric Acid




QUALAB tolerance : 12 %

Uric Acid ($\mu\text{mol/l}$)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	333	2.8	e
2	Beckman	6	100.0	0.0	0.0	345	1.4	e
3	Roche	35	100.0	0.0	0.0	332	2.2	e
4	Siemens	5	100.0	0.0	0.0	352	4.7	e*
5	Autolyser	18	100.0	0.0	0.0	341	4.1	e
6	Selectra Pro	15	93.3	0.0	6.7	334	2.5	e
7	Fuji Dri-Chem	1051	98.7	0.3	1.0	368	2.8	e
8	Spotchem D-Concept	594	99.3	0.5	0.2	314	4.1	e
9	Spotchem SP-4430	97	96.9	2.1	1.0	274	4.1	e
10	Piccolo	34	91.2	2.9	5.9	258	5.0	e
11	Skyla	9	100.0	0.0	0.0	340	5.3	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Urea




QUALAB tolerance : 15 %

Urea (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	10.0	4.2	e
2	Beckman	6	100.0	0.0	0.0	9.9	6.2	e*
3	Roche	35	100.0	0.0	0.0	9.6	3.1	e
4	Siemens	7	100.0	0.0	0.0	9.7	3.5	e
5	Autolyser	17	100.0	0.0	0.0	10.1	4.0	e
6	Selectra Pro	11	100.0	0.0	0.0	10.0	6.3	e*
7	Fuji Dri-Chem	625	98.7	0.2	1.1	10.2	2.3	e
8	Spotchem D-Concept	329	96.4	1.8	1.8	9.8	6.6	e
9	Spotchem SP-4430	50	96.0	4.0	0.0	10.0	6.5	e
10	Piccolo	67	100.0	0.0	0.0	8.7	3.3	e
11	Skyla	5	100.0	0.0	0.0	9.8	1.7	e
12	iStat Chem8	9	100.0	0.0	0.0	13.2	2.7	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Potassium




QUALAB tolerance : 6 %
(< 3.30: +/- 0.20 mmol/l)

Potassium (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	3.58	1.3	e
2	Beckman	6	100.0	0.0	0.0	3.60	1.3	e
3	Roche	38	100.0	0.0	0.0	3.61	1.1	e
4	Siemens	5	100.0	0.0	0.0	3.50	0.0	e
5	Autolyser	18	94.4	5.6	0.0	3.51	3.7	e*
6	Fuji Dri-Chem	1113	98.7	0.3	1.0	3.50	1.8	e
7	Spotchem D-Concept	556	97.7	0.9	1.4	3.20	1.5	e
8	Spotchem EL-SE 1520	78	94.9	1.3	3.8	3.18	2.4	e
9	Piccolo	38	79.0	10.5	10.5	3.70	4.6	e*
10	iStat Chem8	15	100.0	0.0	0.0	3.50	1.1	e
11	Exias	27	100.0	0.0	0.0	3.60	1.0	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)


Creatinine



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	260	2.6	e
2	Beckman	6	100.0	0.0	0.0	261	0.8	e
3	Roche	37	100.0	0.0	0.0	256	2.9	e
4	Siemens	7	85.7	0.0	14.3	263	3.0	e
5	Enzymatic	4	100.0	0.0	0.0	265	3.6	e
6	Autolyser	22	100.0	0.0	0.0	253	3.7	e
7	Selectra Pro	15	100.0	0.0	0.0	254	4.1	e
8	Fuji Dri-Chem	1175	99.1	0.3	0.6	233	4.3	e
9	Spotchem D-Concept	649	99.2	0.2	0.6	127	4.3	e
10	Spotchem SP-4430	143	97.9	0.7	1.4	128	4.1	e
11	Piccolo	71	98.6	0.0	1.4	263	4.3	e
12	Skyla	5	100.0	0.0	0.0	231	6.8	e*
13	EPOC	12	83.3	0.0	16.7	237	6.0	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Creatinine E




QUALAB tolerance : 18 %

Creatinine E (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	iStat Chem8	46	100.0	0.0	0.0	262	2.9	e
2	ABL700/800	15	100.0	0.0	0.0	260	5.6	e

eGFR CKD-EPI




MQ tolerance : 30 %

eGFR CKD-EPI ()

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	57	98.2	1.8	0.0	23	7.3	e
2	Fuji Dri-Chem	393	96.1	0.8	3.1	27	6.8	e
3	Spotchem	262	94.6	2.3	3.1	56	9.6	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

eGFR Cockcroft-Gault




MQ tolerance : 30 %

eGFR Cockcroft-Gault ()

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Fuji Dri-Chem	35	80.0	5.7	14.3	33	19.9	e
2	Spotchem	17	94.1	0.0	5.9	62	11.7	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

LDH




QUALAB tolerance : 18 %

LDH (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	166	6.1	e
2	Beckman	5	100.0	0.0	0.0	173	5.2	e*
3	Roche	36	100.0	0.0	0.0	177	2.3	e
4	Siemens	4	100.0	0.0	0.0	167	5.6	e*
5	Autolyser	7	85.7	14.3	0.0	180	9.8	e*
6	Fuji Dri-Chem	113	97.3	0.9	1.8	146	4.8	e
7	Spotchem D-Concept	42	90.4	4.8	4.8	102	9.5	e
8	Spotchem SP-4430	13	92.3	0.0	7.7	100	9.5	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Magnesium




QUALAB tolerance : 12 %
 (< 0.70: +/- 0.09 mmol/l)

Magnesium (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	14	100.0	0.0	0.0	0.86	2.5	e
2	Beckman	4	100.0	0.0	0.0	0.86	2.5	e
3	Roche	30	96.7	3.3	0.0	0.88	3.6	e
4	Siemens	4	75.0	0.0	25.0	0.88	2.3	e
5	Fuji Dri-Chem	73	93.2	4.1	2.7	0.95	4.7	e
6	Spotchem D-Concept	39	100.0	0.0	0.0	0.70	5.2	e
7	Spotchem SP-4430	4	100.0	0.0	0.0	0.72	3.3	e*

8 additional results were submitted but not published because the method groups were too small. (< results per group)

Sodium




QUALAB tolerance : 6 %

Sodium (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Autolyser	8	100.0	0.0	0.0	144	2.1	e*
2	Abbott	16	93.7	0.0	6.3	139	1.0	e
3	Beckman	6	100.0	0.0	0.0	138	0.8	e
4	Roche	38	100.0	0.0	0.0	139	0.9	e
5	Siemens	5	100.0	0.0	0.0	137	1.4	e
6	Fuji Dri-Chem	1029	98.5	0.9	0.6	142	1.4	e
7	Spotchem D-Concept	481	99.6	0.0	0.4	134	1.1	e
8	Spotchem EL-SE 1520	67	94.0	1.5	4.5	133	2.0	e
9	Piccolo	35	100.0	0.0	0.0	136	1.6	e
10	Exias	28	100.0	0.0	0.0	139	1.0	e
11	iStat Chem8	15	100.0	0.0	0.0	138	0.8	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Phosphate




QUALAB tolerance : 15 %

Phosphate (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	15	100.0	0.0	0.0	1.38	2.3	e
2	Beckman	4	100.0	0.0	0.0	1.41	2.6	e
3	Roche	35	100.0	0.0	0.0	1.38	2.7	e
4	Siemens	4	100.0	0.0	0.0	1.46	1.5	e
5	Fuji Dri-Chem	68	91.1	1.5	7.4	1.45	5.3	e
6	Spotchem D-Concept	14	100.0	0.0	0.0	1.43	4.0	e
7	Piccolo	4	100.0	0.0	0.0	1.73	2.2	e

10 additional results were submitted but not published because the method groups were too small. (< results per group)

Protein total




QUALAB tolerance : 12 %

Protein total (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	14	100.0	0.0	0.0	54.6	2.6	e
2	Beckman	4	100.0	0.0	0.0	54.9	0.8	e
3	Roche	32	100.0	0.0	0.0	54.0	2.6	e
4	Siemens	5	100.0	0.0	0.0	57.0	2.4	e
5	Autolyser	4	100.0	0.0	0.0	54.2	8.1	e*
6	Selectra Pro	8	87.5	0.0	12.5	52.4	4.2	e*
7	Fuji Dri-Chem	183	97.8	0.0	2.2	53.0	3.6	e
8	Spotchem D-Concept	187	92.0	5.3	2.7	55.8	5.9	e
9	Spotchem SP-4430	23	95.7	4.3	0.0	57.7	5.2	e
10	Piccolo	47	100.0	0.0	0.0	54.8	1.6	e
11	Skyla	5	100.0	0.0	0.0	57.0	2.8	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

Aspartate aminotransferase




QUALAB tolerance : 18 %

Aspartate aminotransferase (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Vitros	4	100.0	0.0	0.0	103	10.4	e*
2	Abbott	15	100.0	0.0	0.0	105	4.1	e
3	Beckman	6	100.0	0.0	0.0	106	4.8	e
4	Roche	39	100.0	0.0	0.0	101	3.8	e
5	Siemens	5	100.0	0.0	0.0	100	5.9	e*
6	Autolyser	22	100.0	0.0	0.0	93	4.9	e
7	Selectra Pro	15	93.3	0.0	6.7	103	7.1	e
8	Fuji Dri-Chem	1150	98.7	0.5	0.8	99	4.3	e
9	Spotchem D-Concept	634	99.5	0.0	0.5	70	4.3	e
10	Spotchem SP-4430	135	99.3	0.0	0.7	70	5.4	e
11	Piccolo	73	98.6	1.4	0.0	90	4.0	e
12	Skylla	5	100.0	0.0	0.0	102	4.7	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Alanine aminotransferase




QUALAB tolerance : 18 %

Alanine aminotransferase (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Vitros	5	100.0	0.0	0.0	105	6.0	e*
2	Abbott	13	100.0	0.0	0.0	74	4.5	e
3	Beckman	6	100.0	0.0	0.0	83	4.3	e
4	Roche	39	100.0	0.0	0.0	80	3.3	e
5	Siemens	7	85.7	14.3	0.0	89	12.0	e*
6	Autolyser	22	100.0	0.0	0.0	81	3.7	e
7	Selectra Pro	15	100.0	0.0	0.0	89	6.0	e
8	Fuji Dri-Chem	1169	98.2	0.9	0.9	84	5.5	e
9	Spotchem D-Concept	639	96.9	2.2	0.9	60	7.6	e
10	Spotchem SP-4430	136	99.3	0.0	0.7	69	5.7	e
11	Piccolo	70	97.1	2.9	0.0	73	6.0	e
12	Skylla	5	100.0	0.0	0.0	82	1.8	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Triglycerides




QUALAB tolerance : 18 %

Triglycerides (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	1.98	1.6	e
2	Beckman	5	100.0	0.0	0.0	2.10	3.6	e
3	Roche	31	100.0	0.0	0.0	2.09	2.5	e
4	Siemens	6	100.0	0.0	0.0	2.05	4.2	e
5	Autolyser	23	100.0	0.0	0.0	1.94	4.5	e
6	Selectra Pro	12	100.0	0.0	0.0	2.11	2.8	e
7	Fuji Dri-Chem	963	99.0	0.3	0.7	2.55	4.7	e
8	Spotchem D-Concept	461	98.3	0.2	1.5	1.26	4.5	e
9	Spotchem SP-4430	72	100.0	0.0	0.0	1.23	3.8	e
10	Piccolo	23	100.0	0.0	0.0	2.38	3.5	e
11	Cholestech LDX	256	99.2	0.4	0.4	1.97	4.8	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Lithium




QUALAB tolerance : 15 %

Lithium (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	5	100.0	0.0	0.0	1.29	1.9	e
2	Roche	17	94.1	0.0	5.9	1.32	3.7	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Lactate




QUALAB tolerance : 18 %

Lactate (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	5	100.0	0.0	0.0	3.27	2.1	e
2	Roche	17	100.0	0.0	0.0	3.16	3.2	e
3	Other methods	7	71.4	14.3	14.3	3.15	8.6	e*

HbA1c sample A




QUALAB tolerance : 9 %

HbA1c sample A (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	8.2	2.3	e*
2	Roche, Cobas	21	100.0	0.0	0.0	8.2	2.1	e
3	HPLC	13	100.0	0.0	0.0	8.1	2.4	e
4	Afinion	530	97.6	0.9	1.5	8.0	2.6	e
5	Cobas b101	217	98.6	0.5	0.9	7.6	3.1	e
6	DCA2000/Vantage	139	97.9	1.4	0.7	8.0	3.4	e
7	Celltac chemi	23	87.0	4.3	8.7	7.8	4.9	e
8	NycoCard	5	100.0	0.0	0.0	8.1	4.8	e*
9	Eurolyser	7	100.0	0.0	0.0	8.2	5.3	e*
10	A1c Now	220	87.3	8.6	4.1	7.2	5.3	e
11	AFIAS	128	71.9	20.3	7.8	7.4	7.0	e
12	Others	18	88.9	11.1	0.0	8.1	4.5	c

4 additional results were submitted but not published because the method groups were too small. (< results per group)

HbA1c sample B




QUALAB tolerance : 9 %

HbA1c sample B (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	7.1	2.4	e*
2	Roche, Cobas	18	100.0	0.0	0.0	7.0	3.1	e
3	HPLC	13	100.0	0.0	0.0	6.9	2.4	e
4	Afinion	733	98.8	0.5	0.7	6.7	2.3	e
5	Cobas b101	210	98.5	1.0	0.5	6.6	3.0	e
6	DCA2000/Vantage	189	96.8	1.6	1.6	6.9	3.5	e
7	Celltac chemi	21	100.0	0.0	0.0	6.7	4.1	e
8	NycoCard	4	100.0	0.0	0.0	7.1	3.6	e*
9	Eurolyser	5	80.0	20.0	0.0	6.9	5.5	e*
10	A1c Now	7	57.1	14.3	28.6	6.9	4.9	c
11	AFIAS	191	69.6	21.5	8.9	6.5	6.3	e
12	Others	12	100.0	0.0	0.0	6.9	3.5	c

6 additional results were submitted but not published because the method groups were too small. (< results per group)


pCO2



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Edan	4	100.0	0.0	0.0	3.27	7.9	e*
2	ABL700/800	108	100.0	0.0	0.0	2.85	2.6	e
3	ABL80 FLEX CO-OX / O	4	100.0	0.0	0.0	3.07	6.1	e*
4	ABL90 FLEX / PLUS	129	99.2	0.8	0.0	2.94	2.3	e
5	Cobas b 123	19	94.7	0.0	5.3	3.00	2.3	e
6	Cobas b 221	4	100.0	0.0	0.0	3.04	2.7	e
7	GEM	8	100.0	0.0	0.0	2.80	1.3	e
8	iStat	51	98.0	0.0	2.0	2.50	4.3	e
9	EPOC	55	89.1	9.1	1.8	2.44	6.2	e
10	IL	4	100.0	0.0	0.0	2.80	2.9	e*

5 additional results were submitted but not published because the method groups were too small. (< results per group)

pO2




No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Edan	4	100.0	0.0	0.0	12.50	12.6	a
2	ABL700/800	107	92.6	0.9	6.5	9.93	5.0	e
3	ABL80 FLEX CO-OX / O	4	100.0	0.0	0.0	8.07	6.8	e*
4	ABL90 FLEX / PLUS	131	90.1	1.5	8.4	7.47	4.5	e
5	Cobas b 123	20	90.0	5.0	5.0	9.19	7.6	e
6	GEM	8	87.5	12.5	0.0	9.52	8.7	e*
7	iStat	48	85.4	8.3	6.3	11.57	8.5	e
8	EPOC	55	63.7	21.8	14.5	9.93	11.5	e
9	IL	4	100.0	0.0	0.0	9.55	5.2	e*

6 additional results were submitted but not published because the method groups were too small. (< results per group)

K04 Blood gases

pH




QUALAB tolerance : 1 %

pH ()

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	106	100.0	0.0	0.0	7.57	0.1	e
2	ABL80 FLEX CO-OX / O	4	100.0	0.0	0.0	7.59	0.3	e*
3	ABL90 FLEX / PLUS	131	100.0	0.0	0.0	7.59	0.1	e
4	Cobas b 123	20	100.0	0.0	0.0	7.58	0.1	e
5	Cobas b 221	4	100.0	0.0	0.0	7.57	0.2	e
6	GEM	8	100.0	0.0	0.0	7.63	0.1	e
7	iStat	53	100.0	0.0	0.0	7.65	0.1	e
8	EPOC	55	100.0	0.0	0.0	7.66	0.1	e
9	IL	4	100.0	0.0	0.0	7.64	0.1	e

8 additional results were submitted but not published because the method groups were too small. (< results per group)

Glucose BG




QUALAB tolerance : 9 %

Glucose BG (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	GEM	5	100.0	0.0	0.0	14.6	0.8	e
2	Cobas b 123	12	100.0	0.0	0.0	14.6	2.0	e
3	iStat	10	100.0	0.0	0.0	13.4	1.2	e
4	EPOC	44	95.5	4.5	0.0	14.1	4.3	e
5	ABL700/800	98	99.0	0.0	1.0	14.0	2.4	e
6	ABL90 FLEX / PLUS	119	99.2	0.8	0.0	13.6	2.1	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

Hemoglobin BG




QUALAB tolerance : 9 %

Hemoglobin BG (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	100	100.0	0.0	0.0	189.9	1.6	e
2	ABL90 FLEX / PLUS	119	100.0	0.0	0.0	190.0	0.4	e
3	ABL80 FLEX CO-OX / O	5	100.0	0.0	0.0	186.0	1.7	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Potassium BG




QUALAB tolerance : 6 %

Potassium BG (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	GEM	6	100.0	0.0	0.0	5.7	1.0	e
2	Cobas b 123	21	100.0	0.0	0.0	5.4	0.6	e
3	iStat	18	100.0	0.0	0.0	5.3	0.9	e
4	EPOC	45	97.8	2.2	0.0	4.8	2.5	e
5	ABL700/800	99	100.0	0.0	0.0	5.5	1.2	e
6	ABL90 FLEX / PLUS	123	98.4	0.0	1.6	5.4	0.7	e

9 additional results were submitted but not published because the method groups were too small. (< results per group)

Sodium BG




QUALAB tolerance : 6 %

Sodium BG (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	GEM	7	100.0	0.0	0.0	122.0	0.4	e
2	Cobas b 123	21	100.0	0.0	0.0	122.0	1.3	e
3	iStat	18	100.0	0.0	0.0	124.1	0.5	e
4	EPOC	43	100.0	0.0	0.0	120.3	1.0	e
5	ABL700/800	98	100.0	0.0	0.0	126.7	0.8	e
6	ABL90 FLEX / PLUS	122	100.0	0.0	0.0	126.0	0.3	e

8 additional results were submitted but not published because the method groups were too small. (< results per group)

Chlorid-BG




QUALAB tolerance : 6 %

Chlorid-BG (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas b 123	14	100.0	0.0	0.0	70.3	0.9	e
2	EPOC	19	84.2	15.8	0.0	73.5	3.3	e*
3	ABL700/800	93	95.7	3.2	1.1	67.7	2.9	e
4	ABL90 FLEX / PLUS	118	100.0	0.0	0.0	64.4	1.4	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Calcium-BG




MQ tolerance : 12 %
 (< 2.00: +/- 0.24 mmol/l)

Calcium-BG (mmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 GEM	6	100.0	0.0	0.0	0.28	2.3	e
2 Cobas b123	12	100.0	0.0	0.0	0.18	31.6	e*
3 Roche, Cobas	4	100.0	0.0	0.0	0.22	35.1	e*
4 iStat	15	100.0	0.0	0.0	0.31	1.7	e
5 EPOC	41	97.6	0.0	2.4	0.29	5.1	e
6 ABL700/800	98	100.0	0.0	0.0	0.37	4.5	e
7 ABL90 FLEX / PLUS	120	100.0	0.0	0.0	0.38	1.5	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

FHHb




MQ tolerance : 20 %

FHHb (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL90 FLEX / PLUS	18	100.0	0.0	0.0	20.908	0.2	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Lactate-BG




QUALAB tolerance : 18 %

Lactate-BG (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	GEM	6	100.0	0.0	0.0	11.80	1.3	e
2	Cobas b123	13	100.0	0.0	0.0	10.65	6.7	e
3	IL	4	100.0	0.0	0.0	11.30	4.9	e*
4	EPOC	43	95.3	0.0	4.7	9.84	5.4	e
5	iStat	19	100.0	0.0	0.0	10.71	1.8	e
6	ABL700/800	102	100.0	0.0	0.0	10.31	4.6	e
7	ABL90 FLEX / PLUS	126	100.0	0.0	0.0	10.50	1.8	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

sO2 OR




MQ tolerance : 20 %

sO2 OR (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	iStat	26	100.0	0.0	0.0	98.577	0.6	e
2	ABL700/800	89	100.0	0.0	0.0	70.103	0.1	e
3	ABL90 FLEX / PLUS	109	100.0	0.0	0.0	70.100	0.1	e
4	ABL80 FLEX CO-OX / O	4	75.0	0.0	25.0	70.100	0.2	e

FO2Hb OR




MQ tolerance : 20 %

FO2Hb OR (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	87	100.0	0.0	0.0	48.855	0.3	e
2	ABL90 FLEX / PLUS	110	100.0	0.0	0.0	48.936	0.2	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

FCOHb OR




MQ tolerance : 20 %

FCOHb OR (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	88	100.0	0.0	0.0	20.301	1.1	e
2	ABL90 FLEX / PLUS	110	100.0	0.0	0.0	20.183	0.7	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

FMetHb OR




MQ tolerance : 20 %

FMetHb OR (%)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 ABL700/800	88	98.9	0.0	1.1	10.009	0.9	e
2 ABL90 FLEX / PLUS	110	100.0	0.0	0.0	9.999	1.3	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

FHbF OR




MQ tolerance : 20 %

FHbF OR (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL90 FLEX / PLUS	37	100.0	0.0	0.0	52.243	2.6	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Bilirubin OR




QUALAB tolerance : 18 %

Bilirubin OR (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	5	100.0	0.0	0.0	439.0	2.6	e
2	ABL90 FLEX / PLUS	34	100.0	0.0	0.0	438.6	0.5	e

Troponin I




QUALAB tolerance : 24 %

Troponin I (ng/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Pathfast	24	91.7	0.0	8.3	3021.8	9.3	e
2	Siemens	4	100.0	0.0	0.0	3117.8	2.6	e
3	Vidas	12	100.0	0.0	0.0	6036.8	8.2	e
4	Abbott	8	100.0	0.0	0.0	855.7	5.0	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Troponin T




QUALAB tolerance : 24 %

Troponin T (ng/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas hs	10	100.0	0.0	0.0	569.46	6.9	e
2	Cobas hs STAT	20	100.0	0.0	0.0	498.82	7.7	e

Myoglobin




QUALAB tolerance : 30 %

Myoglobin (µg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas E / Elecsys	8	100.0	0.0	0.0	72.3	4.8	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

CK-MB mass




MQ tolerance : 40 %

CK-MB mass (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	13.4	4.6	e
2	Cobas E / Elecsys	10	100.0	0.0	0.0	15.3	3.8	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

BNP




QUALAB tolerance : 27 %

BNP (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	5	80.0	0.0	20.0	356.7	14.2	a

2 additional results were submitted but not published because the method groups were too small. (< results per group)

NT-proBNP




QUALAB tolerance : 27 %

NT-proBNP (ng/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Pathfast	17	100.0	0.0	0.0	1557.1	11.4	e
2	VIDAS	9	100.0	0.0	0.0	543.0	7.4	e
3	Cobas E / Elecsys	25	100.0	0.0	0.0	541.6	5.5	e
4	Abbott	8	100.0	0.0	0.0	594.9	6.3	e

10 additional results were submitted but not published because the method groups were too small. (< results per group)


TSH



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	32	100.0	0.0	0.0	3.66	3.3	e
2	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	3.32	6.6	e*
3	Abbott	11	100.0	0.0	0.0	2.86	4.2	e
4	VIDAS	16	100.0	0.0	0.0	3.94	8.6	a
5	Dimension	6	100.0	0.0	0.0	3.13	6.0	e*
6	AFIAS	18	83.3	16.7	0.0	5.41	11.3	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

T3




MQ tolerance : 20 %

T3 (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	2.0	5.3	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

T4




MQ tolerance : 20 %

T4 (nmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	5	100.0	0.0	0.0	126	6.1	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

FT3




QUALAB tolerance : 18 %

FT3 (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	28	100.0	0.0	0.0	6.6	2.8	e
2	ADVIA Centaur XP/CP	5	100.0	0.0	0.0	5.6	1.5	e
3	Abbott	9	100.0	0.0	0.0	5.7	3.8	e
4	VIDAS	8	100.0	0.0	0.0	5.0	7.3	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

FT4




QUALAB tolerance : 20 %

FT4 (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	28	100.0	0.0	0.0	23.6	2.9	e
2	ADVIA Centaur XP	8	100.0	0.0	0.0	19.2	5.0	e
3	Abbott	10	100.0	0.0	0.0	17.3	4.9	e
4	VIDAS	9	100.0	0.0	0.0	20.6	4.3	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Testosterone




QUALAB tolerance : 30 %

Testosterone (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	15	100.0	0.0	0.0	9.3	4.1	e
2	Siemens	7	100.0	0.0	0.0	8.6	11.1	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Estradiol




QUALAB tolerance : 30 %

Estradiol (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	14	100.0	0.0	0.0	765	3.3	e
2	Siemens	6	100.0	0.0	0.0	1011	11.8	e*
3	Other methods	5	100.0	0.0	0.0	640	15.3	e*
4	Abbott	4	100.0	0.0	0.0	672	3.1	e

SHBG




MQ tolerance : 30 %

SHBG (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	17	100.0	0.0	0.0	26.8	3.9	e
2	Abbott	5	100.0	0.0	0.0	31.0	15.7	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Cortisol




QUALAB tolerance : 20 %

Cortisol (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	22	100.0	0.0	0.0	349	2.9	e
2	Abbott	5	100.0	0.0	0.0	343	2.8	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Progesteron




MQ tolerance : 30 %

Progesteron (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	9	100.0	0.0	0.0	21.1	2.8	e
2	Abbott	5	100.0	0.0	0.0	16.9	7.8	e
3	Other methods	5	100.0	0.0	0.0	20.8	16.9	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

DHEAS




MQ tolerance : 30 %

DHEAS (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	12	100.0	0.0	0.0	4.56	3.9	e
2	Abbott	4	100.0	0.0	0.0	4.04	13.1	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Luteinizing hormone




QUALAB tolerance : 24 %

Luteinizing hormone (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	15	100.0	0.0	0.0	14.8	5.8	e
2	Siemens	5	100.0	0.0	0.0	12.3	5.2	e
3	Abbott	4	100.0	0.0	0.0	9.7	3.4	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Follicle-stimulating hormone




QUALAB tolerance : 24 %

Follicle-stimulating hormone (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	14	100.0	0.0	0.0	11.8	2.9	e
2	Abbott	7	100.0	0.0	0.0	11.5	2.6	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Prolactin (PRL)




QUALAB tolerance : 24 %

Prolactin (PRL) (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	8.9	5.5	e
2	Cobas/Roche	15	100.0	0.0	0.0	15.2	3.4	e
3	Abbott	5	100.0	0.0	0.0	13.5	9.9	e*

Insulin




MQ tolerance : 25 %

Insulin (pmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	16	100.0	0.0	0.0	318	5.5	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

HGH




MQ tolerance : 25 %

HGH (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	7	100.0	0.0	0.0	8.09	3.2	e
2 Liaison	4	100.0	0.0	0.0	7.43	7.7	e*

Freies Testosteron




MQ tolerance : 25 %

Freies Testosteron (pmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	IDS	6	100.0	0.0	0.0	31.4	10.7	e*


2 additional results were submitted but not published because the method groups were too small. (< results per group)

Androstendion



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	5	100.0	0.0	0.0	3.4	11.9	e*

IGF-1




MQ tolerance : 25 %

IGF-1 (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Liaison	8	100.0	0.0	0.0	88	12.5	e*
2 Other methods	4	100.0	0.0	0.0	63	0.8	e

17-OH-Progesteron




MQ tolerance : 30 %

17-OH-Progesteron (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	4	100.0	0.0	0.0	3.9	19.0	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Troponin T CR




QUALAB tolerance : 24 %

Troponin T CR (ng/l)


No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 Cobas h 232	772	88.4	8.0	3.6	141.92	13.7 e

D-dimer CR



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas h 232	896	87.0	9.5	3.5	0.42	13.6	e
2	Lumira Dx	7	85.7	0.0	14.3	0.46	6.7	e*

CKMB- K8




MQ tolerance : 40 %

CKMB- K8 (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 Cobas h 232	6	100.0	0.0	0.0	15.8	12.0 e*

NT-proBNP CR




QUALAB tolerance : 27 %

NT-proBNP CR (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Lumira Dx	5	100.0	0.0	0.0	374	7.1	e
2	Cobas h 232	542	90.2	4.8	5.0	103	12.0	e

PCO2 CCA




QUALAB tolerance : 12 %

PCO2 CCA (kPa)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 OPTI CCA	11	81.8	0.0	18.2	9.33	3.1	e

PO2 CCA



QUALAB tolerance : 15 %

PO2 CCA (kPa)


No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 OPTI CCA	11	81.8	9.1	9.1	9.48	7.3	e*

pH CCA



No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 OPTI CCA	10	90.0	10.0	0.0	7.15	0.4	e*

Potassium CCA




QUALAB tolerance : 6 %
(< 3.3: +/- 0.2 mmol/l)

Potassium CCA (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	OPTI CCA	5	100.0	0.0	0.0	2.9	1.7	e*

Sodium CCA




QUALAB tolerance : 6 %

Sodium CCA (mmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 OPTI CCA	4	100.0	0.0	0.0	125.4	1.0	e


Ferritin



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Dimension	5	100.0	0.0	0.0	38.00	5.2	e
2	Beckman	5	80.0	20.0	0.0	34.40	16.6	e*
3	Siemens	4	100.0	0.0	0.0	25.10	2.5	e
4	Roche, Cobas	30	100.0	0.0	0.0	37.36	6.3	e
5	Abbott	14	92.9	7.1	0.0	36.06	12.3	e*
6	Mini Vidas	9	88.9	11.1	0.0	26.10	14.1	e*
7	AFIAS	25	100.0	0.0	0.0	28.08	5.4	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Vitamin B12




QUALAB tolerance : 21 %
 (< 200.00: +/- 42.00 pmol/l)

Vitamin B12 (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	5	100.0	0.0	0.0	184.00	9.9	e*
2	Roche, Cobas	22	100.0	0.0	0.0	202.31	6.7	e
3	Abbott	10	90.0	0.0	10.0	192.67	11.3	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Folate




QUALAB tolerance : 24 %
 (< 10.00: +/- 2.40 nmol/l)

Folate (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens	8	100.0	0.0	0.0	4.00	22.0	e*
2	Roche, Cobas	24	100.0	0.0	0.0	4.11	21.6	e*
3	Abbott	10	90.0	0.0	10.0	7.58	10.5	e*

Holotranscobalamine




MQ tolerance : 30 %

Holotranscobalamine (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	35	100.0	0.0	0.0	41.3	3.5	e
2	Abbott	16	100.0	0.0	0.0	51.6	5.2	e
3	Cobas Biotin suppres	4	100.0	0.0	0.0	25.5	3.6	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Bilirubin total Neo




QUALAB tolerance : 18 %

Bilirubin total Neo (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	16	100.0	0.0	0.0	227	4.4	e
2	Dimension	6	100.0	0.0	0.0	273	3.6	e

Bilirubin direct




QUALAB tolerance : 18 %

Bilirubin direct (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	19	94.7	0.0	5.3	144	4.0	e
2	Dimension	4	100.0	0.0	0.0	136	1.4	e

Bilirubin neonatal




QUALAB tolerance : 18 %

Bilirubin neonatal (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ABL700/800	14	100.0	0.0	0.0	274	3.9	e
2	Other methods	12	100.0	0.0	0.0	293	2.9	e

PSA




QUALAB tolerance : 21 %

PSA (µg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	19	100.0	0.0	0.0	7.48	4.9	e
2	Abbott	7	100.0	0.0	0.0	7.02	4.8	e
3	AFIAS	10	100.0	0.0	0.0	7.42	7.2	e

9 additional results were submitted but not published because the method groups were too small. (< results per group)

free PSA



QUALAB tolerance : 21 %


free PSA (µg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	12	100.0	0.0	0.0	1.59	3.6	e
2	Siemens	4	100.0	0.0	0.0	1.74	7.2	e*
3	Abbott	6	100.0	0.0	0.0	1.67	8.2	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

K14 Tumor Markers

CEA




QUALAB tolerance : 21 %

CEA (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	14	100.0	0.0	0.0	5.5	4.4	e
2	Abbott	6	100.0	0.0	0.0	7.3	7.1	e*

7 additional results were submitted but not published because the method groups were too small. (< results per group)

CA 125




MQ tolerance : 25 %

CA 125 (kIU/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	10	100.0	0.0	0.0	57.0	4.3	e
2	Abbott	6	100.0	0.0	0.0	92.2	2.9	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

CA 19-9




MQ tolerance : 25 %

CA 19-9 (kIU/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	9	100.0	0.0	0.0	32.3	4.1	e

11 additional results were submitted but not published because the method groups were too small. (< results per group)

CA 15-3




MQ tolerance : 25 %

CA 15-3 (kIU/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	12	100.0	0.0	0.0	21.0	6.6	e
2	Siemens	4	100.0	0.0	0.0	21.5	5.3	e
3	Abbott	6	100.0	0.0	0.0	24.3	4.5	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

AFP




QUALAB tolerance : 21 %

AFP (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	8	100.0	0.0	0.0	9.7	9.6	e*
2	Abbott	4	100.0	0.0	0.0	8.4	3.4	e

7 additional results were submitted but not published because the method groups were too small. (< results per group)

HCG qn




QUALAB tolerance : 21 %

HCG qn (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	17	100.0	0.0	0.0	33.2	4.9	e
2	VIDAS	11	100.0	0.0	0.0	10.8	10.1	e*
3	Abbott	6	100.0	0.0	0.0	35.3	8.5	d
4	AFIAS	7	100.0	0.0	0.0	55.0	13.5	e*

8 additional results were submitted but not published because the method groups were too small. (< results per group)

HCG intact




QUALAB tolerance : 25 %

HCG intact (U/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Cobas	5	100.0	0.0	0.0	9.7	5.1	e

S100




MQ tolerance : 25 %

S100 (µg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	6	100.0	0.0	0.0	0.56	6.3	e

NSE




MQ tolerance : 25 %

NSE (ng/ml)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	5	100.0	0.0	0.0	6.6	6.6	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

Thyreoglobulin




MQ tolerance : 25 %

Thyreoglobulin (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	5	100.0	0.0	0.0	52.0	6.5	e
2	Other methods	4	100.0	0.0	0.0	87.3	12.5	e*

CK-MB




MQ tolerance : 30 %

CK-MB (U/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Fuji Dri-Chem	19	100.0	0.0	0.0	67.5	6.1	e
2	Cobas/Roche	7	85.7	0.0	14.3	41.0	4.2	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

C-Peptid




MQ tolerance : 25 %

C-Peptid (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	7	85.7	0.0	14.3	3.94	7.1	e
2	Other methods	5	100.0	0.0	0.0	3.43	9.8	e*

ACTH




MQ tolerance : 25 %

ACTH (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	8	100.0	0.0	0.0	25.39	13.7	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Procalcitonin



QUALAB tolerance : 27 %

Procalcitonin (µg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	4	100.0	0.0	0.0	13.46	4.0	e
2	Cobas	19	100.0	0.0	0.0	10.86	5.2	e
3	VIDAS	15	100.0	0.0	0.0	8.18	5.2	e
4	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	14.19	4.9	e
5	Other methods	4	100.0	0.0	0.0	24.50	20.9	a

6 additional results were submitted but not published because the method groups were too small. (< results per group)

EPO




MQ tolerance : 25 %

EPO (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	5	100.0	0.0	0.0	218.9	10.7	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Parathyroid hormone




QUALAB tolerance : 24 %

Parathyroid hormone (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas PTH STAT	9	100.0	0.0	0.0	35.4	5.2	e
2	Cobas	15	100.0	0.0	0.0	28.6	5.5	e
3	Abbott	6	100.0	0.0	0.0	63.4	6.7	e
4	ADVIA Centaur XP/CP	4	100.0	0.0	0.0	65.6	5.9	e*

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Osmolality




QUALAB tolerance : 6 %

Osmolality (mosm/kg)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Cryoskopy	22	100.0	0.0	0.0	349	1.4	e

Potassium-K22




QUALAB tolerance : 6 %

Potassium-K22 (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ISE	11	100.0	0.0	0.0	3.6	1.9	e

Sodium-K22




QUALAB tolerance : 6 %

Sodium-K22 (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	ISE	11	100.0	0.0	0.0	143	0.9	e

Glucose-K22




QUALAB tolerance : 9 %

Glucose-K22 (mmol/l)


No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	11	100.0	0.0	0.0	4.4	3.4	e

Urea-K22



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	11	100.0	0.0	0.0	4.2	7.2	e*

Osmotic Gap




MQ tolerance : 40 %

Osmotic Gap (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Formel 1 (2Na+K+Glu+	8	100.0	0.0	0.0	55.3	9.6	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Digoxin




QUALAB tolerance : 24 %

Digoxin (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Alinity	4	100.0	0.0	0.0	1.38	11.3	e*
2	Other methods	8	100.0	0.0	0.0	1.68	9.7	e*

Phenytoin




MQ tolerance : 25 %

Phenytoin (µmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	4	100.0	0.0	0.0	34	6.2	e*

Phenobarbital




MQ tolerance : 25 %

Phenobarbital (µmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	4	100.0	0.0	0.0	86	6.5	e*

Paracetamol




MQ tolerance : 25 %

Paracetamol (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	4	100.0	0.0	0.0	190.0	5.7	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

Vancomycin




MQ tolerance : 25 %

Vancomycin (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	4	100.0	0.0	0.0	8.7	3.5	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Valproat




MQ tolerance : 24 %

Valproat (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	9	100.0	0.0	0.0	407.3	10.8	a

Carbamazepin




MQ tolerance : 24 %

Carbamazepin (µmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	4	100.0	0.0	0.0	28.8	7.9	e*

Cystatin C




MQ tolerance : 24 %

Cystatin C (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	17	88.2	5.9	5.9	0.76	9.7	e
2	Nephelometry	5	100.0	0.0	0.0	0.74	8.1	e*

Ethanol




QUALAB tolerance : 18 %
 (< 10.0: +/- 1.8 mmol/l)

Ethanol (mmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Others	4	100.0	0.0	0.0	9.0	3.0	e
2 Abbott	8	100.0	0.0	0.0	8.4	8.3	e*
3 Roche, Cobas	24	100.0	0.0	0.0	8.3	9.5	a

Ammonia




QUALAB tolerance : 21 %

Ammonia (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	6	100.0	0.0	0.0	56.3	9.1	e*

6 additional results were submitted but not published because the method groups were too small. (< results per group)

Calprotectin




MQ tolerance : 40 %
 (< 50: +/- 20 µg/g)

Calprotectin (µg/g)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Bühlmann fCALturbo	21	90.5	0.0	9.5	238	13.7	e
2	Bühlmann Quantum Blu	5	80.0	0.0	20.0	82	26.4	a
3	Liaison	12	83.3	0.0	16.7	41	14.1	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Cholesterol total Af/b101




QUALAB tolerance : 10 %

Cholesterol total Af/b101 (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas b101	330	98.8	0.3	0.9	3.62	2.4	e
2	Afinion	451	97.7	0.7	1.6	3.58	3.6	e

Cholesterol HDL Af/b101




QUALAB tolerance : 21 %

Cholesterol HDL Af/b101 (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas b101	327	92.4	0.3	7.3	0.72	6.2	e
2	Afinion	450	92.5	0.2	7.3	0.74	4.5	e

Tryglycerides Af/b101




QUALAB tolerance : 18 %

Tryglycerides Af/b101 (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas b101	327	99.1	0.0	0.9	1.49	3.5	e
2	Afinion	452	99.4	0.2	0.4	1.50	4.6	e

Cholesterol PTS




QUALAB tolerance : 10 %

Cholesterol PTS (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CardioChek	4	75.0	25.0	0.0	3.17	7.1	e*

Cholesterol HDL PTS




QUALAB tolerance : 21 %

Cholesterol HDL PTS (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CardioChek	4	100.0	0.0	0.0	2.16	12.0	e*

Triglycerides




QUALAB tolerance : 18 %

Triglycerides (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	CardioChek	4	75.0	0.0	25.0	1.41	4.0	e

Troponin I AFIAS




QUALAB tolerance : 24 %

Troponin I AFIAS (ng/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 AFIAS	498	87.4	10.0	2.6	617.89	13.9 e

NT-proBNP AFIAS




QUALAB tolerance : 27 %

NT-proBNP AFIAS (ng/l)


No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 AFIAS	384	93.5	4.2	2.3	557.2	12.2	e

Homocystein



No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Beckman	7	100.0	0.0	0.0	20.5	2.4	e
2	all Participants	6	100.0	0.0	0.0	17.3	3.9	e

Lipase




QUALAB tolerance : 18 %

Lipase (U/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche	5	100.0	0.0	0.0	94.0	2.4	e
2	Siemens	6	100.0	0.0	0.0	102.5	4.4	e
3	Abbott	13	100.0	0.0	0.0	88.6	4.6	e
4	Beckman	5	100.0	0.0	0.0	90.0	1.9	e
5	Roche	27	100.0	0.0	0.0	91.6	5.2	e
6	Fuji Dri-Chem	173	98.3	0.0	1.7	79.3	4.6	e

4 additional results were submitted but not published because the method groups were too small. (< results per group)


Fructosamine



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	4	100.0	0.0	0.0	514	5.2	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Bicarbonat




MQ tolerance : 15 %

Bicarbonat (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	8	87.5	12.5	0.0	21.0	8.2	e*
2	Piccolo	4	100.0	0.0	0.0	26.5	5.2	e*

4 additional results were submitted but not published because the method groups were too small. (< results per group)

Glucose CSF




QUALAB tolerance : 9 %
(< 3.30: +/- 0.30 mmol/l)

Glucose CSF (mmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Roche, Cobas	17	100.0	0.0	0.0	1.88	2.7	e
2 Other methods	13	100.0	0.0	0.0	1.86	3.3	e

Lactate CSF




QUALAB tolerance : 18 %

Lactate CSF (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	15	93.3	0.0	6.7	3.97	2.6	e
2	Other methods	9	88.9	0.0	11.1	4.04	1.9	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Protein CSF




QUALAB tolerance : 15 %

Protein CSF (g/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	20	100.0	0.0	0.0	0.93	2.5	e
2	Other methods	7	100.0	0.0	0.0	0.90	6.5	e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Albumine CSF




MQ tolerance : 20 %

Albumine CSF (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	7	85.7	14.3	0.0	340.00	10.9	e*


3 additional results were submitted but not published because the method groups were too small. (< results per group)

CDT



No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Immunoassay	4	100.0	0.0	0.0	1.61	7.0	e*
2	Other methods	6	100.0	0.0	0.0	1.35	14.4	a

Tacrolimus




MQ tolerance : 25 %

Tacrolimus (µg/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	12	100.0	0.0	0.0	16.2	5.3	e

Totalprotein E




MQ tolerance : 12 %

Totalprotein E (g/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	20	95.0	0.0	5.0	70.6	2.4	e

Albumin E




MQ tolerance : 12 %

Albumin E (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	33	100.0	0.0	0.0	62.9	5.7	e

alpha-1-Globuline




MQ tolerance : 30 %

alpha-1-Globuline (%)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	13	100.0	0.0	0.0	2.6	9.1	e
2	capillary electropho	20	100.0	0.0	0.0	4.3	4.3	e

alpha-2-Globuline




MQ tolerance : 30 %

alpha-2-Globuline (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	33	81.8	15.2	3.0	8.6	16.3	e

beta-Globuline




MQ tolerance : 30 %

beta-Globuline (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	10	100.0	0.0	0.0	8.0	8.0	e

Beta-1-Globulin




MQ tolerance : 30 %

Beta-1-Globulin (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	13	76.9	0.0	23.1	7.3	8.9	a

Beta-2-Globulin




MQ tolerance : 30 %

Beta-2-Globulin (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	12	75.0	8.3	16.7	4.6	19.1	e*

gamma-Globuline




MQ tolerance : 30 %

gamma-Globuline (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	20	100.0	0.0	0.0	12.6	11.9	a

Gamma-Globuline+P




MQ tolerance : 30 %

Gamma-Globuline+P (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	13	100.0	0.0	0.0	17.1	15.5	a

Beta-Globuline+P




MQ tolerance : 30 %

Beta-Globuline+P (%)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Electrophoresis	10	100.0	0.0	0.0	11.5	8.6	e

Folate in Erythrocytes




MQ tolerance : 40 %

Folate in Erythrocytes (nmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Siemens	5	80.0	20.0	0.0	580	30.4	a
2	Abbott	11	100.0	0.0	0.0	184	11.1	e
3	Roche, Cobas	26	96.2	0.0	3.8	1212	12.1	e

Gallensäure




MQ tolerance : 30 %

Gallensäure (µmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	7	100.0	0.0	0.0	15.2	12.3	e*
2	Roche	12	100.0	0.0	0.0	17.5	7.0	e
3	Abbott	6	100.0	0.0	0.0	14.2	5.2	e
4	Siemens	6	100.0	0.0	0.0	13.9	4.9	e

BNP




QUALAB tolerance : 27 %
 (< 75.0: +/- 20.0 ng/l)

BNP (ng/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Triage	5	100.0	0.0	0.0	7.0	6.6	e

Troponin I Triage




QUALAB tolerance : 24 %

Troponin I Triage (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Triage high sensitiv	286	95.9	2.4	1.7	510.50	11.5	e


One result was submitted but not published because the method group was too small. (< 4 results per group)

NT-pro BNP



No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Triage	156	91.6	2.6	5.8	180	12.9	e

Vitamin D 25 (OH)




QUALAB tolerance : 27 %

Vitamin D 25 (OH) (nmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 LCMS	4	100.0	0.0	0.0	77.5	3.4	a
2 AFIAS	8	75.0	25.0	0.0	49.3	22.7	e*
3 Cobas	22	100.0	0.0	0.0	64.0	9.1	e
4 VIDAS	5	100.0	0.0	0.0	73.1	7.5	e*
5 Other methods	5	80.0	20.0	0.0	47.1	20.4	e*
6 Abbott	9	100.0	0.0	0.0	69.6	5.1	e
7 ADVIA Centaur XP/CP	5	100.0	0.0	0.0	59.3	8.2	e*

Vitamin D 1,25-(OH)2




MQ tolerance : 27 %

Vitamin D 1,25-(OH)2 (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Other methods	4	100.0	0.0	0.0	70.2	3.8	e

AMH




MQ tolerance : 25 %

AMH (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	17	100.0	0.0	0.0	16.6	3.9	e
2	Beckman	5	100.0	0.0	0.0	18.7	9.9	a

Inhibin B




MQ tolerance : 25 %

Inhibin B (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	7	100.0	0.0	0.0	220.0	28.3	a

Calcitonin




MQ tolerance : 25 %

Calcitonin (pmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Liaison	5	100.0	0.0	0.0	2.1	19.0	a
2	Other methods	7	100.0	0.0	0.0	5.4	7.5	e

IGF-BP3




MQ tolerance : 25 %

IGF-BP3 (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cobas	7	85.7	0.0	14.3	3.30	5.4	e

One result was submitted but not published because the method group was too small. (< 4 results per group)


Renin



No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 Liaison	9	100.0	0.0	0.0	90.4	12.2 e*

One result was submitted but not published because the method group was too small. (< 4 results per group)

Aldosteron




MQ tolerance : 30 %

Aldosteron (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Liaison	6	100.0	0.0	0.0	67.7	13.2	e*

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti Thyreoglobulin




MQ tolerance : 25 %
(< 10: +/- 3 IU/ml)

Anti Thyreoglobulin (IU/ml)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Alinity	6	100.0	0.0	0.0	10	9.2	e*
2 Roche, Cobas	13	100.0	0.0	0.0	95	5.0	e

5 additional results were submitted but not published because the method groups were too small. (< results per group)

Anti TPO




QUALAB tolerance : 25 %

Anti TPO (IU/ml)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	14	92.9	7.1	0.0	42	13.5	e*
2	Abbott	5	100.0	0.0	0.0	70	3.5	e

6 additional results were submitted but not published because the method groups were too small. (< results per group)

TRAK




MQ tolerance : 25 %
 (< 2.00: +/- 0.50 IU/l)

TRAK (IU/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Abbott	6	100.0	0.0	0.0	0.70	0.0	e
2 Roche, Cobas	9	100.0	0.0	0.0	0.80	12.0	e*

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Creatinine WB




QUALAB tolerance : 18 %

Creatinine WB (µmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Statsensor i / Nova	82	81.7	9.8	8.5	274	11.5	e

eGFR CDK-EPI WB




MQ tolerance : 30 %

eGFR CDK-EPI WB ()

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Statsensor i / Nova	4	100.0	0.0	0.0	24	12.9	e*

IL6




MQ tolerance : 30 %

IL6 (ng/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	4	100.0	0.0	0.0	217.0	4.2	e

Pankreas Elastase




MQ tolerance : 40 %

Pankreas Elastase (ug/g)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 Liaison	15	93.3	0.0	6.7	129	18.4	e

Copeptin




MQ tolerance : 30 %

Copeptin (pmol/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV% Type
1 Kryptor	7	100.0	0.0	0.0	4.9	12.5 e*

Occult blood qn




MQ tolerance : 30 %

Occult blood qn (ng/ml)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	OC-Sensor	8	100.0	0.0	0.0	167	34.9	a

Amylase-Urine




MQ tolerance : 25 %

Amylase-Urine (U/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 IFCC	7	100.0	0.0	0.0	183	4.7	e

Pancreatic Amylase-Urine




QUALAB tolerance : 18 %
(< 25.0: +/- 5.0 U/l)

Pancreatic Amylase-Urine (U/l)

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 IFCC	4	100.0	0.0	0.0	5.8	12.0	e*

Calcium-Urine




MQ tolerance : 9 %

Calcium-Urine (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	19	100.0	0.0	0.0	2.64	2.8	e
2	Abbott	11	100.0	0.0	0.0	2.55	2.8	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Chloride-Urine




MQ tolerance : 12 %

Chloride-Urine (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	11	100.0	0.0	0.0	185	1.5	e
2	Roche, Cobas	16	100.0	0.0	0.0	183	1.9	e

2 additional results were submitted but not published because the method groups were too small. (< results per group)

Glucose-Urine




QUALAB tolerance : 9 %

Glucose-Urine (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	27	100.0	0.0	0.0	16.6	2.2	e

Magnesium-Urine




MQ tolerance : 12 %

Magnesium-Urine (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	17	100.0	0.0	0.0	3.57	3.9	e
2	AAS	5	100.0	0.0	0.0	3.70	2.6	e

Osmolality-Urine




MQ tolerance : 10 %

Osmolality-Urine (mosm/kg)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Cryoskopy	21	100.0	0.0	0.0	825	1.3	e

Phosphate-Urine




MQ tolerance : 15 %

Phosphate-Urine (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	29	96.6	3.4	0.0	16.3	5.1	e

Potassium-Urine




MQ tolerance : 15 %

Potassium-Urine (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	39	100.0	0.0	0.0	72	2.7	e

Protein-Urine




QUALAB tolerance : 15 %

Protein-Urine (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Roche, Cobas	24	100.0	0.0	0.0	524.6	2.7	e
2	Dimension	4	100.0	0.0	0.0	717.0	4.9	e*
3	Abbott	16	100.0	0.0	0.0	646.6	4.6	e

Sodium-Urine




MQ tolerance : 15 %

Sodium-Urine (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	39	100.0	0.0	0.0	174	2.6	e

Urea-Urine




MQ tolerance : 15 %

Urea-Urine (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	36	100.0	0.0	0.0	260	4.4	e

One result was submitted but not published because the method group was too small. (< 4 results per group)

Uric Acid-Urine




MQ tolerance : 15 %

Uric Acid-Urine (mmol/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Standard chemistry	26	100.0	0.0	0.0	1.08	3.5	e

3 additional results were submitted but not published because the method groups were too small. (< results per group)

Specific Gravity-Urine




MQ tolerance : 5 %

Specific Gravity-Urine ()

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Refractometer	5	100.0	0.0	0.0	1.022	0.2	e

Ethylglucuronid




MQ tolerance : 25 %

Ethylglucuronid (mg/l)

No.	Method	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	all Participants	5	100.0	0.0	0.0	1.20	10.2	e*

Creatinine U




QUALAB tolerance : 24 %
(< 20.0: +/- 4.8 mg/l)

Creatinine U (mg/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	92.3	0.0	7.7	27.8	3.3	e
2	Roche, Cobas	22	100.0	0.0	0.0	25.3	5.5	e
3	Siemens	4	100.0	0.0	0.0	27.0	7.3	e*
4	Aution	9	100.0	0.0	0.0	10.0	0.0	e
5	AFIAS	13	76.9	15.4	7.7	27.4	16.5	e*
6	Afinion	511	95.6	2.2	2.2	25.1	9.3	e
7	Sysmex U	15	66.7	0.0	33.3	30.0	0.0	e
8	Other methods	4	25.0	0.0	75.0	30.0	0.0	a
9	DCA2000/Vantage	162	90.7	3.1	6.2	25.6	9.4	e
10	Siemens Clinitek	25	68.0	4.0	28.0	30.0	8.0	a

One result was submitted but not published because the method group was too small. (< 4 results per group)

Creatinin Urin




QUALAB tolerance : 21 %

Creatinin Urin (mmol/l)

No.	Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1	Abbott	13	100.0	0.0	0.0	3.5	4.0	e
2	Roche	28	100.0	0.0	0.0	3.7	4.5	e
3	Beckman	4	100.0	0.0	0.0	3.8	4.8	e
4	Siemens	5	100.0	0.0	0.0	3.4	10.0	e*
5	DCA2000/Vantage	160	90.6	3.1	6.3	3.9	7.2	e
6	Afinion	505	97.6	1.0	1.4	3.4	5.7	e
7	Standard chemistry	4	75.0	0.0	25.0	3.7	5.9	a
8	Sysmex U	12	83.3	0.0	16.7	4.4	0.0	e
9	Aution	9	55.6	0.0	44.4	4.4	0.0	e
10	Siemens Clinitek	22	77.3	0.0	22.7	4.4	0.2	a
11	Other methods	5	60.0	0.0	40.0	4.4	0.0	a

2 additional results were submitted but not published because the method groups were too small. (< results per group)

CMV NAT qn




CMV NAT qn (Log10 IU/m)

QUALAB tolerance : +/- 0.50 Log10 IU/m

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	4	100.0	0.0	0.0	4.85	6.5	e*

HBV NAT qn




HBV NAT qn (Log10 IU/m)

QUALAB tolerance : +/- 0.50 Log10 IU/m

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	9	100.0	0.0	0.0	0.00	0.0	e*

HCV NAT qn




HCV NAT qn (Log10 IU/m)

QUALAB tolerance : +/- 0.50 Log10 IU/m

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	10	100.0	0.0	0.0	3.81	4.0	e*

HIV1 NAT qn



HIV1 NAT qn (Log10 cp/m)

QUALAB tolerance : +/- 0.50 Log10 cp/m

No. Methode	Total	% OK	% insuff.	% outlier	Target	CV%	Type
1 all Participants	7	100.0	0.0	0.0	3.69	5.3	e*