

Address for correspondence
European Reference Laboratory
C. Siebelder
Beatrixpark 1
7101 BN Winterswijk
The Netherlands
Telephone: + 31 543 54 47 74
e-mail: c.siebelder@skbwinterswijk.nl
Chamber of Commerce: 09101380

Winterswijk, 3 March 2026

Dear Dr. Fried,

Please find below the IFCC and NGSP certified values of your HbA1c samples. Values have been assigned in duplicate with three IFCC calibrated Secondary Reference Measurement Procedures. NGSP values are derived from the IFCC values using the Master Equation.

Sample ID	Assigned IFCC value mmol/mol	U (k=2)	Assigned NGSP value %	U (k=2)
K3A 2026-1	41,5	1,0	5,95	0,09
K3B 2026-1	60,4	1,0	7,68	0,09

Samples arrived on 27 February and were analysed on 2 March.

Yours sincerely.



Carla Siebelder BSc.
Coordinator ERL